

HemoSphere
Advanced Monitor

Operator's manual



Edwards

Edwards HemoSphere Advanced Monitor Operator's Manual

Because of continuing product improvement, prices and specifications are subject to change without notice. Changes to this manual, either in response to user input or to continuing product improvements, are accomplished through reissue. If, in the normal use of this manual, errors, omissions, or incorrect data are noted, please contact Edwards Technical Support or your local Edwards representative.

Edwards Technical Support

United States and Canada (24 hours) 800.822.9837 or tech_support@edwards.com

Outside the U.S. and Canada (24 hours) 949.250.2222

Europe +8001.8001.801 or techserv_europe@edwards.com

In the UK 0870 606 2040 - Option 4

In Ireland 01 8211012 - Option 4

CAUTION Federal (USA) law restricts this device to sale by or on the order of a physician.

Manufactured by Edwards Lifesciences LLC
One Edwards Way
Irvine, CA 92614

Trademarks Edwards, Edwards Lifesciences, and the stylized E logo are trademarks of Edwards Lifesciences Corporation. Acumen, Acumen AFM, Acumen HPI, Acumen IQ, AFM, BD, CCombo, CCombo V, ClearSight, ClearSight Jr, CO-Set, CO-Set+, FloTrac, FloTrac Jr, FORE-SIGHT, ForeSight, FORE-SIGHT ELITE, ForeSight Jr, HemoSphere, HPI, PediaSat, Physiocal, Swan, Swan-Ganz, Swan-Ganz Jr, Time-In-Target, TruWave, and VitaWave are trademarks of Becton, Dickinson and Company. All other trademarks are the property of their respective owners.

For U.S. Patents that may apply, see bd.com/patents.

©2026 Becton, Dickinson and Company. All rights reserved.

HemoSphere advanced monitor with 8.5 release

Manual Version: 8.5B; Manual Release Date: MARCH 2026; Software Version: 04.05.xxx.xxx

Original Release Date: 9/30/2016



Edwards Lifesciences Services GmbH
Parkring 30
85748 Garching bei München
Germany



Edwards Lifesciences B.V.
Verlengde Poolseweg 16
4818 CL Breda, Netherlands

Using This Manual

Users and/or patients should report any serious incidents to the manufacturer and the Competent Authority of the Member State in which the user and/or patient is established.

The Edwards HemoSphere advanced monitor operator's manual is comprised of fifteen chapters, and eight appendices. Figures in this manual are intended for reference only and may not be an exact replication of the screens as a result of continuous software improvements.

Carefully read these instructions for use, which address the warnings, precautions, and residual risks for this medical device.

WARNING Read this operator's manual carefully before attempting to use the Edwards HemoSphere advanced monitor.

Refer to the instructions for use provided with each compatible accessory before using it with the HemoSphere advanced monitor.

CAUTION Inspect the HemoSphere advanced monitor and all accessories and equipment used with the monitor for damage prior to use. Damage may include cracks, scratches, dents, exposed electrical contacts, or any signs that the housing may be compromised.

WARNING To prevent injury to patient or user, damage to platform, or inaccurate measurements, do not use any damaged or non-compatible platform accessories, components or cables.

Chapter	Description
1	Introduction: Provides an overview of the HemoSphere advanced monitor
2	Safety and Symbols: Includes WARNINGS, CAUTIONS, and NOTES that are found in the manual, as well as illustrations of labels found on the HemoSphere advanced monitor and accessories
3	Installation and Setup: Provides information about setting up the HemoSphere advanced monitor and connections for the first time
4	HemoSphere Advanced Monitor Quick Start: Provides experienced clinicians and users of bedside monitors instructions for immediate monitor use
5	Navigating the HemoSphere Advanced Monitor: Provides information on monitoring screen views
6	User Interface Settings: Provides information about the various display settings including patient information, language and international units, alarm volume, system time, and system date. It also provides instructions for selecting the screen appearance.
7	Advanced Settings: Provides information on advanced settings including alarm targets, graphical scales, serial port setup, and Demo Mode
8	Data Export and Connectivity Settings: Provides information on monitor connectivity for transferring patient and clinical data
9	HemoSphere Invasive Swan-Ganz Module Monitoring: Describes procedures for setup and operation of continuous cardiac output, intermittent cardiac output, and right ventricular end diastolic volume monitoring using the Swan-Ganz module
10	HemoSphere Minimally-Invasive Pressure Cable Monitoring: Describes procedures for setup and operation of vascular pressure monitoring
11	HemoSphere Non-Invasive ClearSight Module Monitoring: Describes the methodology behind ClearSight technology and gives instructions for setup and application of patient monitoring equipment as well as how to measure non-invasive blood pressure, cardiac output, stroke volume, stroke volume variation, and systemic vascular resistance
12	Venous Oximetry Monitoring: Describes procedures for calibration and operation of oximetry (oxygen saturation) measurement
13	Tissue Oximetry Monitoring: Describes procedures for setup and operation of ForeSight tissue oximetry monitoring
14	Advanced Features: Describes the advanced monitoring features that are currently available for upgrade with the HemoSphere advanced monitoring platform
15	Help and Troubleshooting: Describes the Help menu and provides a list of faults, alerts, and messages, with causes and suggested actions

Appendix	Description
A	Specifications
B	Accessories
C	Equations for Calculated Patient Parameters
D	Monitor Settings and Defaults

Appendix	Description
E	<i>Thermodilution Computation Constants</i>
F	<i>Monitor Care, Service and Support</i>
G	<i>Guidance and Manufacturer's Declaration</i>
H	<i>Glossary</i>

Contents

1 Introduction

1.1 Intended Purpose of this Manual	28
1.2 Indications For Use	29
1.2.1 HemoSphere Advanced Monitor with HemoSphere Swan-Ganz Module	29
1.2.2 HemoSphere Advanced Monitor with HemoSphere Oximetry Cable	29
1.2.3 HemoSphere Advanced Monitor with HemoSphere Pressure Cable	29
1.2.4 HemoSphere Advanced Monitor with HemoSphere Technology Module and ForeSight Oximeter Cable	30
1.2.5 HemoSphere Advanced Monitor with HemoSphere ClearSight Module	30
1.2.6 HemoSphere Advanced Monitor with Acumen Assisted Fluid Management Feature and Acumen IQ Sensor	30
1.3 Contraindications For Use	31
1.3.1 HemoSphere Advanced Monitor with ForeSight Oximeter Cable	31
1.3.2 HemoSphere Advanced Monitor with HemoSphere ClearSight Module	31
1.4 Intended Use Statement	31
1.5 Expected Clinical Benefit	37
1.6 HemoSphere Advanced Monitor Hemodynamic Technology Connections	38
1.6.1 HemoSphere Swan-Ganz Module	38
1.6.2 HemoSphere Pressure Cable	39
1.6.3 HemoSphere Oximetry Cable	41
1.6.4 HemoSphere Technology Module	41
1.6.4.1 HemoSphere Technology Module and ForeSight Oximeter Cable	42
1.6.4.2 HemoSphere Technology Module and AFM Cable	42
1.6.5 HemoSphere ClearSight Module	42
1.6.6 Documentation and Training	44
1.7 Manual Style Conventions	44
1.8 Abbreviations Found in This Manual	45

2 Safety and Symbols

2.1 Safety Signal Words Definitions	47
2.1.1 Warning	47
2.1.2 Caution	47
2.1.3 Note	47
2.2 Warnings	48
2.3 Cautions	55
2.4 User Interface Symbols	62
2.5 Symbols on Product Labels	65
2.6 Applicable Standards	68
2.7 HemoSphere Advanced Monitor Essential Performance	68

3 Installation and Setup

3.1 Unpacking	69
3.1.1 Packaging Contents	69
3.1.2 Required Accessories for Platform Modules and Cables	70
3.2 HemoSphere Advanced Monitor Connection Ports	72
3.2.1 Monitor Front	72
3.2.2 Monitor Rear	73
3.2.3 Monitor Right Panel	74
3.2.4 Monitor Left Panel	74
3.3 HemoSphere Advanced Monitor Installation	75
3.3.1 Mounting Options and Recommendations	75
3.3.2 Battery Installation	76
3.3.3 Connecting Power Cord	77
3.3.3.1 Equipotential Connection	77
3.3.4 Connecting and Disconnecting a Hemodynamic Monitoring Module ...	78
3.3.5 Connecting and Disconnecting a Hemodynamic Monitoring Cable. ...	78
3.3.6 Connecting Cables from External Devices	78
3.4 Initial Start Up	79
3.4.1 Start Up Procedure	79
3.4.2 Select Language	79
3.4.3 Select Device ID	80
3.5 Power Off and Power Save Mode	80
4 HemoSphere Advanced Monitor Quick Start	
4.1 HemoSphere Swan-Ganz Module Cardiac Output Monitoring	82
4.1.1 Continuous Cardiac Output Monitoring	83
4.1.2 Intermittent Cardiac Output Monitoring	83
4.1.3 Continuous End Diastolic Volume Monitoring	84
4.2 Monitoring with the HemoSphere Pressure Cable	85
4.2.1 Pressure Cable Setup	85
4.2.2 Zero Pressure Cable	85
4.3 HemoSphere Oximetry Cable Monitoring	87
4.3.1 In Vitro Calibration	87
4.3.2 In Vivo Calibration	88
4.4 HemoSphere Tissue Oximetry Monitoring	90
4.4.1 Connecting the HemoSphere Technology Module	90
4.5 HemoSphere ClearSight Module Monitoring	92
4.5.1 Connecting the HemoSphere Non-Invasive System	92
5 Navigating the HemoSphere Advanced Monitor	
5.1 HemoSphere Advanced Monitor Screen Appearance	96
5.2 Navigation Bar	98
5.3 Monitor Views	101
5.3.1 Parameter Tiles	102
5.3.1.1 Change Parameters	102
5.3.1.2 Change Alarm/Target	103
5.3.1.3 Status Indicators	104

5.3.2 Main Monitoring View	105
5.3.3 Graphical Trend Monitoring View	106
5.3.3.1 Graphical Trend Scroll Mode	107
5.3.3.2 Intervention Events	108
5.3.3.3 Live Blood Pressure Waveform Display	110
5.3.4 Tabular Trends	111
5.3.4.1 Tabular Trend Scroll Mode	111
5.3.5 Graphical/Tabular Trends Split	112
5.3.6 Physiology Screen	113
5.3.6.1 SVV Slope Indicator	114
5.3.7 Cockpit Screen	114
5.3.8 Physio Relationship	115
5.3.8.1 Continuous and Historical Modes	115
5.3.8.2 Parameter Boxes	117
5.3.8.3 Setting Targets and Entering Parameter Values	117
5.3.9 Goal Positioning Screen	118
5.4 Focused Monitoring Format	118
5.4.1 Select Monitoring View	118
5.4.2 Blood Pressure Waveform Tile	119
5.4.3 Focused Parameter Tile	119
5.4.4 Change Parameters	120
5.4.5 Change Alarms /Targets	120
5.4.6 Focused Main Screen	120
5.4.7 Focused Graphical Trend Screen	121
5.4.8 Focused Charting Screen	121
5.5 Clinical Tools	122
5.5.1 Select Monitoring Mode	122
5.5.2 CVP Entry	123
5.5.3 Derived Value Calculator	124
5.5.4 Event Review	124
5.6 Information Bar	129
5.6.1 Device ID	130
5.6.2 Battery	130
5.6.3 Screen Brightness	131
5.6.4 Alarm Volume	131
5.6.5 Screen Capture	131
5.6.6 Lock Screen	131
5.7 Status Bar	132
5.8 Monitor Screen Navigation	132
5.8.1 Vertical Scrolling	132
5.8.2 Navigation Icons	133
6 User Interface Settings	
6.1 Password Protection	135
6.1.1 Changing passwords	136
6.2 Patient Data	137

6.2.1 New Patient	137
6.2.2 Continue Monitoring Patient.....	139
6.2.3 View Patient Data	139
6.3 General Monitor Settings	139
6.3.1 Change Language	139
6.3.2 Change Date and Time Display.....	140
6.3.2.1 Adjust Date or Time	140
6.3.3 Monitoring Screens Settings	141
6.3.4 Time Intervals / Averaging	141
6.3.4.1 Display Parameter Value Change.....	142
6.3.4.2 CO/Pressure Averaging Time	142
6.3.5 Analog Pressure Signal Input.....	143
6.3.5.1 Calibration	145
7 Advanced Settings	
7.1 Alarms / Targets	147
7.1.1 Silence Alarms	148
7.1.1.1 Physiological Alarms.....	148
7.1.1.2 Technical Alarms.....	149
7.1.2 Set Alarm Volume	149
7.1.3 Set Targets	149
7.1.4 Alarms / Targets Setup Screen	150
7.1.5 Configure All Targets.....	151
7.1.6 Configure Targets and Alarms for One Parameter	152
7.2 Adjust Scales	153
7.3 Physiology and Physio Relationship Screen SVV/PPV Parameter Settings ..	155
7.4 CVP Settings	155
7.5 20-Second Flow Parameter Settings	156
7.6 Demo Mode	156
8 Data Export and Connectivity Settings	
8.1 Export Data	158
8.1.1 Data Download	158
8.1.1.1 Monitoring Data	159
8.1.1.2 Case Report.....	159
8.1.1.3 GDT Report.....	160
8.1.2 Diagnostic Export	160
8.2 Wireless Settings	161
8.3 HIS Connectivity	161
8.3.1 Patient Demographic Data	162
8.3.2 Patient Physiological Data.....	163
8.3.3 Physiological Alarms and Device Faults	163
8.4 Viewfinder Hub Connectivity	164
8.4.1 Viewfinder Hub Pairing.....	164
8.4.2 Patient Data	165
8.4.3 Physiological Alarms and Device Faults	165

8.4.4 Software Upgrades	165
8.5 Cyber Security	165
8.5.1 Cybersecurity Updates	165
8.5.2 Vulnerability Management	166
8.5.3 Cybersecurity Incident Response	166
8.5.4 HIPAA	166
9 HemoSphere Swan-Ganz Module Monitoring	
9.1 Connecting the HemoSphere Swan-Ganz Module	167
9.1.1 Patient CCO Cable Test	169
9.1.2 Parameter Selection Menu	170
9.2 Continuous Cardiac Output	171
9.2.1 Connecting the Patient Cables	171
9.2.2 Initiating Monitoring	172
9.2.3 Thermal Signal Conditions	173
9.2.4 CO Countdown Timer	173
9.2.5 STAT CO	174
9.2.6 20-Second Flow Parameters	174
9.2.6.1 PAP Waveform Troubleshooting	174
9.3 Intermittent Cardiac Output	174
9.3.1 Connecting Patient Cables	175
9.3.1.1 Probe Selection	175
9.3.2 Configuration Settings	176
9.3.2.1 Select Injectate Volume	176
9.3.2.2 Select Catheter Size	177
9.3.2.3 Select Computation Constant	177
9.3.2.4 Select Mode	177
9.3.3 Instructions for Bolus Measurement Modes	177
9.3.4 Thermodilution Summary Screen	179
9.4 EDV/RVEF Monitoring	180
9.4.1 Connecting Patient Cables	180
9.4.2 Connecting the ECG Interface Cable	181
9.4.3 Initiating Measurement	182
9.4.4 Active EDV Monitoring	183
9.4.5 STAT EDV and RVEF	184
9.5 SVR	184
9.6 Multiple Technology Monitoring - Acumen Hypotension Prediction Index Software	184
10 Monitoring with the HemoSphere Pressure Cable	
10.1 Pressure Cable Overview	187
10.2 Monitoring Mode Selection	190
10.3 FloTrac Sensor Monitoring	190
10.3.1 Connect FloTrac, FloTrac Jr, or Acumen IQ Sensor	191
10.3.2 Set Averaging Time	191
10.3.3 Zero Arterial Pressure	192

10.3.4 SVR Monitoring	193
10.3.5 Acumen IQ Sensor Parameters Displayed in Invasive Mode	193
10.4 Pressure Cable Monitoring with a TruWave DPT	193
10.4.1 Connect TruWave DPT	194
10.4.2 Zero Intravascular Pressure	194
10.5 Zero & Waveform Screen	196
10.5.1 Select Pressure and Zero Sensor	196
10.5.2 Pressure-Out	196
10.5.3 Waveform Confirmation	198
11 HemoSphere ClearSight Module Non-Invasive Monitoring	
11.1 HemoSphere Non-Invasive System Methodology	199
11.1.1 Volume Clamp Method	199
11.1.2 Physiocal Method	200
11.1.3 Waveform Reconstruction and Hemodynamic Analysis (ClearSight Technology)	200
11.1.4 Heart Reference Sensor	200
11.1.5 Discoloration, Numbness, or Tingling of the Fingertip	201
11.1.6 Single Cuff Monitoring	201
11.1.7 Double Cuff Monitoring	201
11.1.8 Methodology References	202
11.2 Connecting the HemoSphere Non-Invasive System	202
11.2.1 Applying the Pressure Controller	205
11.2.1.1 Wrist Applied Pressure Controller	205
11.2.1.2 Smart Pressure Controller	207
11.2.2 Select Finger Cuff Size	209
11.2.3 Apply Finger Cuff(s)	209
11.2.4 Apply the Heart Reference Sensor	210
11.2.5 Accuracy of ClearSight Technology Blood Pressure Measurements	210
11.2.6 General Troubleshooting of HemoSphere Non-Invasive System Monitoring	211
11.3 Optional HRS	213
11.3.1 Patient Sedated and Stationary	213
11.3.2 Update Offset Value During Monitoring	215
11.3.3 Change Patient Positioning Mode	216
11.4 SQI	216
11.5 Physiocal Method Display	217
11.6 ClearSight Settings and Cuff Options	217
11.6.1 Calibrate the Heart Reference Sensor	218
11.6.2 Cuff Pressure Release Mode	219
11.7 Blood Pressure Calibration	219
11.8 Output Signal to Patient Monitor	221
12 Venous Oximetry Monitoring	
12.1 Oximetry Cable Overview	223
12.2 Venous Oximetry Setup	223

12.3 In Vitro Calibration	225
12.3.1 In Vitro Calibration Error	226
12.4 In Vivo Calibration	226
12.5 Signal Quality Indicator	228
12.6 Recall Venous Oximetry Data	228
12.7 HGB Update	230
12.8 HemoSphere Oximetry Cable Reset	230
12.9 New Catheter	231
13 HemoSphere Tissue Oximetry Monitoring	
13.1 HemoSphere Tissue Oximetry Monitoring	232
13.2 ForeSight Oximeter Cable Overview	233
13.2.1 ForeSight Oximeter Cable Mounting Solutions	234
13.2.2 Installing the Mounting Clip	234
13.2.2.1 Attaching the Mounting Clip Vertically	235
13.2.2.2 Attaching the Mounting Clip Horizontally	235
13.2.3 Removing the Mounting Clip	236
13.3 Connecting the HemoSphere Technology Module and ForeSight Oximeter Cable	238
13.3.1 Attaching Sensors to the Patient	242
13.3.1.1 Selecting a Sensor Site	242
13.3.1.2 Preparing the Sensor Site	243
13.3.1.3 Applying Sensors	244
13.3.1.4 Connecting Sensors to Cables	246
13.3.2 Disconnecting Sensors After Monitoring	247
13.3.3 Monitoring Considerations	248
13.3.3.1 ForeSight Oximeter Cable Use During Defibrillation	248
13.3.3.2 Interference	248
13.3.3.3 Interpreting StO ₂ Values	248
13.3.4 Skin Check Timer	250
13.3.5 Set Averaging Time	251
13.3.6 Signal Quality Indicator	251
13.3.7 Relative Change in Total Hemoglobin – Δ ctHb	251
13.3.8 Tissue Oximetry Physiology Screen	252
14 Advanced Features	
14.1 Acumen Hypotension Prediction Index (HPI) Software Feature	253
14.1.1 Introduction to Acumen Hypotension Prediction Index (HPI) Software in Minimally-Invasive Mode	253
14.1.1.1 Displaying Acumen Hypotension Prediction Index (HPI) Software in Invasive Mode	255
14.1.2 Introduction to Acumen Hypotension Prediction Index (HPI) Software in Non-Invasive Mode	255
14.1.3 Acumen Hypotension Prediction Index Parameters Overview	256
14.1.4 Acumen Hypotension Prediction Index (HPI) Parameter Display	257
14.1.5 HPI as a Key Parameter	258
14.1.6 HPI Alarm	260

14.1.7 HPI on Information Bar	261
14.1.8 Disable HPI Information Bar Indicator	261
14.1.9 HPI High Alert Popup.	261
14.1.10 Hypotension Threshold Setting.	262
14.1.11 HPI Secondary Screen	263
14.1.11.1 Navigate to HPI Secondary Screen	264
14.1.11.2 HPI Relationship View	264
14.1.11.3 HPI Smart Alerts and Smart Trends	266
14.1.12 Clinical Application	269
14.1.13 Additional Parameters	270
14.1.14 Clinical Validation with Hypotension Threshold in Minimally-Invasive Monitored Patients	272
14.1.14.1 Surgical Patients	273
14.1.14.2 Non-Surgical Patients	274
14.1.14.3 Hypotension Threshold Clinical Validation Study Results – Minimal- ly-Invasive Monitoring	275
14.1.15 Clinical Validation with Hypotension Threshold in Non-Invasively Monitored Patients	281
14.1.15.1 Surgical Patients	281
14.1.15.2 Non-Surgical Patients	282
14.1.15.3 Hypotension Threshold Clinical Validation Study Results – Non-In- vasive Monitoring	283
14.1.16 Additional Clinical Data.	289
14.1.16.1 Study Design	289
14.1.16.2 Patient Demographics.	291
14.1.16.3 Study Results	293
14.1.16.4 Study Summary	297
14.1.16.5 Conclusion.	298
14.1.17 References	298
14.2 Assisted Fluid Management	299
14.2.1 Introduction	299
14.2.2 Principle of Operation	299
14.2.3 Help Screens for AFM.	301
14.2.4 Starting or Restarting AFM	301
14.2.5 AFM Dashboard Display.	303
14.2.6 Assisted Fluid Management Settings.	303
14.2.6.1 Fluid Strategy.	303
14.2.6.2 Surgery Mode	304
14.2.6.3 Fluid Tracking.	304
14.2.6.4 Maximum Case Volume	304
14.2.6.5 Live Flow Rate	306
14.2.7 Managing Fluids with the AFM Algorithm.	307
14.2.7.1 Fluid Administration Workflow – Acumen IQ Fluid Meter.	307
14.2.7.2 Fluid Administration Workflow – Manual Mode	312
14.2.7.3 Approaching/Exceeding Maximum Case Volume Workflow	315
14.2.8 Fluid Bolus Information Popup	316
14.2.9 Pausing and Ending an AFM Session.	317

14.2.10	GDT Tracking During an AFM Session	317
14.2.11	Clinical Validation	318
14.2.11.1	Fluid Bolus Activity	320
14.2.12	Fluid Meter Only Mode	323
14.3	Enhanced Parameter Tracking	324
14.3.1	GDT Tracking	324
14.3.1.1	Key Parameter and Target Selection	324
14.3.1.2	Active GDT Tracking	325
14.3.1.3	Historical GDT	326
14.3.2	SV Optimization	326
14.3.3	GDT Report Download	326
14.4	Fluid Responsiveness Test	327
14.4.1	Passive Leg Raise Test	327
14.4.2	Fluid Bolus Test	329
14.4.3	Historical Test Results	330
14.5	Relative Change in Total Hemoglobin – Δ ctHb	330
14.5.1	Δ ctHb Value Display	330
14.5.2	Δ ctHb Trend Display	330
14.5.3	Reset Δ ctHb	331
14.5.4	Validation Methodology and Study Results	331
14.6	Aortic Stenosis Index (ASI) Algorithm	331
14.6.1	Indications For Use	332
14.6.2	Intended Use	332
14.6.3	Aortic Stenosis Index (ASI) Parameter Display	333
14.6.4	Performance Data - Clinical Testing Summary	333
15	Troubleshooting	
15.1	On Screen Help	335
15.2	Monitor Status Lights	336
15.3	Pressure Cable Communication	337
15.4	ForeSight Oximeter Cable Sensor Communication	338
15.5	Pressure Controller Communication	339
15.6	HemoSphere Advanced Monitor Error Messages	340
15.6.1	System/Monitoring Faults/Alerts	340
15.6.2	System/Monitoring Warnings	343
15.6.3	Numeric Keypad Errors	343
15.6.4	Viewfinder Hub Connectivity Errors	344
15.7	HemoSphere Swan-Ganz Module Error Messages	344
15.7.1	CO Faults/Alerts	344
15.7.2	EDV and SV Faults/Alerts	346
15.7.3	iCO Faults/Alerts	347
15.7.4	SVR Faults/Alerts	348
15.7.5	20-Second Parameters Faults/Alerts	348
15.7.6	General Troubleshooting	349
15.8	Pressure Cable Error Messages	350

15.8.1 General Pressure Cable Faults/Alerts	350
15.8.2 Arterial Pressure Faults/Alerts	351
15.8.3 Assisted Fluid Management Faults/Alerts	353
15.8.4 SVR Faults/Alert	354
15.8.5 General Troubleshooting	354
15.9 HemoSphere ClearSight Module Error Messages	355
15.9.1 Faults/Alerts	355
15.10 Venous Oximetry Error Messages	362
15.10.1 Venous Oximetry Faults/Alerts	362
15.10.2 Venous Oximetry Warnings	363
15.10.3 Venous Oximetry General Troubleshooting	364
15.11 Tissue Oximetry Error Messages	364
15.11.1 Tissue Oximetry Faults/Alerts	364
15.11.2 Tissue Oximetry General Troubleshooting	366
Appendix A: Specifications and Device Characteristics	
A.1 Essential Performance Characteristics	368
A.2 HemoSphere Advanced Monitor Characteristics and Specifications	370
A.3 HemoSphere Battery Pack Characteristics and Specifications	372
A.4 HemoSphere Swan-Ganz Module Characteristics and Specifications	373
A.5 HemoSphere Pressure Cable Characteristics and Specifications	375
A.6 HemoSphere Oximetry Cable Characteristics and Specifications	376
A.7 HemoSphere Tissue Oximetry Characteristics and Specifications	377
A.8 HemoSphere ClearSight Module Characteristics and Specifications	378
A.9 Acumen AFM Cable Characteristics and Specifications	380
Appendix B: Accessories	
B.1 Accessories List	381
B.2 Additional Accessories Description	382
B.2.1 Roll Stand.	382
B.2.2 Oximetry Cradle	382
B.2.3 Pressure Controller Cover.	383
Appendix C: Equations for Calculated Patient Parameters	
Appendix D: Monitor Settings and Defaults	
D.1 Patient Data Input Range	390
D.2 Trend Scale Default Limits	390
D.3 Parameter Display and Configurable Alarm/Target Ranges	391
D.4 Alarm and Target Defaults	393
D.5 Alarm Priorities	395
D.6 Language Default Settings	396
Appendix E: Computation Constants	
E.1 Computation Constant Values	397
Appendix F: System Care, Service and Support	
F.1 General Maintenance	399

F.2 Cleaning the Monitor and Modules	400
F.3 Cleaning the Platform Cables	400
F.3.1 Cleaning the HemoSphere Oximetry Cable	401
F.3.2 Cleaning the Patient CCO Cable and Connector	401
F.3.3 Cleaning the HemoSphere Pressure Cable	402
F.3.4 Cleaning the ForeSight Oximeter Cable	402
F.3.5 Cleaning the Heart Reference Sensor and Pressure Controller	403
F.3.5.1 Removing the Pressure Controller Band	404
F.3.6 Cleaning the Smart Pressure Controller	404
F.4 Service and Support	404
F.5 Edwards Lifesciences Regional Headquarters	405
F.6 Monitor Disposal	406
F.6.1 Battery Recycling	406
F.7 Preventive Maintenance	406
F.7.1 Battery Maintenance	406
F.7.1.1 Battery Conditioning	406
F.7.1.2 Battery Storage	406
F.7.2 HemoSphere ClearSight Module Maintenance	407
F.7.3 HRS Preventive Maintenance	407
F.8 Testing of Alarm Signals	407
F.9 Warranty	407
Appendix G: Guidance and Manufacturer's Declaration	
G.1 Electromagnetic Compatibility	409
G.2 Instructions for Use	410
G.3 Wireless Technology Information	418
G.3.1 Quality of Service for Wireless and Wired Technology	420
G.3.2 Wireless Security Measures	421
G.3.3 Troubleshooting Wireless Coexistence Issues	421
G.3.4 Federal Communication Commission (FCC) Interference Statements	421
G.3.5 Industry Canada Statements	422
G.3.6 European Union Radio Equipment Directive (RED) Statements	423
Appendix H: Glossary	

List of Figures

Figure 1-1 HemoSphere advanced monitor hemodynamic technology connections	38
Figure 3-1 HemoSphere advanced monitor front view	72
Figure 3-2 HemoSphere advanced monitor rear view (shown with HemoSphere Swan-Ganz module)	73
Figure 3-3 HemoSphere advanced monitor right panel	74
Figure 3-4 HemoSphere advanced monitor left panel (shown with no modules)	74
Figure 3-5 HemoSphere advanced monitor power entry cover - screw locations	77
Figure 3-6 Startup screen	79
Figure 3-7 Language selection screen	80
Figure 4-1 HemoSphere Swan-Ganz module monitoring connection overview	82
Figure 4-2 Pressure cable connection overview	85
Figure 4-3 Oximetry connection overview	87
Figure 4-4 Tissue oximetry monitoring connection overview	90
Figure 4-5 HemoSphere non-invasive system connection overview (wrist applied pressure controller)	92
Figure 4-6 HemoSphere non-invasive system connection overview (Smart pressure controller)	93
Figure 5-1 HemoSphere advanced monitor screen features	97
Figure 5-2 Navigation bar and icons 1invasive monitoring, 2non-invasive monitoring, 3graphical trend screens	98
Figure 5-3 Example of monitoring screen selection window	102
Figure 5-4 Example of key parameter selection tile configuration menu	103
Figure 5-5 Parameter tile	104
Figure 5-6 Main monitoring view	106
Figure 5-7 Graphical trend screen	107
Figure 5-8 Graphical trend- intervention window	108
Figure 5-9 Graphical trend screen - intervention information balloon	110
Figure 5-10 Tabular trend screen	111
Figure 5-11 Tabular increment popup	111
Figure 5-12 Physiology screen while monitoring with HemoSphere Swan-Ganz module	113
Figure 5-13 Cockpit monitoring screen	114
Figure 5-14 Physio relationship screen while monitoring with HemoSphere Swan-Ganz module	115
Figure 5-15 Historic physio relationship data screen	116
Figure 5-16 Physio relationship parameter boxes	117
Figure 5-17 Physio relationship target/enter popup	117
Figure 5-18 Goal positioning screen	118

Figure 5-19 Focused parameter tile	119
Figure 5-20 Focused parameter tile – parameter and alarm/target selection	120
Figure 5-21 Focused main screen	120
Figure 5-22 Focused Graphical Trend Screen	121
Figure 5-23 Focused Charting Screen	122
Figure 5-24 Focused Charting View – Configure Columns	122
Figure 5-25 Information bar	130
Figure 5-26 Lock screen popup	132
Figure 5-27 Status bar	132
Figure 6-1 New or continuing patient screen	137
Figure 6-2 New Patient Data screen	138
Figure 6-3 General Monitor Settings	140
Figure 7-1 Alarms / Targets configuration	151
Figure 7-2 Set individual parameter alarms and targets	152
Figure 7-3 Graphical trend screen	153
Figure 7-4 Adjust scales	154
Figure 7-5 Tabular increment popup	155
Figure 8-1 Data Download screen	159
Figure 8-2 HIS- Patient query screen	162
Figure 8-3 HIS- New patient data screen	163
Figure 9-1 HemoSphere Swan-Ganz module connection overview	168
Figure 9-2 Patient CCO cable test connections	170
Figure 9-3 HemoSphere Swan-Ganz module key parameter selection window	171
Figure 9-4 CO connection overview	172
Figure 9-5 iCO connection overview	175
Figure 9-6 iCO new set configuration screen	176
Figure 9-7 Thermodilution summary screen	179
Figure 9-8 EDV/RVEF connection overview	181
Figure 9-9 Main monitoring view – Swan-Ganz module monitoring with an Acumen IQ sensor	185
Figure 10-1 HemoSphere pressure cable	188
Figure 10-2 Zero & Waveform screen – Zero sensor and Pressure-Out	196
Figure 10-3 HemoSphere pressure-out cable	197
Figure 11-1 HemoSphere non-invasive system connection overview (wrist applied pressure controller)	203
Figure 11-2 HemoSphere non-invasive system connection overview (Smart pressure controller)	204
Figure 11-3 Wrist applied pressure controller application	206
Figure 11-4 Smart pressure controller application	207
Figure 11-5 Smart pressure controller clip(s)	208

Figure 11-6 Cuff size selection	209
Figure 11-7 Heart reference sensor application	210
Figure 11-8 Patient Positioning Mode Selection – Optional HRS	213
Figure 11-9 Zero & Waveform screen – Vertical Offset Entry	215
Figure 11-10 Heart reference sensor calibration	218
Figure 11-11 HRS calibration screen	218
Figure 11-12 BP Calibration Screen	220
Figure 11-13 Transmit arterial pressure waveform to patient monitor	222
Figure 12-1 Venous oximetry connection overview	224
Figure 13-1 ForeSight oximeter cable front view	233
Figure 13-2 ForeSight oximeter cable rear view	233
Figure 13-3 Mounting clip attachment points	234
Figure 13-4 Cable housing – mounting clip attachment points	234
Figure 13-5 Attaching the mounting clip vertically	235
Figure 13-6 Attaching the mounting clip horizontally	236
Figure 13-7 Removing the mounting clip	237
Figure 13-8 Tissue oximetry monitoring connection overview	238
Figure 13-9 ForeSight oximeter cable status LED	240
Figure 13-10 Removing protective liner from sensor	244
Figure 13-11 Sensor placement (cerebral)	244
Figure 13-12 Sensor placement (non-cerebral)	245
Figure 13-13 Connecting a sensor to the sensor cable connector	246
Figure 13-14 Connecting a sensor to the ForeSight oximeter cable – channel status LED ...	247
Figure 13-15 Tissue Oximetry Physiology Screens	252
Figure 14-1 HPI key parameter tile	259
Figure 14-2 HPI key parameter on cockpit screen	260
Figure 14-3 Information bar with HPI	261
Figure 14-4 HPI high alert popup	262
Figure 14-5 HPI parameter hypotension threshold settings screen	263
Figure 14-6 HPI secondary screen - relationship view	265
Figure 14-7 HPI secondary screen – relationship view with graphical trend value display ...	266
Figure 14-8 HPI smart alert popup	267
Figure 14-9 HPI secondary screen – smart trend display	268
Figure 14-10 HPI secondary screen – smart trend view settings	269
Figure 14-11 Bland-Altman plots for SVV, PPV, and Eadyn	272
Figure 14-12 AFM dashboard	303
Figure 14-13 AFM dashboard with Live Flow Rate enabled	306
Figure 14-14 Enabling Live Flow Rate	306
Figure 14-15 Acumen IQ fluid meter and Acumen AFM cable connection overview	308
Figure 14-16 GDT Menu Screen - Key Parameter Selection	324

Figure 14-17 GDT Menu Screen - Target Selection	324
Figure 14-18 GDT Active Tracking	325
Figure 14-19 Fluid Responsiveness Test – New Test Screen	327
Figure 14-20 Fluid Responsiveness Test – Results Screen	329
Figure 14-21 Example ASI outputs and display user interface	333
Figure 15-1 HemoSphere advanced monitor LED indicators	336
Figure 15-2 Pressure cable LED indicator	337
Figure 15-3 ForeSight oximeter cable LED indicators	338
Figure 15-4 Pressure Controller LED Indicators	339
Figure A-1 Spectral Irradiance and location of light emission aperture	379
Figure B-1 Applying pressure controller cover	383
Figure F-1 Removing pressure controller from band	404

List of Tables

Table 1-1 HemoSphere Swan-Ganz module available parameters list	32
Table 1-2 HemoSphere oximetry cable available parameters list	33
Table 1-3 HemoSphere Swan-Ganz module with oximetry cable available parameters list	33
Table 1-4 HemoSphere Swan-Ganz module with pressure cable available parameters list* ...	33
Table 1-5 HemoSphere pressure cable available parameters list	34
Table 1-6 HemoSphere pressure cable available AFM output list	35
Table 1-7 HemoSphere pressure cable with oximetry cable available parameters list	35
Table 1-8 HemoSphere technology module with ForeSight oximeter cable available parameters list	36
Table 1-9 HemoSphere ClearSight module available parameters list	36
Table 1-10 HemoSphere ClearSight module with oximetry cable available parameters list	37
Table 1-11 HemoSphere Swan-Ganz module parameters description	39
Table 1-12 HemoSphere pressure cable key parameters description	40
Table 1-13 HemoSphere oximetry cable parameters description	41
Table 1-14 HemoSphere technology module with ForeSight oximeter cable parameters description	42
Table 1-15 HemoSphere ClearSight module key parameters description	43
Table 1-16 Operator's manual style conventions	44
Table 1-17 Acronyms, Abbreviations	45
Table 2-1 Monitor display symbols	63
Table 2-2 Symbols on product labels	66
Table 2-3 Applicable standards	68
Table 3-1 HemoSphere advanced monitoring components	69
Table 3-2 Cables and catheters required for monitoring parameters with HemoSphere Swan-Ganz module	70
Table 3-3 Sensor options for monitoring parameters with HemoSphere pressure cable	70
Table 3-5 Catheters required for monitoring parameters with HemoSphere oximetry cable ..	71
Table 3-6 Accessories required for monitoring parameters with HemoSphere technology module	71
Table 3-4 Finger cuff options for monitoring parameters with HemoSphere ClearSight module	71
Table 5-1 Graphical trend scroll rates	107
Table 5-2 Intervention events	109
Table 5-3 Tabular trend scroll rates	112
Table 5-4 CVP value prioritization	124
Table 5-5 Reviewed events	125
Table 5-6 Battery status	131

Table 6-1 HemoSphere advanced monitor password levels	135
Table 6-2 Advanced setup menu navigation and password protection	135
Table 6-3 Export data menu navigation and password protection	136
Table 6-4 CO/pressure averaging time and display update rates – in minimally-invasive monitoring mode	142
Table 6-5 Analog input parameter ranges	145
Table 7-1 Visual alarm indicator colors	147
Table 7-2 Target status indicator colors	150
Table 7-3 Target defaults	150
Table 8-1 Wi-Fi connection status	161
Table 8-2 HIS connectivity status	162
Table 8-3 Viewfinder hub connectivity status	164
Table 9-1 Available HemoSphere Swan-Ganz module parameters and required connections	169
Table 9-2 Unstable thermal signal time lapse for CO alert and fault messages	173
Table 9-3 Acumen IQ sensor parameter availability in Invasive mode	185
Table 10-1 HemoSphere pressure cable configurations and available key parameters	188
Table 11-1 95% Confidence interval (CI) results for repeated blood pressure measurements from the same patient (Bootstrap Re-sampling)	211
Table 11-2 Arterial waveform SQI levels	216
Table 11-3 Physiocal method interval status	217
Table 11-4 BP Calibration performance data	220
Table 12-1 In vitro calibration options	226
Table 12-2 In vivo calibration options	227
Table 12-3 Signal quality indicator levels	228
Table 13-1 Tissue oximetry sensor locations	241
Table 13-2 Sensor selection matrix	243
Table 13-3 StO2 validation methodology	249
Table 13-4 Clinical validation study results for StO2	250
Table 13-5 Signal quality indicator levels	251
Table 14-1 HPI display configurations	257
Table 14-2 HPI value graphical and audible display elements	258
Table 14-3 HPI versus other key parameters: similarities and differences	259
Table 14-4 Parameter status colors for HPI	260
Table 14-5 HPI smart alert parameter default thresholds	268
Table 14-6 dP/dt accuracy comparison of minimally-invasive and non-invasive monitored surgical patients	271
Table 14-7 95% Confidence interval (CI) results for bias and limits of agreement (LoA)	272
Table 14-8 Patient demographics (minimally-invasive monitored surgical patients, N=1141)	273
Table 14-9 Analysis dataset length and number of hypotension events (minimally-invasive monitored surgical patients)	273
Table 14-10 Surgical patient characteristics (minimally-invasive, N=1141)	273

Table 14-11 Patient demographics (minimally-invasive monitored non-surgical patients, N=672)	274
Table 14-12 Analysis dataset length and number of hypotension events (minimally-invasive monitored non-surgical patients)	274
Table 14-13 Non-Surgical patient characteristics (minimally-invasive, N=672)	274
Table 14-14 Clinical validation study results* (minimally-invasive monitored surgical patients)	276
Table 14-15 Clinical validation study results* (minimally-invasive monitored non-surgical patients)	277
Table 14-16 Event rate analysis (surgical minimally-invasive, N=1141)	278
Table 14-17 Event rate analysis (non-surgical minimally-invasive, N=672)	279
Table 14-18 Patient demographics (non-invasively monitored surgical patients, N=927)	281
Table 14-19 Analysis dataset length and number of hypotension events (non-invasively monitored surgical patients)	281
Table 14-20 Surgical patient characteristics (non-invasive, N=927)	281
Table 14-21 Patient demographics (non-invasively monitored non-surgical patients, N=424)	282
Table 14-22 Analysis dataset length and number of hypotension events (non-invasively monitored non-surgical patients)	283
Table 14-23 Non-Surgical patient characteristics (non-invasive, N=424)	283
Table 14-24 Clinical validation study results* (non-invasively monitored surgical patients)	284
Table 14-25 Clinical validation study results* (non-invasively monitored non-surgical patients)	285
Table 14-26 Event rate analysis (surgical non-invasive, N=927)	286
Table 14-27 Event rate analysis (non-surgical non-invasive, N=424)	287
Table 14-28 HPI prospective subject selection criteria	290
Table 14-29 MPOG historical control patient selection criteria	290
Table 14-30 Patient demographics (MPOG study)	291
Table 14-31 Procedure type (HPI)	292
Table 14-32 Surgery type by CPT grouping	293
Table 14-33 Receiver operating characteristics (ROC) for HPI subjects (N=482)*	293
Table 14-34 Mean IOH duration – Primary effectiveness endpoint	294
Table 14-35 Intraoperative hypotension AUC - ITT, pivotal subjects	294
Table 14-36 Effectiveness stratified by MAP level, HPI study versus MPOG historical control	295
Table 14-37 Frequency pattern of subjects and intervention instances by HPI threshold	296
Table 14-39 Length of stay	297
Table 14-38 HPI Study - 30 days postoperative composite endpoint components - CC analysis population (pivotal subjects, n=400)	297
Table 14-40 AFM states	299
Table 14-41 AFM fluid status icons	305
Table 14-42 Subject demographics	318
Table 14-43 AFM response rates by bolus type	319
Table 14-44 AFM performance by bolus volume (mL)	319
Table 14-45 Accuracy results of the AFM feature (bolus level)	320

Table 14-46 Frequency of AFM recommendations per hour**	320
Table 14-47 Complete accounting of fluid boluses	321
Table 14-48 Reasons boluses were discarded (analysis declined) in the per protocol pivotal subjects	321
Table 14-49 Reasons suggestions were declined in the per protocol pivotal subjects	322
Table 14-50 GDT Target Status Indicator Colors	325
Table 14-51 Clinical and blood bench validation study results for trending accuracy of relative change in hemoglobin (Δ ctHb)	331
Table 14-52 General interpretations and recommended actions for ASI algorithm	333
Table 14-53 ASI performance in differentiating severe AS subjects from patients without severe AS	334
Table 15-3 Pressure cable communication light	337
Table 15-1 HemoSphere advanced monitor visual alarm indicator	337
Table 15-2 HemoSphere advanced monitor power light	337
Table 15-4 ForeSight oximeter cable LED communication lights	338
Table 15-5 Pressure controller communication lights*	339
Table 15-6 System faults/alerts	340
Table 15-7 HemoSphere advanced monitor warnings	343
Table 15-8 Numeric keypad errors	343
Table 15-9 Viewfinder hub connectivity errors	344
Table 15-10 HemoSphere Swan-Ganz module CO faults/alerts	344
Table 15-11 HemoSphere Swan-Ganz module EDV and SV faults/alerts	346
Table 15-12 HemoSphere Swan-Ganz module iCO faults/alerts	347
Table 15-13 HemoSphere Swan-Ganz module SVR faults/alerts	348
Table 15-14 HemoSphere Swan-Ganz module 20s parameters faults/alerts	348
Table 15-15 HemoSphere Swan-Ganz module general troubleshooting	349
Table 15-16 HemoSphere pressure cable general faults/alerts	350
Table 15-17 HemoSphere pressure cable ART faults/alerts	351
Table 15-18 HemoSphere pressure cable AFM faults/alerts	353
Table 15-20 HemoSphere pressure cable SVR faults/alerts	354
Table 15-21 HemoSphere pressure cable general troubleshooting	354
Table 15-19 HemoSphere pressure cable AFM warnings	354
Table 15-22 HemoSphere ClearSight module faults/alerts	355
Table 15-23 HemoSphere ClearSight warnings	360
Table 15-24 HemoSphere ClearSight general troubleshooting	361
Table 15-25 Venous oximetry faults/alerts	362
Table 15-26 Venous oximetry warnings	363
Table 15-27 Venous oximetry general troubleshooting	364
Table 15-28 Tissue oximetry faults/alerts	364
Table 15-29 Tissue oximetry general troubleshooting	366

Table A-1 HemoSphere advanced monitor essential performance – transient and non-transient electromagnetic phenomena	369
Table A-2 HemoSphere advanced monitor physical and mechanical characteristics	370
Table A-4 HemoSphere advanced monitor transportation environmental specifications	371
Table A-5 HemoSphere advanced monitor technical characteristics	371
Table A-3 HemoSphere advanced monitor environmental specifications	371
Table A-6 HemoSphere battery pack physical characteristics	372
Table A-9 HemoSphere Swan-Ganz module physical characteristics	373
Table A-10 HemoSphere Swan-Ganz module parameter measurement specifications	373
Table A-7 HemoSphere battery pack environmental specifications	373
Table A-8 HemoSphere battery pack technical characteristics	373
Table A-11 HemoSphere Swan-Ganz module 20-second flow parameter measurement specifications1	374
Table A-12 HemoSphere pressure cable physical characteristics	375
Table A-14 HemoSphere oximetry cable physical characteristics	376
Table A-15 HemoSphere oximetry cable parameter measurement specifications	376
Table A-13 HemoSphere pressure cable parameter measurement specifications	376
Table A-16 HemoSphere technology module physical characteristics	377
Table A-17 ForeSight oximeter cable physical characteristics	377
Table A-19 HemoSphere ClearSight module physical characteristics	378
Table A-20 HemoSphere Smart pressure controller physical characteristics	378
Table A-21 HemoSphere ClearSight module environmental specifications	378
Table A-18 HemoSphere technology module with ForeSight oximeter cable parameter measurement characteristics	378
Table A-22 HemoSphere ClearSight module parameter measurement specifications	379
Table A-23 Edwards finger cuff characteristics	379
Table A-24 Acumen AFM cable physical characteristics	380
Table A-25 Acumen AFM cable operational environmental specifications	380
Table A-26 Acumen AFM cable transportation environmental specifications	380
Table A-27 HemoSphere technology module with Acumen AFM cable parameter measurement specifications	380
Table B-1 HemoSphere advanced monitor components	381
Table C-1 Cardiac and oxygenation profile equations	384
Table D-1 Patient information	390
Table D-2 Graphical trend parameter scale defaults	390
Table D-3 Configurable parameter alarm and display ranges	391
Table D-4 Parameter alarm red zone and target defaults	393
Table D-5 Parameter alarms, faults, and alerts priorities	395
Table D-6 Language default settings	396
Table E-1 Computation constants for bath temperature probe	398
Table E-2 Computation constants for in-line temperature probe	398

Table G-2 Guidance and Manufacturer's Declaration - Immunity to RF wireless communications equipment	411
Table G-1 Electromagnetic emissions	411
Table G-3 Recommended Separation Distances between Portable and Mobile RF Communications Equipment and the HemoSphere advanced monitor	413
Table G-4 In band wireless coexistence – Threshold of Interference (ToI) and Threshold of Communication (ToC) between HemoSphere advanced monitor (EUT) in Invasive mode and external devices	414
Table G-5 In band wireless coexistence – Threshold of Interference (ToI) and Threshold of Communication (ToC) between HemoSphere advanced monitor (EUT) in Non-invasive mode and external devices	415
Table G-6 Electromagnetic Immunity (ESD, EFT, Surge, Dips and Magnetic Field)	416
Table G-7 Electromagnetic Immunity (RF Radiated and Conducted)	417
Table G-8 HemoSphere advanced monitor wireless information	418

Introduction

Contents

Intended Purpose of this Manual	28
Indications For Use	29
Contraindications For Use	31
Intended Use Statement	31
Expected Clinical Benefit	37
HemoSphere Advanced Monitor Hemodynamic Technology Connections	38
Manual Style Conventions	44
Abbreviations Found in This Manual	45

1.1 Intended Purpose of this Manual

This manual describes the features and monitoring options of the Edwards HemoSphere advanced monitor. The HemoSphere advanced monitor is a modular device that displays monitored data obtained through Edwards hemodynamic technologies.

This manual has been prepared for use with the Edwards HemoSphere advanced monitor by trained critical care clinicians, nurses, and physicians in any hospital environment where critical care is administered.

This manual provides the operator of the HemoSphere advanced monitor with setup and operating instructions, device interfacing procedures, and limitations.

NOTE

The following components may have alternative labeling conventions:

ForeSight oximeter cable (FSOC) may also be labeled as FORE-SIGHT ELITE tissue oximeter module (FSM).

HemoSphere technology module may also be labeled as HemoSphere tissue oximetry module.

ForeSight sensors or ForeSight Jr sensors may also be labeled as FORE-SIGHT ELITE tissue oximetry sensors.

NOTE

Not all finger cuffs are supplied with a sizing aid. Refer to the product IFU for detailed instructions on proper finger cuff sizing, if applicable.

1.2 Indications For Use

1.2.1 HemoSphere Advanced Monitor with HemoSphere Swan-Ganz Module

The HemoSphere advanced monitor when used with the HemoSphere Swan-Ganz module and Edwards Swan-Ganz catheters is indicated for use in adult and pediatric critical care patients requiring monitoring of cardiac output (continuous [CO] and intermittent [iCO]) and derived hemodynamic parameters in a hospital environment. Pulmonary artery blood temperature monitoring is used to compute continuous and intermittent CO with thermodilution technologies. It may be used for monitoring hemodynamic parameters in conjunction with a perioperative goal directed therapy protocol in a hospital environment. Refer to the Edwards Swan-Ganz catheter and Swan-Ganz Jr catheter indications for use statement for information on target patient population specific to the catheter being used.

Refer to the Intended Use statement for a complete list of measured and derived parameters available for each patient population.

1.2.2 HemoSphere Advanced Monitor with HemoSphere Oximetry Cable

The HemoSphere advanced monitor when used with the HemoSphere oximetry cable and Edwards oximetry catheters is indicated for use in adult and pediatric critical care patients requiring monitoring of venous oxygen saturation (SvO₂ and ScvO₂) and derived hemodynamic parameters in a hospital environment. Refer to the Edwards oximetry catheter indications for use statement for information on target patient population specific to the catheter being used.

Refer to the Intended Use statement for a complete list of measured and derived parameters available for each patient population.

1.2.3 HemoSphere Advanced Monitor with HemoSphere Pressure Cable

The HemoSphere advanced monitor when used with the HemoSphere pressure cable is indicated for use in adult and pediatric critical care patients in which the balance between cardiac function, fluid status, vascular resistance and pressure needs continuous assessment. It may be used for monitoring of hemodynamic parameters in conjunction with a perioperative goal directed therapy protocol in a hospital environment. Refer to the Edwards FloTrac sensor, FloTrac Jr sensor, Acumen IQ sensor, and TruWave disposable pressure transducer indications for use statements for information on target patient populations specific to the sensor/transducer being used.

The Edwards Acumen Hypotension Prediction Index software feature provides the clinician with physiological insight into a patient's likelihood of future hypotensive events and the associated hemodynamics. The Acumen HPI feature is intended for use in surgical or non-surgical patients receiving advanced hemodynamic monitoring. The Acumen HPI feature is considered to be additional quantitative information regarding the patient's physiological condition for reference only and no therapeutic decisions should be made based solely on the Acumen Hypotension Prediction Index (HPI) parameter.

Refer to the Intended Use statement for a complete list of measured and derived parameters available for each patient population.

1.2.4 HemoSphere Advanced Monitor with HemoSphere Technology Module and ForeSight Oximeter Cable

The non-invasive ForeSight oximeter cable is intended for use as an adjunct monitor of absolute regional hemoglobin oxygen saturation of blood under the sensors in individuals at risk for reduced-flow or no-flow ischemic states. The ForeSight oximeter cable is also intended to monitor relative changes of total hemoglobin of blood under the sensors. The ForeSight oximeter cable is intended to allow for the display of StO₂ and relative change in total hemoglobin on the HemoSphere advanced monitor.

- When used with large sensors, the ForeSight oximeter cable is indicated for use on adults and transitional adolescents ≥ 40 kg.
- When used with medium sensors, the ForeSight oximeter cable is indicated for use on pediatric subjects ≥ 3 kg.
- When used with small sensors, the ForeSight oximeter cable is indicated for cerebral use on pediatric subjects < 8 kg and non-cerebral use on pediatric subjects < 5 kg.

Refer to the Intended Use statement for a complete list of measured and derived parameters available for each patient population.

1.2.5 HemoSphere Advanced Monitor with HemoSphere ClearSight Module

The HemoSphere advanced monitor when used with the HemoSphere ClearSight module, pressure controller, or Smart pressure controller, and a compatible Edwards finger cuff is indicated for adult and pediatric patients in which the balance between cardiac function, fluid status and vascular resistance needs continuous assessment. It may be used for monitoring hemodynamic parameters in conjunction with a perioperative goal directed therapy protocol in a hospital environment. In addition, the noninvasive system is indicated for use in patients with co-morbidities for which hemodynamic optimization is desired and invasive measurements are difficult. The HemoSphere advanced monitor and compatible Edwards finger cuffs noninvasively measures blood pressure and associated hemodynamic parameters. Refer to the ClearSight finger cuff, ClearSight Jr finger cuff, Acumen IQ finger cuff, and Acumen IQ Plus finger cuff indications for use statements for information on target patient population specific to the finger cuff being used.

The Edwards Acumen Hypotension Prediction Index (HPI) feature provides the clinician with physiological insight into a patient's likelihood of future hypotensive events and the associated hemodynamics. The Acumen HPI feature is intended for use in surgical and non-surgical patients receiving advanced hemodynamic monitoring. The Acumen HPI feature is considered to be additional quantitative information regarding the patient's physiological condition for reference only and no therapeutic decisions should be made based solely on the Acumen Hypotension Prediction Index (HPI) parameter.

Refer to the Intended Use statement for a complete list of measured and derived parameters available for each patient population.

1.2.6 HemoSphere Advanced Monitor with Acumen Assisted Fluid Management Feature and Acumen IQ Sensor

The Acumen assisted fluid management (AFM) software feature provides the clinician with physiological insight into a patient's estimated response to fluid therapy and the associated hemodynamics. The Acumen AFM software feature is intended for use in surgical patients ≥ 18 years of age, that require advanced hemodynamic monitoring. The Acumen AFM software feature offers suggestions regarding the patient's physiological condition and estimated response to fluid therapy. Acumen AFM fluid administration

suggestions are offered to the clinician; the decision to administer a fluid bolus is made by the clinician, based upon review of the patient's hemodynamics. No therapeutic decisions should be made based solely on the assisted fluid management suggestions.

The Acumen assisted fluid management software feature may be used with the Acumen AFM cable and Acumen IQ fluid meter.

1.3 Contraindications For Use

The HemoSphere advanced monitor while used with the HemoSphere Swan-Ganz module, oximetry cable or pressure cable has no contraindications for use.

1.3.1 HemoSphere Advanced Monitor with ForeSight Oximeter Cable

The ForeSight/ForeSight Jr sensor is contraindicated for use on patients:

- with a physical site area too limited for proper sensor placement
- with allergic reactions to sensor adhesive
- undergoing an MRI scan because of associate risk of injury

1.3.2 HemoSphere Advanced Monitor with HemoSphere ClearSight Module

The HemoSphere advanced monitor while used with the HemoSphere ClearSight module and compatible finger cuff(s) is contraindicated in some patients with extreme contraction of the smooth muscle in the arteries and arterioles in the lower arm and hand, such as may be present in patients with Raynaud's disease. In these patients, blood pressure measurement can become impossible.

No other contraindications were known at the time this operator's manual was published.

1.4 Intended Use Statement

The HemoSphere advanced monitoring platform is intended to be used by qualified personnel or trained clinicians in a critical care environment in a hospital setting.

The HemoSphere advanced monitoring platform is intended for use with compatible Edwards Swan-Ganz and oximetry catheters, Swan-Ganz Jr catheters, FloTrac sensors, FloTrac Jr sensors, Acumen IQ sensors, TruWave disposable pressure transducers, ForeSight/ForeSight Jr sensors, Acumen IQ fluid meter, and ClearSight/ClearSight Jr/Acumen IQ/Acumen IQ Plus/VitaWave/VitaWave Plus finger cuffs.

A comprehensive list of parameters available while monitoring with the HemoSphere advanced monitor and a connected HemoSphere Swan-Ganz module are listed below in table 1-1. Only iCO, iCI, iSVR, and iSVRI are available to the pediatric patient population.

Table 1-1 HemoSphere Swan-Ganz module available parameters list

Abbreviation	Definition	Sub-system technology used	Patient population	Hospital environment
CO	continuous cardiac output	HemoSphere Swan-Ganz module	adult only	operating room, intensive care unit, emergency room
sCO	STAT cardiac output			
CI	continuous cardiac index			
sCI	STAT cardiac index			
CPO	cardiac power output			
CPI	cardiac power index			
EDV	right ventricular end diastolic volume			
sEDV	STAT right ventricular end diastolic volume			
EDVI	right ventricular end diastolic volume index			
sEDVI	STAT right ventricular end diastolic volume index			
HR _{avg}	averaged heart rate			
LVSWI	left ventricular stroke work index			
PVR	pulmonary vascular resistance			
PVRI	pulmonary vascular resistance index			
RVEF	right ventricular ejection fraction			
sRVEF	STAT right ventricular ejection fraction			
RVSWI	right ventricular stroke work index			
SV	stroke volume			
SVI	stroke volume index			
SVR	systemic vascular resistance			
SVRI	systemic vascular resistance index			
BT	pulmonary artery blood temperature			
iCO	intermittent cardiac output	HemoSphere Swan-Ganz module	adult and pediatric	operating room, intensive care unit, emergency room
iCI	intermittent cardiac index			
iSVR	intermittent systemic vascular resistance			
iSVRI	intermittent systemic vascular resistance index			

A comprehensive list of parameters available for adult and pediatric patient populations while monitoring with the HemoSphere advanced monitor and a connected HemoSphere oximetry cable are listed below in table 1-2.

Table 1-2 HemoSphere oximetry cable available parameters list

Abbreviation	Definition	Sub-system technology used	Patient population	Hospital environment
SvO ₂	mixed venous oxygen saturation	HemoSphere oximetry cable	adult and pediatric	operating room, intensive care unit, emergency room
ScvO ₂	central venous oxygen saturation			

A comprehensive list of parameters available for adult and pediatric patient populations while monitoring with the HemoSphere advanced monitor and both a connected HemoSphere Swan-Ganz module and oximetry cable are listed below in table 1-3.

Table 1-3 HemoSphere Swan-Ganz module with oximetry cable available parameters list

Abbreviation	Definition	Sub-system technology used	Patient population	Hospital environment
DO ₂	oxygen delivery	HemoSphere Swan-Ganz module and HemoSphere oximetry cable	adult and pediatric	operating room, intensive care unit, emergency room
DO ₂ I	oxygen delivery index			
VO ₂	oxygen consumption			
VO ₂ e	estimated oxygen consumption when ScvO ₂ is being monitored			
VO ₂ I	oxygen consumption index			
VO ₂ le	estimated oxygen consumption index when ScvO ₂ is being monitored			

A comprehensive list of parameters available while monitoring with the HemoSphere advanced monitor and both a connected HemoSphere Swan-Ganz module and pressure cable are listed below in table 1-4.

Table 1-4 HemoSphere Swan-Ganz module with pressure cable available parameters list*

Abbreviation	Definition	Sub-system technology used	Patient population	Hospital environment
CO _{20s}	20-second cardiac output	HemoSphere Swan-Ganz module and HemoSphere pressure cable	adult only	operating room, intensive care unit, emergency room
CI _{20s}	20-second cardiac index			
SV _{20s}	20-second stroke volume			
SVI _{20s}	20-second stroke volume index			
*20-second flow parameters are only available if the 20s flow parameter feature is enabled. Please contact your local Edwards representative for more information on enabling this advanced feature. For more information on these parameters, see 20-Second Flow Parameters on page 174.				

A comprehensive list of parameters available while monitoring with the HemoSphere advanced monitor and a connected HemoSphere pressure cable are listed below in table 1-5.

Table 1-5 HemoSphere pressure cable available parameters list

Abbreviation	Definition	Sub-system technology used	Patient population	Hospital environment
CO	continuous cardiac output ¹	HemoSphere pressure cable	adult and pediatric ≥ 12	operating room, intensive care unit, emergency room
CI	continuous cardiac index ¹			
CPO	cardiac power output			
CPI	cardiac power index			
DIA _{ART}	systemic arterial diastolic blood pressure			
MAP	mean arterial blood pressure			
PPV	pulse pressure variation ¹			
PR	pulse rate			
SV	stroke volume ¹			
SVI	stroke volume index ¹			
SVR	systemic vascular resistance ¹			
SVRI	systemic vascular resistance index ¹			
SVV	stroke volume variation ¹			
SYS _{ART}	systemic arterial systolic blood pressure			
CVP	central venous pressure		adult only	
DIA _{PAP}	pulmonary artery diastolic blood pressure			
dP/dt	systolic slope ²			
Ea _{dyn}	dynamic arterial elastance ²			
HPI	Acumen Hypotension Prediction Index ²			
MPAP	mean pulmonary artery blood pressure			
SYS _{PAP}	pulmonary artery systolic blood pressure			
¹ FloTrac parameters are available when using a FloTrac/FloTrac Jr/Acumen IQ sensor and if the FloTrac feature is enabled. ² HPI parameters are available when using an Acumen IQ sensor and if the HPI feature is activated. Activation is only available in certain areas. Please contact your local Edwards representative for more information on enabling this advanced feature.				

A list of Acumen Assisted Fluid Management (AFM) outputs available for surgical patients ≥ 18 years of age while monitoring with the HemoSphere Advanced Monitor and a connected HemoSphere pressure cable are listed below in table 1-6.

Table 1-6 HemoSphere pressure cable available AFM output list

AFM notification/output	Sub-system technology used	Patient population	Hospital environment
Fluid Bolus Suggested	HemoSphere pressure cable	≥18 years of age only	operating room, only
Test Bolus Suggested			
Fluid Not Suggested			
AFM Suggestions Suspended			
Bolus In Progress...			
Bolus Complete			
Bolus Complete; Analyzing Hemodynamic Response...			
Tracked Case Vol.			
Flow Rate			
Bolus Volume			
AFM outputs are available when using an Acumen IQ sensor and if the AFM feature is activated. Flow Rate and Bolus Volume are visible when using automatic fluid tracking mode.			

A comprehensive list of parameters available for adult and pediatric patient populations while monitoring with the HemoSphere advanced monitor and both a connected HemoSphere pressure cable and oximetry cable are listed below in table 1-7.

Table 1-7 HemoSphere pressure cable with oximetry cable available parameters list

Abbreviation	Definition	Sub-system technology used	Patient population	Hospital environment
DO ₂	oxygen delivery	HemoSphere pressure cable and HemoSphere oximetry cable	adult only	operating room, intensive care unit, emergency room
DO ₂ I	oxygen delivery index			
VO ₂	oxygen consumption			
VO ₂ e	estimated oxygen consumption when ScvO ₂ is being monitored			
VO ₂ I	oxygen consumption index			
VO ₂ le	estimated oxygen consumption index when ScvO ₂ is being monitored			

Tissue oxygen saturation, StO_2 , can be monitored with the HemoSphere advanced monitor, a connected HemoSphere technology module, and the ForeSight oximeter cable as listed below in table 1-8.

Table 1-8 HemoSphere technology module with ForeSight oximeter cable available parameters list

Abbreviation	Definition	Sub-system technology used	Patient population	Hospital environment
StO_2	tissue oxygen saturation	ForeSight oximeter cable and HemoSphere technology module	adult and pediatric	operating room, intensive care unit, emergency room
$\Delta ctHb$	relative change in total hemoglobin			

A comprehensive list of parameters available while monitoring with the HemoSphere advanced monitor and a connected HemoSphere ClearSight module are listed below in table 1-9.

Table 1-9 HemoSphere ClearSight module available parameters list

Abbreviation	Definition	Sub-system technology used	Patient population	Hospital environment
CO	continuous cardiac output	HemoSphere ClearSight module	adult and pedatric ≥ 12	operating room and intensive care unit
CI	continuous cardiac index			
CPO	cardiac power output			
CPI	cardiac power index			
DIA _{ART}	arterial diastolic blood pressure			
MAP	mean arterial blood pressure			
PPV	pulse pressure variation			
PR	pulse rate			
SV	stroke volume			
SVI	stroke volume index			
SVR	systemic vascular resistance			
SVRI	systemic vascular resistance index			
SVV	stroke volume variation			
SYS _{ART}	arterial systolic blood pressure			
dP/dt	systolic slope ¹			
Ea _{dyn}	dynamic arterial elastance ¹			
HPI	<u>Acumen Hypotension Prediction Index¹</u>			

¹HPI parameters are available when using an Acumen IQ finger cuff, heart reference sensor (HRS), or Acumen IQ Plus finger cuff and the Smart pressure controller, and if the HPI feature is activated. Activation is only available in certain areas. Please contact your local Edwards representative for more information on enabling this advanced feature.

Note: CO/CI and SV/SVI are measured using a reconstructed brachial arterial waveform. All other monitored parameters use a reconstructed radial arterial waveform. SVR/SVRI are derived from CO/CI and MAP along with an entered or monitored CVP value. For more information, see Waveform Reconstruction and Hemodynamic Analysis (ClearSight Technology) on page 200.

A comprehensive list of parameters available for adult patient populations while monitoring with the HemoSphere advanced monitor and both a connected HemoSphere ClearSight module and oximetry cable are listed below in table 1-10.

Table 1-10 HemoSphere ClearSight module with oximetry cable available parameters list

Abbreviation	Definition	Sub-system technology used	Patient population	Hospital environment
DO ₂	oxygen delivery	HemoSphere ClearSight module and HemoSphere oximetry cable	adult only	operating room and intensive care unit
DO ₂ I	oxygen delivery index			
VO ₂	oxygen consumption			
VO ₂ e	estimated oxygen consumption when ScvO ₂ is being monitored			
VO ₂ I	oxygen consumption index			
VO ₂ le	estimated oxygen consumption index when ScvO ₂ is being monitored			

WARNING Improper use of the HemoSphere advanced monitor could present a hazard to the patient. Carefully read the “warnings” section of this manual, located in chapter 2, before using the platform.

The HemoSphere advanced monitor is intended for use only in patient assessment. This instrument must be used in conjunction with a bedside physiological monitor and/or patient clinical signs and symptoms. If hemodynamic values obtained from the device are not consistent with the clinical presentation of the patient, consider troubleshooting before initiating treatment options.

ECG signal input and all parameters derived from heart rate measurements have not been evaluated for pediatric patients and are therefore not available for that patient population.

1.5 Expected Clinical Benefit

The HemoSphere advanced monitoring platform allows you to see and interact with patient hemodynamic parameters. In conjunction with the compatible sensors and predictive decision support software, the modular HemoSphere platform facilitates proactive clinical decision-making and insight for individualized patient care.

1.6 HemoSphere Advanced Monitor Hemodynamic Technology Connections

The HemoSphere advanced monitor is equipped with three technology expansion module slots (two standard size and one large [L-Tech] size) and two cable ports. Module and cable connection points are located on the left side panel. See figure 1-1.





Figure 1-1 HemoSphere advanced monitor hemodynamic technology connections

Each module/cable is associated with a specific Edwards hemodynamic monitoring technology. Currently available modules include the HemoSphere Swan-Ganz module, introduced below and in detail in chapter 9, *HemoSphere Swan-Ganz Module Monitoring*, and the HemoSphere technology module introduced below and described in detail in chapter 13, *HemoSphere Tissue Oximetry Monitoring* and chapter 14, *Assisted Fluid Management*. Currently available large technology (L-Tech) modules include the HemoSphere ClearSight module, introduced below and in detail in chapter 11, *HemoSphere ClearSight Module Non-Invasive Monitoring*. Currently available cables include the HemoSphere pressure cable, introduced below and described in detail in chapter 10, *Monitoring with the HemoSphere Pressure Cable*, and the HemoSphere oximetry cable, introduced below and described in detail in chapter 12, *Venous Oximetry Monitoring*.

1.6.1 HemoSphere Swan-Ganz Module

The HemoSphere Swan-Ganz module enables continuous cardiac output (CO) and intermittent cardiac output (iCO) monitoring with an Edwards patient CCO cable and compatible Swan-Ganz/Swan-Ganz Jr catheter. Right ventricular end diastolic volume (EDV) monitoring is available with analog




input heart rate (HR_{avg}) data from a bedside patient monitor. The HemoSphere Swan-Ganz module fits into a standard module slot. For more information, see chapter 9, *HemoSphere Swan-Ganz Module Monitoring*. Table 1-11 lists the parameters available while using the HemoSphere Swan-Ganz module.

Table 1-11 HemoSphere Swan-Ganz module parameters description

Parameter	Description	Technology
continuous cardiac output (CO)	continuous assessment through advanced thermodilution technology of the volume of blood pumped by the heart measured in liters per minute	Swan-Ganz CCO and CCombo catheters
continuous cardiac index (CI)	continuous cardiac output relative to body surface area (BSA)	Swan-Ganz CCO and CCombo catheters
intermittent cardiac output (iCO)	intermittent assessment through the bolus thermodilution method of the volume of blood pumped by the heart measured in liters per minute	Swan-Ganz and Swan-Ganz Jr thermodilution catheters
intermittent cardiac index (iCI)	intermittent cardiac output relative to body surface area (BSA)	Swan-Ganz and Swan-Ganz Jr thermodilution catheters
right ventricular ejection fraction (RVEF)	continuous assessment through advanced thermodilution technology and algorithm analysis of the percentage of blood volume ejected from the right ventricle during systole	Swan-Ganz CCombo V catheters with ECG signal input
right ventricular end diastolic volume (EDV)	continuous assessment of the volume of blood in the right ventricle at the end of diastole calculated by dividing stroke volume (mL/beat) by RVEF(%)	Swan-Ganz CCombo V catheters with ECG signal input
stroke volume (SV)	amount of blood ejected from the ventricles with each contraction derived from CO assessment and heart rate ($SV = CO/HR \times 1000$)	Swan-Ganz CCO, CCombo, and CCombo V catheters with ECG signal input
stroke volume index (SVI)	stroke volume relative to body surface area (BSA)	Swan-Ganz CCO, CCombo, and CCombo V catheters with ECG signal input
systemic vascular resistance (SVR)	a derived measure of impedance to blood flow from left ventricle (afterload)	Swan-Ganz CCO and CCombo catheters with MAP and CVP analog pressure signal input
systemic vascular resistance index (SVRI)	systemic vascular resistance relative to body surface area (BSA)	Swan-Ganz CCO and CCombo catheters with MAP and CVP analog pressure signal input

1.6.2 HemoSphere Pressure Cable

The HemoSphere pressure cable enables vascular pressure monitoring with a compatible Edwards pressure transducer/sensor and catheter. A connected FloTrac, FloTrac Jr, or Acumen IQ sensor provides continuous cardiac output (CO) and associated hemodynamic parameters. A connected



TruWave transducer provides location based intravascular pressure. The HemoSphere pressure cable plugs into a monitoring cable port. For more information, see chapter 10, *Monitoring with the HemoSphere Pressure Cable*. Table 1-12 lists the parameters available while using the HemoSphere pressure cable.

Table 1-12 HemoSphere pressure cable key parameters description

Parameter	Description	Technology
continuous cardiac output (CO)	continuous assessment of the volume of blood pumped by the heart measured in liters per minute using the existing arterial pressure waveform and FloTrac system algorithm	FloTrac, FloTrac Jr, or Acumen IQ sensor
continuous cardiac index (CI)	continuous cardiac output relative to body surface area (BSA)	FloTrac, FloTrac Jr, or Acumen IQ sensor
central venous pressure (CVP)	central venous blood pressure	TruWave pressure transducer at central venous catheter line
diastolic blood pressure (DIA _{ART} /DIA _{PAP})	diastolic blood pressure measured at the pulmonary artery (PAP) or a systemic artery (ART)	FloTrac sensor, FloTrac Jr sensor, Acumen IQ sensor, or TruWave pressure transducer
systolic slope (dP/dt)*	maximum upslope of the arterial pressure waveform measured from a peripheral artery*	Acumen IQ sensor
dynamic arterial elastance (Ea _{dyn})*	measure of afterload to the left ventricle by the arterial system (arterial elastance) relative to the left ventricular elastance*	Acumen IQ sensor
Acumen Hypotension Prediction Index (HPI)*	index representing the likelihood that the patient may be trending toward a hypotensive event (MAP < 65 mmHg for at least one minute in duration)*	Acumen IQ sensor
mean arterial pressure (MAP)	averaged systemic blood pressure over one cardiac cycle	FloTrac sensor, FloTrac Jr sensor, Acumen IQ sensor, or TruWave pressure transducer
mean pulmonary artery pressure (MPAP)	averaged pulmonary artery blood pressure over one cardiac cycle	TruWave pressure transducer at pulmonary artery catheter line
pulse pressure variation (PPV)	the percent difference between PPmin and PPmax relative to PPmean where PP = SYS - DIA	FloTrac, FloTrac Jr, or Acumen IQ sensor
pulse rate (PR)	number of arterial blood pressure pulses per minute	FloTrac sensor, FloTrac Jr sensor, Acumen IQ sensor, or TruWave pressure transducer
stroke volume (SV)	volume of blood pumped with each heart beat	FloTrac, FloTrac Jr, or Acumen IQ sensor
stroke volume index (SVI)	stroke volume relative to body surface area (BSA)	FloTrac, FloTrac Jr, or Acumen IQ sensor
systemic vascular resistance (SVR)	a derived measure of impedance to blood flow from left ventricle (afterload)	FloTrac, FloTrac Jr, or Acumen IQ sensor
systemic vascular resistance index (SVRI)	systemic vascular resistance relative to body surface area (BSA)	FloTrac, FloTrac Jr, or Acumen IQ sensor
stroke volume variation (SVV)	the percent difference between SVmin and SVmax relative to SVmean	FloTrac, FloTrac Jr, or Acumen IQ sensor


Table 1-12 HemoSphere pressure cable key parameters description (continued)

Parameter	Description	Technology
systolic pressure (SYS _{ART} /SYS _{PAP})	systolic blood pressure measured at the pulmonary artery (PAP) or a systemic artery (ART)	FloTrac sensor, FloTrac Jr, Acumen IQ sensor, or TruWave pressure transducer
<i>*HPI parameters are available when using an Acumen IQ sensor and if the HPI feature is activated. Activation is only available in certain areas. Please contact your local Edwards representative for more information on enabling this advanced feature.</i>		

NOTE Cardiac output calculated with the HemoSphere pressure cable may differ from that calculated with the HemoSphere Swan-Ganz module due to methodological and algorithmic differences.

1.6.3 HemoSphere Oximetry Cable


The HemoSphere oximetry cable enables mixed venous oxygen saturation (SvO₂) or central venous oxygen saturation (ScvO₂) monitoring with a compatible Edwards oximetry catheter. The HemoSphere oximetry cable plugs into a monitoring cable port and can be used in combination with other hemodynamic monitoring technologies. For more information on oximetry monitoring, see chapter 12, *Venous Oximetry Monitoring*. Table 1-13 lists the parameters available while using the HemoSphere oximetry cable.

**Table 1-13 HemoSphere oximetry cable parameters description**

Parameter	Description
central venous oximetry (ScvO ₂)	venous oxygen saturation as measured in the superior vena cava
mixed venous oximetry (SvO ₂)	venous oxygen saturation as measured in the pulmonary artery
oxygen consumption (VO ₂)	the amount of oxygen used by the body per minute
estimated oxygen consumption (VO _{2e})	an estimate of the amount of oxygen used by the body per minute (ScvO ₂ monitoring only)
oxygen consumption index (VO _{2I})	the amount of oxygen used by the body per minute indexed against body surface area (BSA)
estimated oxygen consumption index (VO _{2Ie})	an estimate of the amount of oxygen used by the body per minute indexed against body surface area (BSA)

1.6.4 HemoSphere Technology Module

The HemoSphere technology module fits into a standard module slot. This module connects with the ForeSight oximeter cable for tissue oximetry (StO₂) and the AFM cable for fluid bolus tracking with the AFM software feature.



1.6.4.1 HemoSphere Technology Module and ForeSight Oximeter Cable

The HemoSphere technology module enables tissue oximetry (StO₂) monitoring with a ForeSight oximeter cable and compatible tissue oximetry sensors. For more information on tissue oximetry monitoring, see chapter 13, *HemoSphere Tissue Oximetry Monitoring*. Table 1-14 lists the parameters available while using the HemoSphere technology module and ForeSight oximeter cable.

NOTE

The following components may have alternative labeling conventions:

ForeSight oximeter cable (FSOC) may also be labeled as FORE-SIGHT ELITE tissue oximeter module (FSM).

HemoSphere technology module may also be labeled as HemoSphere tissue oximetry module.

ForeSight sensors or ForeSight Jr sensors may also be labeled as FORE-SIGHT ELITE tissue oximetry sensors.

Table 1-14 HemoSphere technology module with ForeSight oximeter cable parameters description

Parameter	Description	Technology
tissue oximetry (StO ₂)	absolute tissue oxygen saturation as measured at anatomical surface below sensor location	ForeSight/ForeSight Jr sensor detection of near-infrared light reflection
relative change in total hemoglobin (ΔctHb)	trending value calculated from the sum of relative changes in oxygenated hemoglobin and deoxygenated hemoglobin (ΔO ₂ Hb and ΔHHb)	ForeSight/ForeSight Jr sensor detection of near-infrared light reflection

1.6.4.2 HemoSphere Technology Module and AFM Cable

The HemoSphere technology module enables bolus delivery flow rate tracking in the AFM software feature with a connected AFM cable and compatible fluid meter. For more information on the AFM software feature, which is advanced feature, see *Assisted Fluid Management* on page 299.

1.6.5 HemoSphere ClearSight Module

The HemoSphere ClearSight module with a connected compatible pressure controller and finger cuff(s) enables non-invasive measurement of a patient's arterial pressure waveform and calculation of continuous cardiac output (CO) and associated hemodynamic parameters. The HemoSphere ClearSight module fits into the large technology (L-Tech) slot. For more information, see chapter 11, *HemoSphere ClearSight Module Non-Invasive Monitoring*.




Table 1-15 HemoSphere ClearSight module key parameters description

Parameter	Description	Technology
continuous cardiac output (CO)	continuous assessment of the volume of blood pumped by the heart measured in liters per minute using the monitored arterial pressure waveform and ClearSight algorithm	ClearSight, ClearSight Jr, Acumen IQ, or Acumen IQ Plus cuff
continuous cardiac index (CI)	continuous cardiac output relative to body surface area (BSA)	ClearSight, ClearSight Jr, Acumen IQ, or Acumen IQ Plus cuff
diastolic blood pressure (DIA _{ART})	diastolic blood pressure	ClearSight, ClearSight Jr, VitaWave ¹ , VitaWave Plus ¹ , Acumen IQ, or Acumen IQ Plus cuff
systolic slope (dP/dt) ²	maximum upslope of the arterial pressure waveform measured from a peripheral artery*	Acumen IQ or Acumen IQ Plus cuff
dynamic elastance (Ea _{dyn}) ²	measure of afterload to the left ventricle by the arterial system (arterial elastance) relative to the left ventricular elastance*	Acumen IQ or Acumen IQ Plus cuff
Acumen Hypotension Prediction Index (HPI) ²	index representing the likelihood that the patient may be trending toward a hypotensive event (MAP < 65 mmHg for at least one minute in duration)*	Acumen IQ or Acumen IQ Plus cuff
mean arterial pressure (MAP)	averaged systemic blood pressure over one cardiac cycle	ClearSight, ClearSight Jr, VitaWave ¹ , VitaWave Plus ¹ , Acumen IQ, or Acumen IQ Plus cuff
pulse pressure variation (PPV)	the percent difference between PP _{min} and PP _{max} relative to PP _{mean} where PP = SYS - DIA	ClearSight, ClearSight Jr, Acumen IQ, or Acumen IQ Plus cuff
pulse rate (PR)	number of arterial blood pressure pulses per minute	ClearSight, ClearSight Jr, VitaWave ¹ , VitaWave Plus ¹ , Acumen IQ, or Acumen IQ Plus cuff
stroke volume (SV)	volume of blood pumped with each heart beat	ClearSight, ClearSight Jr, Acumen IQ, or Acumen IQ Plus cuff
stroke volume index (SVI)	stroke volume relative to body surface area (BSA)	ClearSight, ClearSight Jr, Acumen IQ, or Acumen IQ Plus cuff
systemic vascular resistance (SVR)	a derived measure of impedance to blood flow from left ventricle (afterload)	ClearSight, ClearSight Jr, Acumen IQ, or Acumen IQ Plus cuff
systemic vascular resistance index (SVRI)	systemic vascular resistance relative to body surface area (BSA)	ClearSight, ClearSight Jr, Acumen IQ, or Acumen IQ Plus cuff
stroke volume variation (SVV)	the percent difference between SV _{min} and SV _{max} relative to SV _{mean}	ClearSight, ClearSight Jr, Acumen IQ, or Acumen IQ Plus cuff
systolic pressure (SYS _{ART})	systolic blood pressure	ClearSight, ClearSight Jr, VitaWave ¹ , VitaWave Plus ¹ , Acumen IQ, or Acumen IQ Plus cuff

¹Monitoring with a VitaWave cuff or VitaWave Plus cuff is an advanced feature. Activation is only available in certain areas. Please contact your local Edwards representative for more information on enabling this advanced feature.

²HPI parameters are available when using an Acumen IQ finger cuff, heart reference sensor or Acumen IQ Plus and the Smart pressure controller, and if the HPI feature is activated. Activation is only available in certain areas. Please contact your local Edwards representative for more information on enabling this advanced feature.

1.6.6 Documentation and Training

Available documentation and training for the HemoSphere advanced monitor includes:


- HemoSphere Advanced Monitor Operator's Manual
- HemoSphere Advanced Monitor Quick Start Guide
- HemoSphere Pressure-Out Cable Instructions for Use
- Edwards Heart Reference Sensor Instructions for Use
- Edwards Pressure Controller Instructions for Use
- HemoSphere Battery Instructions for Use
- HemoSphere Roll Stand Instructions for Use
- HemoSphere Oximetry Cradle Instructions for Use
- Acumen IQ Fluid Meter Instructions for Use

Instructions for Use are included with HemoSphere Advanced Monitor components. See table B-1, "HemoSphere advanced monitor components," on page 381. For more information on how you can receive training or available documentation for the HemoSphere advanced monitor, contact your local Edwards representative or Edwards Technical Support. See appendix F, *System Care, Service and Support*.

1.7 Manual Style Conventions

Table 1-16 lists the style conventions used in this manual.

Table 1-16 Operator's manual style conventions

Convention	Description
Bold	Bold text indicates a software term. This word or phrase will appear on the screen as shown.
Bold button	A button is a touch screen access point for the option appearing in bold. For example, the Review button appears on-screen as: 
→	An arrow is shown between two on-screen menu options that are selected consecutively by the operator.
	An icon is a touch screen access point for the menu or navigation graphic shown. See table 2-1 on page 63 for full list of menu icons shown on the HemoSphere advanced monitor.
Venous Oximetry Calibration icon 	Bold text with a menu icon indicates an icon that is paired with a software term or phrase appearing on the screen. For example, the Venous Oximetry Calibration icon appears on-screen as: 

1.8 Abbreviations Found in This Manual

Table 1-17 Acronyms, Abbreviations

Abbreviation	Definition
A/D	analog/digital
AFM	Assisted Fluid Management
ART	systemic arterial blood pressure
BMI	body mass index
BSA	body surface area
BT	blood temperature
CaO ₂	arterial oxygen content
CI	cardiac index
CI _{20s}	20-second cardiac index
CO	cardiac output
CO _{20s}	20-second cardiac output
CCO	continuous cardiac output (used when describing certain Swan-Ganz catheters and patient CCO cable)
CPI	cardiac power index
CPO	cardiac power output
CVP	central venous pressure
ΔctHb	relative change in total hemoglobin
DIA _{ART}	systemic arterial diastolic blood pressure
DIA _{PAP}	pulmonary artery diastolic blood pressure
DO ₂	oxygen delivery
DO ₂ I	oxygen delivery index
dP/dt	systolic slope (maximum upslope of the arterial pressure waveform)
DPT	disposable pressure transducer
Ea _{dyn}	dynamic arterial elastance
EDV	end diastolic volume
EDVI	end diastolic volume index
ESV	end systolic volume
ESVI	end systolic volume index
efu	ejection fraction unit
FRT	Fluid Responsiveness Test
FT-CO	FloTrac arterial pressure auto calibrated cardiac output
GDT	goal directed therapy
Hct	hematocrit
HEMPC	pressure controller
HIS	hospital information systems
HGB	hemoglobin
HPI	Acumen Hypotension Prediction Index
HR	heart rate
HR _{avg}	average heart rate
HRS	heart reference sensor

Table 1-17 Acronyms, Abbreviations (continued)

Abbreviation	Definition
IA	Intervention Analysis
iCI	intermittent cardiac index
iCO	intermittent cardiac output
IEC	International Electrotechnical Commission
IT	injectate temperature
LED	light emitting diode
LVSWI	left ventricular stroke work index
MAP	mean arterial pressure
MPAP	mean pulmonary artery pressure
NIBP	non-invasive blood pressure
OR	operating room
PA	pulmonary artery
PAP	pulmonary artery blood pressure
PaO ₂	partial pressure of arterial oxygen
PAWP	pulmonary artery wedge pressure
PC2	pressure controller
PPV	pulse pressure variation
PR	pulse rate
POST	power-on self test
PvO ₂	partial pressure of venous oxygen
PVR	pulmonary vascular resistance
PVRI	pulmonary vascular resistance index
RV	right ventricular
RVEF	right ventricular ejection fraction
RVSWI	right ventricular stroke work index
SaO ₂	oxygen saturation
sCI	STAT cardiac index
sCO	STAT cardiac output
ScvO ₂	central venous oximetry
sEDV	STAT end diastolic volume
sEDVI	STAT end diastolic volume index
SQI	signal quality indicator
sRVEF	STAT right ventricular ejection fraction
ST	surface temperature
STAT	fast estimate of parameter value
StO ₂	tissue oxygen saturation
SV	stroke volume
SV _{20s}	20-second stroke volume
SVI	stroke volume index
SVI _{20s}	20-second stroke volume index
SvO ₂	mixed venous oxygen saturation
SVR	systemic vascular resistance

Table 1-17 Acronyms, Abbreviations (continued)

Abbreviation	Definition
SVRI	systemic vascular resistance index
SVV	stroke volume variation
SYS _{ART}	systemic arterial systolic blood pressure
SYS _{PAP}	pulmonary artery systolic blood pressure
Touch	Interact with the HemoSphere advanced monitor by touching the screen.
TD	thermodilution
USB	Universal Serial Bus
VO ₂	oxygen consumption
VO ₂ I	oxygen consumption index
VO ₂ e	estimation of oxygen consumption
VO ₂ le	estimated oxygen consumption index

Safety and Symbols

Contents

Safety Signal Words Definitions	47
Warnings	48
Cautions	55
User Interface Symbols	62
Symbols on Product Labels	65
Applicable Standards	68
HemoSphere Advanced Monitor Essential Performance	68

2.1 Safety Signal Words Definitions

2.1.1 Warning

A warning advises against certain actions or situations that could result in personal injury or death.

WARNING This is how warnings appear throughout the text of this manual.

2.1.2 Caution

A caution advises against actions or situations that could damage equipment, produce inaccurate data, or invalidate a procedure.

CAUTION This is how cautions appear throughout the text of this manual.

2.1.3 Note

A note draws attention to useful information regarding a function or procedure.

NOTE This is how notes appear throughout the text of this manual.

2.2 Warnings

The following are warnings that are used in the HemoSphere advanced monitor operator's manual. They are introduced in the manual where relevant to the function or procedure being described.

-
- Read this operator's manual carefully before attempting to use the Edwards HemoSphere advanced monitor.
 - Refer to the instructions for use provided with each compatible accessory before using it with the HemoSphere advanced monitor.
 - To prevent injury to patient or user, damage to platform, or inaccurate measurements, do not use any damaged or non-compatible platform accessories, components or cables.
 - Improper use of the HemoSphere advanced monitor could present a hazard to the patient. Carefully read the "warnings" section of this manual, located in chapter 2, before using the platform. (chapter 1)
 - The HemoSphere advanced monitor is intended for use only in patient assessment. This instrument must be used in conjunction with a bedside physiological monitor and/or patient clinical signs and symptoms. If hemodynamic values obtained from the device are not consistent with the clinical presentation of the patient, consider troubleshooting before initiating treatment options. (chapter 1)
 - ECG signal input and all parameters derived from heart rate measurements have not been evaluated for pediatric patients and are therefore not available for that patient population. (chapter 1)
 - Shock hazard! Do not attempt to connect/disconnect system cables while hands are wet. Ensure that hands are dry prior to disconnecting system cables. (chapter 3)
 - Explosion Hazard! Do not use the HemoSphere advanced monitor in the presence of flammable anesthetic mixture with air or with oxygen or nitrous oxide. (chapter 3)
 - This product contains metallic components. Do NOT use in a Magnetic Resonance (MR) environment. (chapter 3)
 - Make sure the HemoSphere advanced monitor is securely positioned or mounted and that all cords and accessory cables are appropriately arranged to minimize the risk of injury to patients, users or the equipment. (chapter 3)
 - Use of this equipment adjacent to or stacked with other equipment should be avoided because it could result in improper operation. If such use is necessary, this equipment and the other equipment should be observed to verify that they are operating normally. (chapter 3)
 - The HemoSphere advanced monitor must be positioned in an upright position to ensure IPX1 ingress protection. (chapter 3)
 - Do not allow any liquids to splash onto the monitoring screen. Liquid buildup may disable the touchscreen functionality. (chapter 3)
 - Do not position the monitor so that it is difficult to access rear panel ports or power cord. (chapter 3)
-

-
- Equipment is rated for use with high-frequency surgical equipment. Inaccurate parameter measurements can be caused by interference from high-frequency surgical equipment. To reduce hazards that can arise from the use of high-frequency surgical equipment, only use undamaged patient cables and accessories connected as specified in this operator's manual. (chapter 3)
 - This system is rated for use with defibrillators. To ensure proper defibrillator-proof operation, only use undamaged patient cables and accessories connected as specified in this operators manual. (chapter 3)
 - All IEC/EN 60950 equipment, including printers, to be positioned no closer than 1.5 meters to the patient's bed. (chapter 3)
 - Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the HemoSphere advanced monitor, including cables specified by the manufacturer. Otherwise, degradation of the performance this equipment could result. (chapter 3)
 - Make sure the battery is fully inserted and the battery door is properly latched. Falling batteries could seriously injure patients or clinicians. (chapter 3)
 - Only use Edwards approved batteries with the HemoSphere advanced monitor. Do not charge the battery pack outside of the monitor. Doing so can damage the battery or injure the user. (chapter 3)
 - To prevent any interruptions to monitoring during power loss, it is recommended to use the HemoSphere advanced monitor with the battery inserted. (chapter 3)
 - In cases of power failure and battery depletion, the monitor will go through a controlled shut off procedure. (chapter 3)
 - Do not use the HemoSphere advanced monitoring platform without an installed power cord entry cover. Failure to do so may result in fluid ingress. (chapter 3)
 - Do not use extension cords or multiple socket devices to connect the power cord. Do not use detachable power cords other than the power cord provided. (chapter 3)
 - To avoid the risk of electric shock, the HemoSphere advanced monitor can only be connected to a supply mains with grounding (protective earth). Do not use three prong to two prong power adaptors. (chapter 3)
 - Grounding reliability can only be achieved when the instrument is connected to a receptacle marked "hospital only", "hospital grade", or its equivalent. (chapter 3)
 - Disconnect the monitor from the AC source by unplugging mains power cable from the AC Mains. The On/Off button on the monitor does not disconnect the system from the AC mains supply. (chapter 3)
 - Only use HemoSphere advanced monitor accessories, cables and or components that have been supplied and labeled by Edwards. Using other unlabeled accessories, cables and or components may affect patient safety and measurement accuracy. (chapter 3)
 - Upon initiation of a new patient session, the default high/low physiological alarm ranges should be checked to ensure that they are appropriate for the given patient. (chapter 6)
 - Perform New Patient or clear the patient data profile whenever a new patient is connected to the HemoSphere advanced monitor. Failure to do so may result in previous patient data in the historical displays. (chapter 6)
-

-
- The analog communication ports of the HemoSphere advanced monitor share a common ground that is isolated from the catheter interface electronics. When connecting multiple devices to the HemoSphere advanced monitor, all devices should be provided with isolated power to avoid compromising the electrical isolation of any of the connected devices. (chapter 6)
 - Risk and Leakage current of the final system configuration must comply with IEC 60601-1:2005/A1:2012. It is the responsibility of the user to ensure compliance. (chapter 6)
 - Accessory equipment connected to the monitor must be certified according to IEC/EN 60950 for data-processing equipment or IEC 60601-1:2005/A1:2012 for electromedical equipment. All combinations of equipment must be in compliance with IEC 60601-1:2005/A1:2012 systems requirements. (chapter 6)
 - When switching to a different bedside monitor, always check that the default values listed are still valid. If necessary, reconfigure the voltage range and corresponding parameter range or calibrate. (chapter 6)
 - Do not turn off the audible alarms in situations in which patient safety could be compromised. (chapter 7)
 - Do not lower the alarm volume to a level that prohibits alarms from being adequately monitored. Failure to do so could result in a situation where patient safety is compromised. (chapter 7)
 - Visual and audible physiological alarms are activated only if the parameter is configured on the screens as a key parameter (1-8 parameters displayed in parameter tiles). If a parameter is not selected and displayed as a key parameter, the audible and visual physiological alarms are not triggered for that parameter. (chapter 7)
 - Make sure that Demo Mode is not activated in a clinical setting to ensure that simulated data is not mistaken for clinical data. (chapter 7)
 - Do not use the HemoSphere advanced monitor as part of a Distributed Alarm System. The HemoSphere advanced monitor does not support remote alarm monitoring/management systems. Data is logged and transmitted for charting purposes only. (chapter 8)
 - Compliance to IEC 60601-1 is only maintained when the HemoSphere Swan-Ganz module (applied part connection, defibrillation proof) is connected to a compatible monitoring platform. Connecting external equipment or configuring the system in a way not described in these instructions will not meet this standard. Failure to use the device as instructed may increase the risk of electrical shock to the patient/operator. (chapter 9)
 - Do not modify, service or alter the product in any way. Servicing, alteration or modification may affect patient/operator safety and/or product performance. (chapter 9)
 - CO monitoring should always be discontinued when blood flow around the thermal filament is stopped. Clinical situations where CO monitoring should be discontinued include, but are not limited to: • Time periods when a patient is on cardiopulmonary bypass • Partial withdrawal of the catheter so that the thermistor is not in the pulmonary artery • Removal of the catheter from the patient (chapter 9)
 - PACEMAKER PATIENTS – Rate meters may continue to count the pacemaker rate during occurrences of cardiac arrest or some arrhythmias. Do not rely entirely upon displayed heart rate. Keep pacemaker patients under close surveillance. See table A-5 on page 371 for disclosure of the pacemaker pulse rejection capability of this instrument. (chapter 9)
-

-
- For patients requiring internal or external pacing support, the HemoSphere advanced monitoring platform should not be used to obtain heart rate and heart rate derived parameters under the following conditions:
 - pacer pulse synch output from bedside monitor includes the pacer pulse, however, the characteristics are outside of the pacemaker pulse rejection capabilities specifications as listed in table A-5
 - pacer pulse synch output characteristics from bedside monitor cannot be determined (chapter 9)
 - Note any discrepancies in heart rate (HRavg) with the patient monitor HR and ECG waveform display when interpreting derived parameters such as SV, EDV, RVEF, and associated index parameters. (chapter 9)
 - Do not resterilize or reuse any FloTrac sensor, FloTrac Jr sensor, Acumen IQ sensor, TruWave transducer, or catheter; refer to the catheter's "directions for use". (chapter 10)
 - Do not use a FloTrac sensor, FloTrac Jr sensor, Acumen IQ sensor, TruWave transducer, or catheter that is wet, damaged, or that has exposed electrical contacts. (chapter 10)
 - Refer to the directions provided with each accessory for specific instructions on placement and use, and for relevant WARNINGS, CAUTIONS, and specifications. (chapter 10)
 - When the pressure cable is not in use, protect the exposed cable connector from fluid. Moisture within the connector may result in the cable malfunctioning or in inaccurate pressure readings. (chapter 10)
 - Compliance to IEC 60601-1 is only maintained when the HemoSphere pressure cable (applied part accessory, defibrillation proof) is connected to a compatible monitoring platform. Connecting external equipment or configuring the system in a way not described in these instructions will not meet this standard. Failure to use the device as instructed may increase the risk of electrical shock to the patient/operator. (chapter 10)
 - Do not use the HemoSphere advanced monitoring platform as a pulse rate or blood pressure monitor. (chapter 10)
 - ClearSight technology use not recommended for patients age < 12 years of age. (chapter 11)
 - Components that are not indicated as APPLIED PARTS should not be placed in a location where the patient may come into contact with the component. (chapter 11)
 - Compliance to IEC 60601-1 is only maintained when the HemoSphere ClearSight module (applied part connection) is connected to a compatible monitoring platform. Connecting external equipment or configuring the system in a way not described in these instructions will not meet this standard. Failure to use the device as instructed may increase the risk of electrical shock to the patient/operator. (chapter 11)
 - Do not modify, service or alter the product in any way. Servicing, alteration or modification may affect patient/operator safety and/or product performance. (chapter 11)
 - Do not sterilize any components of the HemoSphere non-invasive system. The HemoSphere non-invasive system is provided non sterile. (chapter 11)
 - Refer to cleaning instructions. Do not disinfect the instrument by autoclave or gas sterilization. (chapter 11)
 - Refer to the directions provided with each accessory for specific instructions on placement and use, and for relevant WARNINGS, CAUTIONS, and specifications. (chapter 11)
-

-
- Do not use damaged components/sensors or components/sensors with exposed electrical contacts to prevent patient or user shocks. (chapter 11)
 - The HemoSphere non-invasive system monitoring components are not defibrillation proof. Disconnect the system before defibrillating. (chapter 11)
 - Only use compatible Edwards finger cuffs, heart reference sensor and other HemoSphere non-invasive system accessories, cables and or components that have been supplied and labeled by Edwards. Using other unlabeled accessories, cables and or components may affect patient safety and measurement accuracy. (chapter 11)
 - Always remove HemoSphere non-invasive system sensors and components from the patient and completely disconnect the patient from the instrument before bathing the patient. (chapter 11)
 - Do not overtighten the pressure controller band or finger cuff(s). (chapter 11)
 - Do not apply pressure controller band on injured skin as this can cause further injury. (chapter 11)
 - Improper finger cuff placement or sizing can lead to inaccurate monitoring. (chapter 11)
 - Do not use the HemoSphere non-invasive system as a heart rate monitor. (chapter 11)
 - If using the instrument during full body irradiation, keep all HemoSphere non-invasive system monitoring components out of the irradiation field. If a monitoring component is exposed to the irradiation, the readings may be affected. (chapter 11)
 - Strong magnetic fields may cause malfunction of the instrument and burn wounds to the patient. Do not use the instrument during magnetic resonance imaging (MRI) scanning. Induced current could potentially cause burns. The device may affect the MR image, and the MRI unit may affect the accuracy of the measurements. (chapter 11)
 - Compliance to IEC 60601-1 is only maintained when the HemoSphere oximetry cable (applied part accessory, defibrillation proof) is connected to a compatible monitoring platform. Connecting external equipment or configuring the system in a way not described in these instructions will not meet this standard. Failure to use the device as instructed may increase the risk of electrical shock to the patient/operator. (chapter 12)
 - Do not wrap the main body of the oximetry cable in fabric or place directly on the patient's skin. The surface does get warm (up to 45 °C) and needs to dissipate heat to maintain its internal temperature level. A software fault will trigger if the internal temperature exceeds its limits. (chapter 12)
 - Before touching Yes to recall oximetry data, confirm that the displayed data matches the current patient. Recalling incorrect oximetry calibration data and patient demographics will result in inaccurate measurements. (chapter 12)
 - Compliance to IEC 60601-1 is only maintained when the HemoSphere technology module (applied part connection, defibrillation proof) is connected to a compatible monitoring platform. Connecting external equipment or configuring the system in a way not described in these instructions will not meet this standard. Failure to use the device as instructed may increase the risk of electrical shock to the patient/operator. (chapter 13)
-

-
- Inspect all of the ForeSight oximeter cable connections for damage prior to installation. If any damage is noted, the cable must not be used until it has been serviced or replaced. Contact Edwards Technical support. There is a risk that damaged parts could reduce the performance of the cable or present a safety hazard. (chapter 13)
 - To remove any chance of contamination between patients, the ForeSight oximeter cable and cable connections should be cleaned after each case. (chapter 13)
 - To reduce the risk of contamination and cross infection, if the ForeSight oximeter cable or cable connections are grossly contaminated, with blood or other bodily fluids, it should be disinfected. If the ForeSight oximeter cable or cable connections cannot be disinfected, it should be serviced, replaced or discarded. Contact Edwards Technical support. (chapter 13)
 - To reduce the risk of damaging internal elements of the cable assemblies – within the ForeSight oximeter cable housing – avoid excessive pulling, bending or other types of stress on the cable connections. (chapter 13)
 - Do not modify, service or alter the product in any way. Servicing, alteration or modification may affect patient/operator safety and/or product performance. (chapter 13)
 - Sensors are not sterile and therefore should not be applied on abraded, cracked, or lacerated skin. Exercise caution when applying sensors to a site with delicate skin. Applying sensors, tape or pressure to such a site may reduce circulation and/or cause skin deterioration. (chapter 13)
 - Do not place sensor over poorly perfused tissues. Avoid uneven skin surfaces for best adhesion. Do not place sensor over sites with ascites, cellulitis, pneumoencephalus, or edema. (chapter 13)
 - If electrocautery procedures will be performed, sensors and electrocautery electrodes should be placed as far apart as possible to prevent unwanted skin burns; a distance of at least 15 cm (6 in) is recommended. (chapter 13)
 - Use only Edwards supplied accessories with the ForeSight oximeter cable. Edwards accessories ensure patient safety and preserve the integrity, accuracy, and electromagnetic compatibility of the ForeSight oximeter cable. Connecting a non-Edwards sensor will cause an appropriate alert on that channel and no StO₂ values will be recorded. (chapter 13)
 - Sensors are designed for single-patient use, and are not to be reprocessed – re-used sensors present a risk of cross-contamination or infection. (chapter 13)
 - Use a new sensor for each patient and discard it after use. Disposal should follow in accordance with local hospital and institution policies. (chapter 13)
 - If a sensor seems damaged in any way, it must not be used. (chapter 13)
 - Always read the sensor packaging. (chapter 13)
 - Exercise extreme care when applying sensors. Sensor circuits are conductive and must not come into contact with other grounded, conductive parts other than EEG or entropy monitors. Such contact would bridge the patient's isolation and cancel the protection provided by the sensor. (chapter 13)
 - Failure to apply sensors properly may cause incorrect measurements. Misapplied sensors or sensors that become partially dislodged may cause either over- or under-reading of oxygen saturation. (chapter 13)
-

-
- Do not position a sensor under the weight of the patient. Prolonged periods of pressure (such as taping over the sensor or the patient lying on a sensor) transfers weight from the sensor to the skin, which can injure skin and reduce sensor performance. (chapter 13)
 - The sensor site must be inspected at least every 12 hours to reduce the risk of inadequate adhesion, circulation, and skin integrity. If the circulatory condition or skin integrity has deteriorated, the sensor should be applied to a different site. (chapter 13)
 - Do not connect more than one patient to the ForeSight oximeter cable. This may compromise the patient's isolation and cancel the protection provided by the sensor. (chapter 13)
 - The ForeSight oximeter cable has been designed to promote patient safety. All cable parts are "Type BF Defibrillation Proof" and are protected against the effects of the defibrillator discharge and may remain attached to the patient. Cable readings may be inaccurate during defibrillator use and up to twenty (20) seconds thereafter. (chapter 13)
 - No separate actions are required when using this equipment with a defibrillator, but only Edwards-supplied sensors must be used for proper protection against the effects of a cardiac defibrillator. (chapter 13)
 - Do not come into contact with patients during defibrillation, or serious injury or death could result. (chapter 13)
 - If the accuracy of any value displayed on the monitor is questionable, determine the patient's vital signs by alternative means. The functions of the alarm system for patient monitoring must be verified at regular intervals and whenever the integrity of the product is in doubt. (chapter 13)
 - Testing of the ForeSight oximeter cable operation should be done at least once every 6 months, as described in HemoSphere service manual. Failure to comply may lead to injury. If the cable fails to respond, it must not be used until it has been inspected and serviced or replaced. See technical support contact information on inside cover. (chapter 13)
 - The Acumen Hypotension Prediction Index, HPI, should not be used exclusively to treat patients. A review of the patient's hemodynamics is recommended prior to initiating treatment. (chapter 14)
 - The Assisted Fluid Management feature should not be used exclusively to treat the patient. A review of the patient's hemodynamics is recommended throughout the monitoring session to assess fluid responsiveness. (chapter 14)
 - Only use approved HemoSphere advanced monitor accessories, cables and or components that have been supplied and labeled by Edwards. Using unapproved accessories, cables and or components may affect patient safety and measurement accuracy. (appendix B)
 - The HemoSphere advanced monitor contains no user-serviceable parts. Removing the cover or any other disassembly will expose you to hazardous voltages. (appendix F)
 - Shock or fire hazard! Do not immerse the HemoSphere advanced monitor, modules, or platform cables in any liquid solution. Do not allow any fluids to enter the instrument. (appendix F)
 - Do not, under any circumstances, perform any cleaning or maintenance of the ForeSight oximeter cable while the cable is being used to monitor a patient. The cable must be turned off and the HemoSphere advanced monitor power cord disconnected, or the cable must be disconnected from the monitor and the sensors removed from the patient. (appendix F)
-

-
- Before starting cleaning or maintenance of any sort, check the ForeSight oximeter cable, cable connections, ForeSight sensors, and other accessories for damage. Check the cables for bent or broken prongs, cracks, or fraying. If any damage is noted, the cable must not be used until it has been inspected and serviced or replaced. Contact Edwards Technical Support. (appendix F)
 - There is a risk of serious injury or death if this procedure is not followed. (appendix F)
 - Explosion Hazard! Do not open battery, dispose of in fire, store at high temperature or short circuit. It may ignite, explode, leak or get hot, causing serious personal injury or death. (appendix F)
 - Use of accessories, transducers, and cables other than those specified or provided by the manufacturer of this equipment could result in increased electromagnetic emissions or decreased electromagnetic immunity of this equipment and result in improper operation. (appendix G)
 - No modification of the HemoSphere advanced monitor is allowed. (appendix G)
 - Portable and mobile RF communication equipment and other sources of electromagnetic disturbance such as diathermy, lithotripsy, RFID, electromagnetic anti-theft systems and metal detectors can potentially affect all electronic medical equipment, including the HemoSphere advanced monitor. Guidance on maintaining appropriate separation between communications equipment and the HemoSphere advanced monitor is provided in table G-3. The effects of other RF emitters are unknown and may interfere with the function and safety of the HemoSphere monitoring platform. (appendix G)
-

2.3 Cautions

The following are cautions that are used in the HemoSphere advanced monitor operator's manual. They are introduced in the manual where relevant to the function or procedure being described.

-
- Federal (USA) law restricts this device to sale by or on the order of a physician.
 - Inspect the HemoSphere advanced monitor and all accessories and equipment used with the monitor for damage prior to use. Damage may include cracks, scratches, dents, exposed electrical contacts, or any signs that the housing may be compromised.
 - Always grasp the connector, not the cable, when connecting or disconnecting cables. Do not twist or bend the connectors. Confirm that all sensors and cables are connected correctly and completely before use. (chapter 3)
 - To avoid corruption of data on the HemoSphere advanced monitor, always disconnect the patient CCO cable and oximetry cable from the monitor before using a defibrillator. (chapter 3)
 - Do not expose the HemoSphere advanced monitor to extreme temperatures. Refer to environmental specifications in appendix A. (chapter 3)
 - Do not expose the HemoSphere advanced monitor to dirty or dusty environments. (chapter 3)
 - Do not obstruct the HemoSphere advanced monitor ventilation openings. (chapter 3)
 - Do not use the HemoSphere advanced monitor in environments where strong lighting makes the LCD screen difficult to view. (chapter 3)
 - Do not use the monitor as a handheld device. (chapter 3)
-

-
- When moving the instrument, be sure to turn off the power and remove the connected power cord. (chapter 3)
 - Make sure that the HRS is correctly applied so that it can be leveled to the phlebostatic axis. (chapter 4)
 - When connecting the HemoSphere advanced monitor to external devices, refer to the external device's instruction manual for complete instructions. Verify proper operation of the system before clinical use. (chapter 6)
 - Only properly trained personnel should calibrate the HemoSphere advanced monitor analog ports. (chapter 6)
 - The accuracy of continuous SVR while monitoring with the HemoSphere Swan-Ganz module depends upon the quality and accuracy of the MAP and CVP data transmitted from the external monitors. Since MAP and CVP analog signal quality from the external monitor cannot be validated by the HemoSphere advanced monitor, actual values and the values (including all derived parameters) displayed by the HemoSphere advanced monitor may not be consistent. The accuracy of continuous SVR measurement, therefore, cannot be guaranteed. To aid in determining the quality of the analog signals, regularly compare the MAP and CVP values displayed on the external monitor to the values displayed on the physio relationship screen of the HemoSphere advanced monitor. Refer to the external input device operator's manual for detailed information regarding accuracy, calibration, and other variables which may impact the analog output signal from the external monitor. (chapter 6)
 - Use a virus scan on any USB stick before inserting to prevent a virus or malware infection. (chapter 8)
 - Do not force the module into the slot. Apply even pressure to slide and click the module into place. (chapter 9)
 - Inaccurate cardiac output measurements may be caused by:
 - Incorrect placement or position of the catheter
 - Excessive variations in pulmonary artery blood temperature. Some examples that cause BT variations include, but are not limited to:
 - * status post cardiopulmonary bypass surgery
 - * centrally administered cooled or warmed solutions of blood products
 - * use of sequential compression devices
 - Clot formation on the thermistor
 - Anatomical abnormalities (for example, cardiac shunts)
 - Excessive patient movement
 - Electrocautery or electrosurgical unit interference
 - Rapid changes in cardiac output (chapter 9)
 - Inaccurate 20-second flow parameter measurements may be caused by:
 - Incorrect placement or position of the catheter
 - Improperly zeroed and/or leveled transducer
 - Over- or under-damped pressure line
 - Adjustments to the PAP line made after start of monitoring (chapter 9)
 - Refer to Appendix E to ensure computation constant is the same as specified in the catheter package insert. If the computation constant differs, enter the desired computation constant manually. (chapter 9)
 - Sudden changes in PA blood temperature, such as those caused by patient movement or bolus drug administration, may cause an iCO or iCI value to be computed. To avoid falsely triggered curves, inject as soon as possible after the Inject message appears. (chapter 9)
-

-
- Do not use any FloTrac sensor, FloTrac Jr sensor, Acumen IQ sensor, or TruWave transducer past its labeled “Use By Date.” Products used beyond this date may have compromised transducer or tubing performance, or compromised sterility. (chapter 10)
 - Excessive dropping of the HemoSphere pressure cable may result in cable damage and/ or malfunction. (chapter 10)
 - Inaccurate FT-CO measurements can be caused by factors such as: • Improperly zeroed and/or leveled sensor/transducer • Over- or under-damped pressure lines • Excessive variations in blood pressure. Some conditions that cause BP variations include, but are not limited to: * Intra-aortic balloon pumps • Any clinical situation where the arterial pressure is deemed inaccurate or not representative of aortic pressure, including but not limited to: * Extreme peripheral vasoconstriction which results in a compromised radial arterial pressure waveform * Hyperdynamic conditions as seen in post liver transplant • Excessive patient movement • Electrocautery or electrosurgical unit interference Aortic valve regurgitation may cause an over estimation of Stroke Volume / Cardiac Output calculated depending on the amount of valvular disease and the volume lost back into the left ventricle. (chapter 10)
 - Always grasp the connector, not the cable, when connecting or disconnecting the cable. (chapter 10)
 - Do not twist or bend the connectors. (chapter 10)
 - To prevent cable damage, do not apply excessive force to the pressure cable zero button. (chapter 10)
 - Consider the change in performance of the HemoSphere ClearSight module when using a software version of V01.01.000 or later, which displays and analyzes a reconstructed radial arterial waveform. Software versions earlier than V01.01.000 reconstruct brachial arterial pressure from finger arterial pressure. Clinicians should consider this change in waveform reconstruction, especially if they are experienced with viewing the brachial arterial pressure waveform reconstructed in earlier software versions of the HemoSphere ClearSight module. (chapter 11)
 - Do not force the module into the slot. Apply even pressure to slide and click the module into place. (chapter 11)
 - The effectiveness of HemoSphere non-invasive system has not been evaluated in patients under 12 years of age. (chapter 11)
 - Always grasp the connector, not the cable, when connecting or disconnecting cables. Do not twist or bend the connectors. Confirm that all sensors and cables are connected correctly and completely before use. (chapter 11)
 - Do not wrap Smart pressure controller cable. (chapter 11)
 - Do not attach the Smart pressure controller to patient or patient skin. (chapter 11)
 - Make sure that the HRS is correctly applied so that it can be leveled to the phlebostatic axis. (chapter 11)
 - The HemoSphere non-invasive system is not intended for use as an apnea monitor. (chapter 11)
 - In patients with extreme contraction of the smooth muscle in the arteries and arterioles in the lower arm and hand, such as may be present in patients with Raynaud’s disease, blood pressure measurement can become impossible. (chapter 11)
-

-
- Inaccurate non-invasive measurements can be caused by factors such as:
 - Improperly calibrated and/or leveled HRS
 - Excessive variations in blood pressure. Some conditions that cause BP variations include, but are not limited to:
 - * Intra-aortic balloon pumps
 - Any clinical situation where the arterial pressure is deemed inaccurate or not representative of aortic pressure.
 - Poor blood circulation to the fingers.
 - A bent or flattened finger cuff.
 - Excessive patient movement of fingers or hands.
 - Artifacts and poor signal quality.
 - Incorrect placement of finger cuff, position of finger cuff, or finger cuff too loose.
 - Electrocautery or electrosurgical unit interference. (chapter 11)
 - Always disconnect the finger cuff when it is not wrapped around a finger, to prevent damage by accidental over-inflation. (chapter 11)
 - The effectiveness of Edwards compatible finger cuffs has not been established in pre-eclamptic patients. (chapter 11)
 - The pulsations from intra-aortic balloon support can be additive to the pulse rate on the instrument pulse rate display. Verify patient's pulse rate against the ECG heart rate. (chapter 11)
 - The pulse rate measurement is based on the optical detection of a peripheral flow pulse and therefore may not detect certain arrhythmias. The pulse rate should not be used as a replacement or substitute for ECG based arrhythmia analysis. (chapter 11)
 - Monitoring without an HRS may lead to measurement inaccuracies. Ensure patient remains still with accurately measured finger to heart height difference. (chapter 11)
 - Do not place the patient in a non-supine position while monitoring without an HRS. This may lead to an inaccurate vertical offset entry for the HRS and measurement inaccuracies. (chapter 11)
 - Do not perform a BP calibration during monitoring periods when blood pressure appears unstable. This may result in inaccurate blood pressure measurements. (chapter 11)
 - Make sure that the oximetry cable is securely stabilized to prevent unnecessary movement of the attached catheter. (chapter 12)
 - The catheter tip or calibration cup must not get wet before an in vitro calibration is performed. The catheter and the calibration cup must be dry for an accurate oximetry in vitro calibration. Flush the catheter lumen only after the in vitro calibration has been completed. (chapter 12)
 - Performing an in vitro calibration after the oximetry catheter has been inserted into the patient will yield an inaccurate calibration. (chapter 12)
 - The SQI signal is sometimes affected by the use of electrosurgical units. Attempt to distance electrocautery equipment and cables from the HemoSphere advanced monitor and plug the power cords into separate AC circuits if possible. If signal quality problems persist, call your local Edwards representative for assistance. (chapter 12)
 - Do not disconnect the oximetry cable while calibration or data recall are in process. (chapter 12)
 - If the oximetry cable is being transferred from a HemoSphere advanced monitor to another HemoSphere advanced monitor, check that the patient height, weight, and BSA are correct prior to beginning monitoring. Re-enter patient data, if necessary. (chapter 12)
 - Avoid placing the ForeSight oximeter cable where the status LED cannot be easily seen. (chapter 13)
 - Applying too much pressure may break the retaining tab, which may present a risk of the cable falling on the patient, bystander, or operator. (chapter 13)
-

-
- Do not lift or pull the ForeSight oximeter cable by any cable connections, or place the cable in any position that might present a risk that the cable may fall on the patient, bystander or operator. (chapter 13)
 - Avoid placing the ForeSight oximeter cable under sheets, or blanket that could restrict air flow around the cable that may increase the cable's case temperature and present an injury. (chapter 13)
 - Do not force the module into the slot. Apply even pressure to slide and click the module into place. (chapter 13)
 - Sensors should not be placed on high density hair areas. (chapter 13)
 - The sensor must be able to rest flush with clean, dry skin. Any debris, lotion, oil, powder, perspiration, or hair that prevents good contact between the sensor and the skin will affect the validity of the data collected and may result in an alarm message. (chapter 13)
 - When used in settings with LED lighting, sensors may need to be covered with a light blocker prior to connection to the sensor cable, as some high intensity systems can interfere with the sensor's near infrared light detection. (chapter 13)
 - Do not lift or pull the ForeSight oximeter cable by any cable connections, or place the ForeSight oximeter cable in any position that might present a risk that the module may fall on the patient, bystander or operator. (chapter 13)
 - Once patient monitoring has started, do not replace the sensor or disconnect the sensor for more than 10 minutes to avoid restarting the initial StO₂ calculation. (chapter 13)
 - Measurements may be affected in the presence of strong electromagnetic sources such as electrosurgery equipment, and measurements may be inaccurate during use of such equipment. (chapter 13)
 - Elevated levels of carboxyhemoglobin (COHb) or methemoglobin (MetHb) may lead to inaccurate or erroneous measurements, as may intravascular dyes or any substance containing dyes that change usual blood pigmentation. Other factors that may affect measurement accuracy include: myoglobin, hemoglobinopathies, anemia, pooled blood under the skin, interference from foreign objects in the Sensor path, Bilirubinemia, externally applied coloring (tattoos), high levels of HGB or Hct and birthmarks. (chapter 13)
 - When used in settings with LED lighting, sensors may need to be covered with a light blocker prior to connection to the sensor cable, as some high intensity systems can interfere with the sensor's near infrared light detection. (chapter 13)
 - When compared to earlier software versions, a ForeSight oximeter cable with a software version of V3.0.7 or later and used with pediatric sensors (small and medium) is more responsive in the display StO₂ values. Specifically, in the range below 60%, StO₂ measurements could be reported lower than in earlier software versions. Clinicians should consider the faster response and potentially modified StO₂ values when using V3.0.7 software, especially if they are experienced with earlier software versions of the ForeSight oximeter cable. (chapter 13)
 - The effectiveness of the HPI parameter during minimally-invasive monitoring has been established using radial arterial pressure waveform data. The effectiveness of the HPI parameter using arterial pressure from other sites (e.g., femoral) has not been evaluated. (chapter 14)
-

-
- The HPI parameter may not provide advanced notice of a trend towards a hypotensive event in situations where a clinical intervention results in a sudden non-physiological hypotensive event. If this occurs, the HPI feature will provide the following without delay: a high alert popup, a high priority alarm, and an HPI value of 100 will be displayed indicating that the patient is undergoing a hypotensive event. (chapter 14)
 - Exercise caution when using the absolute values of dp/dt . Pressure will change distally due to narrowing of vessels and frictional forces within the vessels. While absolute dp/dt may not be an accurate measure of cardiac contractility, trends may be helpful. (chapter 14)
 - Exercise caution when using dp/dt in patients with severe aortic stenosis, since the stenosis may reduce the coupling between the left ventricle and the afterload. (chapter 14)
 - The dp/dt parameter, although predominantly determined by changes in LV contractility, may be impacted by afterload during periods of vasoplegic states (venoarterial decoupling). During these periods, dp/dt may not reflect changes in LV contractility. (chapter 14)
 - The HPI parameter information provided in table 14-16 and table 14-17 is presented as general guidance and may not be representative of individual experience. A review of the patient's hemodynamics is recommended prior to initiating treatment. (chapter 14)
 - The HPI parameter information provided in table 14-26 and table 14-27 is presented as general guidance and may not be representative of individual experience. A review of the patient's hemodynamics is recommended prior to initiating treatment. (chapter 14)
 - The Assisted Fluid Management software feature relies on information provided by the clinician to accurately assess fluid responsiveness. (chapter 14)
 - Fluid management suggestions provided by the AFM feature can be compromised by factors such as:
 - Inaccurate FT-CO measurements
 - Acute changes in FT-CO measurements secondary to vasoactive medication administration, patient repositioning or surgical interventions
 - Bleeding at rates similar to, or greater than, the rate of fluid delivery
 - Arterial line interference
 Always review patient hemodynamic status before complying with AFM suggestions. (chapter 14)
 - Accurate stroke volume variation (SVV) measurement is necessary for the AFM software feature to make fluid management suggestions. Patients must be:
 - mechanically ventilated
 - have a tidal volume of ≥ 8 mL/kg
 (chapter 14)
 - The presence of confounding factors during bolus delivery may lead to an incorrect fluid recommendation by the AFM software. Therefore, boluses delivered in the presence of confounding factors should be discarded. Potential confounding factors include but are not limited to:
 - Vasoactive agent was administered during bolus administration
 - Additional fluid given after primary bolus administered
 - Subject repositioning
 - Ventilatory changes
 - Surgical manipulation
 - Arterial line interference
 - * External compression (i.e., leaning on A-line)
 - * ABG draw, fast flush
 - * Overdamping of line
 - Vascular clamping
 - Additional line of fluid simultaneously opened during bolus administration
 - Known acute hemorrhage during fluid administration
 - Inaccurate FT-CO measurements
 (chapter 14)
 - Use of any fluids not listed in the specified Fluid Type list or choosing the incorrect fluid type may result in measurement inaccuracies. (chapter 14)
 - Do not rely on the lack of a suspected finding to rule out follow-up. (chapter 14)
-

-
- United States Federal law restricts the sale and use of this device to either by or on the order of a clinician. (chapter 14)
 - This is NOT a diagnosis. The screening results offered by Edwards' ASI algorithm are meant only to provide clinical decision support to the clinician, who may use the result in conjunction with their own evaluation and clinical judgment. The recommendations are not diagnoses. (chapter 14)
 - The quality of the ASI algorithm recommendation depends upon the quality of the inputted blood pressure waveform. Situations which may affect the inputted blood pressure causing invalid waveforms may include, but are not limited to, severe vasoconstriction, patient movement, or Raynaud's disease. (chapter 14)
 - If any of the ForeSight oximeter cable LEDs fail to turn on, the cable must not be used until it has been serviced or replaced. Contact Edwards Technical Support. There is a risk that damaged parts could reduce the performance of the cable. (chapter 15)
 - Do not pinch any heart reference sensor tubes or wires under the pressure controller cover during application. Be careful the only wire between the back mounting notch is the pressure controller cable. (appendix B)
 - Do not lift PCCVR from any other point than the front tab. (appendix B)
 - Clean and store the instrument and accessories after each use. (appendix F)
 - The HemoSphere advanced monitor modules and platform cables are electrostatic discharge (ESD) sensitive. Do not attempt to open cable or module housing or use if the housing has been damaged. (appendix F)
 - Do not pour or spray liquid on any portion of the HemoSphere advanced monitor, accessories, modules, or cables. (appendix F)
 - Do not use any disinfecting solution other than the types specified. (appendix F)
 - DO NOT: • Allow any liquid to come in contact with the power connector • Allow any liquid to penetrate connectors or openings in the monitor case or modules If any liquid does come in contact with any of the above mentioned items, DO NOT attempt to operate the monitor. Disconnect power immediately and call your Biomedical Department or local Edwards representative. (appendix F)
 - Conduct periodic inspections of all cables for defects. Do not coil cables tightly when storing. (appendix F)
 - Do not use any other cleaning agents, spray, or pour cleaning solution directly on platform cables. Do not steam, radiate, or EO sterilize platform cables. Do not immerse platform cables. (appendix F)
 - Do not steam, radiate, or EO sterilize the HemoSphere oximetry cable. Do not immerse the HemoSphere oximetry cable. (appendix F)
 - If any electrolytic solution, for example Ringer's lactate solution, is introduced into the cable connectors while they are connected to the monitor, and the monitor is turned on, the excitation voltage can cause electrolytic corrosion and rapid degradation of the electrical contacts. (appendix F)
 - Do not immerse any cable connectors in detergent, isopropyl alcohol or glutaraldehyde. (appendix F)
-

-
- Do not use a hot air gun to dry cable connectors. (appendix F)
 - Device contains electronics. Handle with care. (appendix F)
 - Do not disinfect the heart reference sensor or pressure controller by autoclave or gas sterilization. (appendix F)
 - Do not immerse the pressure controller, heart reference sensor, or any cable connectors in fluid. (appendix F)
 - Clean and store the heart reference sensor after each use. (appendix F)
 - Do not disinfect the Smart pressure controller by autoclave or gas sterilization. (chapter F)
 - Do not immerse the Smart pressure controller, or any cable connectors in fluid. (chapter F)
 - Do not use any other cleaning agents aside from those listed. (chapter F)
 - Recycle or dispose of the lithium-ion battery in accordance to all federal, state, and local laws. (appendix F)
 - The instrument has been tested and complies with the limits of IEC 60601-1-2. These limits are designed to provide reasonable protection against harmful interference in a typical medical installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to other devices in the vicinity. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to other devices which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:
 - Reorient or relocate the receiving device.
 - Increase the separation between the equipment.
 - Consult the manufacturer for help.(appendix G)
 - The wireless Quality of Service (QoS) may be influenced by the presence of other devices that create radio frequency interference (RFI). Such RFI devices may include electrocautery equipment, cellular telephones, wireless PC and tablets, pagers, RFID, MRI, or other electrically powered devices. When used in the presence of potential RFI devices, consideration should be taken to maximize separation distances and to observe for any potential signs of interference such as loss of communication or reduced Wi-Fi signal strength. (appendix G)
-

2.4 User Interface Symbols

The following are icons that appear on the HemoSphere advanced monitor screen. For more information about screen appearance and navigation, see chapter 5, *Navigating the HemoSphere Advanced Monitor*. Certain icons will only appear while monitoring with a specific hemodynamic technology module or cable, as specified.

Table 2-1 Monitor display symbols

Symbol	Description
Navigation Bar Icons	
	Select Monitoring Mode
	Begin CO monitoring (HemoSphere Swan-Ganz module)
	Stop CO monitoring with CO countdown timer (see <i>CO Countdown Timer</i> on page 173) (HemoSphere Swan-Ganz module)
	Start non-invasive monitoring (HemoSphere ClearSight module)
	Stop non-invasive monitoring (HemoSphere ClearSight module)
	Resume non-invasive monitoring after cuff pressure release (HemoSphere ClearSight module)
	Zero & Waveform
	GDT Tracking
	GDT Tracking/AFM (AFM software feature enabled and Acumen IQ sensor connected)
	Settings menu
	Home (return to main monitoring screen)
	Display pressure waveform
	Hide pressure waveform
	Silence audible alarms
	Alarms paused (silenced) with countdown timer (See <i>Silence Audible Alarms</i> on page 101)
	Resume monitoring with elapsed time from monitoring pause

Table 2-1 Monitor display symbols (continued)

Symbol	Description
	Patient Data (demographics have been entered)
	Patient Data (demographics have been skipped)
Clinical Tools Menu Icons	
	Select Monitoring Mode
	iCO (intermittent cardiac output) (HemoSphere Swan-Ganz module)
	Venous Oximetry Calibration (HemoSphere oximetry cable)
	Enter CVP
	Derived Value Calculator
	Event Review
	Zero & Waveform
	Patient CCO Cable Test (HemoSphere Swan-Ganz module)
	HPI Secondary Screen (advanced feature)
	Fluid Responsiveness Test (advanced feature)
	BP Calibration (ClearSight BP) (HemoSphere ClearSight module)
	HRS Calibration
	ctHb Tools
	Patient Data

Table 2-1 Monitor display symbols (continued)

Symbol	Description
Menu Navigation Icons	
	Return to main monitoring screen
	Return to previous menu
	Cancel
	Scroll to select item on vertical list
	Vertical page scroll
	Horizontal scroll
	Enter
	Keypad enter key
	Keypad backspace key
	Move cursor left by 1 character
	Move cursor right by 1 character
	Keypad cancel key
	Item enabled
	Item not enabled
	Clock/waveform - allows user to view historical data or intermittent data
Parameter Tile Icons	
	Alarms / Targets menu: parameter audible alarm indicator enabled
	Alarms / Targets menu: parameter audible alarm indicator disabled

Table 2-1 Monitor display symbols (continued)

Symbol	Description
	Signal quality indicator bar See <i>Signal Quality Indicator</i> on page 228 (HemoSphere oximetry cable) See <i>SQL</i> on page 216 (HemoSphere ClearSight module)
	AFM dashboard shortcut (SV only)
	SVV Filtering Exceeded Indicator: High degree of pulse rate variability may be impacting SVV values
	Venous Oximetry Calibration (HemoSphere oximetry cable)
	CVP value manually entered (SVR/SVRI only)
	Default CVP value used (SVR/SVRI only)
	ΔctHb value (StO ₂ only) (advanced feature)
Information Bar Icons	
	Viewfinder hub connectivity status icon on information bar See table 8-3 on page 164
	HIS enabled icon on information bar See table 8-2 on page 162
	Snapshot (screen capture)
	Battery life indicator icons on information bar See table 5-6 on page 131
	Screen Brightness
	Alarm Volume
	Lock Screen
	Help menu shortcut
	Event Review
	Beat-to-beat heart rate (HemoSphere Swan-Ganz module with ECG input)

Table 2-1 Monitor display symbols (continued)

Symbol	Description
	Wi-Fi signal See table 8-1 on page 161
	Time until cuff pressure release mode (HemoSphere ClearSight module, see <i>Cuff Pressure Release Mode</i> on page 219)
	Time until conclusion of cuff pressure release mode (HemoSphere ClearSight module, see <i>Cuff Pressure Release Mode</i> on page 219)
Intervention Analysis Icons	
	Intervention analysis button
	Intervention analysis type indicator for custom event (gray)
	Intervention analysis type indicator for positional challenge (purple)
	Intervention analysis type indicator for a fluid challenge (blue)
	Intervention analysis type indicator for intervention (green)
	Intervention analysis type indicator for oximetry (red)
	Intervention analysis type indicator for event (yellow)
	Edit icon on intervention information balloon
	Keyboard icon for entering notes on intervention edit screen
AFM Icons	
	Assisted fluid management (AFM) icon on the navigation bar
	AFM fluid status icons on AFM dashboard. For more information, see table 14-41 on page 305.

Table 2-1 Monitor display symbols (continued)

Symbol	Description
	Start or re-start Assisted Fluid Management (AFM) session
	Pause Assisted Fluid Management (AFM) session
	Decline bolus suggestion
	User initiated bolus (Manual mode only)
	Stop bolus (Manual mode only)
	AFM settings
	Minimize AFM dashboard
	GDT target settings
	AFM context help
	End Assisted Fluid Management (AFM) session
GDT Tracking Icons	
	Add Target button on GDT Tracking Screen
	Target Value button on GDT Tracking Screen
	Exit Target Selection button on GDT Tracking Screen
	Edit Target button on GDT Tracking Screen
	Time-In-Target symbol on GDT Tracking Screen
HPI Icons	
	HPI Secondary Screen shortcut key

2.5 Symbols on Product Labels

This section provides the symbols that are on the HemoSphere advanced monitor and other available HemoSphere advanced monitoring platform accessories.

Table 2-2 Symbols on product labels

Symbol	Description
	Manufacturer
	Date of manufacture
Rx only	Caution: Federal (USA) law restricts this device to sale by, or on the order of a physician.
IPX1	Provides protection against vertically falling water to IPX1 standard
IPX4	Extent of protection against ingress of objects
	Separate collection for electrical and electronic equipment in accordance with EC directive 2012/19/EU.
	Restriction of Hazardous Substances (RoHS) compliance - China only
FCC	Federal Communications Commission (FCC) compliance - USA only
	This device contains a non-ionizing radiation transmitter, which can cause RF interference with other devices near this device.
eifu.edwards.com + 1 888 570 4016	Follow instructions for use on the website
	Instructions for use in electronic form is available by phone or website address.
	Intertek ETL
#	Model number
SN	Serial number
EC REP	Authorized representative in the European Community
	MR unsafe
CE 0123	Conformité Européenne (CE Mark) of TÜV SÜD Product Service GmbH (notified body)
CE	Conformité Européenne (CE Mark)

Table 2-2 Symbols on product labels (continued)

Symbol	Description
LOT	Batch code
PN	Part number
QTY	Quantity
	Lead-free
cULus	Underwriters Laboratories product certification mark
	Recyclable Lithium-Ion
	Technical conformity mark (Japan)
	Do not disassemble
	Do not incinerate
MD	Medical device
UDI	Unique device identifier
	Importer
Connector Identification Labels	
	Equipotential terminal stud
	USB 2.0
SS	USB 3.0
	Ethernet connection
	Analog input 1
	Analog input 2

Table 2-2 Symbols on product labels (continued)

Symbol	Description
	Pressure (DPT) output
	Defibrillation proof type CF applied part or connection
	Defibrillation proof type BF applied part or connection
	Type BF applied part or connection
	Continuous non-invasive arterial blood pressure
	Remove the pressure controller cover from this end
	Do not remove pressure controller cover from this end
	ECG input from external monitor
	High-Definition Multimedia Interface output
	Connector: serial COM output (RS232)
Additional Packaging Labels	
	Keep dry
	Fragile, handle with care

Table 2-2 Symbols on product labels (continued)

Symbol	Description
	This end up
	Do not use if package is damaged and consult instructions for use
	Box made from recyclable cardboard
	Keep away from sunlight
	Temperature limit (X = lower limit, Y = upper limit)
	Humidity limitation (X = lower limit, Y = upper limit)
	Follow instructions for use
	Store in a cool, dry place
	Use-by date
	Environment-friendly use period (EFUP) - China only

NOTE For all accessory product labels, refer to symbol table contained in accessory instructions for use.

2.6 Applicable Standards

Table 2-3 Applicable standards

Standard	Title
IEC 60601-1:2005 / A1:2012	Medical electrical equipment — Part 1: General requirements for basic safety and essential performance + amendment 1 (2012)
IEC 60601-1-2: 2014	Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral standard: Electromagnetic compatibility - Requirements and tests
IEC 60601-2-34: 2011	Medical electrical equipment — Part 2-34: Particular requirements for the basic safety and essential performance of invasive blood pressure monitoring equipment
IEC 60601-2-49:2011/ IEC 80601-2-49:2018	Medical electrical equipment — Part 2-49: Particular requirements for the basic safety and essential performance of multifunction patient monitoring equipment/monitors
IEEE 802.11 b/g/n	Telecommunications and information exchange between systems Local and metropolitan area networks — Specific requirements Part 11: Wireless LAN Medium Access Control (MAC) and Physical Layer (PHY) Specifications

2.7 HemoSphere Advanced Monitor Essential Performance

The platform shall provide display of continuous CO and intermittent CO with a compatible Swan-Ganz catheter according to the specifications provided in appendix A. The platform shall provide display of intravascular blood pressure with a compatible FloTrac, FloTrac Jr, or Acumen IQ sensor or compatible TruWave disposable pressure transducer (DPT) according to the specifications provided in appendix A. The platform shall provide display of SvO₂/ScvO₂ with a compatible oximetry catheter according to the specifications provided in appendix A. The platform shall provide non-invasive measurement of arterial blood pressure with a compatible Edwards finger cuff according to the specifications provided in appendix A. The platform shall provide display of StO₂ with a compatible oximetry module and sensor according to the specifications provided in appendix A. The platform shall provide alarm, alert, indicator, and/or system status when unable to provide accurate measurement of the applicable hemodynamic parameter. For more information, see *Essential Performance Characteristics* on page 368.

Device performance, including functional characteristics, have been verified in a comprehensive series of testing to support the safety and performance of the device for its intended use when used in accordance with the established Instructions For Use.

Installation and Setup

Contents

Unpacking	69
HemoSphere Advanced Monitor Connection Ports	72
HemoSphere Advanced Monitor Installation	75
Initial Start Up	79
Power Off and Power Save Mode	80

3.1 Unpacking

Examine the shipping container for any signs of damage that may have occurred during transit. If any damage is detected, photograph the package and contact Edwards technical support for assistance. Do not use if the package or contents are damaged. Perform a visual inspection of the packaging contents for damage. Damage may include cracks, scratches, dents or any signs that the monitor, modules or cable housing may be compromised. Report any evidence of external damage.

3.1.1 Packaging Contents

The HemoSphere advanced monitoring platform is modular and therefore packaging configurations will vary depending upon the kit ordered. The HemoSphere advanced monitoring system, which is the base kit configuration, contains the HemoSphere advanced monitor, mains power cord, power entry cover, HemoSphere battery pack, two expansion modules, one L-Tech expansion module, a quick start guide and a USB stick containing this operator's manual. See table 3-1. Additional items that may be included and shipped with other kit configurations include the HemoSphere Swan-Ganz module, patient CCO cable, and HemoSphere oximetry cable. Disposable and accessory items may be delivered separately. It is recommended that the user confirm the receipt of all ordered equipment. Refer to appendix B: *Accessories*, for a full list of available accessories.

Table 3-1 HemoSphere advanced monitoring components

HemoSphere advanced monitoring system (base kit)
• HemoSphere advanced monitor • HemoSphere battery pack • mains power cord • power entry cover • L-Tech expansion module • expansion module (2) • quick start guide • operator's manual (on USB stick)

3.1.2 Required Accessories for Platform Modules and Cables

The following tables identify accessories required to display specific monitored and calculated parameters for the specified hemodynamic technology module or cable.

Table 3-2 Cables and catheters required for monitoring parameters with HemoSphere Swan-Ganz module

	Monitored and calculated parameters							
Required cable/catheter	CO	CO _{20s}	EDV	RVEF	SVR	iCO	SV	SV _{20s}
patient CCO cable	•	•	•	•	•	•	•	•
ECG cable or pressure cable monitoring ART**			•	•			•	
analog pressure input cable(s)					•			
injectate temperature probe						•		
Swan-Ganz thermodilution catheter and Swan-Ganz Jr thermodilution catheter						•		
Swan-Ganz CCO catheter or Swan-Ganz CCOMbo catheter	•				•	•	•	
Swan-Ganz CCOMbo V catheter	•	•	•	•	•	•	•	•
TruWave transducer*		•						•
*20-second flow parameters are only available while monitoring with a CCOMbo V catheter (models 777F8 and 774F75) and require a pulmonary artery pressure signal through a HemoSphere pressure cable connection. See 20-Second Flow Parameters on page 157. **Pulse rate (PR) from pressure cable monitoring arterial waveform can be used in place of heart rate (HR) from ECG cable.								

NOTE Not all parameters can be monitored or calculated in pediatric patients. See table 1-1 on page 32 for available parameters.

Table 3-3 Sensor options for monitoring parameters with HemoSphere pressure cable

	Monitored and calculated parameters								
Pressure sensor/transducer options	CO	SV	SVV/PPV	SVR	PR	SYS/DIA/MAP	MPAP	CVP	HPI/dP/dt / Ea _{dyn}
FloTrac sensor or FloTrac Jr sensor	•	•	•	*	•	•			
TruWave transducer					•	•	•	•	
Acumen IQ sensor**	•	•	•	*	•	•			•
*A CVP analog input signal, CVP monitoring, or CVP manual entry is needed to calculate SVR. **The Acumen IQ sensor is required to access the AFM software feature. For more information, see Assisted Fluid Management on page 299.									

Table 3-4 Finger cuff options for monitoring parameters with HemoSphere ClearSight module

Finger cuff options (one required)	Monitored and calculated parameters						
	CO	SV	SVV/PPV	SVR	PR	SYS/DIA/MAP	HPI/dP/dt / Ea _{dyn}
ClearSight finger cuff or ClearSight Jr finger cuff	•	•	•	*	•	•	
Acumen IQ finger cuff	•	•	•	*	•	•	•
Acumen IQ Plus finger cuff	•	•	•	*	•	•	•
VitaWave finger cuff**					•	•	
VitaWave Plus finger cuff**					•	•	
*A CVP analog input signal, CVP monitoring, or CVP manual entry is needed to calculate SVR. **Monitoring with a VitaWave cuff or VitaWave Plus cuff is an advanced feature. Activation is only available in certain areas. Please contact your local Edwards representative for more information on enabling this advanced feature.							

Table 3-5 Catheters required for monitoring parameters with HemoSphere oximetry cable

Required catheter	Monitored and calculated parameters	
	ScvO ₂	SvO ₂
PediaSat oximetry catheter or compatible central venous oximetry catheter	•	
Swan-Ganz oximetry catheter		•

Table 3-6 Accessories required for monitoring parameters with HemoSphere technology module

Required accessory	Tissue oximetry (StO ₂)
ForeSight oximeter cable	•
ForeSight/ForeSight Jr sensor	•

WARNING **Shock hazard!** Do not attempt to connect/disconnect system cables while hands are wet. Ensure that hands are dry prior to disconnecting system cables.

CAUTION Always grasp the connector, not the cable, when connecting or disconnecting cables. Do not twist or bend the connectors. Confirm that all sensors and cables are connected correctly and completely before use.

To avoid corruption of data on the HemoSphere advanced monitor, always disconnect the patient CCO cable and oximetry cable from the monitor before using a defibrillator.

3.2 HemoSphere Advanced Monitor Connection Ports

The following monitor views illustrate the connection ports and other key features of the front, rear, and side panels of the HemoSphere advanced monitor.

3.2.1 Monitor Front




Figure 3-1 HemoSphere advanced monitor front view

3.2.2 Monitor Rear




Figure 3-2 HemoSphere advanced monitor rear view (shown with HemoSphere Swan-Ganz module)

3.2.3 Monitor Right Panel




Figure 3-3 HemoSphere advanced monitor right panel

3.2.4 Monitor Left Panel




Figure 3-4 HemoSphere advanced monitor left panel (shown with no modules)

3.3 HemoSphere Advanced Monitor Installation

3.3.1 Mounting Options and Recommendations

The HemoSphere advanced monitor should be placed on a stable flat surface or securely mounted on a compatible stand, according to your institution's practices. The operator should be positioned in front of the monitor and at close proximity during use. The device is intended to be used by only one user at a time. A roll stand for the HemoSphere advanced monitor is available as an optional accessory. See *Additional Accessories Description* on page 382 for more information. Contact your local Edwards representative for recommendations on additional mounting options.

WARNING **Explosion Hazard!** Do not use the HemoSphere advanced monitor in the presence of flammable anesthetic mixture with air or with oxygen or nitrous oxide.

This product contains metallic components. Do NOT use in a Magnetic Resonance (MR) environment.

Make sure the HemoSphere advanced monitor is securely positioned or mounted and that all cords and accessory cables are appropriately arranged to minimize the risk of injury to patients, users or the equipment.

Use of this equipment adjacent to or stacked with other equipment should be avoided because it could result in improper operation. If such use is necessary, this equipment and the other equipment should be observed to verify that they are operating normally.

The HemoSphere advanced monitor must be positioned in an upright position to ensure IPX1 ingress protection.

Do not allow any liquids to splash onto the monitoring screen. Liquid buildup may disable the touchscreen functionality.

Do not position the monitor so that it is difficult to access rear panel ports or power cord.

Equipment is rated for use with high-frequency surgical equipment. Inaccurate parameter measurements can be caused by interference from high-frequency surgical equipment. To reduce hazards that can arise from the use of high-frequency surgical equipment, only use undamaged patient cables and accessories connected as specified in this operator's manual.

This system is rated for use with defibrillators. To ensure proper defibrillator-proof operation, only use undamaged patient cables and accessories connected as specified in this operators manual.

All IEC/EN 60950 equipment, including printers, to be positioned no closer than 1.5 meters to the patient's bed.

WARNING Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the HemoSphere advanced monitor, including cables specified by the manufacturer. Otherwise, degradation of the performance this equipment could result.

CAUTION Do not expose the HemoSphere advanced monitor to extreme temperatures. Refer to environmental specifications in appendix A.

Do not expose the HemoSphere advanced monitor to dirty or dusty environments.

Do not obstruct the HemoSphere advanced monitor ventilation openings.

Do not use the HemoSphere advanced monitor in environments where strong lighting makes the LCD screen difficult to view.

Do not use the monitor as a handheld device.

3.3.2 Battery Installation

Open the battery door (figure 3-3) and insert the battery into the battery bay, ensuring pack is fully inserted and seated. Close the battery door and ensure that the latch is securely fastened. Follow instructions below to connect the power cord and then fully charge the battery. Do not use a new battery pack as a power source until it has been fully charged.

NOTE To ensure that the battery charge level displayed on the monitor is accurate, please condition the battery before first use. For information on battery maintenance and conditioning, see *Battery Maintenance* on page 406.

The HemoSphere battery pack is intended as a backup power source during power-loss and can only support monitoring for a limited time period.

WARNING Make sure the battery is fully inserted and the battery door is properly latched. Falling batteries could seriously injure patients or clinicians.

Only use Edwards approved batteries with the HemoSphere advanced monitor. Do not charge the battery pack outside of the monitor. Doing so can damage the battery or injure the user.

To prevent any interruptions to monitoring during power loss, it is recommended to use the HemoSphere advanced monitor with the battery inserted.

In cases of power failure and battery depletion, the monitor will go through a controlled shut off procedure.

3.3.3 Connecting Power Cord

Before connecting the power cord to the rear panel of the monitor, ensure that the power entry cover is installed:

- 1 If the power entry cover is already installed, remove the two screws (figure 3-5) that attach the power entry cover to the rear panel of the monitor.
- 2 Connect the detachable power supply cord. Ensure that the plug is seated securely.
- 3 Attach the power cord entry cover over the plug by routing the power cord through the cover opening and then pressing the cover and gasket up against the rear panel of the monitor, aligning the two screw holes.
- 4 Reinsert the screws to fasten the cover onto the monitor.
- 5 Plug power cord into a hospital grade outlet.

WARNING Do not use the HemoSphere advanced monitoring platform without an installed power cord entry cover. Failure to do so may result in fluid ingress.




Figure 3-5 HemoSphere advanced monitor power entry cover - screw locations

3.3.3.1 Equipotential Connection

This monitor **MUST** be grounded during operation (Class I equipment according to IEC 60601-1). If a hospital grade or three-prong receptacle is not available, a hospital electrician must be consulted to ensure proper grounding. An equipotential terminal is provided on the rear panel of the monitor (figure 3-2) to be connected to an equipotential grounding system (equipotential cable).

WARNING Do not use extension cords or multiple socket devices to connect the power cord. Do not use detachable power cords other than the power cord provided.

To avoid the risk of electric shock, the HemoSphere advanced monitor can only be connected to a supply mains with grounding (protective earth). Do not use three prong to two prong power adaptors.

Grounding reliability can only be achieved when the instrument is connected to a receptacle marked “hospital only”, “hospital grade”, or its equivalent.

WARNING Disconnect the monitor from the AC source by unplugging mains power cable from the AC Mains. The On/Off button on the monitor does not disconnect the system from the AC mains supply.

CAUTION When moving the instrument, be sure to turn off the power and remove the connected power cord.

3.3.4 Connecting and Disconnecting a Hemodynamic Monitoring Module

The HemoSphere advanced monitor is shipped with two standard expansion modules and one L-Tech expansion module. Before inserting a new monitoring technology module, remove the expansion module by pressing the release button to unlatch and slide the blank module out.

Inspect the new module for external damage before installation. Insert the desired monitoring module into the open slot by applying even pressure to slide and click the module into place.

3.3.5 Connecting and Disconnecting a Hemodynamic Monitoring Cable

Both monitoring cable ports are equipped with a magnetic latch mechanism. Inspect the cable for damage before connecting. A monitoring cable will snap into place when it is properly seated in the port. To disconnect a cable, hold at the plug to pull it away from the monitor.

3.3.6 Connecting Cables from External Devices

The HemoSphere advanced monitor utilizes analog input monitored data to calculate certain hemodynamic parameters. This includes data from the pressure input data ports and ECG monitor input port. All analog input cable connections are located on the rear panel of the monitor (figure 3-2). See *Required Accessories for Platform Modules and Cables* on page 70 for a list of calculated parameters available with certain cable connections. For more information on configuring the analog pressure ports, see *Analog Pressure Signal Input* on page 143.

IMPORTANT NOTE The HemoSphere advanced monitor is compatible with pressure and ECG analog inputs from any external patient monitor that has analog output ports which meet the signal input specifications identified in appendix A, table A-5 of this operator's manual. These provide a convenient means to utilize information from a patient monitor to calculate additional hemodynamic parameters for display. This is an optional feature that does not impact the HemoSphere advanced monitor's primary function of monitoring cardiac output (with the HemoSphere Swan-Ganz module) or venous oxygen saturation (with the HemoSphere oximetry cable).

WARNING Only use HemoSphere advanced monitor accessories, cables and or components that have been supplied and labeled by Edwards. Using other unlabeled accessories, cables and or components may affect patient safety and measurement accuracy.

3.4 Initial Start Up

3.4.1 Start Up Procedure

To turn on and off the monitor, press the power button located on the front panel. After turning on the monitor, the Edwards screen is displayed followed by the Power-On Self Test (POST) screen. The POST verifies the monitor meets basic operating requirements by exercising critical hardware components and is performed each time the system is turned on. POST status message is displayed on the startup screen along with system information such as serial numbers and software version numbers.




Figure 3-6 Startup screen

NOTE If the diagnostic tests detect an error condition, a system error screen will replace the startup screen. See chapter 15: *Troubleshooting* or appendix F: *System Care, Service and Support*. Otherwise, call your Edwards Lifesciences representative for assistance.

3.4.2 Select Language

Upon initial HemoSphere advanced monitor startup, language options are offered which affect the displayed language, time and date formats, and units of measurement. The language selection screen appears after the software has initialized and POST is complete. Selecting the language also sets the display units and the time and date format to the default settings for that language (see appendix D: *Monitor Settings and Defaults*).

Each of the language-related settings can be changed later in the **Date/Time** screen of the **General Settings** screen and in the language option through **Settings → General**.

When the language selection screen appears, touch the desired language for use.




Figure 3-7 Language selection screen

NOTE Figure 3-6 and figure 3-7 are examples of startup and language selection screens.

3.4.3 Select Device ID

Upon initial HemoSphere advanced monitor startup, the user can select a **Device ID** or name for the monitor on the **New Patient Data** screen. See *New Patient* on page 137. The **Device ID** defaults to the monitor serial number but can be changed to any 20 character name. The **Device ID** is displayed at the center of the information bar. See *Information Bar* on page 129.

The **Device ID** can be changed at any time from the **General Settings** screen through **Settings → General** using a secure user password. All passwords are set during system initialization. Contact your hospital administrator or IT department for password.

3.5 Power Off and Power Save Mode

To power the monitor off, touch the power button. See ② in Figure 3-1 on page 72. The following options will be displayed:

- **End Session:** Touch **Yes** to stop the current monitoring session and put the monitor in **Power Save Mode**. This prevents a full power cycle and the monitor can restart with screen touch activation.
- **Shutdown:** This will power off the monitor.
- **Cancel:** Returns you to the screen displayed prior to touching the power button.

HemoSphere Advanced Monitor Quick Start

Contents

HemoSphere Swan-Ganz Module Cardiac Output Monitoring	82
Monitoring with the HemoSphere Pressure Cable	85
HemoSphere Oximetry Cable Monitoring	87
HemoSphere Tissue Oximetry Monitoring	90
HemoSphere ClearSight Module Monitoring	92

NOTE	This chapter is intended for experienced clinicians. It provides brief instructions for using the HemoSphere advanced monitor. Refer to the manual chapters for more detailed information, warnings, and cautions.

4.1 HemoSphere Swan-Ganz Module Cardiac Output Monitoring

Refer to figure 4-1 for HemoSphere Swan-Ganz module monitoring connections.





Figure 4-1 HemoSphere Swan-Ganz module monitoring connection overview

- 1 Insert the HemoSphere Swan-Ganz module into the monitor. The module will click when properly engaged.
- 2 Press the power button to turn on the HemoSphere advanced monitor. All functions are accessed through the touch screen.
- 3 Select **Continue Same Patient** button or **New Patient** button and enter new patient data.
- 4 Connect the patient CCO cable to the HemoSphere Swan-Ganz module.
- 5 Select the **Invasive** monitoring mode button on the **Monitoring Mode Selection** window.
- 6 Touch **Start Monitoring** to begin monitoring.
- 7 Touch the settings icon  → **Select Screens** tab  to select the desired monitoring screen view.
- 8 Touch inside of a parameter tile to select the desired key parameter from the parameter tile configuration menu.
- 9 Touch within a parameter tile to adjust **Alarms / Targets**.
- 10 Depending on catheter type, continue to step 1 in one of the following sections:
 - section 4.1.1 for CO monitoring
 - section 4.1.2 for iCO monitoring
 - section 4.1.3 for EDV monitoring

4.1.1 Continuous Cardiac Output Monitoring

Follow steps 1-10 in *HemoSphere Swan-Ganz Module Cardiac Output Monitoring* on page 82 before proceeding.

- 1 Attach the thermistor ① and thermal filament ② Swan-Ganz CCO catheter connections (figure 4-1) to the patient CCO cable.
- 2 Verify that the catheter is properly inserted into the patient.
- 3 Touch the start monitoring icon . A countdown clock will appear on the stop monitoring icon  to indicate the time until the first CO value. After approximately 5 to 12 minutes, when sufficient data has been obtained, a CO value will appear in the parameter tile.
- 4 The time until the next CO measurement is displayed below the stop monitoring icon . For shorter time spans between calculations, select STAT CO (sCO) as a key parameter. sCO is a fast estimate of the CO value. 20-second flow parameters (CO_{20s}/CI_{20s} and SV_{20s}/SVI_{20s}) are available when monitoring pulmonary artery pressure with a connected HemoSphere pressure cable and TruWave DPT. For more information, see *20-Second Flow Parameters* on page 174.
- 5 Touch the stop monitoring icon  to stop CO monitoring.

4.1.2 Intermittent Cardiac Output Monitoring


Follow steps 1-10 in *HemoSphere Swan-Ganz Module Cardiac Output Monitoring* on page 82 before proceeding.

- 1 Attach the Swan-Ganz or Swan-Ganz Jr catheter thermistor connection (①, figure 4-1) to the patient CCO cable.
- 2 Connect the injectate temperature probe to the injectate temperature probe connector ③ on the patient CCO cable. The injectate system type (in-line or bath) is automatically detected.
- 3 Touch settings icon  → **Clinical Tools** tab  → iCO icon .
- 4 Select the following settings on the new set configuration screen:
 - **Injectate Volume:** 10 mL, 5 mL, or 3 mL (bath type probe only)
 - **Catheter Size:** 5.5F, 6F, 7F, 7.5F, or 8F
 - **Comp Constant:** Auto, or keypad appears for manual entry when selected

NOTE The computation constant is automatically calculated according to injectate system type, injectate volume and catheter size. If the computation constant is manually entered, the injectate volume and catheter size selections are set to **Auto**.

- **Bolus Mode:** Auto or Manual

- 5 Touch the **Start Set** button.
- 6 If in automatic bolus mode, **Wait** appears highlighted () until the thermal baseline is achieved. If in manual bolus mode, **Ready** () will appear highlighted when the thermal baseline is achieved. Touch the **Inject** button first to start the bolus procedure.

- 7 When **Inject** becomes highlighted (), use a rapid, smooth, continuous method to inject the bolus with the volume amount previously selected.
- 8 **Computing** is highlighted () and then the resultant iCO measurement is displayed.
- 9 Repeat steps 6-8 up to six times as desired.
- 10 Touch the **Review** button and if necessary, edit the bolus series.
- 11 Touch the **Accept** button.

4.1.3 Continuous End Diastolic Volume Monitoring

Follow steps 1-10 in *HemoSphere Swan-Ganz Module Cardiac Output Monitoring* on page 82 before proceeding. To acquire EDV/RVEF parameters, a Swan-Ganz CCO catheter with RVEDV must be used.

- 1 Attach the thermistor ① and thermal filament ② Swan-Ganz volumetric catheter connections (figure 4-1) to the patient CCO cable.
- 2 Verify that the catheter is properly inserted into the patient.
- 3 Connect one end of the ECG interface cable to the rear panel of the HemoSphere advanced monitor and the other end to the bedside monitor's ECG signal output.

OR

Connect a pressure cable monitoring ART to use PR within the EDV/RVEF calculations.

- 4 Touch the start monitoring icon  to begin CO/EDV monitoring.
- 5 A countdown clock will appear on the stop monitoring icon  to indicate the time until the first CO/EDV value. After approximately 5 to 12 minutes, when sufficient data has been obtained, an EDV and/or RVEF value will appear in the configured parameter tile(s).
- 6 The time until the next CO measurement is displayed on the information bar. For longer time spans between calculations, select STAT parameters (sCO, sEDV, and sRVEF) as key parameters. sCO, sEDV, and sRVEF are fast estimates of CO, EDV, and RVEF.
- 7 Touch the stop monitoring icon  to stop CO/EDV monitoring.

4.2 Monitoring with the HemoSphere Pressure Cable




Figure 4-2 Pressure cable connection overview

4.2.1 Pressure Cable Setup

- 1 Connect monitor connection end of the pressure cable to the HemoSphere advanced monitor.
- 2 Press the power button to turn on the HemoSphere advanced monitor. All functions are accessed through the touch screen.
- 3 Select **Continue Same Patient** button or **New Patient** button and enter new patient data.
- 4 Select the **Minimally-Invasive** monitoring mode button on the **Monitoring Mode Selection** window and touch **Start Monitoring**. The **Zero & Waveform** screen will appear.
- 5 Connect the primed pressure sensor to the pressure cable. The pressure cable LED that surrounds the zero button at ③ will flash green indicating that the pressure sensor is detected.
- 6 Follow all instructions found in pressure monitoring catheter IFU for catheter preparation and insertion procedures.

The HemoSphere pressure cable must be zeroed before each monitoring session.

4.2.2 Zero Pressure Cable

- 1 Touch the Zero & Waveform icon  located on the navigation bar or through the Clinical Tools menu.

OR

Press the physical zero button  directly on the pressure cable and hold for three seconds (see figure 4-2).

- 2 Select the type/location of pressure sensor being used next to the displayed **port** of the connected HemoSphere pressure cable. The choices are:
 - **ART**
 - **CVP**
 - **PAP**

This step can be skipped while monitoring with a FloTrac, FloTrac Jr, or Acumen IQ sensor. If a FloTrac, FloTrac Jr, or Acumen IQ sensor is connected **ART** is the only available pressure option and is automatically selected.
- 3 Level the stopcock valve to the patient's phlebostatic axis position according to the instructions for use.
- 4 Open the stopcock valve to measure atmospheric pressure.
- 5 Press and hold the physical zero button  directly on the pressure cable, or touch the zero button  located on the screen. When zeroing is complete, a tone sounds, and the message "Zeroed" appears with the time and date. The zero button LED will stop blinking and turn off once zeroing is completed successfully.
- 6 Confirm stable zero pressure and turn stopcock such that sensor is reading patient intravascular pressure.
- 7 Touch the home icon  to begin monitoring.
- 8 Touch the settings icon  → **Select Screens** tab  to select the desired monitoring screen view.
- 9 Touch inside of a parameter tile to select the desired key parameter from the parameter tile configuration menu.
- 10 Touch within a parameter tile to adjust **Alarms / Targets**.

NOTE

Advanced features available while monitoring with the HemoSphere pressure cable include the Acumen Hypotension Prediction Index (HPI) software feature and the Acumen assisted fluid management (AFM) software feature. The Acumen AFM software feature utilizes an additional cable, the Acumen AFM cable, while in fluid meter mode. For more information on monitoring with these software features see *Acumen Hypotension Prediction Index (HPI) Software Feature* on page 253 and *Assisted Fluid Management* on page 299.

The alarms limits for the Hypotension Prediction Index parameter (HPI) are not adjustable.

4.3 HemoSphere Oximetry Cable Monitoring





Figure 4-3 Oximetry connection overview


- 1 Connect the HemoSphere oximetry cable to the left side of the HemoSphere advanced monitor. See figure 4-3.
- 2 Press the power button to turn on the HemoSphere advanced monitor. All functions are accessed through the touch screen.
- 3 Select **Continue Same Patient** button or **New Patient** button and enter new patient data.
- 4 Select the **Non-Invasive**, **Invasive** or **Minimally-Invasive** monitoring mode button on the Monitoring Mode Selection window as applicable.
- 5 Touch **Start Monitoring**.
- 6 The HemoSphere oximetry cable must be calibrated before each monitoring session. Continue to section 4.3.1 for in vitro calibration instructions and section 4.3.2 for in vivo calibration instructions.

4.3.1 In Vitro Calibration

- 1 Remove a section of the catheter tray lid to expose the optical connector.
- 2 Insert the optical connector of the catheter “TOP” side up into the oximetry cable and snap the enclosure shut.

- 3 Touch the oximetry calibration icon  on the **ScvO₂/SvO₂** parameter tile or, touch the settings icon  → **Clinical Tools** tab  → **Venous Oximetry Calibration** icon .
- 4 Select oximetry type: **ScvO₂** or **SvO₂**.
- 5 Touch **In vitro Calibration** button.
- 6 Enter either the patient's hemoglobin (**HGB**) or hematocrit (**Hct**) value. A default value may be used until the patient's HGB or Hct is available.
- 7 Touch **Calibrate** button.
- 8 When the calibration successfully completes, the following message appears:
In vitro Calibration OK, insert catheter
- 9 Insert the catheter as described in the catheter directions for use.
- 10 Touch **Start** button.
- 11 If **ScvO₂/SvO₂** are not current key parameters, touch the displayed parameter label located inside of any parameter tile to select **ScvO₂/SvO₂** as a key parameter from the parameter tile configuration menu.
- 12 Touch within the **ScvO₂/SvO₂** parameter tile to adjust **Alarms / Targets**.

4.3.2 In Vivo Calibration

- 1 Insert the catheter as described in the catheter directions for use.
- 2 Insert the optical connector of the catheter "TOP" side up into the oximetry cable and snap the enclosure shut.
- 3 Touch the oximetry calibration icon  on the **ScvO₂/SvO₂** parameter tile or, touch the settings icon  → **Clinical Tools** tab  → **Venous Oximetry Calibration** icon .
- 4 Select oximetry type: **ScvO₂** or **SvO₂**.
- 5 Touch **In vivo Calibration** button.

If setup is unsuccessful, one of the following messages will be displayed:

Warning: Wall Artifact or Wedge Detected. Reposition catheter.

OR


Warning: Unstable Signal.

- 6 If a "Wall Artifact or Wedge Detected," or "Unstable Signal" message appears, attempt to troubleshoot the problem as instructed in *Venous Oximetry Error Messages* on page 362 and touch

Recalibrate button to restart the baseline setup.

OR

Touch **Continue** button to proceed to the Draw operation.

- 7 When baseline calibration is successful, touch **Draw** button and then draw the blood sample and send the blood sample to the lab for measured analysis by co-oximeter.
- 8 Enter **HGB** or **Hct** and **ScvO₂/SvO₂** when lab values are received.
- 9 Touch **Calibrate** button.
- 10 Touch the settings icon  → **Select Screens** tab  to select the desired monitoring screen view.
- 11 Touch the displayed parameter label located inside of any parameter tile to select **ScvO₂/SvO₂** as a key parameter from the parameter tile configuration menu.
- 12 Touch within the **ScvO₂/SvO₂** parameter tile to adjust **Alarms / Targets**.

4.4 HemoSphere Tissue Oximetry Monitoring

The HemoSphere technology module is compatible with the ForeSight oximeter cable and ForeSight/ForeSight Jr sensors. The HemoSphere technology module fits into a standard module slot.

NOTE

The following components may have alternative labeling conventions:

ForeSight oximeter cable (FSOC) may also be labeled as FORE-SIGHT ELITE tissue oximeter module (FSM).

HemoSphere technology module may also be labeled as HemoSphere tissue oximetry module.

ForeSight sensors or ForeSight Jr sensors may also be labeled as FORE-SIGHT ELITE tissue oximetry sensors.

4.4.1 Connecting the HemoSphere Technology Module





Figure 4-4 Tissue oximetry monitoring connection overview

- 1 Insert the HemoSphere technology module into the monitor. The module will click when properly engaged.
- 2 Press the power button to turn on the HemoSphere advanced monitor. All functions are accessed through the touch screen.
- 3 Select **Continue Same Patient** button or **New Patient** button and enter new patient data.
- 4 Ensure proper orientation, then plug the ForeSight oximeter cable into the technology module. Up to two ForeSight oximeter cables can be connected to each technology module.

- 5 Connect the compatible ForeSight/ForeSight Jr sensor(s) to the ForeSight oximeter cable. Up to two sensors can be connected to each ForeSight oximeter cable. See *Attaching Sensors to the Patient* on page 242 and refer to the ForeSight and ForeSight Jr sensor instructions for use for proper application directions.
- 6 Select the **Non-Invasive**, **Invasive** or **Minimally-Invasive** monitoring mode button on the **Monitoring Mode Selection** window as applicable.
- 7 Touch **Start Monitoring**.
- 8 If **StO₂** is not a current key parameter, touch the displayed parameter label located inside of any parameter tile to select **StO₂ <Ch>** as a key parameter from the **Select Parameter** tab of the tile configuration menu, where **<Ch>** is the sensor channel. The channel options are **A1** and **A2** for ForeSight oximeter cable A and **B1** and **B2** for ForeSight oximeter cable B.
- 9 The channel will appear in the upper left corner of the parameter tile. Touch the patient figure  on the parameter tile to access the **Sensor Configuration** tab of the tile configuration menu. 
- 10 Select the Patient monitoring mode: adult  or pediatric 
- 11 Select the anatomical location of the sensor. See table 13-1 on page 241 for a list of available sensor locations.
- 12 Touch the home icon  to return to the monitoring window.
- 13 Touch anywhere in the **StO₂** parameter tile → **Sensor Configuration** tab  to adjust **Skin Check Reminder** or **Averaging** for that sensor.
- 14 Touch anywhere in the **StO₂** parameter tile → **Set Targets** tab  to adjust **Alarms / Targets** for **StO₂**.

4.5 HemoSphere ClearSight Module Monitoring

4.5.1 Connecting the HemoSphere Non-Invasive System




Figure 4-5 HemoSphere non-invasive system connection overview (wrist applied pressure controller)




Figure 4-6 HemoSphere non-invasive system connection overview (Smart pressure controller)

- 1 Insert the HemoSphere ClearSight module into the large technology (L-Tech) slot of monitor. The module will click when properly engaged.
- 2 Press the power button to turn on the HemoSphere advanced monitor. All functions are accessed through the touch screen.
- 3 Select **Continue Same Patient** button or **New Patient** button and enter new patient data.
- 4 Select the **Non-Invasive** monitoring mode button on the **Monitoring Mode Selection** window.
- 5 Connect the wrist applied pressure controller (figure 4-5) or Smart pressure controller (figure 4-6) to the HemoSphere ClearSight module.
- 6 Wrap the pressure controller band around the patient's wrist and attach the compatible pressure controller to the band. Either wrist can be used however the non-dominant arm is preferred.

If monitoring with the HemoSphere Smart pressure controller and compatible finger cuff, use the Smart pressure controller cable clip(s) to attach the pressure controller to the patient and/or applicable location(s).

- 7 Select the proper size finger cuff by using the finger cuff sizing aid.
- 8 Place the finger cuff on the patient's finger. Refer to the product IFU for detailed instructions on proper finger cuff placement and actual device illustrations.

NOTE Cuff sizing may not be applicable to all cuffs.

- 9** Connect finger cuff to pressure controller.

NOTE After 8 hours of accumulated monitoring on the same finger, the HemoSphere non-invasive system will stop monitoring and display a warning to place the cuff on another finger if continued monitoring is desired.

- 10** Connect heart reference sensor to the wrist applied pressure controller.

If monitoring with the HemoSphere Smart pressure controller and compatible finger cuff, skip step 10.

NOTE Monitoring without an HRS is available as an advanced feature in sedated and stationary patients only. To enable the Acumen Hypotension Prediction Index (HPI) feature, when using a wrist applied pressure controller, an Acumen IQ finger cuff and HRS are both required. If using a Smart pressure controller, only an Acumen IQ Plus finger cuff is required. See *Optional HRS* on page 213.

- 11** Apply the heart end of the HRS to the patient at phlebostatic axis level by using an HRS clip.

If monitoring with the HemoSphere Smart pressure controller and compatible finger cuff, skip step 11.

CAUTION Make sure that the HRS is correctly applied so that it can be leveled to the phlebostatic axis.

- 12** Attach the other end of the HRS to the finger cuff.

If monitoring with the HemoSphere Smart pressure controller and compatible finger cuff, skip step 12.

- 13** Touch the start monitoring icon  on the navigation bar or on setup help screen to begin monitoring.

- 14** Touch the stop monitoring icon  on the navigation bar to end monitoring at any time.

- 15** Touch the settings icon  → **Select Screens** tab  to select the desired monitoring screen view.

- 16** Touch inside of a parameter tile to select the desired key parameter from the parameter tile configuration menu.

- 17** Touch within a parameter tile to adjust **Alarms / Targets**.

NOTE	The alarms limits for the Hypotension Prediction Index parameter (HPI) are not adjustable.


Navigating the HemoSphere Advanced Monitor

Contents

HemoSphere Advanced Monitor Screen Appearance	96
Navigation Bar	98
Monitor Views	101
Focused Monitoring Format	118
Clinical Tools	122
Information Bar	129
Status Bar	132
Monitor Screen Navigation	132

5.1 HemoSphere Advanced Monitor Screen Appearance

All monitoring functions are initiated by touching the appropriate area on the touch screen. The navigation bar, located on the left side of the screen, includes various controls for stopping and starting monitoring, scrolling and selecting screens, performing clinical actions, adjusting system settings, capturing screen shots, and silencing alarms. The main components of the HemoSphere advanced monitor screen are shown below in figure 5-1. The main window displays the current monitoring view or menu screen. For details on monitoring view types, see *Monitor Views* on page 101. For details on other screen features, see the referenced sections in figure 5-1.



- ① Status Bar (section 5.7)
- ② Information Bars (section 5.6)
- ③ Parameter Tile (section 5.3.1)
- ④ Main Window/Monitor Views (section 5.3)
- ⑤ Navigation Bar (section 5.2)

Figure 5-1 HemoSphere advanced monitor screen features

5.2 Navigation Bar

The navigation bar is present on most screens. Exceptions are the startup screen and screens indicating the HemoSphere advanced monitor has stopped monitoring. The example shown below for figure 5-2 is during invasive monitoring on a graphical trend monitoring screen. All available icons are described in detail below.





Figure 5-2 Navigation bar and icons


¹*invasive monitoring*, ²*non-invasive monitoring*, ³*graphical trend screens*




Select Monitoring Mode. Touch here to switch between monitoring modes. See *Select Monitoring Mode* on page 122.



Start CO Monitoring. While monitoring with the HemoSphere Swan-Ganz module, the start CO monitoring icon allows the user to initiate CO monitoring directly from the navigation bar. See *Continuous Cardiac Output* on page 171.



Stop CO Monitoring. The stop monitoring icon indicates that CO monitoring using the HemoSphere Swan-Ganz module is underway. The user can immediately stop monitoring by touching this icon and then **OK** on the confirmation popup.



Start Non-Invasive Monitoring. While monitoring with the HemoSphere ClearSight module, the start monitoring icon allows the user to initiate non-invasive blood pressure and CO monitoring directly from the navigation bar. See *Connecting the HemoSphere Non-Invasive System* on page 202.



Stop Non-Invasive Monitoring. The stop non-invasive monitoring icon indicates that non-invasive blood pressure and hemodynamic parameter monitoring using the HemoSphere ClearSight module is underway.



Zero & Waveform. This icon allows the user to access the **Zero & Waveform** screen directly from the navigation bar. See *Zero & Waveform Screen* on page 196.



Intervention Analysis. This icon allows the user to access the Intervention Analysis menu. From here clinical interventions can be logged. See *Intervention Events* on page 108.



Display Blood Pressure Waveform. This icon allows the user to display the blood pressure waveform when a HemoSphere pressure cable and compatible sensor is connected or during non-invasive monitoring. See *Live Blood Pressure Waveform Display* on page 110.



Hide Blood Pressure Waveform. This icon allows the user to hide the blood pressure waveform.



GDT Tracking. This icon displays the GDT Tracking Menu. Enhanced parameter tracking allows a user to manage key parameters in the optimal range. See *Enhanced Parameter Tracking* on page 324.



GDT Tracking/AFM. When the AFM software feature is enabled and an Acumen IQ sensor is connected, the AFM icon appears together with the GDT tracking icon in this split view. Touch this icon on the navigation bar and select GDT tracking or AFM to navigate to that feature.



Patient Data (Demographics Entered). This icon appears on the Navigation bar once patient demographics have been entered.




Patient Data (Demographics Skipped). This icon appears on the Navigation bar when patient demographics have been skipped. Touch this icon at any point to enter patient demographics.



Home. This icon returns the user to the main monitoring screen.



Settings. The settings icon provides access to four configuration screens which include:




Clinical Tools. The clinical actions screen provides access to the following clinical tools:

- **Select Monitoring Mode**
- **iCO** (HemoSphere Swan-Ganz module)
- **Zero & Waveform**
- **Venous Oximetry Calibration** (HemoSphere oximetry cable)
- **Enter CVP**
- **Derived Value Calculator**
- **Event Review**
- **Patient CCO Cable Test** (HemoSphere Swan-Ganz module)
- **Fluid Responsiveness Test** (advanced feature – see *Fluid Responsiveness Test* on page 327)
- **Patient Data** (see *Patient Data* on page 137)
- **HPI Secondary Screen** (advanced feature – see *Acumen Hypotension Prediction Index (HPI) Software Feature* on page 253)
- **ctHb Tools** (ForeSight oximeter cable - see *Relative Change in Total Hemoglobin – Δ ctHb* on page 251)
- **BP Calibration** (HemoSphere ClearSight module)
- **HRS Calibration** (HemoSphere ClearSight module - see *Calibrate the Heart Reference Sensor* on page 218)


NOTE

HPI Secondary Screen is available if the Acumen HPI feature is activated. Activation is only available in certain areas. See *Acumen Hypotension Prediction Index (HPI) Software Feature* on page 253. Please contact your local Edwards representative for more information on enabling this advanced feature.

A description of **Select Monitoring Mode**, **Derived Value Calculator**, **Event Review**, and **CVP Entry** can be found in this chapter (see *Clinical Tools* on page 122). For the remaining clinical actions, refer to the specified module or cable chapter for more information.



Select Screens. The select screens tab allows the user to select the desired number of monitored parameters displayed and the type of monitoring view used to display them, which is highlighted in color (see figure 5-3, “Example of monitoring screen selection window,” on page 102). When a monitoring view screen is selected, that monitoring mode is immediately displayed.




Settings. The settings icon provides access to configuration screens which include:

- **General Settings:** See chapter 6: *User Interface Settings*

- **Advanced Setup:** See chapter 7: *Alarms / Targets* , chapter 7: *Adjust Scales* , and chapter 8: *Data Export and Connectivity Settings*
- **Export Data:** See chapter 8: *Data Export and Connectivity Settings*
- **Demo Mode:** See chapter 7: *Demo Mode*
- **ClearSight:** See chapter 11: *ClearSight Settings and Cuff Options*

Advanced Setup and **Export Data** are password protected menu options. See *Password Protection* on page 135.



Help. See chapter 15: *On Screen Help*



Silence Audible Alarms. This icon silences all audio and visual indicator alarms for up to five minutes. The alarm pause interval options are 1, 2, 3, 4 and 5 minutes. New physiological alarms are silenced during the pause period. Alarms will resume sounding after the pause period has elapsed. Faults are silenced until the fault is cleared and re-occurs. If a new fault occurs, the alarm sound will resume.



Audible Alarms Silenced. Indicates that alarms are temporarily silenced. A countdown timer and “**Alarms Paused**” appear. An alarm paused indicator  will appear on any parameter tile that is currently alarming.

Touch the silence audible alarms icon continuously for five seconds to show additional alarm silencing options (below).



Silence All Alarms Permanently. Touch this icon on the alarm expansion menu to silence all alarms indefinitely. Selecting this alarm silence option requires a **Super User** password. See *Password Protection* on page 135.



Non-Pulsatile Mode. Touch this icon to pause CO monitoring and enter **Non-Pulsatile Mode**. A confirmation banner will appear to confirm suspension of CO monitoring operations. Exception: Blood pressure monitoring, tissue oximetry monitoring, and associated alarms will remain active during **Non-Pulsatile Mode**. See table D-3 on page 391 for active parameters. During **Non-Pulsatile Mode**, all blood pressure averaging time defaults to 5 seconds with a 2 second update rate. See table 6-4 on page 142.



Resume Monitoring. After Non-Pulsatile mode confirmation, a resume monitoring icon and elapsed time will appear on the navigation bar. A “**Non-Pulsatile Mode**” banner will be displayed. To return to monitoring, touch the resume monitoring icon.

5.3 Monitor Views

There are eight classic monitoring views: graphical trend, tabular trend, graphical/tabular trend split screen, physiology, cockpit, physio relationship, goal positioning, and the main monitoring view which is a split between the graphical and cockpit view. Depending on the monitoring view selected, up to eight monitored parameters can be displayed.

In addition to these classic monitoring view formats, three additional focused monitoring views are available. These allow the user to see arterial blood pressure values along with three parameters in a streamlined and focused screen layout. See *Focused Main Screen* on page 120, *Focused Graphical Trend Screen* on page 121, and *Focused Charting Screen* on page 121.

To switch between monitoring views, swipe across the screen with three fingers. Or, to select a monitoring view:

- 1 Touch the settings icon  → **Select Screens** tab . The monitor screen selection menu contains icons that are based upon the look of the monitoring screens.



Figure 5-3 Example of monitoring screen selection window

- 2 Touch the circled number, **1**, **2**, **3**, or **4**, that represents the number of key parameters to be displayed on parameter tiles on monitoring screens. Focused screens, shown on the bottom of the selection window, always display 3 key parameters.
- 3 Select and touch a monitor view button to display the key parameters in that screen format.

5.3.1 Parameter Tiles

Parameter tiles are located on the right side of most monitoring screens. The cockpit monitoring view is composed of larger format parameter globes which function identically as described below.

5.3.1.1 Change Parameters

- 1 Touch the displayed parameter label located inside the parameter tile to change it to a different parameter.
- 2 The tile configuration menu will show the selected parameter highlighted in color and other parameters currently being displayed outlined in color. Available parameters appear on the screen without highlights. Figure 5-4 shows the parameter selection tab of the tile configuration menu that will appear while selecting continuous parameters and monitoring with the HemoSphere Swan-Ganz module. The appearance of this window while monitoring with other HemoSphere modules or cables varies from what is shown in figure 5-4.

Parameters are organized into categories. The categories available are based on the current monitoring mode. Categories, listed below, are grouped together on the parameter selection configuration menu. See figure 5-4.

FLOW. Flow parameters measure blood flow from the left heart and include CO, CI, CPO, CPI, SV, SVI, and SVV.

RESISTANCE. Resistance parameters SVR and SVRI are related to systemic resistance to blood flow.

RV FUNCTION. These parameters which include EDV, EDVI, and RVEF are volumetric indicators of the right ventricle (RV).

ACUMEN. Parameters listed here are only available with a connected Acumen IQ sensor and enabled HPI feature. This includes HPI, Ea_{dyn} , and dP/dt .

PRESSURE. These blood pressure parameters include SYS_{ART} , DIA_{ART} , MAP, SYS_{PAP} , DIA_{PAP} , MPAP, PR, CVP, and PPV.

OXIMETRY. Oximetry parameters include venous oximetry ($SvO_2/ScvO_2$) and tissue oximetry (StO_2) when enabled.




Figure 5-4 Example of key parameter selection tile configuration menu

- 3 Touch an available parameter to select the replacement parameter.
- 4 To change the order of any key parameter, touch and hold the parameter tile until the tile appears with a blue outline. Drag and drop the parameter tile to the new desired location to update the order of key parameters.

5.3.1.2 Change Alarm/Target

The **Alarms / Targets** screen lets the user view and set up alarm and target values for the selected parameter or enable/disable the audible alarm and target settings. Additionally, the target settings can be adjusted with a numbered key pad or with the scroll buttons when a minor adjustment is needed. This screen is accessed by touching the parameter value on a parameter tile or through the parameter settings screen. For more information, see *Alarms / Targets* on page 147.

NOTE There is a two minute inactivity timer associated with this menu screen.

The alarms limits and target ranges for the Acumen Hypotension Prediction Index parameter, HPI, are not adjustable.

5.3.1.3 Status Indicators

A parameter tile is outlined in color to indicate the patient's current status. The color changes as the patient's status changes. Items on the tile that appear underlined can be touched to access a configuration menu. The tiles may display additional information.





Figure 5-5 Parameter tile

Status Bar Messages. When a fault, alert, or alarm condition occurs, the message(s) will be displayed on the status bar until the condition is cleared. When there is more than one fault, alert or alarm, the message is cycled every two seconds.

When a fault condition occurs, parameter calculations are stopped, and each affected parameter tile displays the last value, time, and date at which the parameter was measured.

Continuous Change Interval. This indicator displays the percentage of change or absolute value of change, followed by the time period over which it changed. See *Time Intervals / Averaging* on page 141 for configuration options.



↑ 7%
(5 min)



↑ 38% (20 min)

SVV Filtering Exceeded Indicator. The SVV filtering exceeded indicator symbol  appears on the SVV parameter tile if a high degree of pulse rate variability is detected that could affect the SVV value.

SQI Bar. The SQI bar  is a reflection of the signal quality during oximetry or non-invasive monitoring. Signal quality is based on the catheter condition and positioning within the vessel for intra-vascular oximetry or the near-infrared light tissue perfusion index for tissue oximetry. For oximetry indicator levels, see table 12-3, “Signal quality indicator levels,” on page 228. For non-invasive finger cuff monitoring, SQI is based on the quality of the pressure waveform signal from the plethysmograph sensor of the finger cuff. For non-invasive SQI levels, see table 11-2, “Arterial waveform SQI levels,” on page 216.

Target Status Indicators. The colored indicator outlining each monitoring tile indicates the patient’s clinical status. For indicator colors and their clinical indications, see table 7-2, “Target status indicator colors,” on page 150.

NOTE When using the Acumen Hypotension Prediction Index parameter, HPI, the patient status indicators differ from those described. Refer to *Acumen Hypotension Prediction Index (HPI) Software Feature* on page 253 for the patient status indicators available when using the Acumen Hypotension Prediction Index feature.

5.3.2 Main Monitoring View

The main monitoring view displays a combination of the graphical trend monitoring view (see *Graphical Trend Monitoring View* on page 106) and a semicircular variation of the cockpit monitoring view (see *Cockpit Screen* on page 114). The cockpit gauge that appears on the bottom of the main monitoring view utilizes a semicircular gauge area. See figure 5-6. Key parameters displayed on parameter gauges on the bottom of the main monitoring view can be an additional four key parameters beyond those monitored on the graphical

trends and parameter tiles displayed on the screen. Up to eight key parameters can be viewed on the main monitoring view. The position of any key parameter on the screen can be moved by holding down the parameter tile or parameter gauge then dragging and dropping it to the new desired position.



Figure 5-6 Main monitoring view

5.3.3 Graphical Trend Monitoring View

The graphical trend screen displays the current status and history of monitored parameters. The amount of history shown for monitored parameters can be configured by adjusting the time scale.

When the target range for the parameter is enabled, the graph color codes the plot line, green indicating within the target range, yellow indicating the value is outside the target range but within the physiological alarm range, and red indicating the value is outside the alarm range. When the target range is disabled for the parameter the plot line is white. Color plotting can be disabled through general settings. The colors match those of the clinical target indicator (parameter tile outline) on the key parameter tiles in the graphical trend graph when targets are enabled for the parameter. The alarm limits for each parameter are displayed as colored arrows on the graph y-axis.

NOTE The graphical trend for the Acumen Hypotension Prediction Index parameter, HPI, displays as a white trend line when not in alarm range and a red trend line when in alarm range.




Figure 5-7 Graphical trend screen

To change the time scale of the displayed parameters, touch outside of the plot area along the x or y-axis, and a scale popup menu will appear. Touch the value side of the **Graphical Trend Time** button to select a different time period. To move the order of a trend plot, hold the plot down and drag and drop it to a new location. To combine plots, drop the parameter plot onto another graphical trend plot, or touch the combine icon  located between plots. The y-axis values for the second parameter will appear on the right side of the plot. To return to separate graphical trend plots, touch the expand icon .

5.3.3.1 Graphical Trend Scroll Mode



Up to 72 hours of monitored parameter data can be viewed by scrolling back. To start scrolling, swipe to the right/left or touch the appropriate scroll mode button as shown above. Keep touching the scroll mode button to increase the scroll speed. The screen will return to live mode two minutes after the scroll button has been touched, or if the cancel icon  is touched. The scroll rate will appear between the scroll buttons.

Table 5-1 Graphical trend scroll rates

Scroll setting	Description
	Scrolls at two times the current time scale
	Scrolls at the current time scale (one graph width)
	Scrolls at half the current time scale (one-half graph width)

While in scroll mode the user can scroll to data older than the current time scale displays.

NOTE It is not possible to touch past the most recent data or before the oldest data. The graph will scroll only as far as data is available.

5.3.3.2 Intervention Events

While in the graphical trend screen or other monitoring views that display graphical trend plots such as the main monitoring view, selecting the intervention icon  provides a menu of intervention types, details and a notes section.




Figure 5-8 Graphical trend- intervention window

To enter a **New Intervention**:

- 1 Select the **Intervention** type from the **New Intervention** menu on left. Use the vertical scroll arrows to view all available **Intervention** types.
- 2 Select **Detail** from right menu tab. **Unspecified** is set as a default.
- 3 Select the keyboard icon  to enter notes (optional).
- 4 Touch the enter icon .

To enter a previously used **Intervention**:

- 1 Select the **Intervention** from the **Recents** list tab.
- 2 To add, edit, or remove a note, touch the keyboard icon .
- 3 Touch the enter icon .

Table 5-2 Intervention events

Intervention	Indicator	Type
Intervention	 (green)	Inotrope Vasodilator Vasopressor
Positional	 (purple)	Passive Leg Raise Trendelenburg
Fluids	 (blue)	Red Blood Cells Colloid Crystalloid Fluid Bolus*
Oximetry	 (red)	In vitro Calibration* Draw Blood* In vivo Calibration* HGB Update* Recall Oximetry Data*
Event	 (yellow)	PEEP Induction Cannulation CPB Cross Clamp Cardioplegia Pump Flow Circulatory Arrest Warming Cooling Selective Cerebral Perfusion
Custom	 (gray)	Custom Event BP Calibration*
<i>*System generated markers</i>		

NOTE Interventions initiated through the clinical tools menu, such as **Venous Oximetry**, **BP Calibration**, or fluid responsiveness tests, are system generated and cannot be entered through the intervention analysis menu.

After selection of the intervention type, markers indicating the intervention are visually displayed on all graphs. These markers can be selected for more information. Upon touching the marker, an information balloon will appear. See figure 5-9: "Graphical trend screen - intervention information balloon". The information balloon displays the specific intervention, date, time, and notes pertaining to the intervention. Touching the edit button allows the user to edit intervention time, date, and note. Touching the exit button closes the balloon.

NOTE The intervention information balloon has a 2 minute time out.

Intervention Editing. The time, date, and associated note for each intervention can be edited after initial entry:

- 1 Touch the intervention event indicator  associated with the intervention to be edited.
- 2 Touch the edit icon  located on the information balloon.
- 3 To change the time of the selected intervention, touch on **Time Adjust**, and enter the updated time on keypad.
- 4 To change the date, touch on **Date Adjust**, and enter the updated date on keypad.

NOTE The date or time of system generated intervention markers cannot be edited.

- 5 Touch the keyboard icon  to enter or edit notes.
- 6 Touch the enter icon .




Figure 5-9 Graphical trend screen - intervention information balloon

5.3.3.3 Live Blood Pressure Waveform Display

To display the real-time blood pressure waveform, touch the display pressure waveform icon . The display waveform icon appears on the navigation bar while monitoring with the graphical trend or main monitoring screens. A live pressure waveform graph panel will be displayed above the first monitored parameter graph. A numeric reading of the beat to beat systolic, diastolic and mean arterial pressure will be displayed above the first monitored parameter tile. To change the sweep speed (x-axis scale) of the graph, touch the scale area and a popup menu will appear to allow input of a new sweep speed. If multiple monitoring technologies are connected, touch the parameter name on the waveform parameter tile to switch between monitored pressure waveforms.

To stop display of live blood pressure waveform, touch the hide pressure waveform icon .

NOTE If there are 4 key parameters being displayed when the display pressure waveform button is touched, display of the 4th key parameter is temporarily removed and the blood pressure waveform graph is placed at the top of the remaining 3 Key Parameter trend graphs.

5.3.4 Tabular Trends

The tabular trends screen displays selected key parameters and their history in a tabular format.




Figure 5-10 Tabular trend screen

- 1 To change the interval between values, touch inside the table.
- 2 Select a value on the **Tabular Increment** popup.




Figure 5-11 Tabular increment popup

5.3.4.1 Tabular Trend Scroll Mode



Up to 72 hours of data can be viewed by scrolling back. The scroll mode is based on the number of cells. Three scroll speeds are available: 1x, 6x, and 40x.

While the screen scrolls, the date appears above the table. If the time period overlaps two days, both dates will appear on the screen.

- 1 To start scrolling, touch and hold one of the double arrows below the parameter tiles. The scroll rate will appear between the scroll icons.

Table 5-3 Tabular trend scroll rates

Setting	Time	Speed
	one cell	Slow
	six cells	Moderate
	forty cells	Fast

- 2 To exit scroll mode, stop touching the scrolling arrow or touch the cancel icon .

NOTE The screen will return to live mode two minutes after the last touch of the scroll arrow icon or if the cancel icon is touched.

5.3.5 Graphical/Tabular Trends Split

The graphical/tabular trends split screen displays a combination of the graphical trend and tabular trend monitoring views. This display is useful for viewing the current status and history of selected monitored parameters in graphical format and other selected monitored parameters in tabular format at the same time.

If two key parameters are selected, the first key parameter is displayed in graphical trend format, and the second in tabular trend format. Key parameters can be changed by touching the parameter label located on the parameter tile. If more than two key parameters are selected, the first two parameters are displayed in graphical trend format, and the third and fourth — if a fourth is selected — are displayed in tabular trend format. The time scale for data displayed on any key parameter graphical trend view(s) is independent of the time scale displayed on the tabular trend view(s). For more information on the graphical trend view see *Graphical Trend Monitoring View* on page 106. For more on tabular trend view, see *Tabular Trends* on page 111.

5.3.6 Physiology Screen

The Physiology screen is an animation depicting the interaction between the heart, blood, and vascular system. The appearance of this screen varies based on the monitoring technology being used. For example, if the tissue oximetry feature is enabled three additional animations are used to display available tissue oximetry measurement sites along with hemodynamic parameters. See *Tissue Oximetry Physiology Screen* on page 252. Continuous parameter values are displayed in association with the animation.




Figure 5-12 Physiology screen while monitoring with HemoSphere Swan-Ganz module

In the physiology screen the image of the beating heart is a visual representation of the heart rate and is not an exact representation of beats per minute. Key features of this screen are numbered shown in figure 5-12. This example is of the continuous physiology screen during active monitoring with the HemoSphere Swan-Ganz module and analog input ECG, MAP, and CVP signals.

- 1 ScvO₂/SvO₂ parameter data and signal quality indicator (SQI) are displayed here while the HemoSphere oximetry cable is connected and actively monitoring venous oxygen saturation.
- 2 Cardiac output (CO/CI) is indicated on the arterial side of the vascular system animation. The blood flow animation rate will adjust based on the CO/CI value and the low/high target ranges selected for that parameter.
- 3 Systemic Vascular Resistance, indicated in the center of the vascular system animation, is available while monitoring CO/CI and utilizing MAP and CVP analog pressure signal inputs from a connected patient monitor or two HemoSphere pressure cables, as $SVR = [(MAP - CVP) / CO] * 80$. While in minimally-invasive monitoring mode, only CVP is required using the CVP entry screen, CVP monitoring through a HemoSphere pressure cable or through analog input. The level of constriction shown in the vessel will adjust based on the derived SVR value, and the low/high target ranges selected for that parameter.

NOTE

The alarms/targets settings can be adjusted through the Alarms / Targets setting screen (see *Alarms / Targets Setup Screen* on page 150) or by selecting the desired parameter as a key parameter, and accessing the tile configuration menu by touching inside of the parameter tile.

The example shown in figure 5-12 is while monitoring with a HemoSphere Swan-Ganz module. Differences in appearance and parameters will occur with other monitoring

modes. For example, while monitoring within FloTrac sensor, FloTrac Jr sensor, or Acumen IQ sensor monitoring mode, HR_{avg} is replaced by PR, PPV and SVV appear (if configured), and EDV and RVEF are not shown.

5.3.6.1 SVV Slope Indicator

The SVV slope indicator is a visual representation of the Frank-Starling curve used when assessing the stroke volume variation (SVV) value. This appears on the physiology screen while in minimally-invasive and non-invasive monitoring modes. The color of the lantern changes based upon set target ranges. An SVV value of 13% is displayed approximately at the inflection point of the curve. The indicator is displayed on the physiology and historic physiology screens.



The user has the ability to enable or disable the display of the SVV lantern, parameter value, and the SVV filtering exceeded indicator from the monitor settings – monitoring screens settings menu. The default setting is enabled. The system will not show the SVV lantern on the SVV indicator curve when the SVV filtering exceeded indicator is on.

5.3.7 Cockpit Screen

This monitoring screen, shown in figure 5-13, displays large parameter globes with the values of the parameter being monitored. Cockpit parameter globes graphically indicate alarm/target ranges and values, and utilize needle indicators to show where the current parameter value falls. Similar to standard parameter tiles, the value within the globe will flash when the parameter is alarming.




Figure 5-13 Cockpit monitoring screen

The key parameter globes shown on the cockpit screen display a more complex target and alarm indicator than the standard parameter tile. The full display range of the parameter is used to create a gauge from the graphical trends minimum to maximum settings. A needle is used to indicate the current value on the circular gauge scale. When target ranges are enabled, red (alarm zone), yellow (warning target zone), and green (acceptable target zone) are used to indicate the target and alarm regions within the circular gauge. When target ranges are not enabled, the circular gauge area is all gray in color and target or alarm indicators are removed. The value indicator arrow changes to indicate when the values are out of the gauge scale limits.

5.3.8 Physio Relationship

The physio relationship screen depicts the balance between oxygen delivery (DO_2) and oxygen consumption (VO_2). It automatically updates as parameter values change so the values are always current. The connecting lines highlight the relationship of the parameters to each other.

5.3.8.1 Continuous and Historical Modes

The physio relationship screen has two modes: continuous and historical. When in continuous mode, the intermittent and derived values are always displayed as unavailable. HGB is the exception and is displayed as in intermittent parameter in continuous mode with a time stamp of last calculated/entered value.




Figure 5-14 Physio relationship screen while monitoring with HemoSphere Swan-Ganz module

- 1 The vertical lines above and below the parameters appear in the same color as the parameter lantern.
- 2 Vertical lines that directly connect two parameters will appear in the same color as the parameter lantern below (for example, between SVRI and MAP in figure 5-14).
- 3 The horizontal lines are the same color as the line above them.
- 4 The left bar appears after a bolus set is performed. Touch the clock/waveform icon to display historical data when available (see figure 5-14).
- 5 Touch the iCO icon, when available, to open the thermodilution new set configuration screen.

NOTE

The example shown in figure 5-14 is while monitoring with a HemoSphere Swan-Ganz module. Differences in appearance and parameters will occur with other monitoring modes. For example, while monitoring within minimally-invasive monitoring mode, HR_{avg} is replaced by PR, PPV and SVV appear (if configured), and EDV and RVEF are not shown.

NOTE Before a thermodilution set is performed and before any values are entered (see 5.3.8.2 *Parameter Boxes* below) the clock/waveform and iCO icons do not appear. Only the available continuous parameters are displayed.




Figure 5-15 Historic physio relationship data screen

NOTE The historic physio relationship screen displays most of the parameters available on the system at a point in time. The screen displays lines connecting the parameters, highlighting the relationship of the parameters to each other. The historic physio relationship screen displays the configured key parameters (up to eight) on the right hand side of the screen. There is a horizontal tab composite at the top that allows the user to navigate through the database of historic records. The record times correspond to thermodilution bolus sets and derived value calculations.

The historic physio relationship screen allows the user to enter parameters used to calculate derived parameters **DO₂** and **VO₂**, on only the most recent record. The values entered are for the time of the record and not the current time.

The historic physio relationship screen is accessed through the clock/waveform icon on the continuous physio relationship screen. Touch the return icon  to return to the continuous physio relationship screen. There is no 2 minute time-out for this screen.

To calculate **DO₂** and **VO₂**, the partial pressure of arterial (PaO₂) and venous (PvO₂) oxygen is required. For the historic physio relationship screen, a PaO₂ and PvO₂ value of zero (0) is used. To calculate DO₂ and VO₂ using values other than zero (0) for PaO₂ and PvO₂, use the **Derived Value Calculator** (see section 5.5.3 on page 124).

5.3.8.2 Parameter Boxes

Each small parameter box displays:

- Parameter name
- Parameter units
- Parameter value (if available)
- Clinical target status indicator (if a value is available)
- SVV indicator (when applicable)
- Parameter time stamp (for HGB)

If the parameter is in a fault state, the value is blank, indicating it is or was unavailable at the time of the display.




Figure 5-16 Physio relationship parameter boxes

5.3.8.3 Setting Targets and Entering Parameter Values

To change the target settings or enter a value, touch a parameter to bring up the target/enter popup. The physio relationship target/enter popup will be displayed when the following physio relationship small parameter boxes are touched:

- **HGB**
- **SaO₂**
- **SvO₂/ScvO₂** (when no HemoSphere oximetry cable measurement is available)




Figure 5-17 Physio relationship target/enter popup

When the value is accepted, a new time-stamped historic physio relationship record is created. It includes:

- Current continuous parameter data
- The entered value and any derived calculated values.

The historic physio relationship screen is shown with the newly created record; you can then enter the rest of the manually entered values to calculate any derived values.

5.3.9 Goal Positioning Screen

The Goal Positioning screen allows the user to monitor and track the relationship of two key parameters by plotting them against each other on an XY plane.

A single, pulsating blue dot represents the intersection of the two parameters and moves in real time as parameter values change. The additional circles represent the historical parameter trend with the smaller circles indicating older data.

The green target box represents the intersection of the green parameter target zone. The red arrows on the X and Y axis represent the parameter alarm limits.

The first two key parameters selected represent the parameter values plotted on the y and x-axis respectively, as shown in figure 5-18.




Figure 5-18 Goal positioning screen

The following adjustments can be made on this screen:

- To adjust the time interval between the historical trend circles, touch the trend interval icon  displayed on the screen.
- Continue touching the trend interval icon until **Off** appears to turn off historical trend circles.
- To adjust the scale of the X or Y axis, touch along the corresponding axis.
- If the current intersection of parameters moves outside the scale of the X/Y plane, a message will appear indicating this to the user.

5.4 Focused Monitoring Format

The focused monitoring format allows the user to see arterial blood pressure values along with monitored data for up to three key parameters in a streamlined screen layout.

5.4.1 Select Monitoring View


To select a monitoring view in the focused monitoring format, touch settings icon  → **Select Screens**

tab . See figure 5-3 on page 102.

The focused monitoring view has three available monitoring views:



1 Focused Main (see *Focused Main Screen* on page 120)



2 Focused Graphical Trend (see *Focused Graphical Trend Screen* on page 121)



3 Focused Charting (see *Focused Charting Screen* on page 121)

The three focused monitoring formats are displayed across the bottom of the monitoring selection menu, with buttons that are based upon the monitoring screen appearance. Touch a monitor view button to display the key parameters in that screen format.

NOTE If four parameters are selected while monitoring using the formats outlined in *Monitor Views* on page 101 and monitoring is switched to focused monitoring format, only the first three selected parameters are displayed.

5.4.2 Blood Pressure Waveform Tile

All focused monitoring views contain the blood pressure waveform display. See *Live Blood Pressure Waveform Display* on page 110. The focused pressure waveform display uses a format similar to the focused parameter tile, which is described below, to display blood pressure numeric values.

5.4.3 Focused Parameter Tile

The key element in the focused monitoring view is a focused parameter tile. The focused parameter tile displays information similar to the classic parameter tile described in *Parameter Tiles* on page 102. In the focused view, the entire color of the tile changes to match the target status color. For example, the background color of the tile shown figure 5-19 is green; the value is within target range. If monitoring is disabled or paused, the background is black.




Figure 5-19 Focused parameter tile

5.4.4 Change Parameters

To change parameters while in focused monitoring view, touch anywhere above the central line of the parameter tile, where the parameter name is displayed. See figure 5-20.




Figure 5-20 Focused parameter tile – parameter and alarm/target selection


The parameter selection menu will appear. See figure 5-4. On the parameter selection menu, parameters are organized into categories. See *Change Parameters* on page 102 for a description of these categories. Currently selected parameters are highlighted in blue. Other monitored parameters are outlined in blue. Select any available parameter – one that is not highlighted – to actively monitor that parameter.

5.4.5 Change Alarms /Targets

To change the alarms or targets for a key parameter while in focused monitoring view, touch anywhere below the central line of the parameter tile, where the parameter value and units are displayed. The **Alarms / Targets** menu will appear for that parameter. For more information on this menu, see *Alarms / Targets* on page 147.

5.4.6 Focused Main Screen

Within the focused main screen, up to three parameters are displayed in columns and the arterial waveform is displayed across the top of the screen. Each column is titled as the parameter category (for example: **Flow**, **Resistance**, or **Pressure**) and displays a centered parameter tile, the continuous % change or reference value (if enabled), and a vertical target meter on the left side of the column. See figure 5-21.



Vertical target meter on side displays the patient's current parameter value and highlights zone

Figure 5-21 Focused main screen

The vertical meter highlights the target zone of the current value. This will match the color of the parameter tile. To change the parameter value change interval – displayed as a percentage or value – touch the displayed value at the bottom of the parameter column to toggle between interval options (0, 5, 10, 15, 20, 30 mins, or from a reference value when displaying a value change). See *Time Intervals / Averaging* on page 141.

5.4.7 Focused Graphical Trend Screen

The focused graphical trend screen displays a graphical plot of the parameter view over time. The elements of this view match that of the graphical trends view in described in *Graphical Trend Monitoring View* on page 106. See that section for information on Intervention Events and Graphical Trend Scroll Mode.




Figure 5-22 Focused Graphical Trend Screen

The focused graphical trend view is displayed in row format with the parameter category and vertical meter on the left side, the trend plot centered, and the parameter tile on the right side. See figure 5-22. Adjust the time scale or upper/lower display limit of the parameter value by touching anywhere on the x- or y-axis of the parameter trend plot. See *Adjust Scales* on page 153 for information on setting the display ranges for all parameters. Menu options selected through the parameter settings menu affect views in all graphical trend formats – the focused graphical screen and the graphical trend view described in *Graphical Trend Monitoring View* on page 106.

5.4.8 Focused Charting Screen

The focused charting screen displays all available parameters for up to three parameter categories described in *Change Parameters* on page 102. Only the top parameter, which is displayed as a parameter tile, can be configured as a key parameter and display/sound alarms (alarm-able). To change the key parameter, touch the parameter name above the parameter tile line. The parameter selection menu for focused charting view only displays those parameters available within the selected parameter category. The font color of parameter

values displayed below the top parameter tile indicate the color of the current target range. The targets for these unconfigured parameters can be adjusted by touching anywhere on the smaller parameter tile and accessing the **Alarms / Targets** configuration menu for that parameter.




Figure 5-23 Focused Charting Screen

To change the displayed parameter category, touch the currently configured parameter category displayed at the top of the column. A popup menu will appear (figure 5-24). Touch the replacement parameter category.



Figure 5-24 Focused Charting View – Configure Columns

5.5 Clinical Tools

Most options on the clinical actions menu are related to the current monitoring mode (e.g., while monitoring with the HemoSphere Swan-Ganz module). The following clinical actions are available across all monitoring modes.

5.5.1 Select Monitoring Mode

The **Select Monitoring Mode** page allows the user to switch between monitoring modes. This screen will appear before a new monitoring session is started. This screen can also be accessed by:

- a** touching the select monitoring mode icon at the top of the navigation bar




OR

- b** touching the settings icon  → **Clinical Tools** tab  **Clinical Tools** → **Select Monitoring Mode** icon 

From this screen, the user can select from connected monitoring technologies. Oximetry monitoring is available across all monitoring modes.



Minimally-Invasive Monitoring Mode Button. The user can select this button for minimally-invasive hemodynamic monitoring using the HemoSphere pressure cable. Monitoring with a TruWave DPT is also available while in this mode.



Invasive Monitoring Mode Button. The user can select this button for invasive hemodynamic monitoring using a HemoSphere Swan-Ganz module.



Non-Invasive Monitoring Mode Button. The user can select this button for non-invasive hemodynamic monitoring using a HemoSphere ClearSight module.

Touch the home icon  to continue with the selected monitoring mode. The letter “S” (S) will appear on the x-axis of the graphical trends monitoring view at the point in time when the monitoring mode switch occurred.

5.5.2 CVP Entry

The CVP Entry screen allows the user to input a patient’s CVP value to derive continuous SVR/SVRI calculation when MAP data is also available.

- 1 Touch the settings icon  → **Clinical Tools** tab  → **Enter CVP** icon .
- 2 Enter the CVP value.
- 3 Touch the home icon  to return to the main monitoring screen.

NOTE

CVP entry is not available when the HemoSphere pressure cable and a TruWave transducer are monitoring CVP (see table 5-4 on page 124 and *Pressure Cable Monitoring with a TruWave DPT* on page 193).

The default value for CVP when no source is detected is 5 mmHg. If using the default CVP value (5 mmHg), periodically review and update CVP using CVP manual entry as changes are necessary when the actual CVP value differs significantly. This default value can be changed. See *CVP Settings* on page 155.

CVP values can be sourced in the following ways:

- Monitored directly with a TruWave pressure transducer and HemoSphere pressure cable (see *Pressure Cable Monitoring with a TruWave DPT* on page 193).
- Sourced from an external monitoring device with an analog input (see *Analog Pressure Signal Input* on page 143).
- As a static value entered manually by the user (**CVP Entry**).

When multiple sources for CVP are available, the monitor will prioritize the values according to table 5-4 on page 124.

Table 5-4 CVP value prioritization

Priority	CVP value used
1	HemoSphere pressure cable and TruWave pressure transducer
2*	Analog input
3	Manual CVP Entry / default CVP value
*An analog input source for CVP can be switched to manual entry through the CVP Entry screen	

5.5.3 Derived Value Calculator


The **Derived Value Calculator** allows the user to compute certain hemodynamic parameters and provides a convenient way to display these parameters for one-time calculation.

Calculated parameters are based on monitoring mode and may include: CPO/CPI, DO₂/DO₂I, ESV/ESVI, SVI/SV, VO₂/VO₂I, VO₂e/VO₂Ie, SVR/SVRI, LVSWI, RVSWI, and PVR.

- 1 Touch the settings icon  → **Clinical Tools** tab  → **Derived Value Calculator** icon .
- 2 Enter the required values and the derived calculations will automatically display.
- 3 Touch the home icon  to return to the monitoring screen.

5.5.4 Event Review

Use **Event Review** to view parameter-related and system events that occurred during monitoring. This includes the start and end time of any faults, alerts, physiological alarms, or system messages. Up to 72 hours of events and alarm messages are recorded in order with the most recent event at the top.

- 1 Touch the settings icon  → **Clinical Tools** tab  → **Event Review** icon .
- OR
touch the **Event Review** shortcut on the information bar .
- 2 To view system logged events (see table 5-5) select the **Events** tab. To view system generated messages, touch the **Alarms** tab. To scroll up or down on either screen, touch the arrow keys.
- 3 Touch the home icon  to return to the monitoring screen.

The following events are included in the **Events** tab of the event review log.

Table 5-5 Reviewed events

Event	When time logged
Acumen IQ Sensor Zeroed	A connected Acumen IQ sensor is zeroed
AFM - Fluid Bolus #{0} Started (User Bolus)	An AFM session is active and a user-specified bolus is started {0} is the number identifying the bolus within the current AFM session Note: {0}(number) is inclusive of boluses started per recommendation of the AFM algorithm and user-specified boluses
AFM - Fluid Bolus #{0} Stopped ({1} mL, duration: {2} min {3} sec)	An AFM session is active and a bolus is stopped {0} is the number identifying the bolus within the current AFM session {1} is the volume delivered for the bolus {2}, {3} is the time the bolus delivery took in minutes ({2}) and seconds ({3}) Note: {0}(number) is inclusive of boluses started per recommendation of the AFM algorithm and user-specified boluses
AFM - Fluid Bolus #{0} - Analysis Completed	An AFM session is active and a bolus analysis has been completed {0} is the number identifying the bolus within the current AFM session Note: {0}(number) is inclusive of boluses started per recommendation of the AFM algorithm and user-specified boluses
AFM - Fluid Bolus #{0} - Analysis Declined	An AFM session is active and a bolus analysis has been declined {0} is the number identifying the bolus within the current AFM session Note: {0}(number) is inclusive of boluses started per recommendation of the AFM algorithm and user-specified boluses
AFM - Fluid Bolus #{0} - Analysis Started	An AFM session is active and a bolus analysis has begun {0} is the number identifying the bolus within the current AFM session Note: {0}(number) is inclusive of boluses started per recommendation of the AFM algorithm and user-specified boluses
AFM - Fluid Bolus #{0} Started	An AFM session is active and a bolus is started per recommendation of the AFM algorithm {0} is the number identifying the bolus within the current AFM session Note: {0}(number) is inclusive of boluses started per recommendation of the AFM algorithm and user-specified boluses
AFM - Fluid Bolus Suggested	The AFM algorithm is suggesting a bolus
AFM - Fluid Not Suggested	The AFM algorithm is not suggesting a bolus
AFM - Fluid Suggestion Declined	An AFM session is active and the user declines a bolus that was suggested by the AFM algorithm
AFM - Test Bolus Suggested	The AFM algorithm is suggesting a test bolus
AFM Approaching Maximum Case Volume: {0} / {1} mL	An AFM session is active and the AFM bolus is paused by the system as tracked case volume is approaching max case volume {0} is the tracked case volume at the end of AFM session {1} is the current max case volume
AFM Exceeded Maximum Case Volume: {0} / {1} mL	An AFM session is active and the AFM bolus is paused by the system as tracked case volume exceeds max case volume {0} is the tracked case volume at the end of AFM session {1} is the current max case volume
AFM Fluid Strategy Changed: {0}	An AFM session is active and the user changes the fluid strategy {0} is the current fluid strategy

Table 5-5 Reviewed events (continued)

Event	When time logged
AFM Fluid Tracking Mode Changed: {0}	An AFM session is active and the user changes the fluid tracking mode {0} is the current fluid tracking mode
AFM Fluid Type Changed: {0}	An AFM session is active and the user changes the fluid type {0} is the current fluid type
AFM Maximum Case Volume Set: {0} mL	An AFM session is active and the user changes the max case volume (or sets it for the first time) {0} is the current max case volume
AFM Session - Suggestions Taken: {0}, SVV $\leq$ 12%: {1}, Total Tracked Volume: {2} mL	An AFM session is active and the AFM session is stopped {0} is the % of fluid suggestions taken / suggestions given by AFM {1} is the time in target for SVV $\leq$ 12% for AFM session {2} is the total tracked volume at the end of AFM session
AFM Session Paused	An AFM session is active and the AFM session is paused
AFM Session Resumed	An AFM session is active and the AFM session is resumed from previously being paused
AFM Session Started - Fluid Tracking: {0}, Fluid Type: {1}, Surgery Mode: {2}, Fluid Strategy: {3}	The user starts an AFM session with a connected fluid meter {0} is the type of fluid tracking (Fluid Meter) {1} is the current fluid type {2} is the current surgery mode {3} is the current fluid strategy
AFM Session Started - Fluid Tracking: {0}, Surgery Mode: {1}, Fluid Strategy: {2}	The user starts an AFM session {0} is the type of fluid tracking (Manual) {1} is the current surgery mode {2} is the current fluid strategy
AFM Session Stopped	An AFM session is stopped
AFM Surgery Mode Changed: {0}	An AFM session is active and the user changes the surgery mode {0} is the current surgery mode
Arterial Pressure Zeroed	A TruWave pressure transducer is zeroed and the label is ART
Averaging Time – 5 seconds	The CO/pressure averaging time changes to 5 seconds
Averaging Time – 20 seconds	The CO/pressure averaging time changes to 20 seconds
Averaging Time – 5 minutes	The CO/pressure averaging time changes to 5 minutes
BP Calibration Cleared	The existing BP Calibration is cleared
BP Calibration Failed REFERENCE: SYS {0}, DIA {1}	Blood pressure calibration has failed where {0} is the user-entered reference value for SYS and {1} is the user-entered value for DIA
BP Calibration Successful REFERENCE: SYS {0}, DIA {1}	Blood pressure calibration is successfully completed where {0} is the user-entered reference value for SYS and {1} is the user-entered value for DIA
BSA Change	The BSA value changes from the previous BSA value (including when BSA goes to/from blank)
Central Venous Pressure Zeroed	A TruWave pressure transducer is zeroed and the label is CVP
CO Cable Test Passed	When the Patient CCO Cable Test was performed and passed

Table 5-5 Reviewed events (continued)

Event	When time logged
CO Monitoring Started	When CO Monitoring is started
CO Monitoring Stopped	When the user or system stops CO monitoring
ClearSight Monitoring Started	The user begins non-invasive system monitoring
ClearSight Monitoring Started (No HRS; Finger {0} {1} above heart level)	The user begins non-invasive system monitoring without an HRS and the verified height offset of the monitored finger is the specified distance above the heart, where {0} is the value and {1} is the unit of measurement (CM or IN)
ClearSight Monitoring Started (No HRS; Finger {0} {1} below heart level)	The user begins non-invasive system monitoring without an HRS and the verified height offset of the monitored finger is the specified distance below the heart, where {0} is the value and {1} is the unit of measurement (CM or IN)
ClearSight Monitoring Started (No HRS; Finger at heart level)	The user begins non-invasive system monitoring without an HRS and the verified height offset between the monitored finger and heart is zero
ClearSight Monitoring Stopped	The user or system stops non-invasive system monitoring
ClearSight Monitoring Resumed	When monitoring resumes after a cuff pressure release
Continuous monitoring has reached the 72 hour limit.	Non-invasive system monitoring has stopped due to 72 hour limit
Cuff 1 Monitoring	Cuff 1 monitoring begins
Cuff 2 Monitoring	Cuff 2 monitoring begins
Cuff Pressure Released	A cuff pressure release has occurred
Cuff Pressure Release Acknowledged	The Acknowledge button is touched on Pressure Release notification popup
CVP Cleared	The user has cleared the manually entered CVP value
CVP Entered <value><units>	A CVP value has been manually entered with the shown value and units
[IA#N] Draw Blood	The Draw option is selected in the In vivo Calibration Draw screen. This is logged as an intervention analysis where #N is the enumeration of interventions for this patient.
FloTrac Sensor Zeroed	The FloTrac, FloTrac Jr, or Acumen IQ sensor is zeroed
FRT Start Baseline	An FRT baseline measurement is started
FRT End Baseline	An FRT baseline measurement is completed with a valid measurement
FRT Cancel Baseline	An FRT baseline measurement is canceled
FRT Unstable Baseline	An FRT baseline measurement is stopped with a valid measurement however the measurement is unstable
FRT Start Challenge	An FRT challenge measurement is started
FRT End Challenge	An FRT challenge measurement is stopped with a valid measurement. This occurs at the end of the challenge duration or when the user touches END NOW .
FRT Cancel Challenge	An FRT measurement is canceled
FRT Insufficient Data	An FRT measurement is stopped and invalid
GDT Session Started: #nn	A GDT Tracking Session is started. 'nn' is the GDT tracking session number for the current patient.
GDT Session Stopped: #nn	A GDT Tracking Session is stopped. 'nn' is the tracking session number for the current patient.
GDT Session Paused: #nn	A GDT Tracking Session is paused. 'nn' is the tracking session number for the current patient.

Table 5-5 Reviewed events (continued)

Event	When time logged
GDT Session Resumed: #nn	A GDT Tracking Session is resumed. 'nn' is the tracking session number for the current patient.
GDT Session Targets Updated: #nn; <pppp>:<qqq><uuu>,<...>	GDT Tracking Session targets are updated. 'nn' is the tracking session number for the current patient, <pppp> is the parameter whose target range <qqq> with units <uuu> was updated. <...> additional targets were updated.
HPI Alert	Acumen Hypotension Prediction Index, HPI, alert becomes active. [HPI only]
HPI Alert Acknowledged*	Acumen Hypotension Prediction Index, HPI, alert is acknowledged*. [HPI only]
HPI Alert Cleared (Acknowledged*)	Acumen Hypotension Prediction Index, HPI, alert is cleared as the HPI value was lower than 75 for the last two consecutive 20-second updates. The HPI high alert popup was acknowledged* prior to the alert clearing. [HPI only]
HPI Alert Cleared (Not Acknowledged*)	Acumen Hypotension Prediction Index, HPI, alert is cleared as the HPI value was lower than 75 for the last two consecutive 20-second updates. The HPI high alert popup was not acknowledged* prior to the alert clearing. [HPI only]
iCO Bolus Performed	When an iCO bolus is performed
In vitro Calibration	When oximetry cable update completes following the In vitro calibration process
In vivo Calibration	When oximetry cable update completes following In vivo calibration process
[IA#N] <sub-type> <detail> <note>	An intervention analysis is performed where #N is the enumeration of interventions for this patient <sub-type> is the intervention sub-type selected (for general Intervention: Inotrope, Vasodilator, or Vasopressor; for Fluid analysis: Red Blood Cells, Colloid, or Crystalloid; for Position Challenge: Passive Leg Raise or Trendelenburg; for Event: PEEP, Induction, Cannulation, CPB, Cross Clamp, Cardioplegia, Pump Flow, Circulatory Arrest, Warming, Cooling, Selective Cerebral Perfusion) <detail> is the selected detail <note> is note added by user
[IA#N] ΔctHb Reset Initiated	The Reset ΔctHb button is touched on the ctHb Tools screen
[IA#N] HGB Update	Oximetry cable update completes following the HGB update process
[IA#N] Custom <detail> <note>	A Custom intervention analysis is performed where #N is the enumeration of interventions for this patient <detail> is the selected detail <note> is note added by user
[IA#N Updated] Note: <updated note>	The note associated with the Nth intervention was edited but the time and date were not edited. Logged when the Accept button on Edit Intervention popup is enabled and touched. N is the enumeration of the original intervention.
[IA#N Updated] Time: <Updated date> - <Updated Time>	The date or time associated with the Nth intervention was edited but the note was not edited. Logged when the Accept button on Edit Intervention popup is enabled and touched. N is the enumeration of the original intervention.
[IA#N Updated] Time: <Updated date> - <Updated Time>; Note: <updated note>	The (time OR date) AND note associated with the Nth intervention were edited. Logged when the Accept button on Edit Intervention popup is enabled and touched. N is the enumeration of the original intervention.
Light Out of Range	When the Oximetry Light Range Fault occurs
Monitoring Mode Switched from {0} to {1}	The user switches between the two specified monitoring modes, where {0} and {1} are Minimally-Invasive mode (with FloTrac/FloTrac Jr/Acumen IQ sensor or TruWave disposable pressure transducer), Invasive mode (with Swan-Ganz/Swan-Ganz Jr catheter), or Non-Invasive mode (with ClearSight/ClearSight Jr/Acumen IQ finger cuff)

Table 5-5 Reviewed events (continued)

Event	When time logged
Monitoring Stopped as Single Cuff Use Has Exceeded 8 Hours	Monitoring for 8 continuous hours on a single cuff has occurred
Non-Pulsatile Mode Entered	Active CO monitoring paused to prevent audible alarms and parameter monitoring. Blood pressure and tissue oximetry monitoring and alarms continued.
Non-Pulsatile Mode Exited	Normal CO monitoring resumed. Audible alarms and parameter monitoring were activated.
Oximetry Cable Disconnected	An oximetry cable disconnection is detected
Positioning Mode: <mode>	The user has started non-invasive system monitoring and the positioning mode is selected as <Patient Sedated and Stationary> or <Variable Patient Positioning>
Postpone Pressure Release	Monitoring is extended to delay a finger cuff pressure release
Pulmonary Artery Pressure Zeroed	A TruWave pressure transducer is zeroed and the label is PAP
[IA#N] Recall Venous Oximetry Data	When recalled oximetry calibration data is accepted by the user
System Restart Recovery	When the system has resumed monitoring without being prompted following a power cycle
Switched Cuff – Restarting	Monitoring is switched from one cuff to the other during non-invasive double cuff monitoring
Time Change	The system clock is updated
Vertical Offset Updated: Finger <position>	The finger to height offset is updated by the user during Patient Sedated and Stationary positioning mode where <position> is the verified height offset between the monitored finger and heart.
* Acknowledgment is logged when the user touches either button on the HPI High Alert popup.	

5.6 Information Bar

The information bar appears on all active monitoring screens and most clinical tools screens. It displays the Device ID, current time, date, battery status, screen brightness menu shortcut, alarm volume menu shortcut, help screen shortcut, event review shortcut, and the lock screen symbol. For information on switching the monitoring mode, see *Select Monitoring Mode* on page 122. While monitoring with the HemoSphere Swan-Ganz module, the parameter information bar may display blood temperature and heart rate from an analog input. While monitoring with the HemoSphere pressure cable, in minimally-invasive monitoring mode, the parameter information bar may display CO/pressure averaging time and HPI parameter values. For more information on the Acumen Hypotension Prediction Index feature (HPI), which is an advanced feature, see *Acumen Hypotension Prediction Index (HPI) Software Feature* on page 253. While monitoring in non-invasive monitoring mode, the information bar may display HPI parameter values and a cuff pressure release countdown clock. See *Cuff Pressure Release Mode* on page 219. When the monitor has an activated HIS, Wi-Fi, or Viewfinder hub connection, the status will be displayed. See table 8-1 on page 161 for Wi-Fi status symbols, table 8-2 on page 162 for HIS connectivity status symbols, and table 8-3 on page 164 for Viewfinder hub connectivity status symbols. Figure 5-25 shows an example of an information bar while monitoring with the HemoSphere Swan-Ganz module with averaged ECG heart rate data from an analog input.




Figure 5-25 Information bar

① alarms silenced	⑧ snapshot	⑮ heart rate ¹
② sensor technology	⑨ screen brightness	⑯ blood temperature ¹
③ Device ID	⑩ alarm volume	⑰ averaging time ²
④ Viewfinder hub status	⑪ lock screen	⑱ HPI parameter ²
⑤ HIS status	⑫ date/time	⑲ cuff release countdown ³
⑥ Wi-Fi status	⑬ help menu	
⑦ battery status	⑭ event review	

¹ invasive HemoSphere Swan-Ganz module monitoring

² minimally-invasive HemoSphere pressure cable monitoring

³ non-invasive HemoSphere ClearSight module monitoring

NOTE Figure 5-25 is an example of an information bar with selected language standard defaults. To see the defaults for all languages, see table D-6, “Language default settings,” on page 396.

5.6.1 Device ID

The Device ID serves as a device identifier on the Viewfinder network. For more information see *Select Device ID* on page 80 and *Viewfinder Hub Connectivity* on page 164.

5.6.2 Battery

The HemoSphere advanced monitor allows for uninterrupted monitoring during power loss when the HemoSphere battery pack is installed. Battery life is indicated on the information bar by the symbols shown in table 5-6. For more information on battery installation, see *Battery Installation* on page 76. To ensure that

the battery charge status displayed on the monitor is correct, it is recommended to perform periodic checks of battery health through battery conditioning. For information on battery maintenance and conditioning, see *Battery Maintenance* on page 406.

Table 5-6 Battery status

Battery symbol	Indication
	The battery has greater than 50% charge remaining.
	The battery has less than 50% charge remaining.
	The battery has less than 20% charge remaining.
	The battery is charging and connected to mains power.
	The battery is fully charged and connected to mains power.
	The battery is not installed.

WARNING To prevent any interruptions to monitoring during power loss, always use the HemoSphere advanced monitor with the battery inserted.

In cases of power failure and battery depletion, the monitor will go through a controlled shut off procedure.

5.6.3 Screen Brightness

To adjust the screen brightness, touch the shortcut located on the information bar .

5.6.4 Alarm Volume

To adjust the alarm volume, touch the shortcut located on the information bar .

5.6.5 Screen Capture

The snapshot icon captures an image of the screen at the current time. A USB stick attached to one of the two USB ports (rear and right panels) of the HemoSphere advanced monitor is required to save the image. Touch the snapshot icon located on the information bar .

5.6.6 Lock Screen

If the monitor is being cleaned or moved, lock the screen. For cleaning instructions refer to *Cleaning the Monitor and Modules* on page 400. The screen will automatically unlock once the internal timer has counted down.

- 1 Touch the lock screen icon .

- 2 Touch the time that the screen will remain locked on the **Lock Screen** popup.




Figure 5-26 Lock screen popup

- 3 A red lock icon will appear on the information bar.
- 4 To unlock the screen, touch the red lock icon  and touch **Unlock Screen** on the **Lock Screen** menu.

5.7 Status Bar

The status bar appears at the top of all active monitoring screens below the information bar. It displays faults, alarms, alerts, some warnings and notifications. When there is more than one fault, alert or alarm, the message is cycled every two seconds. The message number out of total messages is displayed on the left. Touch this to toggle through the current messages. Touch the question icon to access the help screen for non-physiological alarm messages.



Figure 5-27 Status bar

5.8 Monitor Screen Navigation

There are several standard navigational procedures on the screen.

5.8.1 Vertical Scrolling

Some screens will have more information than fits on the screen at one time. If vertical arrows appear on a review list, touch the up or down arrow to see the next set of items.



If selecting from a list, the vertical scroll arrows move up or down one item at a time.



5.8.2 Navigation Icons

There are some buttons that always perform the same function:



Home. The home icon takes you to the most recently viewed monitoring screen and stores any modification made to data on the screen.



Return. The return icon takes you to the previous menu screen and stores any modification made to data on the screen.



Enter. The enter icon stores any modification made to data on the screen and returns to the monitoring screen or brings up the next menu screen.



Cancel. The cancel icon causes any entries to be discarded.

On some screens, for example Patient Data, there is no cancel button. As soon as a patient's data is entered, it is stored by the system.

List buttons. Some of the screens have buttons that appear next to menu text.



In these cases, touching anywhere on the button reveals a list of selectable items associated with the menu text. The button displays the current selection.

Value button. Some screens have square buttons as shown below. Touch the button to display a keypad.




Toggle button. When an option exists between two choices, such as on/off, a toggle button appears.




Touch on the opposite side of the button to switch the choice.

Keypad. Touch the keys on the keypad to enter numeric data.



Keyboard. Touch the keys on the keyboard to enter alphanumeric data.



User Interface Settings

Contents

Password Protection.....	135
Patient Data	137
General Monitor Settings.....	139

6.1 Password Protection

The HemoSphere advanced monitor has three levels of password protection.

Table 6-1 HemoSphere advanced monitor password levels

Level	Digits required	User description
Super User	four	Clinicians
Secure User	eight	Hospital authorized personnel
Edwards User	rolling password	internal Edwards use only

Any settings or features described in this manual that require a password are **Super User** features. The **Super User** and **Secure User** passwords require a reset during system initialization the first time a password screen is accessed. Contact your hospital administrator or IT department for passwords. If a password is entered incorrectly ten times, the password keypad will become locked for a certain time period. Monitoring will remain active. In the event of forgotten passwords, contact your local Edwards representative.

Two settings menu options are password protected: **Advanced Setup** and **Export Data**.

To access the **Advanced Setup** features described below in table 6-2,


touch settings icon  → Settings tab  → **Advanced Setup** button.

Table 6-2 Advanced setup menu navigation and password protection

Advanced setup menu selection	Sub-menu selection	Super User	Secure User	Edwards User
Parameter Settings	Alarms / Targets	•	•	•
	Adjust Scales	•	•	•
	SVV/PPV	•	•	•
	20-Second Flow Settings	•	•	•
	CVP Entry	•	•	•
GDT Settings		•	•	•
Analog Input		•	•	•

Table 6-2 Advanced setup menu navigation and password protection (continued)

Advanced setup menu selection	Sub-menu selection	Super User	Secure User	Edwards User
Setting Profile		no access	•	•
System Reset	Restore Factory Defaults	no access	•	•
	Data Wipe	no access	•	•
	Decommission Monitor	no access	no access	•
Connectivity	Wireless	no access	•(if enabled)	•
	Serial Port Setup	no access	•	•
	HL7 Setup	no access	•(if enabled)	•
	Viewfinder Hub Setup	no access	•(if enabled)	•
Service	Manage Features	no access	•	•
	System Status	no access	•	•
	Software Update	no access	•	•
Change Passwords		no access	•	•
Engineering	Alarm Settings	no access	•	•
	Tissue Oximetry	no access	•	•
	AFM	no access	•	•
	Viewfinder Hub Settings	no access	•	•

To access the **Export Data** features described below in table 6-3,

touch settings icon  → **Settings** tab  **Settings** → **Export Data** button.

Table 6-3 Export data menu navigation and password protection

Export data menu selection	Super User	Secure User	Edwards User
Diagnostics Export	•	•	•
Data Download	•	•	•
Manage Clinical Data	no access	•(if enabled)	•
Export Service Data	•	•	•

6.1.1 Changing passwords

Changing passwords requires **Secure User** access. Contact your hospital administrator or IT department for password. To change passwords:

- 1 Touch settings icon  → **Settings** tab  **Settings** → **Advanced Setup** button.
- 2 Enter the **Secure User** password.
- 3 Touch **Change Passwords** button.
- 4 Enter the new **Super User** and/or **Secure User** password digits in both value boxes until the green check mark appears. A check mark confirms that the minimum digit requirement has been met and both entries of the desired password are identical.

- 5 Touch the **Confirm** button.

6.2 Patient Data

After the system is turned on, the user has the option to either continue monitoring the last patient or to start monitoring a new patient. See figure 6-1 below.

NOTE If data for the last patient monitored is 12 hours or older, the only option is to start a new patient.




Figure 6-1 New or continuing patient screen

6.2.1 New Patient

Starting a new patient clears all previous patient data. The alarm limits and continuous parameters are set to their default values.

WARNING Upon initiation of a new patient session, the default high/low physiological alarm ranges should be checked to ensure that they are appropriate for the given patient.

The user has the option of entering a new patient, with or without specific demographics, upon initial startup of the system or while the system is running.

WARNING Perform **New Patient** or clear the patient data profile whenever a new patient is connected to the HemoSphere advanced monitor. Failure to do so may result in previous patient data in the historical displays.

- 1 After turning on the monitor, the new or continuing patient screen appears (figure 6-1). Touch **New Patient** and continue to step 6.

OR


Touch **Skip** to start monitoring without inputting the patient's demographics and continue to step 15.

OR

If the monitor is already on, touch the settings icon



→ **Clinical Tools** tab



and continue to step 2.

NOTE

If the user skips entering patient demographics, only the following limited parameters can be monitored: StO_2 , ΔctHb , SYS_{ART} , SYS_{PAP} , DIA_{ART} , DIA_{PAP} , MAP, PR, MPAP, CVP

- 2 Touch **Patient Data** icon .
- 3 Touch **End Session** button.
- 4 Touch **Yes** button on the confirmation screen to start a new patient.
- 5 The **New Patient Data** screen appears. See figure 6-2.




Figure 6-2 New Patient Data screen

- 6 Touch the enter key  on the keypad/keyboard to save each patient demographic selection value and return to the patient data screen.
- 7 Touch **Patient ID** button and use the keyboard to enter the patient's hospital ID. Entering this information is optional, and is only available when either the Clinical Data Advanced Feature or Viewfinder connectivity is enabled.
- 8 Touch **Height** button and use the keypad to enter the patient's height. The unit default for your language is at the upper right of the keypad. Touch it to change the unit of measurement.
- 9 Touch **Age** and use the keypad to enter the patient's age.
- 10 Touch **Weight** and use the keypad to enter the patient's weight. The unit default for your language is at the upper right of the keypad. Touch it to change the unit of measurement.
- 11 Touch **Gender** and select **Male** or **Female**.
- 12 The **BSA** is calculated from the height and weight using the DuBois formula.
- 13 If desired, enter the **Room** and **Bed** for the patient. Entering this information is optional.

- 14 Touch the **Next** button.


NOTE The **Next** button is disabled until all patient data is entered.

- 15 Select the appropriate monitoring mode on the **Monitoring Mode Selection** window. See *Select Monitoring Mode* on page 122. Refer to instructions for starting monitoring with the desired hemodynamic monitoring technology.

6.2.2 Continue Monitoring Patient

If the last patient's data is less than 12 hours old, the patient's demographics and patient ID will be displayed when the system is turned on. When monitoring of the last patient is continued, the patient's data is loaded and the trend data is retrieved. The most recently viewed monitoring screen is displayed. Touch **Continue Patient**.

6.2.3 View Patient Data

- 1 Touch settings icon  → **Clinical Tools** tab .
- 2 Touch **Patient Data** icon  to see patient data. The screen will also include an **End Session** button.
- 3 Touch the return icon  to return to the settings screen. The patient demographic popup screen will appear. If returning to the same patient, review the patient demographics and press **Yes** if they are correct.

6.3 General Monitor Settings


The General Monitor Settings are those that affect every screen. These are the display language, units used, alarm volume, snapshot sound, date/time settings, screen brightness, Device ID, and monitoring screen display settings.

The HemoSphere advanced monitor interface is available in several languages. A language selection screen appears the first time the HemoSphere advanced monitor is started. See figure 3-7, "Language selection screen," on page 80. The language screen will not appear again, but the display language can be changed at any time.

The selected language determines the default time and date format. These can also be changed independently of the language selected.

NOTE If power is lost and restored to the HemoSphere advanced monitor, the system settings prior to the power loss, including alarm settings, alarm volume, target settings, monitoring screen, parameter configuration, language and unit selection, are automatically restored to last configured settings.

6.3.1 Change Language

- 1 Touch the settings icon  → **Settings** tab .

- 2 Touch the **General** button.




Figure 6-3 General Monitor Settings


- 3 Touch the value section of the **Language** button and select the desired display language.
- 4 Touch the home icon  to return to the monitoring screen.

NOTE See appendix D for all language default settings.

6.3.2 Change Date and Time Display

English (US) dates default to **MM/DD/YYYY**, and the time defaults to a **12 Hour** clock.


When an international language is selected, the date defaults to the format found in appendix D: *Monitor Settings and Defaults*, and the time defaults to a 24 hour clock.

- 1 Touch the settings icon  → **Settings** tab .
- 2 Touch **General** button.
- 3 Touch the value section of the **Date Format** button and touch the desired format.
- 4 Touch the value section of the **Time Format** button and touch the desired format.
- 5 Touch the value section of the **Time Zone** button to select the desired time zone.
- 6 The monitor time setting can adjust for daylight savings. Select **On** next to “**Automatically adjust for daylight savings**” to enable this adjustment.
- 7 Touch the home icon  to return to the monitoring screen.

6.3.2.1 Adjust Date or Time

The system time can be reset if necessary. When the time or date is changed, trended data is updated to reflect the change. Any retained data is updated to reflect the time change.

NOTE Adjustments to the date or time are disabled when the monitor is paired to the Viewfinder hub and time synchronization is configured.


- 1 Touch the settings icon  → **Settings** tab .
 - 2 Touch **General** button.
 - 3 To change the date, touch the value section of the **Date Adjust** button and enter the date on the keypad.
 - 4 To change the time, touch the value section of the **Time Adjust** button and enter the time.
-

NOTE The time and date can also be adjusted by touching the date/time directly on the information bar.

- 5 Touch the home icon  to return to the monitoring screen.

6.3.3 Monitoring Screens Settings

From the **General Settings** screen, the user can also set physiology and physio relationship monitoring screen and graphical trend monitoring screen options.

- 1 Touch the settings icon  → **Settings** tab .
- 2 Touch the **General** button.
- 3 Select the **Indexed or Non-Indexed** toggle for parameters in the physiology and physio relationship screens.
- 4 Next to **Plot Trends using target colors** select **On** or **Off** to display target colors on graphical trend monitoring screens.

6.3.4 Time Intervals / Averaging

The **Time Intervals / Averaging** screen lets the user select the continuous change % or value interval. During FloTrac/FloTrac Jr sensor monitoring mode, the user can also change the CO/pressure averaging time.

NOTE The screen will return to the monitoring view after two minutes of inactivity.

The **CO/Pressure Averaging Time** value button is only available in FloTrac/FloTrac Jr sensor monitoring mode.

- 1 Touch within a parameter tile to access the parameter configuration menu.
- 2 Touch the **Intervals / Averaging** tab.

6.3.4.1 Display Parameter Value Change

The change in value or percent change in value of a key parameter over a selected time interval can be displayed on a parameter tile.

- 1 Touch the **Change Display** menu button to select the format for which the change interval is displayed: **% Changed** or **Value Difference**.
- 2 Touch the **Change Interval** value button and select one of the following time interval options:
 - None
 - Reference
 - 1 min
 - 3 min
 - 5 min
 - 10 min
 - 15 min
 - 20 min
 - 30 min

If **Reference** is selected, the change interval will be calculated from the start of monitoring. The **Reference Value** can be adjusted on the **Intervals/Averaging** tab of the tile configuration menu.

6.3.4.2 CO/Pressure Averaging Time

Touch the right side of the **CO/Pressure Averaging Time** value button and touch one of the following interval options:

- 5 sec
- 20 sec (default and recommended time interval)
- 5 min

The **CO/Pressure Averaging Time** selection affects the averaging time and display update rate of CO and other additional parameters while in minimally-invasive monitoring mode. See table 6-4 below for details of which parameter averaging and update rates are affected based on menu selection.

Table 6-4 CO/pressure averaging time and display update rates – in minimally-invasive monitoring mode

CO/Pressure Averaging Time menu selection	Parameter update rate		
	5 sec*	20 sec	5 min*
Cardiac Output (CO)	2 sec	20 sec	20 sec
Stroke Volume (SV)	2 sec	20 sec	20 sec
Systolic Pressure (SYS)	2 sec	20 sec [^]	20 sec [^]
Diastolic Pressure (DIA)	2 sec	20 sec [^]	20 sec [^]
Mean Arterial Pressure (MAP)	2 sec	20 sec [^]	20 sec [^]
Pulse Rate (PR)	2 sec	20 sec [^]	20 sec [^]
Central Venous Pressure (CVP)	2 sec [†]	n/a [†]	n/a [†]
Mean Pulmonary Artery Pressure (MPAP)	2 sec [†]	n/a [†]	n/a [†]
Stroke Volume Variation (SVV)	20 sec**	20 sec	20 sec

Table 6-4 CO/pressure averaging time and display update rates – in minimally-invasive monitoring mode

	Parameter update rate		
Pulse Pressure Variation (PPV)	20 sec**	20 sec	20 sec
<i>*When a FloTrac IQ/Acumen IQ sensor is connected and the HPI feature is activated, all parameters will be available with 20 second averaging interval / 20 second update rate only. This includes Acumen parameters: HPI, Ea_{dyn}, and dP/dt.</i> <i>^When using a TruWave transducer or while in Non-Pulsatile mode (except PR), only 5 second averaging with 2 second update rate is available.</i> <i>†Parameter averaging time is always 5 seconds with an update rate of 2 seconds for CVP and MPAP.</i> <i>**When this averaging interval is selected, SVV and PPV are only available with 20 second averaging and a 20 second update rate.</i>			

NOTE For real-time blood pressure waveform displayed on the blood pressure waveform display (see *Live Blood Pressure Waveform Display* on page 110) or on the Zero & Waveform screen (see *Zero & Waveform Screen* on page 196), the update rate is always 2 seconds.

Touch home icon  to return to the monitoring screen.

6.3.5 Analog Pressure Signal Input

While performing CO monitoring, the HemoSphere advanced monitor can also calculate SVR by utilizing analog pressure signal input from a connected patient monitor.

NOTE Connecting to external input devices allows additional information to be displayed. For example, while monitoring with the HemoSphere Swan-Ganz module and when MAP and CVP are available continuously from a bedside monitor, SVR is displayed if configured in a parameter tile. MAP and CVP are displayed on the physio relationship and physiology monitoring screens.

WARNING The analog communication ports of the HemoSphere advanced monitor share a common ground that is isolated from the catheter interface electronics. When connecting multiple devices to the HemoSphere advanced monitor, all devices should be provided with isolated power to avoid compromising the electrical isolation of any of the connected devices.

Risk and Leakage current of the final system configuration must comply with IEC 60601-1:2005/A1:2012. It is the responsibility of the user to ensure compliance.

Accessory equipment connected to the monitor must be certified according to IEC/EN 60950 for data-processing equipment or IEC 60601-1:2005/A1:2012 for electro-medical equipment. All combinations of equipment must be in compliance with IEC 60601-1:2005/A1:2012 systems requirements.


CAUTION When connecting the HemoSphere advanced monitor to external devices, refer to the external device's instruction manual for complete instructions. Verify proper operation of the system before clinical use.

Once the bedside monitor has been configured for the desired parameter output, connect the monitor via an interface cable to the selected analog input port on the HemoSphere advanced monitor.

NOTE A compatible bedside monitor must provide an analog output signal.

Please contact your local Edwards representative to obtain the correct HemoSphere advanced monitor analog input interface cable for your bedside monitor.

The following procedure describes how to configure the analog input ports of the HemoSphere advanced monitor.

- 1 Touch the settings icon  → **Settings** tab .
- 2 Touch **Advanced Setup** button and enter the required password. All passwords are set during system initialization. Contact your hospital administrator or IT department for password.
- 3 Touch **Analog Input** button.
- 4 If monitoring with the HemoSphere Swan-Ganz module, select **MAP** from the **Parameter** list button for the numbered analog port where MAP is connected (1 or 2). The default setting values for MAP will be displayed.

NOTE While in FloTrac/FloTrac Jr sensor monitoring mode, MAP data through analog input is not available.

If an analog signal is not detected on the selected port, “**Not Connected**” will be displayed below the **Port** list button.

When an analog input connection or disconnection is first detected, a brief notification message will be displayed on the status bar.

- 5 Select **CVP** in the **Parameter** list button for the numbered analog port where CVP is connected. The default setting values for CVP will be displayed.
-

NOTE The same parameter may not be configured on more than one analog input at the same time.

While in FloTrac sensor monitoring mode, and a TruWave DPT monitoring CVP is connected, CVP data through analog input is not available.

- 6 If the default values are correct for the bedside monitor being used, touch the home icon .

If the default values are not correct for the bedside monitor being used (refer to the bedside monitor operator's manual), the user can modify the voltage range, full scale range, or perform the calibration option described in section 6.3.5.1 of this chapter.

Touch the **Full Scale Range** value button to change the displayed full scale signal value. Table 6-5 below shows the allowable input values for full scale range based on the selected parameter.

Table 6-5 Analog input parameter ranges

Parameter	Full Scale Range
MAP	0 to 510 mmHg (0 kPa to 68 kPa)
CVP	0 to 110 mmHg (0 kPa to 14.6 kPa)

NOTE A voltage reading of zero is automatically set to a minimum pressure reading of 0 mmHg (0 kPa). The **Full Scale Range** represents the full scale signal or maximum pressure reading for the selected **Voltage Range**.

Touch the **Voltage Range** list button to change the displayed voltage range. The selectable voltage ranges available for all of the parameters are:

- 0 - 1 volts
- 0 - 5 volts
- 0 - 10 volts
- Custom (see 6.3.5.1: Calibration)


WARNING When switching to a different bedside monitor, always check that the default values listed are still valid. If necessary, reconfigure the voltage range and corresponding parameter range or calibrate.

6.3.5.1 Calibration

The calibration option is required when default values are incorrect, or the voltage range is not known. The calibration process configures the HemoSphere advanced monitor with the analog signal received from the bedside monitor.

NOTE If the default values are correct, do not calibrate.

CAUTION Only properly trained personnel should calibrate the HemoSphere advanced monitor analog ports.

- 1 Touch the settings icon  → Settings tab .

- 2 Touch **Advanced Setup** button and enter the required password. All passwords are set during system initialization. Contact your hospital administrator or IT department for password.
- 3 Touch **Analog Input** button.
- 4 Select the desired port number (1 or 2) from the **Port** list button and corresponding parameter (**MAP** or **CVP**) from the **Parameter** list button.
- 5 Select **Custom** from the voltage value popup screen. The **Analog Input Custom Settings** screen will appear.
- 6 Simulate a full scale signal from the bedside monitor to the selected analog input port on the HemoSphere advanced monitor.
- 7 Set the maximum parameter value equal to the full-scale signal value.
- 8 Touch **Calibrate Maximum** button. The **Maximum A/D** value will appear on the **Analog Input Custom Settings** screen.

NOTE If an analog connection is not detected, the **Calibrate Maximum** and **Calibrate Minimum** buttons will be disabled and the Maximum A/D value will be displayed as **Not Connected**.

- 9 Repeat the process to calibrate the minimum parameter value.
- 10 Touch the **Accept** button to accept the displayed custom settings and return to the Analog Input screen.
- 11 Repeat steps 4-10 to calibrate another port if needed, or touch the home icon  to return to the monitoring screen.

CAUTION The accuracy of continuous SVR while monitoring with the HemoSphere Swan-Ganz module depends upon the quality and accuracy of the MAP and CVP data transmitted from the external monitors. Since MAP and CVP analog signal quality from the external monitor cannot be validated by the HemoSphere advanced monitor, actual values and the values (including all derived parameters) displayed by the HemoSphere advanced monitor may not be consistent. The accuracy of continuous SVR measurement, therefore, cannot be guaranteed. To aid in determining the quality of the analog signals, regularly compare the MAP and CVP values displayed on the external monitor to the values displayed on the physio relationship screen of the HemoSphere advanced monitor. Refer to the external input device operator's manual for detailed information regarding accuracy, calibration, and other variables which may impact the analog output signal from the external monitor.

Advanced Settings

Contents

Alarms / Targets.....	147
Adjust Scales.....	153
Physiology and Physio Relationship Screen SVV/PPV Parameter Settings.....	155
CVP Settings.....	155
20-Second Flow Parameter Settings.....	156
Demo Mode.....	156

7.1 Alarms / Targets

There are two types of alarms on the HemoSphere advanced monitor intelligent alarm system:

- Physiological alarms: These are set by the clinician and signify the upper and/or lower alarm ranges for configured key continuous parameters.
- Technical alarms: This alarm signifies a device fault or alert.

Physiological alarms occur with either Medium or High priority. Only parameters that are displayed on tiles (key parameters) will have active visual and audible alarms.

Among technical alarms, faults are of medium or high priority, and will halt operation of the related monitoring activity. Alerts are of low priority and will not halt any monitoring activity.

All alarms have an associated text displayed on the status bar. The intelligent alarm system will actively cycle through every active alarm text on the status bar. In addition, alarms will generate the visual alarm indicator shown in table 7-1 below. For additional information, see table 15-1 on page 337.

Table 7-1 Visual alarm indicator colors

Alarm priority	Color	Light pattern
High	red	Flashing ON/OFF
Medium	yellow	Flashing ON/OFF
Low	yellow	Solid ON

The visual alarm indicator will indicate the highest active alarm priority. Alarm messages displayed on the status bar are outlined in the alarm priority color indicated in table 7-1. The audible tone associated with the highest priority active alarm will be played. Where the priority levels are the same, physiological alarms take priority over faults and alerts. All technical alarms are generated once detected by the system; there is no inherent delay in alarms from the point of detection. For physiological alarms, the delay is the amount of time it takes to calculate the next physiological parameter after the parameter is out of range continuously for five or more seconds:

- HemoSphere Swan-Ganz module continuous CO and associated parameters: varies, but is typically around 57 seconds (see *CO Countdown Timer* on page 173)
- HemoSphere pressure cable continuous CO and associated FloTrac/FloTrac Jr sensor measured parameters: varies based on CO/pressure averaging time menu selection and associated update rate (see table 6-4 on page 142)
- HemoSphere pressure cable arterial blood pressure parameters (SYS/DIA/MAP) while arterial waveform is displayed: 2 seconds
- HemoSphere ClearSight module continuous CO and associated hemodynamic parameters: 20 seconds
- HemoSphere ClearSight module arterial blood pressure parameters (SYS/DIA/MAP) while arterial waveform is displayed: 5 heartbeats
- HemoSphere pressure cable with TruWave pressure transducer measured parameters: 2 seconds
- Oximetry: 2 seconds

NOTE Arterial blood pressure (ART) related physiological and technical alarms will only sound after ART is zeroed and mean arterial pressure (MAP) has 10 continuous readings above 10 mmHg.

All alarms are logged and stored for the given patient and can be accessed via the Data Download function (see *Data Download* on page 158). The Data Download log is cleared when initiating a new patient (see *New Patient* on page 137). The current patient can be accessed from up to 12 hours following a system power-off.

WARNING Do not use alarm settings/presets that differ from the same or similar equipment in any single area, e.g. an intensive care unit or cardiac operating theater. Conflicting alarms can affect patient safety.

7.1.1 Silence Alarms

7.1.1.1 Physiological Alarms

Physiological alarms can be silenced directly from the monitoring screen by touching the silence audible

alarms icon . The physiological alarm audio tone is silenced for a user selected alarm pause time period.

No audio tone or LED visual alarm indicator (blinking yellow or red) for any physiological alarm, medium or high priority, will be emitted during this alarm pause period, including new physiological alarms triggered during this time. If a technical alarm is generated during this alarm pause time period, the audio silence will be cleared, allowing alarm audio tones to resume. The user can also manually clear the alarm pause period by pressing the alarm silence button again. Once the alarm pause period has elapsed, active physiological alarms will resume audio sound.

For information on physiological alarm priorities, see *Alarm Priorities* on page 395.

NOTE Physiological parameters can be configured to have no alarms. See sections 7.1.5 and 7.1.6.

WARNING Do not turn off the audible alarms in situations in which patient safety could be compromised.

7.1.1.2 Technical Alarms

During an active technical alarm, the user can silence the alarm and clear the visual alarm indicator (medium and low priority) by touching the silence audible alarms icon . The visual alarm indicator and audio tone will remain inactive unless another technical or physiological alarm condition triggers, or the original technical alarm resolves and re-triggers.

7.1.2 Set Alarm Volume

The alarm volume ranges from low to high with a default of medium. It applies to physiological alarms, technical faults, and alerts. Alarm volume can be changed at any time.

- 1 Touch the settings icon  → **Settings** tab  **Settings**.
- 2 Touch **General** button.
- 3 Touch the right side of the **Alarm Volume** list button to select the desired volume.
- 4 Touch the home icon  to return to the monitoring screen.

WARNING Do not lower the alarm volume to a level that prohibits alarms from being adequately monitored. Failure to do so could result in a situation where patient safety is compromised.

7.1.3 Set Targets

Targets are visual indicators set by the clinician to indicate if the patient is in the ideal target zone (green), warning target zone (yellow), or alarm zone (red). Target colors are displayed as a shaded outline around parameter tiles (see figure 5-5). The use of target zone ranges can be enabled or disabled by the clinician. Alarms (high/low) differ from target zones in that the alarm parameter value flashes and has an audible alarm.

Parameters that can “Alarm” are indicated by a bell icon  in the **Alarms / Targets** settings screen. High/low alarms by default also become the ranges for the red caution zone for that parameter. Parameters which DO NOT have the ability to set a high/low alarm will not have a bell icon in the **Alarms /Targets** settings screen for that parameter but can still have target ranges set.

Target behavior and range of HPI are described in *HPI on Information Bar* on page 261.

Table 7-2 Target status indicator colors

Color	Indication
Green	Acceptable – Green target zone is considered an ideal range for parameter as set by the clinician.
Yellow	Yellow target zone is considered a warning range and visually indicates that the patient has exited the ideal range but has not entered the alarm or caution range as set by the clinician.
Red	Red alarm and/or target zones can be considered “Alarm” parameters indicated by a bell icon in the Alarms / Targets settings screen. High/low alarms by default also become the range for the red caution zone for that parameter. Parameters which DO NOT have the ability to set a high/low alarm will not have a bell icon in the Alarms / Targets settings screen for that parameter but can still have target ranges set. Ranges for the alarm and/or target zone are to be set by the clinician.
Gray	If a target is not set, the status indicator appears as gray.

7.1.4 Alarms / Targets Setup Screen

The **Alarms / Targets** setup screen allows the clinician to view and set up alarms and targets for each key parameter. From the **Alarms / Targets** screen, located within the **Advanced Setup** settings menu, the user can adjust targets and enable/disable audible alarms. Any features accessed through the **Advanced Setup** settings menu are password protected and should only be altered by experienced clinicians. The settings for each key parameter are displayed in a parameter box. The currently configured key parameters are the first set of key parameters displayed. The remaining key parameters are displayed in a defined order. The parameters also indicate what the target ranges are based on: Custom Default, Edwards Default, and Modified.

Table 7-3 Target defaults

Default name	Description
Custom Default	A custom default target range was set for the parameter and the parameter target range has not been modified from that default.
Edwards Default	The parameter target range has not been changed from the original settings.
Modified	Parameter target range was changed for this patient.

NOTE Visual and audible alarm settings are only applicable to parameters being displayed.

To modify **Alarms / Targets**:

- 1 Touch the settings icon  → **Settings** tab  **Settings**.
- 2 Touch **Advanced Setup** button and enter the required password.
- 3 Touch **Parameter Settings** button → **Alarms / Targets** button.

- 4 Touch anywhere in a parameter box to display the **Alarms / Targets** menu for the parameter.




Figure 7-1 Alarms / Targets configuration

NOTE There is a 2 minute inactivity timer associated with this screen.

The red, yellow and green rectangles are fixed shapes, and don't change size/shape.

7.1.5 Configure All Targets

Alarms / Targets can easily be configured or changed all at the same time. From the **Configure All** screen, the user can:

- Restore all parameter alarm and target settings to Custom Defaults.
- Restore all parameter alarm and target settings to Edwards Defaults.
- Enable or disable audible physiological alarms for all applicable parameters.
- Enable or disable all audible alarms.

- 1 Touch the settings icon  → **Settings** tab .
- 2 Touch **Advanced Setup** button and enter the required **Secure User** password.
- 3 Touch **Parameter Settings** button → **Alarms / Targets** button.
- 4 Touch the **Configure All** button.
 - To enable or disable all audible physiological alarms for all parameters, touch the **Disabled/Enabled** toggle button for **Targets** within the **Audible Alarm** box.
 - To enable or disable all audible technical alarms for all parameters, touch the **Disabled/Enabled** toggle button for **All Alarms** within the **Audible Alarm** box.
 - To restore all settings to the custom defaults, touch **Restore All to Custom Defaults**. The message, “**This action will restore ALL Alarms and Targets to the Custom Defaults.**” appears. Touch **Continue** button on the confirmation popup to confirm the restore.

- To restore all settings to the Edwards defaults, touch **Restore All to Edwards Defaults**. The message, “**This action will restore ALL Alarms and Targets to the Edwards’ Defaults.**” appears. Touch **Continue** button on the confirmation popup to confirm the restore.

7.1.6 Configure Targets and Alarms for One Parameter

The **Alarms / Targets** menu lets the user set up alarm and target values for the selected parameter. The user can also enable or disable the audible and LED visual alarm. Adjust the target settings by using the numbered keypad or by using the scroll buttons when a minor adjustment is needed.

- 1 Touch inside a tile to open the alarms/targets menu for that parameter. The alarms/targets menu is also available on the physio relationship screen by touching a parameter box.
- 2 To disable the audible and LED visual alarm for the parameter, touch the **Audible Alarm** icon  at the top right of the menu.

NOTE Parameters that DO NOT have the ability to set a high/low alarm will not have an **Audible Alarm** icon  on the **Alarms / Targets** menu.

The alarms limits for the Acumen Hypotension Prediction Index, HPI, are not adjustable. Target behavior and range of HPI are described in *HPI Alarm* on page 260.

- 3 To disable visual targets for the parameter, touch the **Target** enabled icon  at the top left of the menu. The target indicator for that parameter will appear gray.
- 4 Use the arrows to adjust the zone settings or touch the value button to open a numeric keypad.




Figure 7-2 Set individual parameter alarms and targets

- 5 When the values are correct, touch the enter icon .
- 6 To cancel, touch the cancel icon .

WARNING Visual and audible physiological alarms are activated only if the parameter is configured on the screens as a key parameter (1-8 parameters displayed in parameter tiles). If a parameter is not selected and displayed as a key parameter, the audible and visual physiological alarms are not triggered for that parameter.

7.2 Adjust Scales

The graphical trend data fills the graph from left to right with the most recent data at the right. The parameter scale is on the vertical axis with the time scale on the horizontal.





Figure 7-3 Graphical trend screen

The scales setup screen allows the user to set up both the parameter and time scales. The key parameters are at the top of the list. Use the horizontal scroll buttons to see additional parameters.

- 1 Touch the settings icon  → **Settings** tab .
- 2 Touch **Advanced Setup** button and enter the required password.

- 3 Touch **Parameter Settings** button → **Adjust Scales** button.




Figure 7-4 Adjust scales

NOTE The screen will return to the monitoring view after two minutes of inactivity.

- 4 For each parameter, touch the **Lower** button to enter the minimum value to appear the vertical axis. Touch the **Upper** button to enter the maximum value. Use the horizontal scroll icons   to view additional parameters.
- 5 Touch the right side of the **Graphical Trend Time** value button to set the total amount of time displayed on the graph. The options are:
- 3 minutes • 1 hour • 12 hours
 - 5 minutes • 2 hours (default) • 18 hours
 - 10 minutes • 4 hours • 24 hours
 - 15 minutes • 6 hours • 48 hours
 - 30 minutes
- 6 Touch the right side of the **Tabular Increment** value icon to set the amount of time to each tabbed value. The options are:
- 1 Minute (default) • 30 Minutes
 - 5 Minutes • 60 Minutes
 - 10 Minutes




Figure 7-5 Tabular increment popup

- 7 To advance to the next set of parameters, touch the arrow at the bottom left.
- 8 Touch the home icon  to return to the monitoring screen.

7.3 Physiology and Physio Relationship Screen SVV/PPV Parameter Settings

- 1 Touch the settings icon  → **Settings** tab .
- 2 Touch the **Advanced Setup** button and enter the required password.
- 3 Touch **Parameter Settings** button → **SVV/PPV** button.
- 4 To turn the SVV indicator **On** or **Off**, touch the **SVV: Physiology and Physio Relationship Screens** toggle button.
- 5 To turn PPV data **On** or **Off**, touch the **PPV: Physiology and Physio Relationship Screens** toggle button.

7.4 CVP Settings

CVP values can be sourced in the following ways:


- Monitored directly with a TruWave pressure transducer and HemoSphere pressure cable (see *Pressure Cable Monitoring with a TruWave DPT* on page 193).
- Sourced from an external monitoring device with an analog input (see *Analog Pressure Signal Input* on page 143).
- As a static value entered manually by the user (see *CVP Entry* on page 123)..

When none of these sources is detected or entered, the monitor will assign a default value for CVP. The monitor's configured default value is used for all patient monitoring sessions. To change this default CVP value:

- 1 Touch the settings icon  → **Settings** tab .
- 2 Touch the **Advanced Setup** button and enter the required password.
- 3 Touch **Parameter Settings** button → **CVP Settings** button.
- 4 Touch on the value button for **Default CVP Entry** to enter a CVP value (mmHg).

7.5 20-Second Flow Parameter Settings

This parameter setting automatically switches the display of 20-second flow parameters (CO_{20s} , CI_{20s} , SV_{20s} , SVI_{20s}) to the standard averaged equivalent (CO , CI , SV , and SVI) when the PA pressure signal is poor. For more information on the 20-second flow parameters, see *20-Second Flow Parameters* on page 174.

- 1 Touch the settings icon  → **Settings** tab .
- 2 Touch **Advanced Setup** button and enter the required password.
- 3 Touch **Parameter Settings** button → **20-Second Flow Settings** button.
- 4 Touch the toggle button to switch the setting to **On** or **Off**.


NOTE 20-second flow parameters are available when monitoring with the HemoSphere Swan-Ganz module and a PA (pulmonary artery) pressure signal is also monitored through a connected HemoSphere pressure cable, TruWave DPT, and CCombo V catheter (models 777F8 and 774F75). In addition, the 20-second flow parameter feature must be activated. Please contact your local Edwards representative for more information on enabling this advanced feature.

7.6 Demo Mode

Demonstration Mode is used to display simulated patient data to assist in training and demonstration.

Demonstration mode displays data from a stored set and continually loops through a predefined data set. During **Demo Mode**, the HemoSphere advanced monitoring platform user interface retains the same functionality as a fully operational platform. Simulated patient demographics must be entered to demonstrate the selected monitoring mode functions. The user can touch the controls as if a patient was being monitored.

When **Demo Mode** is entered, trended data and events are cleared from being displayed and saved for return to patient monitoring.

- 1 Touch settings icon  → **Settings** tab .
- 2 Touch the **Demo Mode** button.

NOTE When the HemoSphere advanced monitoring platform runs in **Demo Mode**, all audible alarms are disabled.

- 3 Select the demonstration monitoring mode:

Invasive: See chapter 9: *HemoSphere Swan-Ganz Module Monitoring* for details on monitoring with the HemoSphere Swan-Ganz module and **Invasive** monitoring mode.

Minimally-Invasive: See chapter 10: *Monitoring with the HemoSphere Pressure Cable* for details on monitoring with the HemoSphere pressure cable and **Minimally-Invasive** monitoring mode.

Non-Invasive: See chapter 11: *HemoSphere ClearSight Module Non-Invasive Monitoring* for details on monitoring with the HemoSphere ClearSight module and Non-Invasive monitoring mode.

NOTE Selecting Minimally-Invasive demo mode simulates using an Acumen IQ sensor when the HPI feature has been activated.

- 4 Touch **Yes** on the **Demo Mode** confirmation screen.
- 5 The HemoSphere advanced monitoring platform must be restarted prior to monitoring a patient.

WARNING Make sure that **Demo Mode** is not activated in a clinical setting to ensure that simulated data is not mistaken for clinical data.

Data Export and Connectivity Settings

Contents

Export Data	158
Wireless Settings	161
HIS Connectivity	161
Viewfinder Hub Connectivity	164
Cyber Security	165

8.1 Export Data

The **Export Data** screen lists a number of data export features of the HemoSphere advanced monitor. This screen is password protected. From this screen clinicians can export diagnostic reports, delete monitoring sessions, or export monitoring data reports. For more on exporting monitoring data reports, see below.

8.1.1 Data Download

The **Data Download** screen allows the user to export monitored patient data to a USB device in Windows Excel XML 2003 format.

NOTE The screen will return to the monitoring view after two minutes of inactivity.

- 1 Touch the settings icon  → **Settings** tab .
- 2 Touch the **Export Data** button.
- 3 Enter password when prompted in **Export Data Password** popup window. All passwords are set during system initialization. Contact your hospital administrator or IT department for password.
- 4 Ensure a USB device has been inserted.

NOTE When exceeding 4GB of data the USB storage device should not use FAT32 formatting.

CAUTION Use a virus scan on any USB stick before inserting to prevent a virus or malware infection.

- 5 Touch the **Data Download** button.

8.1.1.1 Monitoring Data

To generate a spreadsheet of monitored patient data:

- 1 Touch the value side of the Interval button and select the frequency of the data to download. The shorter the frequency, the greater the amount of data. Options are:
 - 20 seconds (default)
 - 1 minute
 - 5 minutes
- 2 Select the case to download
- 3 Touch the **Start Download** button.




Figure 8-1 Data Download screen

NOTE The Data Download Screen can only be accessed once patient monitoring has ended.

NOTE All alarms are logged and stored for the given patient and can be accessed via the **Monitoring Data** download. Alarm data logging discards older data when the log becomes full. This log also contains timestamped alarm conditions and the system power-off time.

8.1.1.2 Case Report

To generate a report of key parameters:

- 1 Touch the **Case Report** button.

- 2 Select desired parameters from the case report popup menu. A maximum of three parameters can be selected.
- 3 Check **De-identify**  to exclude patient demographic data.
- 4 Touch the enter icon  to export PDF.

8.1.1.3 GDT Report.

To generate a report of GDT tracking sessions:

- 1 Touch the **GDT Report** button.
- 2 Select desired GDT tracking session(s) from the GDT Report popup menu. Use the scroll buttons to select older tracking sessions.
- 3 Check **De-identify**  to exclude patient demographic data.
- 4 Touch the enter icon  to export PDF.

NOTE Do not disconnect the USB device until the “**Download complete. Remove the USB drive.**” message appears.

If a message appears stating that the USB device is out of space, insert a different USB device and restart the download.

All monitored patient data may be cleared by the user. Touch the **Clear All** button and confirm to clear.

8.1.2 Diagnostic Export

The capturing of all events, alerts, alarms and monitoring activity is logged if investigations or detailed troubleshooting is needed. A **Diagnostics Export** option within the **Export Data** settings menu is provided where this information can be downloaded for diagnostic purposes. This information may be requested by Edwards service personnel to help troubleshoot issues. In addition, this engineering section provides detailed software revision information of connected platform components.

- 1 Touch the settings icon  → **Settings** tab .
- 2 Touch the **Export Data** button.
- 3 Enter the **Super User** password. All passwords are set during system initialization. Contact your hospital administrator or IT department for password.
- 4 Touch **Diagnostics Export** button.
- 5 Insert an Edwards approved USB flash drive into one of the available monitor USB ports.
- 6 Allow the diagnostic export to complete as indicated on the screen.

The diagnostic data will be located in a folder labeled with the monitor serial number on the USB flash drive.

8.2 Wireless Settings

The HemoSphere advanced monitor can connect to available wireless networks. For information on connecting to a wireless network contact your local Edwards representative.

Wi-Fi connection status is indicated on the information bar by the symbols shown in table 8-1.

Table 8-1 Wi-Fi connection status

Wi-Fi Symbol	Indication
	very high signal strength
	medium signal strength
	low signal strength
	very low signal strength
	no signal strength
	no connection

8.3 HIS Connectivity

The HemoSphere advanced monitor has the ability to interface with the Hospital Information Systems (HIS) to send and receive patient demographics and physiological data. The HemoSphere advanced monitor supports Health Level 7 (HL7) messaging standard and implements Integrating Healthcare Enterprise (IHE) profiles. HL7's version 2.6 messaging standard is the most commonly used means for electronic data exchange in the clinical domain. Use a compatible interface to access this feature. The HemoSphere advanced monitor HL7 communication protocol, also referred to as HIS Connectivity, facilitates the following types of data exchanges between the HemoSphere advanced monitor and external applications and devices:



- Sending of physiological data from the HemoSphere advanced monitor to the HIS and/or medical devices
- Sending of physiological alarms and device faults from the HemoSphere advanced monitor to the HIS
- HemoSphere advanced monitor retrieval of patient data from the HIS.

HIS Connection Status should only be queried through the Monitor Settings menu after the HL7 connectivity feature has been configured and tested by the facility network administrator. If HIS Connection Status is queried while the feature setup is incomplete, the **Connection Status** Screen will remain open for 2 minutes before timing out.




Figure 8-2 HIS- Patient query screen

HIS connectivity status is indicated on the information bar by the symbols shown in Table 8-2.

Table 8-2 HIS connectivity status

HIS symbol	Indication
	Connection to all configured HIS actors are good.
	Unable to establish communication with configured HIS actors.
	Patient ID is set to "Unknown" in all outbound HIS messages.
	Intermittent errors are occurring in communications with configured HIS actors.
	Persistent errors are occurring in communications with configured HIS actors.

8.3.1 Patient Demographic Data

The HemoSphere advanced monitor with HIS Connectivity enabled, can retrieve patient demographics data from enterprise application. Once the HIS Connectivity feature is enabled, touch the **Query** button. The **Patient Query** screen allows the user to search for a patient based on name, patient ID or room and bed information. The **Patient Query** screen can be used to retrieve patient demographics data when starting a new patient or to associate the patient physiological data being monitored on the HemoSphere advanced monitor with a patient record retrieved from HIS.

NOTE Stopping an incomplete patient query may result in a connection error. If encountered, close error window and restart the query.

Once a patient is selected from the query results, patient demographics data is displayed in the **New Patient Data** screen.

To complete the query, the configured HIS must have patient gender values of either 'Male,' 'Female,' or blank. If the query exceeds the maximum duration defined in HIS configuration file, an error message will be displayed to prompt manual entry of patient data.

Figure 8-3 HIS- New patient data screen

The user can enter or edit patient height, weight, age, gender, room and bed information on this screen. The selected or updated patient data can be saved by touching the home icon . Once patient data is saved, the HemoSphere advanced monitor generates unique identifiers for the selected patient and sends out this information in outbound messages with physiological data to the enterprise applications.

8.3.2 Patient Physiological Data

The HemoSphere advanced monitor can send monitored and calculated physiological parameters in outbound messages. Outbound messages can be sent to one or more configured enterprise applications. Continuously monitored and calculated parameters with the HemoSphere advanced monitor can be sent to the enterprise application.

8.3.3 Physiological Alarms and Device Faults

The HemoSphere advanced monitor can send physiological alarms and device faults to configure HIS. Alarms and faults can be sent to one or more configured HIS. Statuses of individual alarms including change in states are sent out to the enterprise application.

For more information on how to receive access to HIS Connectivity, contact your local Edwards representative or Edwards Technical Support.

WARNING Do not use the HemoSphere advanced monitor as part of a Distributed Alarm System. The HemoSphere advanced monitor does not support remote alarm monitoring/management systems. Data is logged and transmitted for charting purposes only.

8.4 Viewfinder Hub Connectivity

The HemoSphere advanced monitor has the ability to interface with the Viewfinder hub to send a patient's monitoring data to the Viewfinder remote mobile application. The hub needs to be correctly installed and provisioned before it can be paired to the HemoSphere advanced monitor. The Viewfinder hub can be configured with EMR integration to deliver a more comprehensive account of the patient. Certain features may not be available in all regions. For questions on installation of the Viewfinder hub, contact your Edwards representative. Refer to your local Edwards representative for more information.

8.4.1 Viewfinder Hub Pairing

The HemoSphere advanced monitor needs to be paired with Viewfinder hub to enable Viewfinder hub connectivity.

- 1 Touch the settings icon  → **Settings** tab .
- 2 Touch the **Advanced Setup** button.
- 3 Enter the Secure User password when prompted in **Advanced Setup Password** popup window. All passwords are set during system initialization. Contact your hospital administrator or IT department for password.
- 4 Touch the **Connectivity** button → **Viewfinder Hub Setup** button.
- 5 Enter the **Address** and **Port** for the Viewfinder hub. Touch the next arrow .
- 6 A unique pairing code will be provided along with an approval URL. Use this code and the monitor device ID to register the monitor in the Viewfinder devices application.
- 7 After successful pairing, a green arrow and linkage symbol will be displayed on the Viewfinder hub connection screen and on the information bar . For troubleshooting on potential pairing issues, see *Viewfinder Hub Connectivity Errors* on page 344.
To unpair the monitor from the Viewfinder hub, touch the **Unpair** button.

Viewfinder hub connectivity status is indicated on the information bar by the symbols shown in table 8-3.

For help with this process, contact your Viewfinder hub technical admin or technical supervisor, or your Edwards representative.

Table 8-3 Viewfinder hub connectivity status

Info bar symbol	Connection status	Indication
	Not Paired	The HemoSphere advanced monitor is not paired with the Viewfinder hub
	Pending	Pairing of the HemoSphere advanced monitor and Viewfinder hub is pending approval on the server side
	Paired	The HemoSphere advanced monitor is successfully paired with the Viewfinder hub
	Failure	A connection failure occurred during or after attempting to pair the HemoSphere advanced monitor with Viewfinder hub. The Viewfinder hub may be unreachable.

8.4.2 Patient Data

The HemoSphere advanced monitor can send continuously monitored and calculated physiological parameters to the Viewfinder hub. This data is near real time and buffered for re-transmission after loss of connectivity. Up to 72 hours of buffered data is retained per patient.

8.4.3 Physiological Alarms and Device Faults

The HemoSphere advanced monitor sends physiological alarms and device faults to a paired Viewfinder hub. Statuses of individual alarms including change in states are sent out. All alarm and target settings are configured on the HemoSphere advanced monitor.

8.4.4 Software Upgrades

When connected to the Viewfinder hub, the HemoSphere advanced monitor can receive remote software updates. If this feature is enabled, available software updates may be displayed on the power save mode screen. See *Power Off and Power Save Mode* on page 80. For more information on this feature, contact your Edwards representative.

8.5 Cyber Security

This chapter outlines ways in which patient data can be transferred to and from the HemoSphere advanced monitor. It is important to note that any facility using the HemoSphere advanced monitor must take measures to protect the privacy of a patient's personal information in accordance with country-specific regulations, and consistent with the facility's policies for managing this information. Steps that can be taken to safeguard this information and the general security of the HemoSphere advanced monitor include:

- **Physical Access:** Limit use of the HemoSphere advanced monitor to authorized users. The HemoSphere advanced monitor has password protection for certain configuration screens. Passwords should be protected. See *Password Protection* on page 135 for more information.
- **Active Use:** Users of the monitor should take measures to limit patient data storage. Patient data should be removed from the monitor after a patient is discharged and patient monitoring has ended.
- **Network Security:** The facility must take measures to ensure the security of any shared network to which the monitor may be connected to.
- **Device Security:** Users should only use Edwards approved accessories. In addition, ensure that any connected device is free of malware.

The use of any HemoSphere advanced monitor interface outside of its intended purpose could pose cyber security risks. No HemoSphere advanced monitor connections are meant to control the operations of another device. All available interfaces are shown in *HemoSphere Advanced Monitor Connection Ports* on page 72 and specifications for these interfaces are listed in table A-5, "HemoSphere advanced monitor technical characteristics," on page 371.

8.5.1 Cybersecurity Updates

When a cybersecurity update to the HemoSphere monitor is required, Edwards will issue and provide Emergency patches to customers within 60 days after the identification of a cybersecurity incident and Cybersecurity patches within 120 days after the identification of a cybersecurity incident. All other vulnerabilities will be addressed in routine updates and communicated to customers upon request. To maintain device security, it is recommended that cybersecurity controls are implemented such as, but not limited to, internal hardening methodologies, role-based access control (RBAC), and adding the

HemoSphere monitor into a subnet dedicated to medical devices. For additional recommendations on maintaining devices security please contact your local Edwards representative or Edwards Technical Support.

8.5.2 Vulnerability Management

Vulnerability scans are performed on the monitor by Edwards on a routine basis to ensure HemoSphere monitor software remains in a secure state. If a critical and/or highly-exploitable vulnerability is discovered, customers will be directly notified by Edwards via email within 30 days and a patch will be provided as applicable. Additionally, customers can access Edwards' Product Security website at <https://www.edwards.com/devices/support/product-security> to review cybersecurity bulletins. For additional inquiries, please contact your local Edwards representative or Edwards Technical Support.

8.5.3 Cybersecurity Incident Response

If there is or has been a suspected cybersecurity incident(s) that has affected the HemoSphere monitor, please contact your local Edwards representative or Edwards Technical Support. It is recommended that an internal cybersecurity incident response plan be in place which includes – but is not limited to – an incident response policy, incident response procedures, short and long term goals for the organization, and metrics for measuring the success of the plan. Along with mitigation recommendations from Edwards, these actions should return the product to secure operability.

8.5.4 HIPAA

The Health Insurance Portability and Accountability Act of 1996 (HIPAA), introduced by the U.S. Department of Health and Human Services, outlines important standards to protect individually identifiable health information. If applicable, these standards should be followed during monitor use.

HemoSphere Swan-Ganz Module Monitoring

Contents

Connecting the HemoSphere Swan-Ganz Module.....	167
Continuous Cardiac Output.....	171
Intermittent Cardiac Output	174
EDV/RVEF Monitoring.....	180
SVR.....	184

9.1 Connecting the HemoSphere Swan-Ganz Module

The HemoSphere Swan-Ganz module is compatible with all approved Edwards Swan-Ganz pulmonary artery catheters. The HemoSphere Swan-Ganz module acquires and processes signals to and from a compatible Edwards Swan-Ganz catheter for CO, iCO and EDV/RVEF monitoring. This section provides an overview of the HemoSphere Swan-Ganz module connections. See figure 9-1.

WARNING Compliance to IEC 60601-1 is only maintained when the HemoSphere Swan-Ganz module (applied part connection, defibrillation proof) is connected to a compatible monitoring platform. Connecting external equipment or configuring the system in a way not described in these instructions will not meet this standard. Failure to use the device as instructed may increase the risk of electrical shock to the patient/operator.

Do not modify, service or alter the product in any way. Servicing, alteration or modification may affect patient/operator safety and/or product performance.




Figure 9-1 HemoSphere Swan-Ganz module connection overview

NOTE Appearance of catheters and injectate systems shown in this chapter are for example only. Actual appearance may vary depending on catheter and injectate system models.

Pulmonary artery catheters are TYPE CF defibrillation proof APPLIED PARTS. Patient cables that attach to the catheter, such as the patient CCO cable, are not intended to be applied parts but may come into contact with the patient and meet the relevant applied part requirements per IEC 60601-1.

- 1 Insert the HemoSphere Swan-Ganz module into the HemoSphere advanced monitor. The module will click when properly engaged.

CAUTION Do not force the module into the slot. Apply even pressure to slide and click the module into place.

- 2 Press the power button to turn on the HemoSphere advanced monitor and follow steps for entering patient data. See *Patient Data* on page 137. Connect the patient CCO cable to the HemoSphere Swan-Ganz module.

- 3 Connect the compatible Swan-Ganz catheter to the patient CCO cable. See table 9-1 below for available parameters and required connections.

Table 9-1 Available HemoSphere Swan-Ganz module parameters and required connections

Parameter	Required connection	See
CO	thermistor and thermal filament connection	<i>Continuous Cardiac Output</i> on page 171
CO _{20s} , CI _{20s} , SV _{20s} , SVI _{20s}	thermistor and thermal filament connection *PAP signal from HemoSphere pressure cable	<i>20-Second Flow Parameters</i> on page 174
iCO	thermistor and injectate (bath or in-line) probe	<i>Intermittent Cardiac Output</i> on page 174
EDV/RVEF (SV)	thermistor and thermal filament connection *HR analog input to HemoSphere advanced monitor	<i>EDV/RVEF Monitoring</i> on page 180
SVR	thermistor and thermal filament connection *MAP and CVP analog input to HemoSphere advanced monitor	<i>SVR</i> on page 184
CPO/CPI	MAP analog to HemoSphere advanced monitor or ART signal from pressure cable	N/A


NOTE Pulmonary artery pressure data is available with a HemoSphere pressure cable connection. See *Pressure Cable Monitoring with a TruWave DPT* on page 193 for more information.

- 4 Follow the necessary directions for monitoring. See *Continuous Cardiac Output* on page 171, *Intermittent Cardiac Output* on page 174 or *EDV/RVEF Monitoring* on page 180.

9.1.1 Patient CCO Cable Test

To test the integrity of the Edwards patient CCO cable, perform a cable integrity test. It is recommended to test the integrity of the cable as part of a troubleshooting process. This does not test the injectate temperature probe connection of the cable.

To access the patient CCO cable test window, touch the settings icon  → **Clinical Tools** tab



→ **Patient CCO Cable Test** icon .

Refer to figure 9-2 for numbered connections.




Figure 9-2 Patient CCO cable test connections

- 1 Attach the patient CCO cable to the inserted HemoSphere Swan-Ganz module ①.
- 2 Attach the patient CCO cable thermal filament connector ③ and thermistor connector ② to their mating test ports on the HemoSphere Swan-Ganz module.
- 3 Touch the **Start** button to begin the cable test. A progress bar will appear.
- 4 If the patient CCO cable fails, reconnect and perform the patient CCO cable test again. Replace the patient CCO cable if it fails the cable test repeatedly.
- 5 Touch the enter icon  when the cable has passed. Disconnect the patient cable thermal filament connector and thermistor connector from the HemoSphere Swan-Ganz module.

9.1.2 Parameter Selection Menu

Parameter categories while monitoring with a Swan-Ganz module are **Flow** (see *Continuous Cardiac Output* on page 171), **Resistance** (see *SVR* on page 184), and **RV Function** (*EDV/RVEF Monitoring* on page 180). **Oximetry** is also available if an oximetry cable or tissue oximetry module is connected (see *Venous Oximetry Monitoring* on page 223). Touch parameter buttons that display an arrow () to view additional monitoring

options for that parameter based on the display update rate and averaging time. See *STAT CO* on page 174, *STAT EDV and RVEF* on page 184, and *20-Second Flow Parameters* on page 174. Touch the blue arrow () to see definitions of these monitoring options or the help icon () for more information.




Figure 9-3 HemoSphere Swan-Ganz module key parameter selection window

9.2 Continuous Cardiac Output

The HemoSphere advanced monitor measures cardiac output continuously by introducing small pulses of energy into the blood stream and measuring blood temperature via a pulmonary artery catheter. The maximum surface temperature of the thermal filament used to release these pulses of energy within the blood is 48 °C. Cardiac output is computed using proven algorithms derived from the conservation of heat principles, and indicator dilution curves that are obtained by cross-correlation of energy input and blood temperature waveforms. After initialization, the HemoSphere advanced monitor continuously measures and displays the cardiac output in liters per minute without operator calibration or intervention.

9.2.1 Connecting the Patient Cables

- 1 Connect the patient CCO cable to the inserted HemoSphere Swan-Ganz module as previously described in section 9.1.
- 2 Attach the catheter end of the patient cable to the thermistor and thermal filament connectors on the Swan-Ganz CCO catheter. These connections are emphasized as numbers ② and ③ in figure 9-4 on page 172.

- 3 Verify that the CCO catheter is properly inserted into the patient.




Figure 9-4 CO connection overview

9.2.2 Initiating Monitoring

WARNING CO monitoring should always be discontinued when blood flow around the thermal filament is stopped. Clinical situations where CO monitoring should be discontinued include, but are not limited to:

- Time periods when a patient is on cardiopulmonary bypass
- Partial withdrawal of the catheter so that the thermistor is not in the pulmonary artery
- Removal of the catheter from the patient

When the system is properly connected, touch the start monitoring icon  to begin CO monitoring. The CO countdown timer will appear on the stop monitoring icon. After approximately 5 to 12 minutes, when sufficient data has been obtained, a CO value will appear in the parameter tile. The CO value displayed on the screen will be updated approximately every 60 seconds.

NOTE No CO value will be displayed until sufficient time-averaged data is available.

9.2.3 Thermal Signal Conditions

In some situations where patient conditions create large changes in pulmonary artery blood temperature over several minutes, the monitor may take longer than 6 minutes to obtain an initial CO measurement. When CO monitoring is in progress, updating of the CO measurement may also be delayed by unstable pulmonary artery blood temperature. The last CO value and measurement time will be displayed in place of an updated CO value. Table 9-2 shows the alert/fault messages that appear on the screen at different time points while the signal stabilizes. Refer to table 15-10, “HemoSphere Swan-Ganz module CO faults/alerts,” on page 344 for more information on CO faults and alerts.

Table 9-2 Unstable thermal signal time lapse for CO alert and fault messages

Condition	Notification	Alert CO		Fault CO
	Cardiac Output calculation in progress	Signal Adapting — Continuing	Unstable Blood Temp — Continuing	Thermal Signal Loss
Monitoring Commencing: time from commencement without CO measurement	3 ½ minutes	6 minutes	15 minutes	30 minutes
Monitoring in Progress: time from last CO update	5 seconds from expiry of CO countdown timer	na	6 minutes	20 minutes

A fault condition terminates monitoring. A fault condition could result from migration of the catheter tip into a small vessel preventing the thermistor from accurately sensing the thermal signal. Check catheter position and reposition the catheter, if necessary. After verifying patient status and catheter position, CO monitoring may be resumed by touching the start monitoring icon .

CAUTION

Inaccurate cardiac output measurements may be caused by:

- Incorrect placement or position of the catheter
- Excessive variations in pulmonary artery blood temperature. Some examples that cause BT variations include, but are not limited to:
 - * status post cardiopulmonary bypass surgery
 - * centrally administered cooled or warmed solutions of blood products
 - * use of sequential compression devices
- Clot formation on the thermistor
- Anatomical abnormalities (for example, cardiac shunts)
- Excessive patient movement
- Electrocautery or electrosurgical unit interference
- Rapid changes in cardiac output

9.2.4 CO Countdown Timer

The CO countdown timer is located on the stop monitoring icon . This timer alerts the user as to when the next CO measurement will take place. The time to the next CO measurement varies from 60 seconds to 3 minutes or longer. A hemodynamically unstable thermal signal may delay CO calculations.

9.2.5 STAT CO

For longer time spans between CO measurements, the STAT CO is available. The STAT CO (sCO) is a fast estimate of the CO value and is updated every 60 seconds. Select sCO as a key parameter to view STAT CO values. Select CO and sCO as key parameters while viewing the graphical/tabular trends split screen and CO monitored data is graphically plotted alongside tabular/numerical data for STAT values of sCO. See *Graphical/Tabular Trends Split* on page 112.

9.2.6 20-Second Flow Parameters

The 20-second flow parameters are available when monitoring with the HemoSphere Swan-Ganz module and a PA (pulmonary artery) pressure signal is also monitored through a connected HemoSphere pressure cable, TruWave DPT, and CCombo V catheter (models 777F8 and 774F75). A pulse contour analysis of the pulmonary artery pressure signal is used in combination with the CCO thermodilution algorithm to obtain a faster parameter calculation for CO, CI, SV and SVI. The 20-second flow parameters are labeled with “20s” (CO_{20s}, CI_{20s}, SV_{20s}, SVI_{20s}). These parameters are only available if the 20s flow parameter feature is enabled. Please contact your local Edwards representative for more information on enabling this advanced feature. For more information on PA monitoring, see *Pressure Cable Monitoring with a TruWave DPT* on page 193.

CAUTION Inaccurate 20-second flow parameter measurements may be caused by:

- Incorrect placement or position of the catheter
- Improperly zeroed and/or leveled transducer
- Over- or under-damped pressure line
- Adjustments to the PAP line made after start of monitoring

9.2.6.1 PAP Waveform Troubleshooting

The calculation of 20-second flow parameters is highly dependent on a good pulmonary artery pressure waveform. Use the **Zero & Waveform**  screen to view and evaluate the PAP waveform. The features of a good waveform include:

- Dicrotic notch with minimal dip between systole and diastole
- Clean signal without noise or high-frequency artifacts
- Minimal “whip” artifacts caused by catheter tip movement in the right ventricle
- Sharp waveform morphology and minimal over-damping due to bubbles or kinking in tubing

PAP waveforms that do not display the above listed features have not been validated. These waveforms may result in a loss of 20-second flow parameter calculation.

9.3 Intermittent Cardiac Output

The HemoSphere Swan-Ganz module measures cardiac output intermittently using the bolus thermodilution technique. With this technique, a small amount of sterile physiological solution (e.g., saline or dextrose) at a known volume and temperature — cooler than blood temperature — is injected through the catheter injectate port, and the resultant decrease in blood temperature is measured by the thermistor in the pulmonary artery (PA). Up to six bolus injections can be completed in one series. The average value of

the injections in the series is displayed. The results of any series may be reviewed, and the user can remove individual iCO (bolus) measurements that may have been compromised (e.g., patient movement, diathermia, or operator error).

9.3.1 Connecting Patient Cables

- 1 Connect the patient CCO cable to the inserted HemoSphere Swan-Ganz module as previously described in section 9.1.
- 2 Attach the catheter end of the patient CCO cable to the thermistor connector on the Swan-Ganz or Swan-Ganz Jr iCO catheter as shown by ② in figure 9-5.
- 3 Verify that the catheter is properly inserted into the patient.




Figure 9-5 iCO connection overview

9.3.1.1 Probe Selection


An injectate temperature probe senses injectate temperature. The selected probe is connected to the patient CCO cable (figure 9-5). Either of two probes may be used:

- An in-line probe is connected to the flow-thru housing on the CO-Set/CO-Set+ injectate delivery system.
- A bath probe measures the temperature of the injectate solution. Bath probes are intended to measure the temperature of a sample solution that is kept at the same temperature as the sterile solution used for injectate when calculating bolus cardiac output.

Connect the injectate temperature probe (in-line or bath) to the injectate temperature probe connector on the patient CCO cable illustrated by ③ in figure 9-5.

9.3.2 Configuration Settings

The HemoSphere advanced monitor provides the operator with the choice of entering a specific computation constant, or configuring the HemoSphere Swan-Ganz module to allow it to automatically determine the computation constant by selecting the injectate volume and catheter size. The operator can also select the parameter display type and bolus mode.

Touch settings icon  → **Clinical Tools** tab  → **iCO** icon .




Figure 9-6 iCO new set configuration screen

CAUTION Refer to Appendix E to ensure computation constant is the same as specified in the catheter package insert. If the computation constant differs, enter the desired computation constant manually.

NOTE The HemoSphere Swan-Ganz module will automatically sense the type of temperature probe in use (ice bath or in-line). The module will use this information to determine the computation constant.

If an injectate temperature (IT) probe is not detected by the monitor, the message “**Fault: iCO – Check Thermistor Connection**” is displayed.

9.3.2.1 Select Injectate Volume

Select a value from the **Injectate Volume** list button. The available choices are:

- 10 mL
- 5 mL
- 3 mL (bath type probe only)

When a value is chosen, the computation constant is automatically set.

9.3.2.2 **Select Catheter Size**

Select a catheter size from the **Catheter Size** list button. The available choices are:

- **5.5F**
- **6F**
- **7F**
- **7.5F**
- **8F**

When a value is chosen, the computation constant is automatically set.

9.3.2.3 **Select Computation Constant**

To manually enter a computation constant, touch the **Comp Constant** value button and enter a value on the keypad. If a computation constant is manually entered, injectate volume and catheter size are automatically set, and value entry is set to **Auto**.

9.3.2.4 **Select Mode**

Select **Auto** or **Manual** from the **Mode** list button. The default mode is **Auto**. In the **Auto** mode, the HemoSphere advanced monitor automatically highlights an **Inject** message upon achieving a baseline blood temperature. The **Manual** mode operation is similar to the **Automatic** mode except that the user must touch the **Inject** button prior to each injection. The following section provides instructions for both of these bolus modes.

9.3.3 **Instructions for Bolus Measurement Modes**

The HemoSphere Swan-Ganz module factory default setting for bolus measurement is **Auto** mode. In this mode, the HemoSphere advanced monitor highlights an **Inject** message upon achieving a baseline blood temperature. During **Manual** mode, the operator will initiate when to inject by touching the **Inject** button. When an injection is complete, the module computes a value and is ready to process another bolus injection. Up to six bolus injections can be completed in one series.

The following provides step-by-step instructions for performing bolus cardiac measurements starting from the iCO new set configuration screen.

- 1 Touch the **Start Set** button at the bottom of the iCO new set configuration screen after selecting thermodilution configuration settings.

The button is disabled if:

- The injectate volume is invalid or not selected
- Injectate temperature (IT) is not connected
- Blood temperature (BT) is not connected
- An iCO fault is active

If continuous CO measurements are active, a popup window will appear to confirm the suspension of CO monitoring. Touch the **Yes** button.

NOTE During bolus CO measurements, any parameters calculated using an ECG input signal (HR_{avg}) are unavailable.

- 2 The iCO new set screen appears with **Wait** highlighted ().
- 3 When in auto mode and the thermal baseline is established **Inject** becomes highlighted on the screen (), signifying when to begin the bolus injection series.

OR


If in manual mode, **Ready** () will appear highlighted on the screen when the thermal baseline is established. Touch the **Inject** button when ready to inject and then **Inject** becomes highlighted on the screen.

- 4 Use a rapid, smooth, continuous method to inject the bolus with the volume amount previously selected.

CAUTION Sudden changes in PA blood temperature, such as those caused by patient movement or bolus drug administration, may cause an iCO or iCI value to be computed. To avoid falsely triggered curves, inject as soon as possible after the **Inject** message appears.

Once a bolus is injected, the thermodilution washout curve appears on the screen, **Computing** is highlighted () and the resultant iCO measurement is displayed.

- 5 When the thermal washout curve is complete the HemoSphere advanced monitor will highlight **Wait** and then **Inject** – or **Ready** during manual mode – when a stable thermal baseline is reached again. Repeat steps 2 through 4 up to six times as desired. The highlighted messages are repeated as follows:



NOTE When the bolus mode is set to **Auto**, the maximum time allowed between the appearance of the **Inject** message and injection of the bolus is four minutes. If no injection is detected within this time interval, the **Inject** message will disappear and the **Wait** message will reappear.

While in **Manual** bolus mode, the operator has a maximum of 30 seconds in which to make a bolus injection after touching the **Inject** button. If no injection is detected within the time interval, the **Inject** button is enabled again and the Inject message disappears.

If a bolus measurement is compromised, as indicated by an alert message, an  will

appear in place of the CO/CI value displayed on screen.

To discontinue iCO (bolus) measurements, touch the cancel icon .

- 6 After the desired number of bolus injections has been performed, review the set of washout curves by touching the **Review** button.
- 7 Remove any of the six injections in the set by touching on it on the review screen.



A red “X” appears over the waveform removing it from the averaged CO/CI value.

Waveforms that are irregular or questionable will have an  next to the waveform data set.

If desired, touch the cancel icon  to delete the bolus set. Touch the **Yes** button to confirm.

- 8 Touch the **Accept** button after completing the review of bolus injections to use the averaged CO/CI value or touch the return icon  to resume the series and add additional bolus injections (up to six) for averaging.

9.3.4 Thermodilution Summary Screen

After the set has been accepted, the set summary will be displayed as a time stamped tab on the thermodilution summary screen. This screen can be accessed anytime by touching the historical thermodilution icon  from certain monitoring screens or by touching

the settings icon  → **Clinical Tools** tab  → **iCO** icon .

The following actions are available to the operator on the thermodilution summary screen:




Figure 9-7 Thermodilution summary screen

New Set. Touch the return icon  or the **New** tab to perform another thermodilution set. The previous CO/CI average value and associated washout curves will be saved as a tab in the thermodilution summary screen.

Review. Review the thermal washout curves from the bolus set. Touch any tab to review the thermal washout curves from other bolus sets.

CO Monitoring. If the system is properly connected for continuous CO monitoring, touch the start monitoring icon  to begin CO monitoring at any time.

9.4 EDV/RVEF Monitoring

Right ventricular end diastolic volume (EDV) monitoring is available in conjunction with CO monitoring mode when using a Swan-Ganz CCombo V catheter and ECG signal input. Pulse rate (PR) from a pressure cable monitored arterial waveform can be used in place of an ECG heart rate (HR) signal, if available. During EDV monitoring, the HemoSphere advanced monitor continuously displays EDV and right ventricular ejection fraction (RVEF) measurements. EDV and RVEF are time-averaged values that can be numerically displayed in parameter tiles, and graphically trended over time in the graphical trend view.

In addition, estimates of EDV and RVEF values at approximately 60 second intervals are calculated and displayed by selecting sEDV and sRVEF as key parameters.

9.4.1 Connecting Patient Cables

- 1 Connect the patient CCO cable to the inserted HemoSphere Swan-Ganz module as previously described in section 9.1.
- 2 Attach the catheter end of the patient cable to the thermistor and thermal filament connectors on the Swan-Ganz CCombo V catheter. These connections are emphasized by ② and ③ in figure 9-8.

- 3 Verify that the catheter is properly inserted into the patient.




Figure 9-8 EDV/RVEF connection overview

9.4.2 Connecting the ECG Interface Cable

Connect the ECG interface cable's 1/4 inch miniature phone plug to the ECG monitor input on the rear panel of the HemoSphere advanced monitor.



Connect the other end of the interface cable to the bedside monitor's ECG signal output. This will provide an average heart rate (HR_{avg}) measure to the HemoSphere advanced monitor for EDV and RVEF measurements. Pulse rate (PR) from a pressure cable monitored arterial waveform can be used in place of an ECG heart rate (HR) signal, if available. For compatible ECG cables, contact your local Edwards representative.

IMPORTANT NOTE

The HemoSphere advanced monitor is compatible with an ECG analog input from any external patient monitor that has an analog output port which meets the ECG signal input specifications identified in appendix A, table A-5 of this operator's manual. The ECG signal is used to derive heart rate which is then used to calculate additional hemodynamic parameters for display. This is an optional feature that does not impact the HemoSphere advanced monitor's primary function of monitoring cardiac output

(with the HemoSphere Swan-Ganz module) and venous oxygen saturation (with the HemoSphere oximetry cable). Device performance testing was conducted using ECG input signals.

WARNING **PACEMAKER PATIENTS** – Rate meters may continue to count the pacemaker rate during occurrences of cardiac arrest or some arrhythmias. Do not rely entirely upon displayed heart rate. Keep pacemaker patients under close surveillance. See table A-5 on page 371 for disclosure of the pacemaker pulse rejection capability of this instrument.

For patients requiring internal or external pacing support, the HemoSphere advanced monitoring platform should not be used to obtain heart rate and heart rate derived parameters under the following conditions:

- pacer pulse synch output from bedside monitor includes the pacer pulse, however, the characteristics are outside of the pacemaker pulse rejection capabilities specifications as listed in table A-5
- pacer pulse synch output characteristics from bedside monitor cannot be determined

Note any discrepancies in heart rate (HR_{avg}) with the patient monitor HR and ECG waveform display when interpreting derived parameters such as SV, EDV, RVEF, and associated index parameters.

ECG signal input and all parameters derived from heart rate measurements have not been evaluated for pediatric patients and are therefore not available for that patient population.

NOTE When an ECG input connection or disconnection is first detected, a brief notification message will be displayed on the status bar.

SV is available with any compatible Swan-Ganz catheter and an ECG signal input. For EDV/RVEF monitoring, a Swan-Ganz CCombo V catheter is required.

9.4.3 Initiating Measurement

WARNING CO monitoring should always be discontinued when blood flow around the thermal filament is stopped. Clinical situations where CO monitoring should be discontinued include, but are not limited to:

- Time periods when a patient is on cardiopulmonary bypass
- Partial withdrawal of the catheter so that the thermistor is not in the pulmonary artery
- Removal of the catheter from the patient

When the system is properly connected, touch the start monitoring icon  to begin CO

monitoring. The CO countdown timer will appear on the stop monitoring icon. After approximately 5 to 12 minutes, when sufficient data has been obtained, an EDV and/or RVEF value will appear in the configured parameter tiles. The EDV and RVEF values displayed on the screen will be updated approximately every 60 seconds.

NOTE No EDV or RVEF value will be displayed until sufficient time-averaged data is available.

In some situations where patient conditions create large changes in pulmonary artery blood temperature over several minutes, the monitor may take longer than 9 minutes to obtain an initial EDV or RVEF measurement. In these cases, the following alert message will appear 9 minutes after monitoring has commenced:

Alert: EDV – Signal Adapting — Continuing

The monitor will continue to function and no user action is required. When continuous EDV and RVEF measurements are obtained, the alert message will be removed and the current values will be displayed and plotted.

NOTE CO values may still be available even when EDV and RVEF are not.

9.4.4 Active EDV Monitoring

When EDV monitoring is in progress, updating of the continuous EDV and RVEF measurement may be delayed by unstable pulmonary artery blood temperature. If the values are not updated for 8 minutes, the following message will appear:

Alert: EDV – Signal Adapting — Continuing

In cases when the average heart rate goes out of range (i.e., less than 30 bpm or greater than 200 bpm) or when no heart rate is detected, the following message will appear:

Alert: EDV – Heart Rate Signal Missing

Continuous EDV and RVEF monitoring values will no longer be displayed. This condition could result from physiologic changes in the patient's status or the loss of the ECG analog signal. Check the ECG interface cable connections and reconnect if necessary. After verifying patient status and cable connections, EDV and RVEF monitoring will automatically be resumed.

NOTE SV, EDV, and RVEF values are dependent on accurate heart rate calculations. Care should be taken that accurate heart rate values are being displayed, and that double counting should be avoided, especially in case of AV pacing.

If the patient has an atrial or atrial-ventricular (AV) pacer, the user should assess for the presence of double sensing (for accurate HR determinations, only one pacer spike or one contraction per cardiac cycle should be sensed). In the event of double sensing, the user should:

- Reposition the reference lead to minimize atrial spike sensing

- Select appropriate lead configuration to maximize HR triggers and minimize atrial spike sensing, and
- Assess appropriateness of milliamperage (mA) pacing levels.

The accuracy of continuous EDV and RVEF determinations is dependent upon a consistent ECG signal from the bedside monitor. For additional troubleshooting, see table 15-11, “HemoSphere Swan-Ganz module EDV and SV faults/alerts,” on page 346 and table 15-15, “HemoSphere Swan-Ganz module general troubleshooting,” on page 349.

If EDV monitoring is stopped, by touching the stop monitoring icon , the parameter tile target indicator for EDV and/or RVEF will become gray, and a time stamp will be placed below the value indicating the time that the last value was measured.

NOTE Pressing the stop monitoring icon  will stop EDV, RVEF and CO monitoring.

If EDV monitoring is resumed, a gap will appear in the plotted line of the trend graph indicating the time period when continuous monitoring was interrupted.

9.4.5 STAT EDV and RVEF

A hemodynamically unstable thermal signal may delay the HemoSphere advanced monitor from displaying an EDV, EDVI and/or RVEF value after monitoring has been initiated. The clinician may use the STAT values, which presents estimates of EDV or EDVI, and RVEF values updated approximately 60 seconds. Select sEDV, sEDVI, or sRVEF as a key parameter to view STAT values. EDV, EDVI, and RVEF values can be graphically trended over time alongside numerical values of sEDV, sEDVI, and sRVEF using the graphical/tabular trends split screen monitoring view. Up to two parameters can be viewed in tabular format on this screen. See *Graphical/Tabular Trends Split* on page 112.

9.5 SVR

While performing CO monitoring, the HemoSphere advanced monitor can also calculate SVR by utilizing MAP and CVP analog pressure signal inputs from a connected patient monitor. See *Analog Pressure Signal Input* on page 143. See *CVP Entry* on page 123 for additional CVP sources and system prioritization.

9.6 Multiple Technology Monitoring - Acumen Hypotension Prediction Index Software

To display Acumen Hypotension Prediction Index software parameters in Invasive monitoring mode, connect a pressure cable and Acumen IQ sensor. With an Acumen IQ sensor, an additional five key parameters can be displayed: stroke volume variation (SVV), dynamic arterial elastance ($E_{a_{dyn}}$), systolic slope (dP/dt), pulse pressure variation (PPV), and Acumen Hypotension Prediction Index (HPI). These five parameters are labeled as “Acumen IQ” parameters and can be configured on any monitor screen. Blood pressure parameters monitored with a pressure cable in Invasive mode are always selectable as key parameters. Similarly, Acumen IQ sensor arterial pressure parameters are selectable as key parameters. See table 9-3 on page 185 for availability of Acumen IQ sensor parameters in Invasive mode.

Table 9-3 Acumen IQ sensor parameter availability in Invasive mode

Minimally-Invasive mode monitored parameter	Invasive mode viewing status
SVV*	selectable as key parameter
Ea _{dyn} *	
dP/dt*	
HPI*	
DIA _{ART}	
SYS _{ART}	
MAP	
PR	
PPV*	
CO	not available
CI	
SV	
SVI	
*Acumen IQ parameter only	

- 1 Connect HemoSphere pressure cable and Acumen IQ sensor. Follow instructions outlined in *FloTrac Sensor Monitoring* on page 190.
- 2 After successfully zeroing the Acumen IQ sensor, the HPI value will appear on the information bar. **HPI 40 / 100**
- 3 Touch inside of parameter tile to select the desired HPI parameter from the parameter tile configuration. For more information, see *Change Parameters* on page 120.

**Figure 9-9 Main monitoring view – Swan-Ganz module monitoring with an Acumen IQ sensor**

NOTE	Smart alerts and smart trends are not available while in Swan-Ganz module (Invasive) monitoring mode. If HPI parameter alarms, the HPI parameter high alert popup will appear. See figure 14-4 on page 262. Touching the Review button will take you to the Acumen IQ parameter dashboard.

Monitoring with the HemoSphere Pressure Cable

Contents

Pressure Cable Overview	187
Monitoring Mode Selection	190
FloTrac Sensor Monitoring	190
Pressure Cable Monitoring with a TruWave DPT	193
Zero & Waveform Screen	196

10.1 Pressure Cable Overview

The HemoSphere pressure cable is a reusable device that connects with the HemoSphere monitor on one end ④ and any approved single Edwards disposable pressure transducer (DPT) or sensor on the other end ①. See figure 10-1 on page 188. The HemoSphere pressure cable acquires and processes a single pressure signal from a compatible Edwards DPT, such as the TruWave DPT, or a FloTrac sensor. A FloTrac, FloTrac Jr, or Acumen IQ sensor connects to an existing arterial catheter to provide minimally-invasive hemodynamic parameters. A TruWave transducer can connect to any compatible pressure monitoring catheter to provide location based intravascular pressure. Refer to the directions for use provided with each catheter for specific instructions on catheter placement and use, and for relevant warnings, cautions and notes. The HemoSphere pressure cable can be monitored through two technology monitoring modes based on the paired sensor/transducer: FloTrac, FloTrac Jr, or Acumen IQ sensor monitoring mode or Swan-Ganz catheter monitoring mode. The monitoring mode appears on the top of the navigation bar (see figure 5-2 on page 98). The appearance and connection points for the HemoSphere pressure cable are shown in figure 10-1.

Pressure Type Color Insert. If desired, the appropriate color insert can be used on the pressure cable to indicate the monitored pressure type. See ③ in figure 10-1 below. The colors are as follows:

- Red for arterial pressure (AP)
- Blue for central venous pressure (CVP)
- Yellow for pulmonary artery pressure (PAP)
- Green for cardiac output (CO)




Figure 10-1 HemoSphere pressure cable

Table 10-1 HemoSphere pressure cable configurations and available key parameters

Available key parameters	Pressure cable configuration					
	FloTrac/ FloTrac Jr/ Acumen IQ sensor	FloTrac/ FloTrac Jr/ Acumen IQ sensor with CVP entry or analog input CVP signal	FloTrac/ Acumen IQ sensor with CVP entry or analog input CVP signal and oximetry cable	TruWave transducer connected to arterial line	TruWave transducer connected to central line	TruWave transducer connected to pulmonary artery catheter
CO/CI	•	•	•			
CPO/CPI	•	•	•			
SV/SVI	•	•	•			
SVV/PPV	•	•	•			
SVR/SVRI		•	•			
SvO ₂ /ScvO ₂			•			
PR	•	•	•	•		
SYS _{ART}	•	•	•	•		
DIA _{ART}	•	•	•	•		
MAP	•	•	•	•		
MPAP						•
SYS _{PAP}						•
DIA _{PAP}						•
CVP		•	•		•	
HPI*	•	•	•			

Table 10-1 HemoSphere pressure cable configurations and available key parameters (continued)

Available key parameters	Pressure cable configuration					
	FloTrac/ FloTrac Jr/ Acumen IQ sensor	FloTrac/ FloTrac Jr/ Acumen IQ sensor with CVP entry or analog input CVP signal	FloTrac/ Acumen IQ sensor with CVP entry or analog input CVP signal and oximetry cable	TruWave transducer connected to arterial line	TruWave transducer connected to central line	TruWave transducer connected to pulmonary artery catheter
dP/dt*	•	•	•			
Ea _{dyn} *	•	•	•			

***NOTE** The Acumen Hypotension Prediction Index parameter, HPI, is an advanced feature that must be activated using a Acumen IQ sensor connected to a radial arterial catheter. See *Acumen Hypotension Prediction Index (HPI) Software Feature* on page 253 for more information.

WARNING Do not resterilize or reuse any FloTrac sensor, FloTrac Jr sensor, Acumen IQ sensor, TruWave transducer, or catheter; refer to the catheter's "directions for use".

Do not use a FloTrac sensor, FloTrac Jr sensor, Acumen IQ sensor, TruWave transducer, or catheter that is wet, damaged, or that has exposed electrical contacts.

Do not modify, service or alter the product in any way. Servicing, alteration or modification may affect patient/operator safety and/or product performance.

Refer to the directions provided with each accessory for specific instructions on placement and use, and for relevant WARNINGS, CAUTIONS, and specifications.

When the pressure cable is not in use, protect the exposed cable connector from fluid. Moisture within the connector may result in the cable malfunctioning or in inaccurate pressure readings.

Compliance to IEC 60601-1 is only maintained when the HemoSphere pressure cable (applied part accessory, defibrillation proof) is connected to a compatible monitoring platform. Connecting external equipment or configuring the system in a way not described in these instructions will not meet this standard. Failure to use the device as instructed may increase the risk of electrical shock to the patient/operator.

CAUTION Do not use any FloTrac sensor, FloTrac Jr sensor, Acumen IQ sensor, or TruWave transducer past its labeled "Use By Date." Products used beyond this date may have compromised transducer or tubing performance, or compromised sterility.

Excessive dropping of the HemoSphere pressure cable may result in cable damage and/or malfunction.

10.2 Monitoring Mode Selection

The primary monitoring mode for the HemoSphere pressure cable is minimally-invasive monitoring mode with a connected FloTrac or Acumen IQ sensor. The pressure cable can also be used to collect intravascular pressure data (CVP and/or PAP) while in any monitoring mode using a connected TruWave pressure transducer. See *Select Monitoring Mode* on page 122 for more information on switching between monitoring modes.

10.3 FloTrac Sensor Monitoring

The HemoSphere pressure cable serves as an Edwards FloTrac sensor connecting cable for the HemoSphere advanced monitoring platform. The HemoSphere pressure cable with a connected FloTrac, FloTrac Jr, or Acumen IQ sensor uses the patient's existing arterial pressure waveform to continuously measure cardiac output (FloTrac arterial pressure autocalibrated cardiac output [FT-CO]). With the input of patient height, weight, age, and gender, a specific vascular compliance is determined. The FloTrac algorithm's automatic vascular tone adjustment recognizes and adjusts for changes in vascular resistance and compliance. Cardiac output is displayed on a continuous basis by multiplying the pulse rate and calculated stroke volume as determined from the pressure waveform. The FloTrac, FloTrac Jr, or Acumen IQ sensor measures variations of arterial pressure proportional to stroke volume.

The HemoSphere pressure cable and FloTrac, FloTrac Jr, or Acumen IQ sensor, use the patient's existing arterial pressure waveform to continuously measure stroke volume variation (SVV). SVV is a sensitive indicator of the patient's preload responsiveness when the patient is 100% mechanically ventilated with a fixed rate and tidal volume, and no spontaneous breaths. SVV is always used best in conjunction with stroke volume or cardiac output assessment.

When using the Acumen IQ sensor, the patient's existing arterial pressure waveform is used to continuously measure systolic slope (dp/dt) and dynamic arterial elastance (Ea_{dyn}). Ea_{dyn} is a measure of the afterload to the left ventricle by the arterial system (arterial elastance) relative to left ventricular elastance (dynamic arterial elastance). See *Acumen Hypotension Prediction Index (HPI) Software Feature* on page 253 for more information on the Acumen IQ sensor and the Acumen Hypotension Prediction Index (HPI) feature. Activation of the Acumen HPI feature is only available in certain areas. Please contact your local Edwards representative for more information on enabling this advanced feature.

Available parameters using FloTrac technology include cardiac output (CO), cardiac index (CI), cardiac power output (CPO), cardiac power index (CPI), stroke volume (SV), stroke volume index (SVI), stroke volume variation (SVV), systolic pressure (SYS), diastolic pressure (DIA), mean arterial pressure (MAP), and pulse rate (PR). When using a Acumen IQ sensor and the Acumen HPI feature is activated, additional available parameters include dynamic arterial elastance (Ea_{dyn}), systolic slope (dp/dt), pulse pressure variation (PPV), and Acumen Hypotension Prediction Index parameter (HPI). When the FloTrac, FloTrac Jr, or Acumen IQ sensor is paired with the patient's central venous pressure (CVP), systemic vascular resistance (SVR) and systemic vascular resistance index (SVRI) are also available.

CAUTION

Inaccurate FT-CO measurements can be caused by factors such as:

- Improperly zeroed and/or leveled sensor/transducer
- Over- or under-damped pressure lines
- Excessive variations in blood pressure. Some conditions that cause BP variations include, but are not limited to:
 - * Intra-aortic balloon pumps
- Any clinical situation where the arterial pressure is deemed inaccurate or not representative of aortic pressure, including but not limited to:

* Extreme peripheral vasoconstriction which results in a compromised radial arterial pressure waveform

* Hyperdynamic conditions as seen in post liver transplant

- Excessive patient movement
- Electrocautery or electrosurgical unit interference

Aortic valve regurgitation may cause an over estimation of Stroke Volume / Cardiac Output calculated depending on the amount of valvular disease and the volume lost back into the left ventricle.

10.3.1 Connect FloTrac, FloTrac Jr, or Acumen IQ Sensor

- 1 Connect one end of the pressure cable to the HemoSphere advanced monitor.
- 2 To de-air and prime I.V. bag and FloTrac, FloTrac Jr, or Acumen IQ sensor: Invert normal saline I.V. bag (anticoagulation per institution policy). Spike I.V. bag with fluid administration set, keeping drip chamber upright. While keeping I.V. bag inverted, gently squeeze air out of bag with one hand while pulling flush tab (Snap-Tab) with the other hand until air is emptied from I.V. bag and drip chamber is filled half-way.
- 3 Insert I.V. bag into the Pressure Bag and hang on I.V. pole (DO NOT INFLATE).
- 4 With gravity only (no pressure in Pressure Bag), flush FloTrac/FloTrac Jr sensor holding pressure tubing in upright position as the column of fluid raises through the tubing, pushing air out of the pressure tubing until the fluid reaches the end of the tubing.
- 5 Pressurize the Pressure Bag until it reaches 300 mmHg.
- 6 Fast-flush the FloTrac/FloTrac Jr sensor and tap on tubing and stopcocks to remove any residual bubbles.
- 7 Use a straight in or out motion to connect the green connector of the primed FloTrac/FloTrac Jr sensor. The pressure cable LED that surrounds the zero button (see ② in figure 10-1) will flash green indicating that the pressure sensor is detected. A yellow light indicates a fault condition. If this occurs refer to the status bar for specific fault condition details.
- 8 Connect tubing to arterial catheter, then aspirate and flush system to assure no residual bubbles remain.
- 9 Use routine transducer calibration procedures (according to institutional policy) to ensure proper pressure signals are being transmitted. Refer to the FloTrac, FloTrac Jr, or Acumen IQ sensor's instructions for use.
- 10 Follow steps for entering patient data. See *Patient Data* on page 137.
- 11 Follow the instructions below for zeroing the FloTrac, FloTrac Jr, or Acumen IQ sensor.

CAUTION Always grasp the connector, not the cable, when connecting or disconnecting the cable.

Do not twist or bend the connectors.

10.3.2 Set Averaging Time

- 1 Touch within a parameter tile to access the tile configuration menu.
- 2 Touch the **Intervals / Averaging** tab

- 3 Touch the **CO/Pressure Averaging Time** value button and select one of the following interval options:
 - 5 sec
 - 20 sec (default and recommended time interval)
 - 5 min

For more information on **CO/Pressure Averaging Time** menu choices, see *Time Intervals / Averaging* on page 141.

- 4 Touch the return icon .

10.3.3 Zero Arterial Pressure

The FloTrac, FloTrac Jr, or Acumen IQ sensor must be zeroed to atmospheric pressure to ensure accurate monitoring.

- 1 Touch the Zero & Waveform icon  located on the navigation bar or through the Clinical Tools menu.

OR

Press the physical zero button  directly on the pressure cable and hold for three seconds (see figure 10-1).

CAUTION To prevent cable damage, do not apply excessive force to the pressure cable zero button.

- 2 The current arterial pressure waveform is displayed and continually updated on the screen. This is to confirm the zero operation is successful.
- 3 Select **ART** (arterial) next to the listed port for which the active pressure cable is connected. Up to two pressure cables can be connected at once.
- 4 Make sure the sensor is leveled to the patient's phlebostatic axis position according to the instructions for use.

NOTE It is important to keep the FloTrac, FloTrac Jr, or Acumen IQ sensor level to the phlebostatic axis at all times to ensure accuracy of cardiac output.

- 5 Open the FloTrac/FloTrac Jr sensor stopcock valve to measure atmospheric air. The pressure should display as a flat line.
- 6 Press the physical zero button  directly on the pressure cable and hold for three seconds, or touch the zero button  located on the screen. When zeroing is complete, a tone sounds, and “Zeroed” appears along with the current time and date above the waveform plot for the connected pressure cable port.
- 7 Confirm stable zero pressure value and turn stopcocks such that sensors are reading patient intravascular pressure.

- 8 If desired, output the pressure signal to a connected patient monitor. See *Pressure-Out* on page 196 for more information on this option.
- 9 Touch the home icon  to begin CO monitoring. When the next CO value is calculated, it is displayed and updates will continue as determined by the **CO/Pressure Averaging Time**.

Once CO monitoring is initiated, the blood pressure waveform can also be viewed using the blood pressure waveform display. See *Live Blood Pressure Waveform Display* on page 110. When unplugging the HemoSphere pressure cable from a compatible monitor or sensors from the pressure cable, always pull at the connection site. Do not pull from cables or use tools to disconnect.

10.3.4 SVR Monitoring

When paired with the FloTrac, FloTrac Jr, or Acumen IQ sensor, the HemoSphere pressure cable can monitor systemic vascular resistance (SVR) and systemic vascular resistance index (SVRI) with an analog input CVP pressure signal, pressure cable monitored CVP, or if the user manually enters the patient's CVP value. For information on utilizing the analog signal from a compatible bedside monitor, see *Analog Pressure Signal Input* on page 143. For information on monitoring CVP with a connected pressure cable, see *Pressure Cable Monitoring with a TruWave DPT* on page 193. For information on CVP source prioritization, see table 5-4 on page 124. To manually input the patient's CVP:

- 1 Touch the settings icon  → **Clinical Tools** tab  → **Enter CVP** icon .
- 2 Enter a CVP value.
- 3 Touch the home icon .

When no source of CVP is detected, the default value assigned is 5 mmHg. To change the default value, see *CVP Settings* on page 155. When using the Acumen Hypotension Prediction Index (HPI) feature, SVR is available on the **HPI Secondary Screen**.

10.3.5 Acumen IQ Sensor Parameters Displayed in Invasive Mode

Acumen HPI software parameters can be displayed in Swan-Ganz module (Invasive) monitoring mode with a connected pressure cable and Acumen IQ sensor. With an Acumen IQ sensor, an additional five parameters can be displayed: stroke volume variation (SVV), dynamic arterial elastance ($E_{a_{dyn}}$), systolic slope (dP/dt), pulse pressure variation (PPV), and Acumen Hypotension Prediction Index (HPI). These five parameters are labeled as “Acumen IQ” parameters and can be configured on any monitor screen. The HPI parameter smart alerts and smart trends feature is not available while in Invasive monitoring mode. For more information, see *Multiple Technology Monitoring - Acumen Hypotension Prediction Index Software* on page 184.

10.4 Pressure Cable Monitoring with a TruWave DPT

The HemoSphere pressure cable connects to a single TruWave pressure transducer to provide location based intravascular pressure. Available pressures measured by a TruWave DPT include central venous pressure (CVP) when monitored from a central venous line, diastolic pressure (DIA_{ART}), systolic pressure (SYS_{ART}), mean arterial pressure (MAP), and pulse rate (PR) when monitored from an arterial line, and mean pulmonary arterial pressure (MPAP), diastolic pressure (DIA_{PAP}), and systolic pressure (SYS_{PAP}) when monitored from a pulmonary arterial line. See table 10-1.

While in HemoSphere Swan-Ganz module monitoring mode, the pressure cable can be connected to a TruWave DPT on a pulmonary artery line. Monitoring of PAP while monitoring with a HemoSphere Swan- Ganz module also enables monitoring of 20-second parameter values. See *20-Second Flow Parameters* on page 174.

10.4.1 Connect TruWave DPT

- 1 Connect one end of the pressure cable to the HemoSphere advanced monitor.
- 2 To de-air and prime I.V. flush bag and TruWave transducer: Invert normal saline bag (anticoagulation per institution policy). Spike I.V. bag with fluid administration set, keeping drip chamber upright. While keeping I.V. bag inverted, gently squeeze air out of bag with one hand while pulling flush tab (Snap-Tab) with the other hand until air is emptied from I.V. bag and drip chamber is filled to desired level (½ or full).
- 3 Insert flush bag into pressure infuser bag (DO NOT INFLATE) and hang on IV pole at least 2 ft (60 cm) above the transducer.
- 4 With gravity only (no pressure in Pressure Bag), flush TruWave transducer holding pressure tubing in upright position as the column of fluid raises through the tubing, pushing air out of the pressure tubing until the fluid reaches the end of the tubing (flushing under pressure creates turbulence and increased occurrence of bubbles).
- 5 Pressurize the pressure bag until it reaches 300 mmHg.
- 6 Fast-flush transducer tubing while tapping on tubing and stopcocks to remove any residual bubbles.
- 7 Use a straight in or out motion to connect the TruWave DPT to the HemoSphere pressure cable. The pressure cable LED that surrounds the zero button (see ② in figure 10-1) will flash green indicating that the pressure sensor is detected. A yellow light indicates a fault condition. If this occurs refer to the status bar for specific fault condition details.
- 8 Connect tubing to catheter, and then aspirate and flush system to assure catheter is intra-vascular and remove residual bubbles.
- 9 Use routine transducer calibration procedures (according to institutional policy) to ensure proper pressure signals are being transmitted. Refer to the TruWave pressure transducer's instructions for use.
- 10 Follow steps for entering patient data. See *Patient Data* on page 137.
- 11 Follow the instructions below for zeroing the transducer.

10.4.2 Zero Intravascular Pressure

The TruWave DPT must be zeroed to atmospheric pressure to ensure accurate monitoring.

- 1 Touch the Zero & Waveform icon  located on the navigation bar.

OR

Press the physical zero button  directly on the pressure cable and hold for three seconds (see figure 10-1).

CAUTION To prevent cable damage, do not apply excessive force to the pressure cable zero button.

- 2 The current intravascular pressure waveform is displayed and continually updated on the screen. This is to confirm the zero operation is successful.
- 3 Use the pressure type button for the connected pressure cable port (1 or 2) to select the type/location of pressure sensor being used. The waveform color will match the pressure type selected. The choices for **Pressure Transducer** are:
 - **ART** (red)
 - **CVP** (blue)
 - **PAP** (yellow)

While using multiple pressure cables, the pressure type configured for the first cable is not an available selection choice for the second pressure cable.

- 4 Level the stopcock valve (vent port) just above the TruWave transducer to the patient's phlebostatic axis position according to the instructions for use.
- 5 Open the stopcock valve to measure atmospheric conditions. The pressure should display as a flat line.
- 6 Press the physical zero button  directly on the pressure cable and hold for three seconds, or touch the zero button  located on the screen. When zeroing is complete, a tone sounds, and the message “Zeroed” appears along with the current time and date above the waveform plot for the connected pressure cable port.
- 7 Confirm stable zero pressure value and turn stopcocks such that sensors are reading patient intravascular pressure.
- 8 If desired, output the pressure signal to a connected patient monitor. See *Pressure-Out* on page 196 for more information on this option.
- 9 Touch the home icon  to begin monitoring. See table 10-1 for which key parameters are available based on the type of configuration.

Once pressure cable monitoring is initiated, the blood pressure waveform can also be viewed using the blood pressure waveform display. See *Live Blood Pressure Waveform Display* on page 110.

Parameter values monitored using the TruWave DPT are averaged over a 5 second interval, and displayed every 2 seconds. See table 6-4 on page 142.

10.5 Zero & Waveform Screen




Figure 10-2 Zero & Waveform screen – Zero sensor and Pressure-Out

This screen is accessed through the navigation bar or the clinical actions menu and provides three primary functions:

- 1 Select pressure and zero the sensor
- 2 Output pressure signal. See *Pressure-Out* on page 196
- 3 Check waveform

NOTE Zero & Waveform screen functionality, while monitoring with a pressure cable is accessed through the invasive pressure tab. If ClearSight technology is also connected, it can be accessed through the ClearSight tab.

10.5.1 Select Pressure and Zero Sensor

As previously described, the primary function of the **Zero & Waveform** screen is to allow the user to zero the attached pressure sensor/transducer. The user is required to zero the sensor before monitoring is initiated with the pressure cable.

10.5.2 Pressure-Out

The **Zero & Waveform** screen allows the user to output the pressure waveform to a connected patient monitor. The HemoSphere pressure-out cable is a reusable accessory that enables the user to output pressure monitored by the HemoSphere advanced monitoring platform to a compatible patient monitor for standard pressure monitoring. Up to three waveforms can be transmitted: arterial pressure (AP, red), central venous pressure (CVP, blue), and pulmonary artery pressure (PAP, yellow). Device performance, including

functional characteristics, have been verified in a comprehensive series of testing to support the safety and performance of the device for its intended use when used in accordance with the established Instructions For Use.




Figure 10-3 HemoSphere pressure-out cable


- 1 Plug the 18-pin connector of the HemoSphere pressure-out cable (see ① in figure 10-3) into the rear panel of the monitor at the pressure out port denoted by the analog out symbol . See ⑨ in figure 3-2 on page 73.
- 2 Use the two jack screws to properly secure the pressure-out cable plug in place. See ② in figure 10-3.
- 3 Connect the desired pressure signal plug into a compatible patient monitor:
 - arterial pressure (AP, red, ③)
 - pulmonary artery pressure (PAP, yellow, ④)
 - central venous pressure (CVP, blue, ⑤)

Ensure that the selected connector is fully engaged. Refer to the patient monitor instructions for use.

- 4 Zero the patient monitor.
- 5 Confirm that a value of 0 mmHg is displayed on the patient monitor.
- 6 Toggle to the **Transmit Waveform** icon  to begin pressure signal output to the patient monitor. A “**Sending Waveform Started:**” message with the timestamp is displayed when the live waveform is being transmitted to the connected patient monitor. See figure 10-2 on page 196.

10.5.3 Waveform Confirmation

The Zero & Waveform screen displays the blood pressure waveform. Use this screen or the continuous, real-time blood pressure waveform display (see *Live Blood Pressure Waveform Display* on page 110) to assess the quality of the arterial waveform in response to “Fault: Check Arterial Waveform”. This fault is generated when the arterial pressure signal quality has been poor for too long.



The vertical axis is auto-scaled to the Average BP value ± 50 mmHg.

Monitoring PAP in Invasive Monitoring Mode. The Zero & Waveform is also utilized to monitor the pulmonary artery pressure (PAP) when using the HemoSphere Swan-Ganz module in combination with the pressure cable. While monitoring PAP, touch the **Reference** button to view a waveform screen displaying example waveforms of various catheter tip positions and confirm correct placement in the pulmonary artery.

WARNING Do not use the HemoSphere advanced monitoring platform as a pulse rate or blood pressure monitor.

HemoSphere ClearSight Module Non-Invasive Monitoring

Contents

HemoSphere Non-Invasive System Methodology	199
Connecting the HemoSphere Non-Invasive System	202
Optional HRS	213
SQI	216
Physiocal Method Display	217
ClearSight Settings and Cuff Options	217
Blood Pressure Calibration	219
Output Signal to Patient Monitor	221

11.1 HemoSphere Non-Invasive System Methodology


The HemoSphere non-invasive system is composed of the HemoSphere advanced monitor with ClearSight module and connected pressure controller, heart reference sensor, and compatible Edwards finger cuff(s). See system connections in figure 11-1 on page 203. Accurate measurement of the patient's blood pressure and key hemodynamic parameters is based on the Volume Clamp method, Physiocal method and ClearSight algorithm.

11.1.1 Volume Clamp Method

The ClearSight, ClearSight Jr, Acumen IQ, and Acumen IQ Plus finger cuffs use the Volume Clamp method developed by Czech physiologist J. Peňáz (Penaz J 1973)¹. The finger cuff is equipped with a plethysmograph sensor, which is a combination of a light source and light receiver, to continuously monitor changes in finger arterial blood volume. An inflatable bladder within the cuff rapidly adjusts to this change in volume to equilibrate the pressure of the cuff with the pressure inside of the artery. The artery is therefore clamped at its “un-stretched” volume and the pressure of the cuff is equal to that of the finger arterial pressure at all times.

11.1.2 Physioal Method

The Physioal method, developed by K.H. Wesseling (K.H. Wesseling et al. 1995)², is short for physiological calibration. The Physioal method adjusts for changes in the “un-stretched” volume during a normal measurement period. Cuff pressure is kept constant for one or more heart beats and blood pressure measurement is momentarily interrupted to observe the physiological properties of the finger artery. Early in the measurement period, these interruptions occur regularly. If the properties of the artery are sufficiently constant over time, the interval between Physioal method adjustments will be increased up to 70 heart beats, with higher intervals representing increased measurement stability.



11.1.3 Waveform Reconstruction and Hemodynamic Analysis (ClearSight Technology)

The arterial blood pressure waveform is known to change between the arm and finger arteries due to physiological reasons. ClearSight technology uses advanced processing methods to reconstruct the finger pressure waveform into a radial arterial pressure waveform. Waveform reconstruction yields beat-to-beat values of systolic (SYS), diastolic (DIA) and mean (radial) arterial (MAP) non-invasive pressures. Arterial pulse pressure variation (PPV) is also available. Waveform hemodynamic analysis yields values for pulse rate (PR) using an advanced pulse contour method. Advanced algorithms are used to compute stroke volume variation (SVV) to evaluate dynamic fluid responsiveness.

CAUTION Consider the change in performance of the HemoSphere ClearSight module when using a software version of V01.01.000 or later, which displays and analyzes a reconstructed radial arterial waveform. Software versions earlier than V01.01.000 reconstruct brachial arterial pressure from finger arterial pressure. Clinicians should consider this change in waveform reconstruction, especially if they are experienced with viewing the brachial arterial pressure waveform reconstructed in earlier software versions of the HemoSphere ClearSight module.

ClearSight technology uses advanced processing methods to reconstruct the finger pressure waveform which yields values for cardiac output (CO), cardiac index (CI), stroke volume (SV), and stroke volume index (SVI) using an advanced pulse contour method.

Cardiac power output (CPO) and cardiac power input (CPI) are derived using MAP and CO. Systemic vascular resistance (SVR), and systemic vascular resistance index (SVRI) are derived using MAP and CO when a central venous pressure (CVP) value is entered or monitored.

All non-invasive parameters selected as a key parameter (see table 1-9, “HemoSphere ClearSight module available parameters list,” on page 36) are averaged and have an update rate of 20 seconds.

If an Acumen IQ or Acumen IQ Plus finger cuff and HRS, if applicable, are connected and the Acumen Hypotension Prediction Index feature is activated, the Hypotension Prediction Index, HPI, systolic slope (dP/dt), and dynamic elastance (Ea_{dyn}) can be monitored as key parameters. For more information on setup and usage, see *Acumen Hypotension Prediction Index (HPI) Software Feature* on page 253.

11.1.4 Heart Reference Sensor

The heart reference sensor (HRS) takes into account differences in pressure between the finger and heart. The hydrostatic pressure changes due to difference in height between the finger and heart are compensated by the HRS. One end of the HRS is placed on the finger at the cuff level, and the other end is placed at heart level.

11.1.5 Discoloration, Numbness, or Tingling of the Fingertip

The Volume Clamp methodology places a continual pressure on the finger which never fully occludes the arteries, but inhibits venous return and causes some venous congestion in the fingertip distal to the cuff. As a result, the patient's fingertip may often experience discoloration (blue or red coloring) after a few minutes of monitoring. After longer periods of monitoring (approximately 30 minutes - 2 hours), some patients may experience some tactile sensations (tingling or numbness) in the fingertip. Immediately after removing the cuff, the middle phalanx often shows a slightly decreased volume and may show some reactive hyperemia or swelling. All of these phenomena generally subside within a few minutes of relieving the cuff pressure. Keeping the fingers and hand warm during the measurement improves the arterialization of the fingertip, which can improve coloration and reduce the rate of occurrence of tactile numbing.

11.1.6 Single Cuff Monitoring

A single compatible Edwards finger cuff can be used for accumulated monitoring of the same patient for up to 8 hours on a single finger. During single cuff monitoring, the HemoSphere non-invasive system will automatically release the pressure in the cuff at regular user selected intervals (30 minutes, 2 hours, and 4 hours). See *Cuff Pressure Release Mode* on page 219.

NOTE	After 8 hours of accumulated monitoring on the same finger, the HemoSphere non-invasive system will stop monitoring and display a warning to place the cuff on another finger if continued monitoring is desired.

11.1.7 Double Cuff Monitoring

For monitoring periods lasting longer than 8 hours, the HemoSphere non-invasive system enables two compatible Edwards finger cuffs to be connected simultaneously on separate fingers. In this configuration, the system switches active monitoring between the two cuffs at a user selected interval – 15, 30, or 60 minutes – to allow for minimally interrupted continuous monitoring. During cuff switching, there may be up to a minute pause in monitoring. See *ClearSight Settings and Cuff Options* on page 217.

NOTE	The HemoSphere non-invasive system does not continuously monitor a single finger for more than 60 minutes when two cuffs are used. The double cuff monitoring feature allows for minimum interruptions to monitoring for durations of up to 72 hours. Continuous monitoring cannot be extended beyond 60 minutes on a single finger during double cuff monitoring.

When using the double cuff configuration, ensure that each finger is sized separately. It is not uncommon for patients to have two different sized fingers requiring two different sized compatible Edwards finger cuffs. Failure to select the correct finger cuff can result in measurement inaccuracy.

Cuff sizing may not be applicable to all cuffs.

If an Acumen IQ or Acumen IQ Plus finger cuff and HRS, if applicable, are connected and the Acumen Hypotension Prediction Index feature is activated, the Hypotension Prediction Index feature, HPI parameter, arterial pulse pressure variation (PPV), systolic slope (dP/dt), and dynamic arterial elastance ($E_{a_{dyn}}$) can be monitored as key parameters.

For more information on setup and usage, see *Acumen Hypotension Prediction Index (HPI) Software Feature* on page 253.

When using the double cuff configuration, both finger cuffs must be an Acumen IQ finger cuff to enable HPI.

Upon starting a measurement, the finger cuff will expire after 72 hours for a single patient.

Double cuff monitoring is only applicable with the wrist applied pressure controller.

11.1.8 Methodology References

- 1 Penaz J (1973), “Photoelectric measurement of blood pressure, volume and flow in the finger” *Digest of the 10th Int Conf Med Biol Engng, Dresden*, p. 104.
- 2 Wesseling KH, et al. (1995), “Physiocal, calibration finger vascular physiology for Finapres” *Homeostasis* **36** (2-3), pp. 67-82.

11.2 Connecting the HemoSphere Non-Invasive System

The HemoSphere ClearSight module is compatible with all approved Edwards finger cuffs. See figure 11-1 for an overview of the HemoSphere non-invasive system connections.

- 1 Align and insert the HemoSphere ClearSight module into an large technology (L-Tech) module slot on the left panel of the HemoSphere advanced monitor. The module will click in place when properly engaged.

CAUTION Do not force the module into the slot. Apply even pressure to slide and click the module into place.

- 2 Press the power button to turn on the HemoSphere advanced monitor and follow steps for entering patient data. See *Patient Data* on page 137.

WARNING ClearSight technology use not recommended for patients age < 12 years of age.

- 3 Follow instructions below on how to apply the pressure controller, select finger cuff size and apply the finger cuff(s) to the patient.

NOTE Cuff sizing may not be applicable to all cuffs.




Figure 11-1 HemoSphere non-invasive system connection overview (wrist applied pressure controller)




Figure 11-2 HemoSphere non-invasive system connection overview (Smart pressure controller)

NOTE Components indicated by * in figure 11-1 and figure 11-2 legend are APPLIED PARTS as defined in IEC 60601-1 that in normal use necessarily come into physical contact with the patient for the HemoSphere non-invasive system to perform its function.

WARNING Components that are not indicated as APPLIED PARTS should not be placed in a location where the patient may come into contact with the component.

Compliance to IEC 60601-1 is only maintained when the HemoSphere ClearSight module (applied part connection) is connected to a compatible monitoring platform. Connecting external equipment or configuring the system in a way not described in these instructions will not meet this standard. Failure to use the device as instructed may increase the risk of electrical shock to the patient/operator.

Do not modify, service or alter the product in any way. Servicing, alteration or modification may affect patient/operator safety and/or product performance.

Do not sterilize any components of the HemoSphere non-invasive system. The HemoSphere non-invasive system is provided non sterile.

Refer to cleaning instructions. Do not disinfect the instrument by autoclave or gas sterilization.

Refer to the directions provided with each accessory for specific instructions on placement and use, and for relevant WARNINGS, CAUTIONS, and specifications.

Do not use damaged components/sensors or components/sensors with exposed electrical contacts to prevent patient or user shocks.

The HemoSphere non-invasive system monitoring components are not defibrillation proof. Disconnect the system before defibrillating.

Only use compatible Edwards finger cuffs, heart reference sensor and other HemoSphere non-invasive system accessories, cables and or components that have been supplied and labeled by Edwards. Using other unlabeled accessories, cables and or components may affect patient safety and measurement accuracy.

Always remove HemoSphere non-invasive system sensors and components from the patient and completely disconnect the patient from the instrument before bathing the patient.

CAUTION The effectiveness of HemoSphere non-invasive system has not been evaluated in patients under 12 years of age.

Always grasp the connector, not the cable, when connecting or disconnecting cables. Do not twist or bend the connectors. Confirm that all sensors and cables are connected correctly and completely before use.

11.2.1 Applying the Pressure Controller

11.2.1.1 Wrist Applied Pressure Controller

The pressure controller kit (PC2K or HEMPC2K) consists of a pressure controller (PC2 or HEMPC) and accompanying band (PC2B). A pressure controller cover is available as an accessory. The pressure controller cover secures the heart reference sensor into the pressure controller. See *Pressure Controller Cover* on page 383. Device performance, including functional characteristics, have been verified in a comprehensive series of

testing to support the safety and performance of the device for its intended use when used in accordance with the established Instructions for Use. The pressure controller is worn on the patient's wrist and connects to the HemoSphere ClearSight module, HRS and finger cuff(s). See figure 11-3.




Figure 11-3 Wrist applied pressure controller application

- 1 Wrap the pressure controller band around the patient's wrist. The non dominant hand is preferred for monitoring in awake patients. (figure 11-3, left)
- 2 Snap the pressure controller into the plastic sleeve of the band, making sure that the cuff connectors are facing towards the fingers.
- 3 Attach the pressure controller cable to the HemoSphere ClearSight module. (figure 11-1)
- 4 Remove the plastic connector caps in order to connect the finger cuff (s) and heart reference sensor.

NOTE It is recommended that the cuff connector caps be kept and used to protect the pressure controller against the ingress of water and dirt when only a single cuff is used.

WARNING Do not overtighten the pressure controller band or finger cuff(s).
Do not apply pressure controller band on injured skin as this can cause further injury.

11.2.1.2 Smart Pressure Controller

The Smart pressure controller is clipped near the patient and connects to the ClearSight module and compatible finger cuff. See figure 11-4.




Figure 11-4 Smart pressure controller application

- 1 Insert the Smart pressure controller cable into the HemoSphere ClearSight module port (figure 11-2).

- 2 Use the Smart pressure controller cable clip(s) to place the Smart pressure controller near the patient's forearm. See figure 11-5 (preferred location)




Figure 11-5 Smart pressure controller clip(s)

NOTE Do not clip cable directly next to patient's skin.

- 3 Remove the plastic connector plug in order to connect the finger cuff.

NOTE It is recommended that the cuff connector caps be kept and used to protect the pressure controller against the ingress of water and dirt when the device is not in use.

- 4 Connect the finger cuff to the Smart pressure controller according to the instructions for use supplied with the finger cuff unit.

CAUTION Do not wrap Smart pressure controller cable.

CAUTION Do not attach the Smart pressure controller to patient or patient skin.

11.2.2 Select Finger Cuff Size

Not all finger cuffs are supplied with a sizing aid. Refer to the product IFU for detailed instructions on proper finger cuff sizing, if applicable.




Figure 11-6 Cuff size selection

- 1 Size the finger(s) that will be used for monitoring by using the finger cuff sizing aid. Best results are obtained from the middle, ring or index finger. The cuff is not intended to be placed on the thumb or previously fractured fingers.
- 2 Wrap the sizing aid around the middle phalanx of the finger by pulling the color coded smaller end through the slot to create a snug fit.
- 3 The black arrows indicate suitable cuff size. Match the indicated color with the correct finger cuff size.

WARNING Improper finger cuff placement or sizing can lead to inaccurate monitoring.

11.2.3 Apply Finger Cuff(s)

Refer to the product IFU for detailed instructions on proper compatible Edwards finger cuff placement and actual device illustrations.

Single Patient Use. The ClearSight, ClearSight Jr, Acumen IQ, and Acumen IQ Plus finger cuffs are designed for single patient use. Upon starting a measurement, the finger cuff will expire after 72 hours for a single patient.

Double Cuff Monitoring. The HemoSphere non-invasive system allows two compatible Edwards finger cuffs to be connected simultaneously to alternate the measurement between two fingers. This feature allows for minimum interruptions to monitoring for durations of up to 72 hours and is required for measurements that take longer than 8 hours. This feature can also be used to increase patient comfort. This is only applicable for the wrist applied pressure controller.

11.2.4 Apply the Heart Reference Sensor

The Heart Reference Sensor (HRS) should always be used in conscious patients, freely moving patients or those patients that will be frequently re-positioned during the case, if applicable. Follow the on-screen prompts or the steps below to connect the HRS. This is only applicable if monitoring with the wrist applied pressure controller and compatible finger cuff.




Figure 11-7 Heart reference sensor application

CAUTION Make sure that the HRS is correctly applied so that it can be leveled to the phlebostatic axis.

- 1 Connect the HRS to the wrist applied pressure controller. See ① in figure 11-7.
- 2 Place the pressure controller cover on the wrist applied pressure controller. (optional - see *Pressure Controller Cover* on page 383)
- 3 Apply the heart end of the HRS to the patient at phlebostatic axis level by using an HRS clip. See ② in figure 11-7.

NOTE If the patient is rotated or moved, the phlebostatic axis will rotate or move with the patient. If necessary, be sure to reapply the heart end of the HRS to ensure that it is still at the same vertical level as the heart in the patient's new position.

- 4 Attach the other end of the HRS to the finger cuff. See ③ in figure 11-7.
- 5 Touch the start monitoring icon  on the navigation bar or on setup help screen to begin monitoring.
- 6 Touch the stop monitoring icon  on the navigation bar to end monitoring at any time.
- 7 If ClearSight non-invasive blood pressure measurements vary from a reference measurement, assess the integrity of the HRS by performing an HRS calibration. An HRS calibration must be performed as part of the troubleshooting process. See *Calibrate the Heart Reference Sensor* on page 218.

11.2.5 Accuracy of ClearSight Technology Blood Pressure Measurements

Precaution. Correlation of blood pressure measurements to the reference arterial line may be affected during initial system startup and following a system restart.

Table 11-1 provides a summary of repeated measurements from the same patient to provide accuracy of ClearSight technology non-invasive blood pressure outputs.

Table 11-1 95% Confidence interval (CI) results for repeated blood pressure measurements from the same patient (Bootstrap Re-sampling)

Pediatric ≥ 12 years of age	Bias [95% CI]	Precision [95% CI]
SYS (mmHg)	-9.55 [-10.1, -9.02]	13.1 [10.8, 15.4]
MAP (mmHg)	-7.95 [-8.36, -7.55]	9.35 [7.65, 11.1]
DIA (mmHg)	-5.90 [-6.30, -5.50]	9.22 [7.55, 10.9]
Adult	Bias [95% CI]	Precision [95% CI]
SYS (mmHg)	-2.74 [-4.95, -0.72]	6.15 [4.25, 7.82]
MAP (mmHg)	-1.29 [-2.33, -0.22]	3.14 [2.15, 4.14]
DIA (mmHg)	-1.07 [-2.26, 0.21]	3.71 [2.43, 5.29]

11.2.6 General Troubleshooting of HemoSphere Non-Invasive System Monitoring

Listed below are common issues that may occur during normal monitoring and some troubleshooting steps.

- If ClearSight non-invasive blood pressure measurements vary from a reference measurement, assess the integrity of the HRS by performing an HRS calibration. An HRS calibration must be performed as part of the troubleshooting process. See *Calibrate the Heart Reference Sensor* on page 218.
- If waveform does not appear within minutes after monitoring is initiated, check the status bar for any faults or alerts that may indicate there is a problem. Touch the question icon for more information on a displayed message or see table 15-22, “HemoSphere ClearSight module faults/alerts,” on page 355.
- During measurement, the tip of the finger being monitored by the cuff may show some coloring. This is normal and will disappear within a few minutes of cuff removal.
- During measurement, a conscious patient may notice slight pulsations in the finger to which the cuff is applied. These pulsations will stop momentarily during Physiocal. The patient should be made aware that these irregularities are normal and not caused by the patient’s heart.
- If the patient is responsive, instruct the patient to keep the hand relaxed and not tense the muscles or overstretch the hand.
- Make sure that the blood flow to the hand is not (partially) obstructed, e.g. because the wrist is pressing on a hard surface.
- Some situations, such as cold hands, may make it difficult to start monitoring. If the patient has cold hands, try to warm the hand.

WARNING Do not use the HemoSphere non-invasive system as a heart rate monitor.

If using the instrument during full body irradiation, keep all HemoSphere non-invasive system monitoring components out of the irradiation field. If a monitoring component is exposed to the irradiation, the readings may be affected.

Strong magnetic fields may cause malfunction of the instrument and burn wounds to the patient. Do not use the instrument during magnetic resonance imaging (MRI) scanning. Induced current could potentially cause burns. The device may affect the MR image, and the MRI unit may affect the accuracy of the measurements.

CAUTION The HemoSphere non-invasive system is not intended for use as an apnea monitor.

In patients with extreme contraction of the smooth muscle in the arteries and arterioles in the lower arm and hand, such as may be present in patients with Raynaud's disease, blood pressure measurement can become impossible.

CAUTION Inaccurate non-invasive measurements can be caused by factors such as:

- Improperly calibrated and/or leveled HRS
- Excessive variations in blood pressure. Some conditions that cause BP variations include, but are not limited to:
 - * Intra-aortic balloon pumps
- Any clinical situation where the arterial pressure is deemed inaccurate or not representative of aortic pressure.
- Poor blood circulation to the fingers.
- A bent or flattened finger cuff.
- Excessive patient movement of fingers or hands.
- Artifacts and poor signal quality.
- Incorrect placement of finger cuff, position of finger cuff, or finger cuff too loose.
- Electrocautery or electrosurgical unit interference.

Always disconnect the finger cuff when it is not wrapped around a finger, to prevent damage by accidental over-inflation.

The effectiveness of Edwards compatible finger cuffs has not been established in pre-eclamptic patients.

The pulsations from intra-aortic balloon support can be additive to the pulse rate on the instrument pulse rate display. Verify patient's pulse rate against the ECG heart rate.

The pulse rate measurement is based on the optical detection of a peripheral flow pulse and therefore may not detect certain arrhythmias. The pulse rate should not be used as a replacement or substitute for ECG based arrhythmia analysis.

11.3 Optional HRS

Optional HRS is an advanced feature that must be enabled. If this feature is enabled, steps vary from how described previously in *Heart Reference Sensor* on page 200. The HemoSphere non-invasive system finger cuff algorithm must account for differences in pressure due to the change in vertical level of the monitored finger relative to the heart. This can be performed in one of two ways on the **Patient Positioning Mode Selection** window (see figure 11-8):





Figure 11-8 Patient Positioning Mode Selection – Optional HRS

Manually enter height differences. Use this method to account for height differences only in stationary and sedated patients. After entering patient data, touch on the **Patient Sedated and Stationary** icon and proceed with steps outlined below in *Patient Sedated and Stationary* on page 213.



Use Heart Reference Sensor (HRS). The HRS must be used in patients where the vertical level of the finger relative to the heart may change at any time during monitoring. After entering patient data, touch the **Variable Patient Positioning** button and proceed with steps outlined in the *Heart Reference Sensor* on page 200.



11.3.1 Patient Sedated and Stationary

This mode could be chosen for those patients under general anesthesia with limited or no re-positioning needs anticipated. The HRS can be used during this mode, but is not required.

- 1 Touch **Patient Sedated and Stationary** button to highlight and select this mode.
- 2 Touch **OK**.

CAUTION Monitoring without an HRS may lead to measurement inaccuracies. Ensure patient remains still with accurately measured finger to heart height difference.

Do not place the patient in a non-supine position while monitoring without an HRS. This may lead to an inaccurate vertical offset entry for the HRS and measurement inaccuracies.

NOTE If the Acumen Hypotension Prediction Index feature is enabled, the alert “**HRS and Acumen IQ Cuff(s) are required for HPI features**” will be displayed. Touch **Acknowledge** button if the Acumen HPI feature is not desired for the current monitoring session.

To enable HPI, an Acumen IQ finger cuff and HRS is required.

If an HRS is connected a popup screen with the message “**Alert: HRS Detected**” is displayed. To start monitoring with the HRS, touch **Yes** and proceed to step 2 under *Apply the Heart Reference Sensor* on page 210. To monitor without an HRS, disconnect the HRS and touch **No** and proceed with the steps below.




Figure 11-9 Zero & Waveform screen – Vertical Offset Entry

- 3 The **Zero & Waveform** screen in this mode (shown in figure 11-9) will depict a vertical scale bar to represent the offset of the hand relative to the heart; the heart level is set at zero. A positive offset signifies a patient position where the hand is above the heart. Select the units of the scale bar: **CM** or **IN**.
- 4 Use the slider to move the vertical level of the hand and set the offset between the hand and heart.
- 5 Touch the next arrow .
- 6 A confirmation screen will appear. If the displayed offset is correct for the current patient position, touch **Start Monitoring** to begin monitoring. If the displayed offset value is incorrect, touch **Cancel** and adjust the offset value as needed.
- 7 Touch the stop monitoring icon  on the navigation bar to end monitoring at any time.

Two alerts will cycle through on the information bar with the texts “**Alert: No HRS Connected – Verify Patient Positioning**” and “**Alert Current Offset: Finger <position>**” where <position> is the verified height offset between the monitored finger and heart. The offset value must be updated each time a patient is re-positioned in this mode. In addition, if monitoring is stopped for more than one minute, the vertical offset must be verified again upon restarting monitoring.

11.3.2 Update Offset Value During Monitoring

To update the finger to heart vertical offset value:

- 1 Touch the Zero & Waveform icon  located on the navigation bar or through the Clinical Tools menu.
- 2 Touch the **Update Offset** button on the **Zero & Waveform (Vertical Offset Entry)** screen.
- 3 Use the slider to move the vertical level of the hand to set the offset value to match the new patient position.
- 4 Touch the next arrow .
- 5 A confirmation screen will appear. If the displayed offset is correct for the current patient position, touch **Confirm Offset** to begin monitoring. If the displayed offset value is incorrect, touch **Cancel** and adjust the offset value as needed.

11.3.3 Change Patient Positioning Mode

To change the patient position mode between **Patient Sedated and Stationary** and **Variable Patient Positioning**:

- 1 Touch settings icon  → **Clinical Tools** tab .
- 2 Touch **Patient Data** icon .
- 3 Touch the **Positioning Mode** list button to access the **Patient Positioning Mode Selection** screen.
- 4 Touch and highlight the desired patient positioning mode: **Patient Sedated and Stationary** or **Variable Patient Positioning**.
- 5 Touch the OK button and follow steps outlined in *Patient Sedated and Stationary* on page 213 for **Patient Sedated and Stationary** or *Heart Reference Sensor* on page 200 for **Variable Patient Positioning**.

NOTE While monitoring with an HRS and switching to **Variable Patient Positioning** from **Patient Sedated and Stationary**, monitoring will stop. Touch the start monitoring

icon  to restart monitoring after touching the enter icon.

11.4 SQI

A signal quality indicator (SQI) is present on all non-invasive parameter tiles during HemoSphere non-invasive system monitoring. SQI level is calculated with each parameter update every 20 seconds. See table 11-2 below for a description of arterial waveform SQI levels. SQI levels of one and two are typically associated with alert conditions. An SQI level of zero is shown when monitoring is initializing (starting or resuming). A zero SQI value can also be associated with a fault condition. See table 15-22 on page 355 for a list of finger cuff faults and alerts.

Table 11-2 Arterial waveform SQI levels

Appearance	Level	Indication
	4	Normal
	3	Intermediate (moderately compromised)
	2	Poor (possible alert status causing limited signal)
	1	Unacceptable (possible alert status causing extremely limited or no signal; see table 15-22 on page 355 for a list of finger cuff alerts)
	0	Pressure waveform unavailable (see table 15-22 on page 355 for a list of finger cuff faults)

11.5 Physiological Method Display

The Physiological method is an automatic calibration of the arterial waveform which occurs at regular intervals during non-invasive monitoring. The Physiological method can be observed on the live pressure waveform display as a stepwise increase in pressure upon startup and as brief interruptions throughout monitoring. The interval between Physiological method calibrations is displayed on the arterial waveform graph in parenthesis next to the Physiological method interval icon (see table 11-3). To accurately account for changes in the finger artery characteristics throughout monitoring, Physiological method is performed at regular intervals resulting in momentary interruptions to the arterial waveform.

Table 11-3 Physiological method interval status

Appearance	Physiological method beats interval	Indication
 (60)	≥30	Normal measurement stability
 (20)	<30	Frequent Physiological method interruptions; variable physiological artery properties and decreased measurement stability
 (--)	--	Physiological method being performed or status not available

11.6 ClearSight Settings and Cuff Options

The ClearSight settings screen allows the user to select the time interval between cuff pressure release and the switching time interval for double cuff monitoring. Sensor status and information for connected finger cuff(s) and HRS are displayed, if applicable, and HRS calibration is also performed from this screen.

NOTE Allow for at least 10 minutes of monitoring before reviewing sensor status information.

- 1 Touch settings icon  → **Settings** tab  → **ClearSight** button.
- 2 Touch the **Options** tab to view monitoring settings. All selection options on this settings screen are not available during active monitoring or during cuff pressure release mode.

Single Cuff. For single cuff monitoring, select a cuff pressure release time interval from the available option list. At the end of the cuff pressure time release interval, the pressure will be released from the cuff for a duration indicated by the countdown timer on the information bar. See *Cuff Pressure Release Mode* on page 219.

Double Cuff. For double cuff monitoring, select a switching time interval from the available option list. This is only applicable for the wrist applied pressure controller.

Optional HRS. The optional heart reference sensor (HRS) feature can be **enabled** or **disabled** from this toggle button. This menu option is an advanced feature and must be enabled. If the **Optional HRS** feature is enabled, the user has the option of manually entering a vertical offset value between the hand and heart instead of using an HRS. See *Optional HRS* on page 213.

- 3 Touch the **Sensor Status** tab to view connected finger cuff(s) and HRS status and information.
- 4 Touch the **HRS Calibration** tab to calibrate the HRS.

11.6.1 Calibrate the Heart Reference Sensor

The Heart Reference Sensor (HRS), if applicable, should be calibrated to ensure optimal performance.





Figure 11-10 Heart reference sensor calibration

- 1 Navigate to the **HRS Calibration** screen by touching the settings icon  → **Settings** tab

 **Settings** → **ClearSight** button → **HRS Calibration** tab

OR

Touching the settings icon  → **Clinical Tools** tab  → **HRS Calibration**

icon .

- 2 Connect the HRS to the wrist applied pressure controller. See ① in figure 11-10.
- 3 Vertically align both ends of the HRS and touch the **Calibrate** button. See ② in figure 11-10.
- 4 Wait for the indication that the HRS has been calibrated.




Figure 11-11 HRS calibration screen

11.6.2 Cuff Pressure Release Mode



During single cuff monitoring, the HemoSphere non-invasive system will automatically release pressure from the finger cuff at regular intervals.



When ≤ 5 minutes remain until **Cuff Pressure Release Mode**, a white countdown timer icon will appear on the information bar along with the time remaining until pressure release. A notification popup will indicate that the countdown clock has been initiated. The user has the option to extend the countdown time until cuff pressure release by touching **Postpone** on the notification popup. Continuous monitoring will not be extended beyond the 8 hour cumulative monitoring limit on a single finger. Refer to *Single Cuff Monitoring* on page 201 and *Double Cuff Monitoring* on page 201.



At the end of the cuff pressure time release interval, pressure will be released from the cuff and monitoring will be temporarily suspended. A notification will appear on the screen to indicate that finger cuff pressure has been released. The cuff pressure release icon will appear yellow and the timer will indicate time until monitoring is automatically resumed.



During **Cuff Pressure Release Mode**, a countdown clock appears on the navigation bar. A **Pressure Release Active** popup menu will appear on the screen. This menu can also be accessed by touching the navigation or information bar countdown clocks. Menu options on this popup include: **Postpone Release** and **Stop Monitoring**.

NOTE

Cuff pressure release intervals can only be changed when monitoring is stopped. Avoid frequent changes to cuff release intervals during a patient monitoring session.

11.7 Blood Pressure Calibration

The **BP Calibration** screen allows the user to calibrate ClearSight finger cuff monitored blood pressure values with reference blood pressure monitored values. Both brachial oscillometric cuff or radial arterial line reference values can be used.

NOTE

BP Calibration is not available during double cuff monitoring.

BP Calibration is recommended in pediatric patients.


CAUTION

Do not perform a BP calibration during monitoring periods when blood pressure appears unstable. This may result in inaccurate blood pressure measurements.

- 1 Touch the settings icon



→ **Clinical Tools** tab



- **BP Calibration** icon



- 2 Touch **Add Measurement** to enter the reference BP values.

NOTE Once the **Add Measurement** button is touched, the current ClearSight technology BP values are displayed and the user has five minutes to enter reference BP values. If more than five minutes are needed, the **Add Measurement** button can be touched again to reset the five minute timer.




Figure 11-12 BP Calibration Screen

- 3** Enter a **Reference SYS** and **Reference DIA** value.
- 4** Touch **Calibrate** to complete the calibration process. The abbreviation of calibration (**CAL**) will appear above the parameter name on the BP tile to indicate that ClearSight technology BP has been calibrated.
- 5** To clear the last entered BP reference values, touch **Clear BP Calibration**.

NOTE The current **BP calibration** will be cleared if monitoring is paused for more than 10 minutes.

If monitoring without an HRS, BP Calibration will be disabled for one minute after updating the HRS vertical offset entry.

Table 11-4 provides bias and precision performance data for each parameter of the ClearSight system, comparing BP calibrated with radial line monitored patients and BP Calibration with brachial oscillometric cuff monitored patients.

Table 11-4 BP Calibration performance data

Parameter (units)	Calibration reference	Bias	Precision
Pediatric ≥ 12 years of age			
SYS (mmHg)	Radial	0.18 [0.01, 0.36]	3.98 [3.61, 4.35]
	Brachial	0.86 [0.11, 1.61]	5.86 [4.62, 7.11]
DIA (mmHg)	Radial	-0.29 [-0.43, -0.16]	2.91 [2.64, 3.18]
	Brachial	-1.22 [-2.16, -0.28]	5.20 [4.46, 5.94]

Table 11-4 BP Calibration performance data (continued)

Parameter (units)	Calibration reference	Bias	Precision
MAP (mmHg)	Radial	-0.50 [-0.66, -0.34]	3.54 [3.11, 3.98]
	Brachial	-0.87 [-1.63, -0.12]	5.16 [4.05, 6.26]
Parameter (units)	Calibration reference	Bias	Precision
Adult			
SYS (mmHg)	Radial	2.2 [1.3, 3.1]	2.8 [2.0, 3.5]
	Brachial	3.4 [1.1, 5.5]	5.1 [3.2, 7.0]
DIA (mmHg)	Radial	1.1 [0.4, 1.8]	2.1 [1.6, 2.6]
	Brachial	1.6 [0.3, 2.9]	3.0 [1.6, 4.3]
MAP (mmHg)	Radial	1.3 [0.4, 2.3]	2.8 [2.1, 3.6]
	Brachial	2.0 [0.4, 3.6]	3.7 [2.0, 5.5]
CO (L/min)*	Radial	-0.1 [-0.1, -0.1]	0.6 [0.5, 0.6]
	Brachial	-0.1 [-0.2, -0.0]	0.5 [0.3, 0.6]
SVV (%)	Radial	-0.5 [-0.6, -0.5]	1.3 [1.1, 1.4]
	Brachial	-0.7 [-0.9, -0.4]	1.1 [0.8, 1.4]
PPV (%)	Radial	0.2 [0.1, 0.3]	1.7 [1.6, 1.9]
	Brachial	0.0 [-0.3, 0.3]	1.2 [0.8, 1.5]
Ea _{dyn} (none)	Radial	0.1 [0.1, 0.1]	0.2 [0.1, 0.2]
	Brachial	0.1 [0.0, 0.1]	0.1 [0.1, 0.1]
dP/dt (mmHg/s)	Radial	21.1 [15.0, 27.3]	124.0 [107.0, 141.1]
	Brachial	20.8 [-4.8, 46.3]	105.4 [73.5, 137.3]
HPI (none)	Radial	-0.9 [-1.6, -0.1]	15.8 [14.6, 16.9]
	Brachial	-0.3 [-2.1, 1.4]	5.9 [4.1, 7.7]
PR (bpm) RMSE	Radial	0.59 [0.23, 0.91]	N/A
	Brachial	0.27 [0.10, 0.44]	N/A
<i>*Note: The bias and precision measurements for the reported parameters are in reference to FloTrac (minimally-invasive) derived measurements and may not represent the performance of the ClearSight (NIBP) system compared to appropriate reference measurements for CO (e.g., multiple averaged bolus thermodilution measurements).</i>			

11.8 Output Signal to Patient Monitor

The **Zero & Waveform** screen provides the user with the option to send the arterial waveform signal to a bedside patient monitor.

- 1 Touch the Zero & Waveform icon  located on the navigation bar or through the Clinical Tools menu.
- 2 Plug the HemoSphere pressure-out cable into the rear panel of the monitor at the pressure out port. See © in figure 3-2 on page 73. For more detailed connection instructions see *Pressure-Out* on page 196.

- 3 Connect the arterial pressure (AP, red) pressure signal plug into a compatible patient monitor. Ensure that the selected connector is fully engaged. Refer to the patient monitor instructions for use.
- 4 Zero patient monitor and confirm 0 mmHg is displayed. See ② in figure 11-13. Refer to the patient monitor instructions for use.
- 5 Toggle to the **Transmit Waveform** icon  to begin pressure signal output to the patient monitor. See ③ in figure 11-13.
- 6 A “**Sending Waveform Started:**” message with the timestamp is displayed when the live waveform is being transmitted to the connected patient monitor. See ③ in figure 11-13.

NOTE Normal interruptions to arterial waveform monitoring, such as during PhysioCal, cuff switching, or cuff pressure release mode, can trigger an alert on the patient monitor.




Figure 11-13 Transmit arterial pressure waveform to patient monitor

Venous Oximetry Monitoring

Contents

Oximetry Cable Overview	223
Venous Oximetry Setup	223
In Vitro Calibration	225
In Vivo Calibration	226
Signal Quality Indicator	228
Recall Venous Oximetry Data	228
HGB Update	230
HemoSphere Oximetry Cable Reset	230
New Catheter	231

12.1 Oximetry Cable Overview

The HemoSphere oximetry cable is a reusable device that connects with HemoSphere advanced monitor on one end and any approved Edwards oximetry catheter on the other end. The HemoSphere oximetry cable is a non-contact device and should not touch the patient during normal use. The oximetry cable continuously measures venous oxygen saturation by reflectance spectrophotometry. LEDs within the oximetry cable transmit light fiber optically to the distal end of the catheter. The amount of light absorbed, refracted, and reflected depends on the relative amounts of oxygenated and deoxygenated hemoglobin in the blood. This optical intensity data is gathered by the oximetry catheter, processed by the HemoSphere oximetry cable and displayed on a compatible monitoring platform. Parameter output is mixed venous oxygen saturation (SvO₂) or central venous oxygen saturation (ScvO₂).

12.2 Venous Oximetry Setup

Refer to the directions for use provided with each catheter for specific instructions on catheter placement and use, and for relevant warnings, cautions and notes.

Precaution. Unwind the cable carefully while removing it from its packed configuration. Do not pull at the cable to uncoil it. Check that the enclosure door at catheter connection point of the oximetry cable moves freely and latches properly. Do not use the oximetry cable if the door is damaged, open, or missing. If the door becomes damaged, contact Edwards technical support.

The HemoSphere oximetry cable must be calibrated before monitoring. For information on tissue oximetry monitoring, see *HemoSphere Tissue Oximetry Monitoring* on page 232.

- 1 Connect the HemoSphere oximetry cable to the HemoSphere advanced monitor. The following message will appear:

Oximetry Cable Initializing, Please Wait

- 2 If the HemoSphere advanced monitor is not on, turn on the power switch and follow steps for entering patient data. See *Patient Data* on page 137.
- 3 Remove a section of the catheter tray lid to expose the optical connector.
- 4 Insert the optical connector of the catheter “TOP” side up into the oximetry cable and snap the enclosure shut.




Figure 12-1 Venous oximetry connection overview

NOTE

Appearance of catheter shown in figure 12-1 is for example only. Actual appearance may vary depending on catheter model.

When unplugging the HemoSphere oximetry cable from the HemoSphere advanced monitor or catheters from the oximetry cable, always pull at the connection site. Do not pull from cables or use tools to disconnect.

Pulmonary artery and central venous catheters are TYPE CF defibrillation proof APPLIED PARTS. Patient cables that attach to the catheter, such as the HemoSphere oximetry cable, are not intended to be applied parts but may come into contact with the patient and meet the relevant applied part requirements per IEC 60601-1.

CAUTION Make sure that the oximetry cable is securely stabilized to prevent unnecessary movement of the attached catheter.

WARNING Compliance to IEC 60601-1 is only maintained when the HemoSphere oximetry cable (applied part accessory, defibrillation proof) is connected to a compatible monitoring platform. Connecting external equipment or configuring the system in a way not described in these instructions will not meet this standard. Failure to use the device as instructed may increase the risk of electrical shock to the patient/operator.

Do not wrap the main body of the oximetry cable in fabric or place directly on the patient's skin. The surface does get warm (up to 45 °C) and needs to dissipate heat to maintain its internal temperature level. A software fault will trigger if the internal temperature exceeds its limits.

Do not modify, service or alter the product in any way. Servicing, alteration or modification may affect patient/operator safety and/or product performance.

12.3 In Vitro Calibration

In vitro calibration is performed before the catheter is inserted into the patient, using the calibration cup provided in the catheter packaging.

NOTE Once an oximetry cable has been in vitro or in vivo calibrated, faults or alerts can be generated if monitoring venous oximetry without a connected patient catheter.

CAUTION The catheter tip or calibration cup must not get wet before an in vitro calibration is performed. The catheter and the calibration cup must be dry for an accurate oximetry in vitro calibration. Flush the catheter lumen only after the in vitro calibration has been completed.

Performing an in vitro calibration after the oximetry catheter has been inserted into the patient will yield an inaccurate calibration.

- 1 Touch the oximetry calibration icon  on the **ScvO₂/SvO₂** parameter tile or, touch the settings icon  → **Clinical Tools** tab  **Clinical Tools** → **Venous Oximetry Calibration** icon .
- 2 At the top of the **Venous Oximetry Calibration** screen, select oximetry type: **ScvO₂** or **SvO₂**.
- 3 Touch **In vitro Calibration** button.

- 4 On the **In vitro Calibration** screen, enter either the patient's hemoglobin (**HGB**) or hematocrit (**Hct**). Hemoglobin may be entered in either g/dL or mmol/L on the keypad. See table 12-1 for acceptable ranges.

Table 12-1 In vitro calibration options

Option	Description	Selection range
HGB (g/dL)	Hemoglobin	4.0 to 20.0
HGB (mmol/L)		2.5 to 12.4
Hct (%)	Hematocrit	12 to 60

- 5 Touch **Calibrate** button to start the calibration process.
- 6 When the calibration successfully completes, the following message appears:
In vitro Calibration OK, insert catheter
- 7 Insert the catheter as described in the catheter instructions for use.
- 8 Touch **Start** button.

12.3.1 In Vitro Calibration Error

If the HemoSphere advanced monitor is unable to perform an in vitro calibration, an error popup screen appears.

Touch **In vitro Calibration** button to repeat the oximetry calibration process.

OR

Touch **Cancel** button to return to the **Venous Oximetry Calibration** menu.

12.4 In Vivo Calibration

Use in vivo calibration to perform a calibration after the catheter has been inserted into the patient.

NOTE This process requires approved personnel to draw waste blood (clearing volume) and a blood sample for laboratory processing. A measured oximetry value must be obtained from a co-oximeter.

For optimal accuracy, in vivo calibration should be performed at least every 24 hours.

Signal quality is displayed during in vivo calibration. It is recommended that calibration be performed only when the SQI level is 3 or 4. See *Signal Quality Indicator* on page 228.

- 1 Touch the oximetry calibration icon  on the **ScvO₂/SvO₂** parameter tile or, touch the settings icon  → **Clinical Tools** tab  → **Venous Oximetry Calibration** icon .
- 2 At the top of the **Venous Oximetry Calibration** screen, select oximetry type: **ScvO₂** or **SvO₂**.

- 3 Touch **In vivo Calibration** button.

If setup is unsuccessful, one of the following messages will be displayed:

Warning: Wall Artifact or Wedge Detected. Reposition catheter.

OR

Warning: Unstable Signal.

- 4 If a “Wall Artifact or Wedge Detected,” or “Unstable Signal” message appears, attempt to troubleshoot the problem as instructed in table 15-26, “Venous oximetry warnings,” on page 363 and touch **Recalibrate** button to restart the baseline setup.

OR

Touch **Continue** button to proceed to the draw operation.

- 5 When baseline calibration is successful, touch **Draw** button and then draw the blood sample.
- 6 Draw the blood sample slowly (2 mL or 2 cc over 30 seconds) and send the blood sample to the lab for measured analysis by co-oximeter.
- 7 When lab values are received, touch **HGB** button to enter the patient’s hemoglobin and touch g/dL or mmol/L or **Hct** button to enter the patient’s hematocrit. See table 12-2 for acceptable ranges.

Table 12-2 In vivo calibration options

Option	Description	Selection range
HGB (g/dL)	Hemoglobin	4.0 to 20.0
HGB (mmol/L)		2.5 to 12.4
Hct (%)	Hematocrit	12 to 60

NOTE When an HGB or Hct value is entered, the system automatically calculates the other value. If both values are selected, the last value entered is accepted.

- 8 Enter the lab oximetry value (**ScvO₂** or **SvO₂**).
- 9 Touch **Calibrate** button.

12.5 Signal Quality Indicator



Signal quality indicator (SQI) is a reflection of the signal quality based on the catheter condition and position within the vessel. The SQI bar boxes fill based on the level of oximetry signal quality. The SQI level is updated every two seconds after oximetry calibration is complete and will display one of four signal levels as described in table 12-3.

Table 12-3 Signal quality indicator levels

SQI symbol	Bars filled	Level	Description
	four	normal	All aspects of the signal are optimal
	three	intermediate	Indicates a moderately compromised signal
	two	poor	Indicates poor signal quality
	one	unacceptable	Indicates a severe problem with one or more aspects of signal quality

Signal quality may be compromised by the following during intravascular oximetry:

- Pulsatility (for example, the catheter tip is wedged)
- Signal Intensity (for example, the catheter is kinked, a blood clot, hemodilution)
- Intermittent vessel wall contact by the catheter

Signal quality is displayed during in vivo calibration and HGB update functions. It is recommended that calibration be performed only when the SQI level is 3 or 4. When SQI is 1 or 2, see *Venous Oximetry Error Messages* on page 362 to determine and resolve the issue.

CAUTION The SQI signal is sometimes affected by the use of electrosurgical units. Attempt to distance electrocautery equipment and cables from the HemoSphere advanced monitor and plug the power cords into separate AC circuits if possible. If signal quality problems persist, call your local Edwards representative for assistance.

12.6 Recall Venous Oximetry Data


Recall Venous Oximetry Data can be used to recall data from the oximetry cable after a patient has been transported away from the HemoSphere advanced monitor. This allows the patients last calibration to be recalled along with the patients demographic data for immediate oximetry monitoring. Calibration data within the oximetry cable must be less than 24 hours old to use this function.

NOTE If patient data has already been entered into the HemoSphere advanced monitor, only system calibration information is recalled. The HemoSphere oximetry cable is updated with current patient data.

- 1 With the catheter connected to the HemoSphere oximetry cable, unplug the cable from the HemoSphere advanced monitor and transport it with the patient. The catheter should not be disconnected from the oximetry cable.
- 2 If the oximetry cable is being connected to another HemoSphere advanced monitor, make sure that previous patient data is cleared.
- 3 Once the patient has been transferred, reconnect the oximetry cable to the HemoSphere advanced monitor and turn it on.
- 4 Touch the gray oximetry calibration icon  on the **ScvO₂/SvO₂** parameter tile or, touch the settings icon  → **Clinical Tools** tab  → **Venous Oximetry Calibration** icon .
- 5 Touch **Recall Venous Oximetry Data** button.
- 6 If the oximetry cable data is less than 24 hours old, touch **Yes** button to start oximetry monitoring using the recalled calibration information.
OR
Touch **No** button and perform an in vivo calibration.

WARNING Before touching **Yes** to recall oximetry data, confirm that the displayed data matches the current patient. Recalling incorrect oximetry calibration data and patient demographics will result in inaccurate measurements.

CAUTION Do not disconnect the oximetry cable while calibration or data recall are in process.

- 7 From the oximetry calibration menu, touch **In vivo Calibration** button to recalibrate the cable. To review patient data that was transported with the oximetry cable, touch the settings icon  → **Clinical Tools** tab  → **Patient Data** icon .


CAUTION If the oximetry cable is being transferred from a HemoSphere advanced monitor to another HemoSphere advanced monitor, check that the patient height, weight, and BSA are correct prior to beginning monitoring. Re-enter patient data, if necessary.

NOTE Keep the time and date of all HemoSphere advanced monitors current. If the date and/or time of the HemoSphere advanced monitor being transported “from” differs from the HemoSphere advanced monitor being transported “to” the following message may appear:
“**Patient data in oximetry cable more than 24 hours old - Recalibrate.**”

If the system needs to be recalibrated, a 10 minute warm up period for the oximetry cable may be required.

12.7 HGB Update

Use the **HGB Update** option to adjust the HGB or Hct value of a previous calibration. The update function can be used only if a previous calibration has been performed, or if the calibration data has been recalled from the oximetry cable.


- 1 Touch the gray oximetry calibration icon  on the **ScvO₂/SvO₂** parameter tile or, touch settings icon  → **Clinical Tools** tab  → **Venous Oximetry Calibration** icon .
- 2 Touch **HGB Update** button.
- 3 You can use the displayed HGB and Hct values or touch **HGB** or **Hct** buttons to enter a new value.
- 4 Touch **Calibrate** button.
- 5 To stop the calibration process, touch the cancel icon .

NOTE To achieve optimal accuracy, we recommended you update the HGB and Hct values when there is a change of 6% or greater in Hct or of 1.8 g/dL (1.1 mmol/L) or greater in HGB. A change in hemoglobin may also affect SQI. Use **HGB Update** to resolve signal quality problems.

12.8 HemoSphere Oximetry Cable Reset

Use HemoSphere oximetry cable reset when the SQI level is continuously low. An oximetry cable reset may stabilize the signal quality. It should be performed only after attempting other actions to resolve the low SQI as defined in Troubleshooting.

NOTE The HemoSphere advanced monitor will not permit an oximetry cable reset before performing a calibration or recalling calibration from the oximetry cable.

- 1 Touch the gray oximetry calibration icon  on the **ScvO₂/SvO₂** parameter tile or, touch settings icon  → **Clinical Tools** tab  → **Venous Oximetry Calibration** icon .
- 2 Touch **Oximetry Cable Reset** button.
- 3 A progress bar will appear. Do not disconnect the oximetry cable.

12.9 New Catheter

Use the **New Catheter** option any time a new catheter is used for a patient. After **New Catheter** is confirmed, oximetry must be re-calibrated. Refer to the directions for use provided with each catheter for specific instructions on catheter placement, calibration type, and use, and for relevant warnings, cautions and notes.

- 1 Touch the gray oximetry calibration icon  on the **ScvO₂/SvO₂** parameter tile or, touch settings icon  → **Clinical Tools** tab  → **Venous Oximetry Calibration** icon .
- 2 Touch **New Catheter** button.
- 3 Touch **Yes** button.

HemoSphere Tissue Oximetry Monitoring

Contents

HemoSphere Tissue Oximetry Monitoring.	232
ForeSight Oximeter Cable Overview	233
Connecting the HemoSphere Technology Module and ForeSight Oximeter Cable.	238

13.1 HemoSphere Tissue Oximetry Monitoring

The HemoSphere technology module is an interface module intended to be used with the ForeSight oximeter cable to display continuous monitoring of blood oxygen saturation in the tissue (StO₂). The ForeSight oximeter cable is a non-invasive device that measures absolute tissue oxygen saturation. It operates on the principle that blood contains hemoglobin in two primary forms – oxygenated hemoglobin (HbO₂) and de-oxygenated hemoglobin (Hb) – which absorb near-infrared light in different, measurable ways.

Tissue oxygen saturation (StO₂) levels are determined by the ratio of oxygenated hemoglobin to total hemoglobin at the microvascular level (arterioles, venules, and capillaries) in the region to which the sensor is applied:

$$\%StO_2 = \frac{\text{Oxygenated Hemoglobin}}{\text{Total Hemoglobin}} = \frac{HbO_2}{HbO_2 + Hb} \times 100$$

The ForeSight oximeter cable incorporates Edwards technology to project harmless near-infrared light (in five precise wavelengths) through the overlying tissue (e.g. scalp and skull) and into the underlying tissue (e.g. brain) via a disposable sensor on the patient's skin. Reflected light is captured by detectors positioned on the sensor for optimal signal collection. After analyzing the reflected light, the cable provides the tissue oxygen saturation level to the HemoSphere technology module and advanced monitor as an absolute number and provides a graphical representation of historical values.

A pulse oximeter only reflects arterial blood oxygen saturation (SpO₂) and requires pulsations to operate; whereas the ForeSight oximeter cable measures even in pulseless conditions and displays the balance of oxygen supply and demand in a target tissue (StO₂), e.g., brain, abdomen, limb muscle. Thus, HemoSphere advanced monitor StO₂ values indicate overall tissue oxygenation state, which provides direct feedback for guiding care interventions.

NOTE

The following components may have alternative labeling conventions:

ForeSight oximeter cable (FSOC) may also be labeled as FORE-SIGHT ELITE tissue oximeter module (FSM).

HemoSphere technology module may also be labeled as HemoSphere tissue oximetry module.

ForeSight sensors or ForeSight Jr sensors may also be labeled as FORE-SIGHT ELITE tissue oximetry sensors.

13.2 ForeSight Oximeter Cable Overview

The following diagrams provide an overview of the ForeSight oximeter cable's physical features.



Figure 13-1 ForeSight oximeter cable front view

① module connector	③ cable housing	⑤ sensor cables
② module cable	④ LED display	⑥ sensor connectors

NOTE

The technology module and sensor cables are shown cut; see table A-17 on page 377. For a description of LED status indicators, see *ForeSight Oximeter Cable Sensor Communication* on page 338.

CAUTION

Avoid placing the ForeSight oximeter cable where the status LED cannot be easily seen.




Figure 13-2 ForeSight oximeter cable rear view

NOTE Images of cable housing rear view in this manual are shown without labeling for clarity.

13.2.1 ForeSight Oximeter Cable Mounting Solutions

The ForeSight oximeter cable is packaged with a mounting clip.

Figure 13-3 and figure 13-4 identify attachment points on the mounting clip and cable housing.




Figure 13-3 Mounting clip attachment points




Figure 13-4 Cable housing – mounting clip attachment points

13.2.2 Installing the Mounting Clip

The mounting clip can be attached to the ForeSight oximeter cable either vertically (typical for a bed rail – see figure 13-5) or horizontally (typical for a pole mount – see figure 13-6).

13.2.2.1 Attaching the Mounting Clip Vertically

To attach the mounting clip vertically:

- 1 On the rear of the cable housing, position the mounting clip with the slot facing the mounting clip slide.
- 2 Slide the mounting clip towards the top of the cable housing, until the mounting clip retaining tab locks in to the vertical mounting clip retaining recess.

NOTE The mounting clip is not designed to be attached with the opening facing up.




Figure 13-5 Attaching the mounting clip vertically

13.2.2.2 Attaching the Mounting Clip Horizontally

To attach the mounting clip horizontally:

- 1 Position the mounting clip with the mounting clip retaining tab facing away from the cable housing, from either the left or right.
- 2 Slide the mounting clip across the rear of the cable housing, until the mounting clip retaining tab locks in to the one of horizontal mounting clip retaining recesses.

NOTE You may attach the mounting clip with the opening facing the left or right side.




Figure 13-6 Attaching the mounting clip horizontally

13.2.3 Removing the Mounting Clip

To remove the mounting clip from the rear of the cable housing (see figure 13-7 on page 237):

- 1 Gently lift the mounting clip retaining tab until it disengages from its recess.

CAUTION Applying too much pressure may break the retaining tab, which may present a risk of the cable falling on the patient, bystander, or operator.

NOTE For information on replacement parts, technical support numbers are located on inside cover. See table B-1 on page 381 for approved parts and accessories.

- 2 Slide the mounting clip in the direction of the mounting clip retaining tab until the mounting clip is free from the mounting clip slide.




Figure 13-7 Removing the mounting clip

- 3 Remove the mounting clip from the rear of the cable housing.

CAUTION Do not lift or pull the ForeSight oximeter cable by any cable connections, or place the cable in any position that might present a risk that the cable may fall on the patient, bystander or operator.

Avoid placing the ForeSight oximeter cable under sheets, or blanket that could restrict air flow around the cable that may increase the cable's case temperature and present an injury.

13.3 Connecting the HemoSphere Technology Module and ForeSight Oximeter Cable

The HemoSphere technology module is compatible with a ForeSight oximeter cable and ForeSight/ForeSight Jr sensors. The HemoSphere technology module fits into a standard module slot.

NOTE The following components may have alternative labeling conventions:
 ForeSight oximeter cable (FSOC) may also be labeled as FORE-SIGHT ELITE tissue oximeter module (FSM).
 HemoSphere technology module may also be labeled as HemoSphere tissue oximetry module.
 ForeSight sensors or ForeSight Jr sensors may also be labeled as FORE-SIGHT ELITE tissue oximetry sensors.




Figure 13-8 Tissue oximetry monitoring connection overview

NOTE ForeSight/ForeSight Jr sensors are TYPE BF defibrillation proof APPLIED PARTS. Patient cables that attach to the sensors, such as the ForeSight oximeter cable, are not intended to be applied parts but may come into contact with the patient and meet the relevant applied part requirements per IEC 60601-1.

The ForeSight oximeter cable can remain connected to the patient during cardiac defibrillation.

The HemoSphere technology module is shipped with ESD covers for the ForeSight

oximeter cable connection ports. After removing them when using the system for the first time, it is recommended that they be kept and used to protect the electrical connection points when the ports are not in use.

WARNING Compliance to IEC 60601-1 is only maintained when the HemoSphere technology module (applied part connection, defibrillation proof) is connected to a compatible monitoring platform. Connecting external equipment or configuring the system in a way not described in these instructions will not meet this standard. Failure to use the device as instructed may increase the risk of electrical shock to the patient/operator.

Inspect all of the ForeSight oximeter cable connections for damage prior to installation. If any damage is noted, the cable must not be used until it has been serviced or replaced. Contact Edwards Technical support. There is a risk that damaged parts could reduce the performance of the cable or present a safety hazard.

To remove any chance of contamination between patients, the ForeSight oximeter cable and cable connections should be cleaned after each case.

To reduce the risk of contamination and cross infection, if the ForeSight oximeter cable or cable connections are grossly contaminated, with blood or other bodily fluids, it should be disinfected. If the ForeSight oximeter cable or cable connections cannot be disinfected, it should be serviced, replaced or discarded. Contact Edwards Technical support.

To reduce the risk of damaging internal elements of the cable assemblies – within the ForeSight oximeter cable housing – avoid excessive pulling, bending or other types of stress on the cable connections.

Do not modify, service or alter the product in any way. Servicing, alteration or modification may affect patient/operator safety and/or product performance.

CAUTION Do not force the module into the slot. Apply even pressure to slide and click the module into place.

- 1 Press the power button to turn on the HemoSphere advanced monitor. All functions are accessed through the touch screen.
 - 2 Ensure proper orientation, then plug the ForeSight oximeter cable into the technology module. Up to two ForeSight oximeter cables can be connected to each technology module.
-

NOTE The ForeSight oximeter cable only connects one way to the HemoSphere technology module. If at first the connection does not go in, rotate the connector and try inserting it again.

Do not pull on any part of the ForeSight oximeter cable connections when unplugging

it from the HemoSphere technology module. If it is necessary to remove the HemoSphere technology module from the monitor, press the release button to unlatch and slide module out.

Once the ForeSight oximeter cable connection has been made to the technology module, the channel 1 and channel 2 status LEDs should turn on. The group status LED will also turn on, indicating the module channels are group A (connected to port A on inserted technology module) or group B (connected to port B on inserted technology module).




Figure 13-9 ForeSight oximeter cable status LED

- 3** Select **Continue Same Patient** button or **New Patient** button and enter new patient data.
- 4** Connect the compatible ForeSight/ForeSight Jr sensor(s) to the ForeSight oximeter cable. Up to two sensors can be connected to each ForeSight oximeter cable. Available sensor locations are listed in table 13-1. See *Attaching Sensors to the Patient* on page 242 and refer to the ForeSight sensor and ForeSight Jr sensor instructions for use for proper sensor application directions.
- 5** Select the **Non-Invasive**, **Invasive** or **Minimally-Invasive** monitoring mode button on the **Monitoring Mode Selection** window as applicable.

6 Touch **Start Monitoring**.

Table 13-1 Tissue oximetry sensor locations

Symbol (right)*	Symbol (left)*	Adult (≥40 kg) anatomical location* (sensor size)	Pediatric (<40 kg) anatomical location* (sensor size)
		brain (large)	brain (medium/small)
		shoulder (large)	n/a
		arm (large)	n/a
		flank/abdomen (large)	flank/abdomen (medium/small)
		n/a	abdomen (medium/small)
		leg – quadriceps (large)	leg – quadriceps (medium)
		leg – calf (gastrocnemius or tibialis, large)	leg – calf (gastrocnemius or tibialis, medium)
*Symbols are color coded based on ForeSight oximeter cable group channel: green for channel A and blue (shown) for channel B			

- If **StO₂** is not a current key parameter, touch the displayed parameter label located inside of any parameter tile to select **StO₂ <Ch>** as a key parameter from the tile configuration menu, where <Ch> is the sensor channel. The channel options are **A1** and **A2** for ForeSight oximeter cable A and **B1** and **B2** for ForeSight oximeter cable B.
- The channel will appear in the upper left corner of the parameter tile. Touch the patient figure  on the parameter tile to access the **Sensor Configuration** window. 
- Select the Patient monitoring mode: adult  or pediatric .

NOTE The sensor mode selection is automatically selected based on the patient's entered body weight. Adult sensor mode is configured for any body weight ≥40 kg.

- Select the anatomical location of the sensor. See table 13-1 for a list of available sensor locations. The sensor locations are color coded based on the HemoSphere technology module connection port:

- **Green:** Sensor locations for a ForeSight oximeter cable connected to port A on HemoSphere technology module
- **Blue:** Sensor locations for a ForeSight oximeter cable connected to port B on HemoSphere technology module

11 Touch the home icon  to return to the monitoring screen.

13.3.1 Attaching Sensors to the Patient

The following sections describe how to prepare the patient for monitoring. For additional information on how to apply a sensor to the patient, see the instructions included in the ForeSight/ForeSight Jr sensor packaging.

13.3.1.1 Selecting a Sensor Site

To ensure patient safety and proper data collection, consider the following items when selecting a sensor site.

WARNING Sensors are not sterile and therefore should not be applied on abraded, cracked, or lacerated skin. Exercise caution when applying sensors to a site with delicate skin. Applying sensors, tape or pressure to such a site may reduce circulation and/or cause skin deterioration.

Do not place sensor over poorly perfused tissues. Avoid uneven skin surfaces for best adhesion. Do not place sensor over sites with ascites, cellulitis, pneumoencephalus, or edema.

If electrocautery procedures will be performed, sensors and electrocautery electrodes should be placed as far apart as possible to prevent unwanted skin burns; a distance of at least 15 cm (6 in) is recommended.

CAUTION Sensors should not be placed on high density hair areas.

The sensor must be able to rest flush with clean, dry skin. Any debris, lotion, oil, powder, perspiration, or hair that prevents good contact between the sensor and the skin will affect the validity of the data collected and may result in an alarm message.

NOTE Skin pigmentation does not affect the validity of collected data. The ForeSight oximeter cable compensates automatically for skin pigmentation.

In the event that the location of the selected tissues cannot be palpated or visualized, confirmation by ultrasound or X-ray is recommended.

Table 13-2 provides sensor selection guidelines based on patient monitoring mode, patient weight, and body location.

Table 13-2 Sensor selection matrix

Patient Mode	Sensor	Weight	Body Location				
			Brain	Flank	Abdomen	Legs	Arms/ Deltoids
Adult	Large	≥40 kg	•	•		•	•
Pediatric	Medium	≥3 kg	•	•	•	•	
Pediatric neonatal	Small	<8 kg	•				
		<5 kg	•	•	•		
Pediatric neonatal	Small, non-adhesive	<8 kg	•				
		<5 kg	•	•	•		

NOTE If you connect a sensor that is sized inappropriately for the current patient monitoring mode, that channel displays an alert on the status bar. If this is the only sensor connected, you may be prompted to switch modes (adult or pediatric).

If you connect a sensor that is sized inappropriately for the selected body location, that channel displays an alert on the status bar. If this is the only sensor connected, you may be prompted to select a different body location or use a different sensor size.

WARNING Use only Edwards supplied accessories with the ForeSight oximeter cable. Edwards accessories ensure patient safety and preserve the integrity, accuracy, and electromagnetic compatibility of the ForeSight oximeter cable. Connecting a non-Edwards sensor will cause an appropriate alert on that channel and no StO₂ values will be recorded.

Sensors are designed for single-patient use, and are not to be reprocessed – re-used sensors present a risk of cross-contamination or infection.

Use a new sensor for each patient and discard it after use. Disposal should follow in accordance with local hospital and institution policies.

If a sensor seems damaged in any way, it must not be used.

Always read the sensor packaging.

13.3.1.2 Preparing the Sensor Site

To prepare the patient's skin for sensor placement:

- 1 Verify that the skin area where the sensor is to be placed is clean, dry, intact, and free of powder, oil, or lotion.
- 2 If necessary, shave hair from skin at the chosen site.

- 3 Use an appropriate cleanser to gently clean the intended sensor site.
You may use Tegaderm or Mepitel under the sensor in patients with delicate skin or edema.
- 4 Allow the skin to dry completely before applying the sensors.

13.3.1.3 Applying Sensors

- 1 Select the appropriate sensor (see table 13-2 on page 243) and remove it from the package.
- 2 Remove and discard the protective liner from the sensor (figure 13-10).




Figure 13-10 Removing protective liner from sensor

NOTE

When using the non-adhesive small sensor, you must size and cut the sensor band length to fit the patient.

- Shorten the sensor band away from the patient. Do not cut the sensor band while on the patient, and do not cut any other part of the sensor.
- Attach the sensor band to the patient with the print facing out.
- Do not over-tighten the sensor band, as pressure can be transferred to the baby.

- 3 Affix the sensor to the patient in the chosen location.

Cerebral Use (figure 13-11): Select the site on the forehead above the eyebrow and just below the hairline where the sensors will be linearly aligned.





Figure 13-11 Sensor placement (cerebral)

Non-Cerebral Use (figure 13-12): Select the site that provides the ideal access to the desired skeletal muscle tissue (if muscle cannot be palpated, too much adipose or edema may be present).

- Arm: Position sensor over the deltoid (shoulder), biceps (upper arm), or brachioradialis muscle.

- Leg: Position sensor over the quadriceps (upper leg), gastrocnemius (calf), or tibialis (calf) muscle. Apply the sensor with the connector towards the feet.
- Flank/Abdomen: Position sensor over the Latissimus dorsi (flank) or external oblique (abdomen) muscle.



① adult

② pediatric/neonatal

③ non-adhesive small sensor

Figure 13-12 Sensor placement (non-cerebral)

NOTE

When monitoring muscle tissue, place the sensor centrally over the selected muscle bed (e.g., middle of upper half of the lower leg as diagrammed).

A muscle bed with significant atrophy may not provide enough tissue for monitoring.

When monitoring for the effects of vascular obstruction in a limb, place a sensor on both the limb of concern and in the same location on the opposing limb.

WARNING Exercise extreme care when applying sensors. Sensor circuits are conductive and must not come into contact with other grounded, conductive parts other than EEG or entropy monitors. Such contact would bridge the patient's isolation and cancel the protection provided by the sensor.

Failure to apply sensors properly may cause incorrect measurements. Misapplied sensors or sensors that become partially dislodged may cause either over- or under-reading of oxygen saturation.

Do not position a sensor under the weight of the patient. Prolonged periods of pressure (such as taping over the sensor or the patient lying on a sensor) transfers weight from the sensor to the skin, which can injure skin and reduce sensor performance.

The sensor site must be inspected at least every 12 hours to reduce the risk of inadequate adhesion, circulation, and skin integrity. If the circulatory condition or skin integrity has deteriorated, the sensor should be applied to a different site.

13.3.1.4 Connecting Sensors to Cables

- 1 Be sure that ForeSight oximeter cable is connected to the technology module and that sensors are placed correctly on the patient's skin.
- 2 Use the clips on the sensor cable to secure and prevent the cable from being pulled away from the patient.

WARNING Do not connect more than one patient to the ForeSight oximeter cable. This may compromise the patient's isolation and cancel the protection provided by the sensor.

CAUTION When used in settings with LED lighting, sensors may need to be covered with a light blocker prior to connection to the sensor cable, as some high intensity systems can interfere with the sensor's near infrared light detection.

Do not lift or pull the ForeSight oximeter cable by any cable connections, or place the ForeSight oximeter cable in any position that might present a risk that the module may fall on the patient, bystander or operator.

- 3 Position the sensor connector in front of the sensor cable connector and align the marks on each (figure 13-13).




Figure 13-13 Connecting a sensor to the sensor cable connector

- 4 Gently push the sensor connector straight into the sensor cable connector until it snaps into place.
- 5 Gently pull back on the sensor to verify the sensor is fully inserted into the connector.

- 6 Verify that the channel status LED indicator on the ForeSight oximeter cable changes from white to green when the sensor is fully connected. See figure 13-14.




Figure 13-14 Connecting a sensor to the ForeSight oximeter cable – channel status LED

CAUTION Once patient monitoring has started, do not replace the sensor or disconnect the sensor for more than 10 minutes to avoid restarting the initial StO₂ calculation.

NOTE If the ForeSight oximeter cable cannot read sensor data properly after starting a new patient, a message to verify the sensors are properly applied to the patient may be displayed.

Confirm that sensors are properly adhered to the patient and dismiss the message and begin monitoring.

When displaying the parameter value change or percent change, the StO₂ parameter value from the start of monitoring is used as a Reference value. See *Display Parameter Value Change* on page 142. If replacing or repositioning a sensor, it is recommended to update the Reference value.

13.3.2 Disconnecting Sensors After Monitoring

Once you are done monitoring a patient, you need to remove the sensors from the patient and disconnect the sensors from the sensor cable as described in the instructions included in the ForeSight/ForeSight Jr sensor packaging.

13.3.3 Monitoring Considerations

13.3.3.1 ForeSight Oximeter Cable Use During Defibrillation

WARNING The ForeSight oximeter cable has been designed to promote patient safety. All cable parts are “Type BF Defibrillation Proof” and are protected against the effects of the defibrillator discharge and may remain attached to the patient. Cable readings may be inaccurate during defibrillator use and up to twenty (20) seconds thereafter.

No separate actions are required when using this equipment with a defibrillator, but only Edwards-supplied sensors must be used for proper protection against the effects of a cardiac defibrillator.

Do not come into contact with patients during defibrillation, or serious injury or death could result.

13.3.3.2 Interference

CAUTION Measurements may be affected in the presence of strong electromagnetic sources such as electrosurgery equipment, and measurements may be inaccurate during use of such equipment.

Elevated levels of carboxyhemoglobin (COHb) or methemoglobin (MetHb) may lead to inaccurate or erroneous measurements, as may intravascular dyes or any substance containing dyes that change usual blood pigmentation. Other factors that may affect measurement accuracy include: myoglobin, hemoglobinopathies, anemia, pooled blood under the skin, interference from foreign objects in the Sensor path, Bilirubinemia, externally applied coloring (tattoos), high levels of HGB or Hct and birthmarks.

When used in settings with LED lighting, sensors may need to be covered with a light blocker prior to connection to the sensor cable, as some high intensity systems can interfere with the sensor’s near infrared light detection.

13.3.3.3 Interpreting StO₂ Values

WARNING If the accuracy of any value displayed on the monitor is questionable, determine the patient’s vital signs by alternative means. The functions of the alarm system for patient monitoring must be verified at regular intervals and whenever the integrity of the product is in doubt.

WARNING Testing of the ForeSight oximeter cable operation should be done at least once every 6 months, as described in HemoSphere service manual. Failure to comply may lead to injury. If the cable fails to respond, it must not be used until it has been inspected and serviced or replaced. See technical support contact information on inside cover.

CAUTION When compared to earlier software versions, a ForeSight oximeter cable with a software version of V3.0.7 or later and used with pediatric sensors (small and medium) is more responsive in the display StO₂ values. Specifically, in the range below 60%, StO₂ measurements could be reported lower than in earlier software versions. Clinicians should consider the faster response and potentially modified StO₂ values when using V3.0.7 software, especially if they are experienced with earlier software versions of the ForeSight oximeter cable.

NOTE For patients experiencing complete bilateral external carotid artery (ECA) occlusion, measurements may be lower than expected.

Table 13-3 and table 13-4 summarize the validation methodology and study results associated with the ForeSight oximeter cable.

Table 13-3 StO₂ validation methodology

Patient population	ForeSight sensor	Cerebral reference	Non-cerebral reference	Type measurement	Subject weight range
Adult	Large	Co-oximetry of jugular bulb and arterial blood samples	Co-oximetry of central venous and arterial blood samples	Single point	≥40 kg
Pediatric – adolescents, children, infants, and neonates	Medium	Co-oximetry of internal jugular vein and arterial blood samples	Co-oximetry of central venous and arterial blood samples	Single point	≥3 kg
Pediatric – adolescents, children, infants, and neonates	Small	Co-oximetry of internal jugular vein and arterial blood samples	Co-oximetry of central venous and arterial blood samples	Single point	3 to 8 kg

Table 13-3 StO₂ validation methodology (continued)


Patient population	ForeSight sensor	Cerebral reference	Non-cerebral reference	Type measurement	Subject weight range
Pediatric – neonates (term, premature, low birth weight, very low birth weight)	Small	FORE-SIGHT MC3010 ¹	Co-oximetry of umbilical venous and pulse oximetry samples	StO ₂ data averaged in two-minute windows ²	<5 kg
¹ Unlike the other ForeSight validation studies, this cerebral validation study did not include invasive measurements because of the challenge for medical centers to obtain consent to insert an internal jugular venous catheter in very small subjects. ² StO₂ data was averaged in two-minute windows for term, premature low birth weight (LBW), and very low birth weight (VLBW) neonates for the following reasons: 1) to reduce the influence of acute changes in StO₂ due to changes in body position or touch as the hemodynamics in premature LBW and VLBW neonates are not as stable compared to normal birth weight neonates, and 2) to enable measurements for both FORE-SIGHT MC3010 and ForeSight sensors or across multiple abdominal locations at nominally the same time for the smallest neonates for which only one sensor can be fitted on the head or specific abdominal location at a time.					

Table 13-4 Clinical validation study results for StO₂

StO ₂ measurement location	Sensor size	Accuracy (Bias ± Precision)*
Cerebral StO ₂	large	46% to 88%: -0.06 ± 3.25% at 1 SD
		46% to 88%: -0.06 ± 3.28% at 1 SD†
	medium	44% to 91%: 0.97 ± 5.43% at 1 SD
		44% to 91%: 1.21 ± 5.63% at 1 SD† 44% to 91%: 1.27 ± 4.93% at 1 SD‡
	small	44% to 90%: -0.74 ± 5.98% at 1 SD
Non-cerebral StO ₂ (somatic)	large	51% to 92%: -0.12 ± 4.15% at 1 SD 51% to 92%: -0.12 ± 4.17% at 1 SD†
	medium	52% to 88%: -0.14 ± 5.75% at 1 SD
	small	66% to 96%: 2.35 ± 5.25% at 1 SD
*Not determined outside of the listed ranges †Dependent Data Bland-Altman Analysis (DDBA) ‡Brain StO₂ values averaged versus REF CX bias and precision Note: StO₂ accuracy is determined based on 30:70% (arterial: venous) reference measurement for REF CX. The method of evaluation for all StO₂ sensor size accuracy measurements was under human clinical evaluation studies.		

13.3.4 Skin Check Timer


Tissue oximetry sensor sites must be inspected at least every 12 hours to reduce the risk of inadequate adhesion, circulation, and skin integrity. The **Skin Check Reminder** displays a reminder every 12 hours, by default. The time interval for this reminder can be modified:

- 1 Touch anywhere in the StO₂ parameter tile → **Sensor Configuration** tab .
- 2 Touch the value button for **Skin Check Reminder** to select a time interval between skin check notifications. The options are: **2 Hours**, **4 Hours**, **6 Hours**, **8 Hours** or **12 Hours** (default).

- 3 To reset the timer, select **Reset** from the **Skin Check Reminder** value button.

13.3.5 Set Averaging Time

StO₂ is measured constantly and parameter display is updated every 2 seconds. The averaging time used to smooth monitored data points can be adjusted. Faster averaging times will limit the filter of irregular or noisy data points.

- 1 Touch anywhere in the **StO₂** parameter tile → **Sensor Configuration** tab .
- 2 Touch the value button for **Averaging** to select a time interval between skin check notifications. The options are:
 - **Slow (24 seconds)**: Higher number of samples gives slower response.
 - **Normal (16 seconds)**
 - **Fast (8 seconds)**: Smaller numbers of samples gives a faster response. This is the default setting for Adult Mode.
 - **None**: Displays values at measurement update rate of 2 seconds. This is the default setting for Pediatric Mode.

13.3.6 Signal Quality Indicator



The signal quality indicator (SQI), displayed on parameter tiles configured for tissue oximetry is a reflection of the signal quality based on the amount of near-infrared light tissue perfusion. The SQI bar boxes fill based on the level of oximetry signal quality. The update rate for StO₂ and SQI level is two seconds. SQI will display one of four signal levels as described in table 13-5.

13.3.7 Relative Change in Total Hemoglobin – Δ ctHb

Table 13-5 Signal quality indicator levels

SQI symbol	Bars filled	Level	Description
	four	normal	All aspects of the signal are optimal
	three	intermediate	Indicates a moderately compromised signal
	two	poor	Indicates poor signal quality
	one	unacceptable	Indicates a severe problem with one or more aspects of signal quality

The relative change in total hemoglobin (Δ ctHb) is a sub-parameter of StO₂. A trending value, Δ ctHb is calculated from the sum of relative changes in oxygenated hemoglobin and deoxygenated hemoglobin (Δ O₂Hb and Δ HHb). Each connected tissue oximetry sensor site StO₂ measurement has its own Δ ctHb sub-parameter. Δ ctHb parameters are only available if the Δ ctHb parameter feature is enabled. Please contact your local Edwards representative for more information on enabling this advanced feature. For additional information, see *Relative Change in Total Hemoglobin – Δ ctHb* on page 330.

13.3.8 Tissue Oximetry Physiology Screen

While monitoring with a ForeSight oximeter cable, three additional physiology screens are available to display the interaction between location specific tissue oximetry values and the cardiovascular system. These three views are shown below in figure 13-15. The default physiology screen while monitoring with the oximeter cable is the tissue oximetry view, which is shown first in figure 13-15. Touch the heart to view the main physiology screen described in *Physiology Screen* on page 113. To return to the tissue oximetry view, touch the magnifying glass.




Figure 13-15 Tissue Oximetry Physiology Screens

Tissue Oximetry. This view displays monitored tissue oximetry values, including cerebral sensor sites, and any of the monitored cardiovascular parameters displayed on the main physiology screen described in *Physiology Screen* on page 113. Touch on the magnifying glass to return to this screen when viewing other physiology screens.

Cerebral Oximetry/Cardiovascular. This view is similar to the main physiology screen with the addition of monitored cerebral oximetry values, if available. Touch between the heart and brain on the tissue oximetry physiology screen to display this view.

Cerebral Oximetry. The cerebral oximetry view displays tissue oximetry values for cerebral configured sensors. Touch on the brain on the tissue oximetry physiology screen to display this view.

Advanced Features

Contents

Acumen Hypotension Prediction Index (HPI) Software Feature	253
Assisted Fluid Management	299
Enhanced Parameter Tracking	324
Fluid Responsiveness Test	327
Relative Change in Total Hemoglobin – Δ ctHb	330
Aortic Stenosis Index (ASI) Algorithm	331

14.1 Acumen Hypotension Prediction Index (HPI) Software Feature

The Acumen Hypotension Prediction Index (HPI) software can be activated in **Minimally-Invasive** monitoring mode, with an Acumen IQ sensor connected, or in **Non-Invasive** monitoring mode, with an Acumen IQ cuff and heart reference sensor (HRS) connected. If using a Smart pressure controller, only an Acumen IQ Plus finger cuff is required. Due to the differences in performance and indications for use depending upon chosen sensor technology, the Acumen Hypotension Prediction Index (HPI) software feature is introduced below based on monitoring technology. Unless otherwise stated, such as the introduction sections below, content in this HPI advanced feature section applies to both monitoring technologies.

14.1.1 Introduction to Acumen Hypotension Prediction Index (HPI) Software in Minimally-Invasive Mode

Acumen Hypotension Prediction Index (HPI) software, when activated and using an Acumen IQ sensor connected to a radial arterial catheter, provides the clinician with information regarding the likelihood of a patient trending towards a hypotensive event and the associated hemodynamics. A hypotensive event is defined as mean arterial pressure (MAP) < 65 mmHg for at least one minute. A target hypotension threshold of 65 mmHg is a default target setting and can be adjusted to 55, 60, 70, 75, 80 or 85 mmHg to fit the hemodynamic profile of an individual patient. The accuracy of the presented measurements is based upon several factors: the arterial line is reliable (not damped), the connected arterial line pressure sensor is well aligned and properly zeroed, and patient demographics (age, gender, height, and weight) have been accurately entered into the device.

CAUTION The effectiveness of the HPI parameter during minimally-invasive monitoring has been established using radial arterial pressure waveform data. The effectiveness of the HPI parameter using arterial pressure from other sites (e.g., femoral) has not been evaluated.

The Acumen HPI feature is intended for use in surgical and non-surgical patients receiving advanced hemodynamic monitoring. The additional quantitative information provided by using the Acumen HPI feature is for reference only and no therapeutic decisions should be made based solely on the Acumen Hypotension Prediction Index (HPI) parameter.

Precaution. If in the clinician's judgment, a mean arterial pressure (MAP) value of < 55, 60, 65, 70, 75, 80, or 85 mmHg would not be meaningful for an individual patient, the clinician may choose to disable the HPI feature completely from the parameter settings menu, or if the information available on the secondary screen is useful, may choose to silence the HPI alarm from the Alarms / Targets screen.

When enabled, the HPI smart alerts and smart trends feature can assist clinicians in the identification of potential underlying mechanism(s) that may be possible targets for intervention to prevent or treat hypotension based on review of the patient's complete hemodynamic state before treatment. These mechanisms include preload, contractility, and afterload. See *HPI Smart Alerts and Smart Trends* on page 266 for more information. When HPI alarms, the HPI high alert popup and smart trends screen display smart alerts for linked parameters.

NOTE When using both HPI smart alerts and AFM simultaneously, it is important to consider that HPI smart alert behaviors are based upon identification of potential underlying mechanism(s) to prevent or treat hypotension, while AFM fluid recommendation behavior is based upon a prediction of fluid responsiveness. As such, these two software features are considering different targets and patient hemodynamic conditions, and should be considered independently. Current patient hemodynamics should be reviewed prior to determining the most appropriate course of action. See *Assisted Fluid Management* on page 299 for more information on that feature.

CAUTION Inaccurate FT-CO measurements can be caused by factors such as:

- Improperly zeroed and/or leveled sensor/transducer
- Over- or under-damped pressure lines
- Excessive variations in blood pressure. Some conditions that cause BP variations include, but are not limited to:
 - * Intra-aortic balloon pumps
- Any clinical situation where the arterial pressure is deemed inaccurate or not representative of aortic pressure, including but not limited to:
 - * Extreme peripheral vasoconstriction which results in a compromised radial arterial pressure waveform
 - * Hyperdynamic conditions as seen in post liver transplant
- Excessive patient movement
- Electrocautery or electrosurgical unit interference

Aortic valve regurgitation may cause an over estimation of Stroke Volume / Cardiac Output calculated depending on the amount of valvular disease and the volume lost back into the left ventricle.

14.1.1.1 Displaying Acumen Hypotension Prediction Index (HPI) Software in Invasive Mode

Acumen HPI software parameters can be displayed in Swan-Ganz module (Invasive) monitoring mode with a connected pressure cable and Acumen IQ sensor. With an Acumen IQ sensor, an additional five parameters can be displayed: stroke volume variation (SVV), dynamic arterial elastance ($E_{a_{dyn}}$), systolic slope (dP/dt), pulse pressure variation (PPV), and Acumen Hypotension Prediction Index (HPI). These five parameters are labeled as “Acumen IQ” parameters and can be configured on any monitor screen. The HPI smart alerts and smart trends feature is not available while in Invasive monitoring mode. For more information, see *Multiple Technology Monitoring - Acumen Hypotension Prediction Index Software* on page 184.

14.1.2 Introduction to Acumen Hypotension Prediction Index (HPI) Software in Non-Invasive Mode

The Edwards Acumen Hypotension Prediction Index (HPI) feature provides the clinician with physiological insight into a patient’s likelihood of future hypotensive events (defined as mean arterial pressure < 55, 60, 65, 70, 75, 80, or 85 mmHg for at least one minute in duration) and the associated hemodynamics. The Acumen HPI feature is intended for use in surgical and non-surgical patients receiving advanced hemodynamic monitoring. The Acumen HPI feature is considered to be additional quantitative information regarding the patient’s physiological condition for reference only and no therapeutic decisions should be made based solely on the Acumen Hypotension Prediction Index (HPI) parameter.

The accuracy of the Acumen Hypotension Prediction Index (HPI) software, when activated and using an Acumen IQ finger cuff and heart reference sensor (HRS), is based upon several factors: the finger cuff has been properly sized and placed, the HRS has been properly calibrated and positioned, and patient demographics (age, gender, height, and weight) have been accurately entered into the device.

NOTE Cuff sizing may not be applicable to all cuffs.

Precaution. If in the clinician’s judgment, a mean arterial pressure (MAP) value of < 55, 60, 65, 70, 75, 80, or 85 mmHg would not be meaningful for an individual patient, the clinician may choose to disable the HPI feature completely from the parameter settings menu, or if the information available on the secondary screen is useful, may choose to silence the HPI alarm from the Alarms/Targets screen.

Clinical validation studies (see *Clinical Validation with Hypotension Threshold in Non-Invasively Monitored Patients* on page 281) demonstrate that ClearSight (NIBP) HPI is accurate and hence useful across the typical range of variation of patient hemodynamics and clinical practice for surgical procedures. The surgery types, surgical characteristics, and non-surgical patient conditions studied are identified in table 14-20 on page 281 to inform clinicians of the patient populations studied.

When enabled, the HPI smart alerts and smart trends feature can assist clinicians in the identification of potential underlying mechanism(s) that may be possible targets for intervention to prevent or treat hypotension based on review of the patient’s complete hemodynamic state before treatment. These mechanisms include preload, contractility, and afterload. See *HPI Smart Alerts and Smart Trends* on page 266 for more information. When HPI alarms, the HPI high alert popup and smart trends screen display smart alerts for linked parameters.

NOTE When using both HPI smart alerts and the AFM algorithm simultaneously, it is important to consider that HPI smart alert behaviors are based upon identification of potential underlying mechanism(s) to prevent or treat hypotension, while AFM algorithm fluid recommendation behavior is based upon a prediction of fluid respon-

siveness. As such, these two software features are considering different targets and patient hemodynamic conditions, and should be considered independently. Current patient hemodynamics should be reviewed prior to determining the most appropriate course of action. See *Assisted Fluid Management* on page 299 for more information on that feature.

-
- CAUTION** Inaccurate non-invasive measurements can be caused by factors such as:
- Improperly calibrated and/or leveled HRS
 - Excessive variations in blood pressure. Some conditions that cause BP variations include, but are not limited to:
 - * Intra-aortic balloon pumps
 - Any clinical situation where the arterial pressure is deemed inaccurate or not representative of aortic pressure.
 - Poor blood circulation to the fingers.
 - A bent or flattened finger cuff.
 - Excessive patient movement of fingers or hands.
 - Artifacts and poor signal quality.
 - Incorrect placement of finger cuff, position of finger cuff, or finger cuff too loose.
 - Electrocautery or electrosurgical unit interference.
-

14.1.3 Acumen Hypotension Prediction Index Parameters Overview

The Acumen Hypotension Prediction Index parameter, HPI, which can be configured as a key parameter on all monitoring screens, displays as an integer value ranging from 0 to 100, with higher values indicating a higher likelihood of a hypotensive event. In addition, the Acumen Hypotension Prediction Index (HPI) software provides three additional configurable parameters, dP/dt , Ea_{dyn} , and PPV, which together with SVV, provide decision support based upon preload responsiveness [SVV or PPV], contractility [dP/dt], and afterload [Ea_{dyn}]. Refer to *Acumen Hypotension Prediction Index (HPI) Parameter Display* on page 257, *HPI Secondary Screen* on page 263, and *Clinical Application* on page 269, for additional information regarding SVV, dP/dt and Ea_{dyn} .

To activate the Acumen HPI software, the platform requires entry of a password to access the Manage Features screen, where an activation key must be entered. Please contact your local Edwards Representative for more information on enabling this Advanced Feature.

Like other monitored parameters, the HPI value updates every 20 seconds. When the HPI value exceeds 85, a high priority alarm is initiated. If the HPI value exceeds 85 for two consecutive readings (total of 40 seconds), an HPI High Alert popup appears on the screen recommending a review of the patient hemodynamics. Hemodynamic information associated with hypotension is available for the user on the HPI Secondary Screen. That information includes several key parameters (MAP, CO, SVR, PR, and SV), as well as more advanced indicators of preload, contractility, and afterload (SVV or PPV, dP/dt , Ea_{dyn}). Additionally, the patient hemodynamics may also be assessed by review of currently configured key parameters, as for example, SVV, PPV, CO and SVR.

Once the Acumen HPI feature is activated, the user can choose to configure Acumen Hypotension Prediction Index (HPI) as a key parameter, display it on the Information Bar, or choose not to display it. dP/dt , Ea_{dyn} , and PPV can also be configured as key parameters.

Refer to the HPI as a Key Parameter and HPI in the Information Bar sections for information about configuring the parameter. See *HPI as a Key Parameter* on page 258 and *HPI on Information Bar* on page 261.

The alarm and alert functions for HPI will differ with the chosen display option for HPI as described in table 14-1.

Table 14-1 HPI display configurations

Display option	Audible and visual alarm	Alert popup
Key Parameter	Yes	Yes
Information Bar	No	Yes
Not displayed	No	No

Unlike other monitored parameters, the HPI alarm limits are not adjustable, as HPI is not a physiologic parameter with a selectable target range (as with cardiac output, for example), but rather a likelihood of physiological state. The alarm limits are displayed to the user in the software, but the controls to change the alarm limits are disabled. The alarm limit for the HPI parameter (> 85 for red alarm range) is a fixed value that may not be modified. The yellow target limit for the HPI parameter ($50 < \text{HPI} \leq 85$ for yellow target range) is also a fixed value that may not be modified.

The visual and audible cues available to the user when the HPI value is > 85 (red alarm range) result from the analysis of multiple variables from an arterial pressure waveform and patient demographic information, and application of a data-driven model developed from retrospectively annotating hypotensive and non-hypotensive episodes. The HPI alarm limit is provided in Table 14-2 on page 258 and in Table D-4 on page 393. The algorithm performance characteristics for the alarm threshold of 85 are provided in table 14-14 and table 14-15 (minimally invasive), and table 14-24 and table 14-25 (non-invasive), included in the clinical validation section.

The parameters dP/dt , Ea_{dyn} , and PPV can be configured as key parameters. PPV and dP/dt behave as other monitored parameters, however Ea_{dyn} is not an alarmable parameter. Alarm/target ranges are unavailable for Ea_{dyn} and target status indicators appear white at all times. A dashed line appears at a value of 0.8 on the Ea_{dyn} graphical trend plot for reference.

14.1.4 Acumen Hypotension Prediction Index (HPI) Parameter Display

The HPI value will update every 20 seconds and displays as a value equating to the likelihood that a hypotensive event may occur on a scale from 0 to 100. The higher the value, the higher the likelihood that a hypotensive event (predefined as mean arterial pressure $< 55, 60, 65, 70, 75, 80$, or 85 mmHg for at least one minute) will occur.

The HPI parameter uses data from the first ten minutes of monitoring to establish a 'base value.' Device performance during these first ten minutes may differ as a result. Table 14-2 provides a detailed explanation and interpretation of HPI graphical display elements (trendline, dial segment [cockpit display], audible alarms, and parameter value [tile display]) and recommended user action when HPI is configured as a key parameter.

WARNING The Acumen Hypotension Prediction Index, HPI, should not be used exclusively to treat patients. A review of the patient's hemodynamics is recommended prior to initiating treatment.

Table 14-2 HPI value graphical and audible display elements

HPI value	Graphical display elements	Audible	General interpretation	Recommended user action
HPI ≤ 50	White	None	Patient hemodynamics indicate that there is a low to moderate likelihood of a hypotensive event occurring. A low HPI value does not exclude a hypotensive event from occurring for surgical patients in the next 5-15 minutes or for non-surgical patients in the next 20-30 minutes regardless of MAP value.	Continue monitoring patient hemodynamics. Remain vigilant with respect to changing patient hemodynamics using the primary monitoring screen, HPI Secondary Screen, HPI, and trends in parameters and vital signs.
50 < HPI ≤ 85	Yellow	None		
HPI > 85	Red (flashing)	High priority alarm tone	Surgical patient has a high likelihood of experiencing a hypotensive event within 15 minutes Non-surgical patient has a high likelihood of experiencing a hypotensive event within 20 minutes	Check patient hemodynamics using the secondary screen and other primary screen parameters in order to investigate the potential cause of the high likelihood of hypotension in order to inform a potential course of action
HPI > 85 and persists for two continuous readings (40 seconds)	Red (flashing) Popup	High priority alarm tone	Surgical patient has a high likelihood of experiencing a hypotensive event within 15 minutes Non-surgical patient has a high likelihood of experiencing a hypotensive event within 20 minutes	Acknowledge popup by chosen method Check patient hemodynamics using the secondary screen and other primary screen parameters in order to investigate the potential cause of the high likelihood of hypotension in order to inform a potential course of action
HPI = 100	Red (flashing) Popup	High priority alarm tone	Patient is hypotensive	Acknowledge popup by chosen method Check patient hemodynamics using the secondary screen and other primary screen parameters in order to investigate the potential cause of the hypotension in order to inform a potential course of action

NOTE If HPI is displayed on the Information Bar the graphical display element changes will not change color nor alarm. Instead the user will only be notified when HPI exceeds 85 for consecutive updates by displaying the HPI High Alert popup.

14.1.5 HPI as a Key Parameter

Once the Acumen HPI feature is activated, the user can configure HPI as a key parameter using the steps described in *Change Parameters* on page 102.

The display of HPI differs in several ways from other key parameters. Display of other key parameters is described in *Status Indicators* on page 104.

Table 14-3 describes the similarities and differences between HPI and other key parameters.

Table 14-3 HPI versus other key parameters: similarities and differences

Similarities	Differences
• Values update every 20 seconds • Audible alarm when > alarm limit • Visual alarm when > alarm limit • Can display % change, if configured • Audible alarm can be disabled	• HPI key parameter tile does not have target color in colored font, depending on clinical/alarm indicator status • HPI key parameter tile has shortcut key in top-right corner to provide direct access to HPI Secondary Screen • HPI will display Alert popup when HPI exceeds high alarm limit for two consecutive updates or HPI value is 100 • HPI only available as key parameter if activation key entered • HPI alarm or yellow limit is not adjustable • HPI does not have a target, green-shaded region with red arrows at the upper and lower limits when displayed as a trend on the main monitoring screen because it is not a physiologic parameter with a target range. Instead HPI is a quantitative indication of physiological status used to inform users of patient likelihood of trending toward a hypotensive event. Specifically: * When HPI is less than or equal to 50, the graphic elements (displayed number trend line or dial segment) are white and clinician should continue monitoring patient hemodynamics using the primary monitoring screen, HPI Secondary Screen, HPI, and trends in parameters and vital signs. * When HPI is greater than 50 and less than or equal to 85, the graphic elements (displayed number trend line or dial segment) are yellow and clinician should continue monitoring patient hemodynamics using the primary monitoring screen, HPI secondary screen, HPI, and trends in parameters and vital signs. * When HPI exceeds 85, the graphical elements (displayed number, trend line, or dial segment) appear red indicating the user should check patient hemodynamics using the secondary screen and other monitoring screen parameters in order to investigate the potential cause of the high likelihood of hypotension (or hypotension if HPI = 100) in order to inform a potential course of action • HPI has four parameter status colors: gray, white, yellow, and red. See table 14-4.




Figure 14-1 HPI key parameter tile

HPI will be displayed as shown in figure 14-1 when configured as a key parameter in all screens except the cockpit screen (figure 14-2). For more information about the cockpit screen, see *Cockpit Screen* on page 114.




Figure 14-2 HPI key parameter on cockpit screen

On all monitoring screens, there is a shortcut icon  in the top-left corner of the HPI key parameter tile. If touched, this shortcut button will display the HPI Secondary Screen (Figure 14-6 on page 265).

On all monitoring screens except the cockpit screen, the font color of the parameter value denotes parameter status as shown in table 14-4. On the cockpit screen, HPI has the same alarm and target ranges, but it is displayed as shown in figure 14-2.

Table 14-4 Parameter status colors for HPI

Parameter status color	Lower limit	Upper limit
Gray	Fault condition	
White	0	50
Yellow	51	85
Red/Gray Flashing	86	100

14.1.6 HPI Alarm

When HPI is configured as a key parameter and exceeds the upper threshold of 85, a high priority alarm will activate which indicates to the user that the patient may be trending towards a hypotensive event. This includes an alarm tone, red parameter status color, and flashing parameter value. The alarm limit of HPI shown in table 14-4 divides the display range into areas of lower and higher likelihood of hypotension. HPI uses features extracted from Acumen IQ measurements, some compared to an initial base value determined over the first 10 minutes of the patient monitoring session, to a data-driven model developed from retrospective analysis of an arterial waveform database collected from ICU and surgical patients containing annotated hypotensive (predefined as mean arterial pressure < 55, 60, 65, 70, 75, 80, or 85 mmHg for at least 1 minute) and non-hypotensive events. HPI is displayed as an integer value between 0 and 100. The assessment of hypotension likelihood using HPI should consider both the displayed value along the range from 0 to 100 and the associated parameter color (white/red). As with other available alarms on the HemoSphere advanced monitoring platform, the volume of the HPI available alarm is adjustable. See *Alarms / Targets* on page 147 for information about silencing the alarm and configuring the alarm volume. Occurrence of HPI alarm will be logged in the data download file following an update with HPI exceeding the alarm limit.

CAUTION The HPI parameter may not provide advanced notice of a trend towards a hypotensive event in situations where a clinical intervention results in a sudden non-physiological hypotensive event. If this occurs, the HPI feature will provide the following without delay: a high alert popup, a high priority alarm, and an HPI value of 100 will be displayed indicating that the patient is undergoing a hypotensive event.

14.1.7 HPI on Information Bar

When HPI is not configured as a key parameter, the parameter value is still computed and displayed on the information bar as shown in figure 14-3.




Figure 14-3 Information bar with HPI

14.1.8 Disable HPI Information Bar Indicator

To disable the HPI information bar indicator:

- 1 Navigate to the HPI Secondary Screen (see *Navigate to HPI Secondary Screen* on page 264).
- 2 Touch the settings icon .
- 3 Disable the **Always Display HPI** option button. See figure 14-10 on page 269.

The HPI feature remains available even when HPI is not displayed on the screen. If HPI is configured as a key parameter, the parameter will alarm and alert as described in *HPI Alarm* on page 260.

14.1.9 HPI High Alert Popup

When HPI exceeds 85 for two consecutive 20-second updates or reaches 100 at any time, the HPI high alert popup becomes active. See figure 14-4. This popup recommends a review of patient hemodynamics and displays either when HPI is configured as a key parameter or appears on the information bar.

WARNING The Acumen Hypotension Prediction Index, HPI, should not be used exclusively to treat patients. A review of the patient's hemodynamics is recommended prior to initiating treatment.

To review patient hemodynamics on the HPI secondary screen (see *HPI Secondary Screen* on page 263) and acknowledge the HPI high alert popup, touch the **Review** button. To acknowledge the HPI high alert popup without reviewing patient hemodynamics on the HPI Secondary Screen, touch the X icon .



Figure 14-4 HPI high alert popup

Upon acknowledgment of the popup, the following will occur:


- The popup will be removed from the display.
- The HPI alarm tone will be silenced for as long as the alert is active.
- The HPI high alert is acknowledged.

The **Review** button is enabled when any monitoring screen is displayed. If the **Review** button on the HPI high alert popup is touched, the HPI Secondary Screen is displayed. When the **Review** button is disabled, the HPI Secondary Screen can still be accessed as described in *HPI Secondary Screen* on page 263.

To disable the HPI alert popup, see *Disable HPI Information Bar Indicator* on page 261.

14.1.10 Hypotension Threshold Setting

To change the MAP threshold used to determine the Acumen Hypotension Prediction Index, HPI, parameter value, navigate to the HPI settings screen by:

- 1 Touching anywhere in the **HPI** parameter tile → **Set Targets** tab  → value under **Hypotension Threshold** → select Hypotension Threshold parameter value.




Figure 14-5 HPI parameter hypotension threshold settings screen

- 2 Select one of the following Hypotension Threshold value options: **55, 60, 65, 70, 75, 80, or 85 mmHg**. The default value is **65 mmHg**. The selected value will be displayed on the parameter tile and on the high alert notification.

NOTE If a new patient is selected, the Acumen Hypotension Prediction Index, HPI, will reset to the default threshold value of **65 mmHg**.

14.1.11 HPI Secondary Screen

The HPI Secondary Screen provides hemodynamic information about the patient. It may be a useful tool to quickly review the patient hemodynamics related to hypotension. This screen may be accessed at any time during hemodynamic monitoring with a Acumen IQ sensor, Acumen IQ cuff, or Acumen IQ Plus cuff.

The HPI Secondary Screen has two viewing modes:

-  HPI relationship view screen
-  HPI smart trend screen




To toggle between these views, touch the toggle icon at the top of the screen.

The HPI Secondary Screen, along with other key parameters on the monitoring screen, can be used to provide potential insight into the cause of a high hypotension likelihood or hypotension when such an event occurs.

14.1.11.1 Navigate to HPI Secondary Screen

To access the HPI Secondary Screen, touch one of the following:

- **Review** button  on the **HPI High Alert** popup
or
• **Review Smart Trends** button  (**Smart Trends** enabled) on the **HPI High Alert** popup.
- HPI information bar indicator button .
- HPI key parameter shortcut icon .
- **Settings** icon  → **Clinical Tools** tab 
→ **HPI Secondary Screen** icon .

NOTE The HPI Secondary Screen is also accessible if the HPI feature is activated and an Acumen IQ sensor or Acumen IQ/Acumen IQ Plus cuff is not connected.

14.1.11.2 HPI Relationship View

The parameters displayed on the HPI Secondary Screen include the following key parameters:



- cardiac output (CO) / cardiac index (CI)
- pulse rate (PR)
- mean arterial pressure (MAP)
- stroke volume (SV) / stroke volume index (SVI)
- systemic vascular resistance (SVR) / systemic vascular resistance index (SVRI)

Additional advanced parameters are arranged visually on the screen by preload, contractility, and afterload. These advanced parameters are:

- stroke volume variation (SVV) or pulse pressure variation (PPV)
- systolic slope (dP/dt)
- dynamic arterial elastance ($E_{a_{dyn}}$)

To toggle between display of PPV or SVV on the relationship view screen, touch the currently displayed parameter name (PPV or SVV) on the HPI Secondary Screen. To toggle between display indexed and non-indexed parameters (CO/CI, SV/SVI, or SVR/SVRI), select the desired parameter as a key parameter. For all of the parameters on the HPI Secondary Screen, the percent change and direction of change (via up/


down arrow) over a user-selectable time interval and small graphical trend plots are displayed. The arterial blood pressure waveform is also displayed. All parameter boxes are outlined in the current target status color, matching visual indicator functionality of parameter tiles.



Figure 14-6 HPI secondary screen - relationship view

The displayed trend graph parameter value scales match the currently configured scales on the graphical trend monitoring screen. See *Adjust Scales* on page 153. The time scale matches the currently selected **% Change** value. The current change interval value is displayed at the top of the HPI Secondary Screen. Configure the change interval directly on the HPI Secondary Screen by touching the settings icon .

The displayed trend graphs can be turned off by touching the **Mini Trends** toggle button. When turned off, the parameter values appear larger and replace the trend plots. See figure 14-7.



Touch any parameter graph to view a larger graphical trend plot. The selected parameter graphical trend plot will appear in place of the blood pressure waveform plot. See figure 14-7. Touch anywhere on the HPI Secondary Screen to exit the enlarged trend graph plot. The graphical trend plot has a thirty second timeout.

For parameter derivations, see table C-1 in appendix C, *Equations for Calculated Patient Parameters*.




Figure 14-7 HPI secondary screen – relationship view with graphical trend value display

14.1.11.3 HPI Smart Alerts and Smart Trends

The HPI smart alerts and smart trends feature can assist clinicians in the identification of potential underlying mechanism(s) that may be possible targets for intervention to prevent or treat hypotension based on review of the patient's complete hemodynamic state before treatment. These mechanisms include preload, contractility, and afterload. The smart alerts algorithm considers the value and % change in value of parameters in relation to user defined thresholds to assist the user in determining the most appropriate course of action. The clinician can link parameters to each of the three physiological mechanisms (preload, contractility, afterload) and customize factors that affect when the category is triggered.

To disable HPI smart alerts, touch the settings icon  on the upper right corner of the HPI Secondary Screen and touch and disable the **Smart Alert** option button .



$E_{a_{dyn}}$ parameter value, MAP parameter value, and the HPI trend plot are displayed on this screen along with one parameter related to each of the following mechanisms:

Mechanism	Related parameter choice
PRELOAD	pulse pressure variation (PPV) stroke volume variation (SVV) stroke volume index (SVI)
CONTRACTILITY	systolic slope (dP/dt) cardiac index (CI)
AFTERLOAD	systemic vascular resistance (SVR)

NOTE The CVP value required for SVR calculation can come from an analog input CVP pressure signal, pressure cable monitored CVP, or a user entered CVP value. For information on CVP source prioritization, see table 5-4, “CVP value prioritization,” on page 124. When no source of CVP is detected, the default value assigned is 5 mmHg. To change the default value, see *CVP Settings* on page 155.

With HPI **Smart Alert** enabled, an HPI smart alert popup appears when HPI alarms. The categories are triggered based on the linked parameter’s state, which includes the parameter’s value and its trend over a user-defined time interval in comparison to defined thresholds.




Figure 14-8 HPI smart alert popup

Triggers for smart alerts are defined by changes in a parameter value beyond a pre-selected parameter target value, and/or % change threshold (10%, 15% or 20%) over a pre-set time interval (5, 10, 15, or 30 minutes) in accordance with user-configurable settings set on the HPI settings screen.

For each parameter, there are specific thresholds that are relevant to the HPI smart alerts decisions. See table 14-5 on page 268. Pre-selected parameter target values are set on the parameter Alarms/Targets screen. See *Alarms / Targets* on page 147. The hard threshold target values listed below are the Edwards default thresholds for parameter warning (yellow) ranges.

Table 14-5 HPI smart alert parameter default thresholds

Parameter	Default threshold
SVV & PPV (%)	≥ 13
SVI (mL/beat/m ²)	≤ 30
CI (L/min/m ²)	≤ 2
dP/dt (mmHg/s)	≤ 480
SVR (dyne-s/cm ⁵)	≤ 1970 /BSA
MAP (mmHg)*	≤ 72
*Note: Hypotension Threshold + 10% (Not configurable) ≤ 72	

A smart alert condition is displayed as a shaded region on the trend graph for that parameter. Smart alert settings (% change value and time interval) are configured by the user.



Figure 14-9 HPI secondary screen – smart trend display

Touch the settings icon  on the upper right corner of the HPI Secondary Screen to access the settings menu.

% Change Threshold (%) (10%, 15%, or 20%). This value determines the change in value over the **% Change Time Interval** at which a parameter displays smart alerts.

% Change Time Interval (Min) (5, 10, 15 or 30 minutes). This interval determines the time frame in which the **% Change Threshold (%)** is evaluated for each displayed parameter.

Parameter Selection. Select a **Preload Parameter** (PPV, SVV, or SVI) and **Contractility Parameter** (dP/dt or CI).




Figure 14-10 HPI secondary screen – smart trend view settings

14.1.12 Clinical Application

The Acumen Hypotension Prediction Index parameter, HPI, can be configured as a key parameter on the monitoring screen, or it can be displayed only in the Information Bar at the bottom right of the monitoring screen, as described in *Acumen Hypotension Prediction Index (HPI) Software Feature* on page 253.

When HPI is displayed in the Information Bar:

- After a second consecutive HPI value exceeds 85, High Alert popup appears.
- Check patient hemodynamics using the HPI Secondary Screen and other primary screen parameters in order to investigate the potential cause of the high likelihood of hypotension in order to inform a potential course of action.

When HPI is configured as a key parameter, HPI and trend graph appear on the monitoring screen:

- Alarm occurs when HPI exceeds 85.
- Trend line and parameter tile outline appears yellow (warning target zone) when HPI is greater than 50 and less than or equal to 85.
- When HPI is less than or equal to 50:
 - * The trend line and value appear white.
 - * Continue monitoring patient hemodynamics. Remain vigilant with respect to changing patient hemodynamics using the primary monitoring screen, HPI Secondary Screen, HPI, and trends in parameters and vital signs.
- When HPI exceeds 85, check patient hemodynamics using the HPI Secondary Screen and other primary screen parameters in order to investigate the potential cause of the high likelihood of hypotension in order to inform a potential course of action.
- Once mean arterial pressure remains below 65 mmHg for three consecutive readings, indicating the occurrence of a hypotensive event:
 - * HPI displays 100.

- * Check patient hemodynamics using the HPI Secondary Screen and other primary screen parameters in order to investigate the potential cause of the hypotension in order to inform a potential course of action.

14.1.13 Additional Parameters

- Stroke Volume Variation (SVV) and Pulse Pressure Variation (PPV) – sensitive dynamic measures of fluid responsiveness, which predict whether the preload is increased – by giving more fluid or by reducing the venous unstressed volume via compensatory control mechanisms or drugs – the heart will respond with an increase in stroke volume [1]. Low values of SVV or PPV are an indicator that a patient is not fluid responsive; high values are an indicator that a patient is fluid responsive; and there is a gray zone in between [6].
- Systolic slope (dP/dt) – the maximum upslope of the arterial pressure waveform from a peripheral artery. The arterial pressure dP/dt (by nature of its computation during outflow) will have absolute values lower than the isovolumic LV pressure dP/dt -max, but their changes correlate strongly [1, 2].

NOTE dP/dt measured from the peripheral artery has not been studied as a measure of left ventricular contractility in all patient populations.

- Dynamic arterial elastance ($E_{a_{dyn}}$) – a measure of the afterload to the left ventricle by the arterial system (arterial elastance), relative to the left ventricular elastance, computed as the ratio between PPV and SVV [8]. The arterial elastance is an integrative arterial load parameter that incorporates systemic vascular resistance (SVR), total arterial compliance (C) and systolic and diastolic time intervals [9, 10].

The correlation of these parameters to physiological status and their relationship to clinical outcome has been well-studied with a large body of clinical literature.

Most interventions to treat SV (or SVI) and MAP, impact primarily SV and its determinants preload, contractility, afterload. Decision support for treatment decisions should integrally provide information on all three aspects, since they often inter-relate.



SVV is limited as preload measure to patients that are mechanically ventilated with stable ventilation frequency and tidal volumes and that do not have intra-abdominal insufflation [6, 7]. SVV is best used in conjunction with stroke volume or cardiac output assessment.

The trending change in dP/dt is helpful as decision support to assess change in contractility of the left ventricle in conjunction with stroke volume variation and stroke volume or cardiac output assessment.

Table 14-6 demonstrates the improved bias and precision of the trended percentage change of dP/dt when compared to absolute values of dP/dt.

Table 14-6 dP/dt accuracy comparison of minimally-invasive and non-invasive monitored surgical patients

Intra-patient bias $\pm$ precision of absolute value dP/dt	Bias $\pm$ precision of percentage changes of dP/dt	Concordance of percentage changes of dP/dt
-3.6 [-58.9, 51.7], mmHg/s	0.02 [-0.00, 0.04] %	88.9% [82.7%, 93.6%]
$\pm$ 83.6 [69.9, 97.4], mmHg/s	$\pm$ 1.35 [1.34, 1.37] %	

CAUTION Exercise caution when using the absolute values of dP/dt. Pressure will change distally due to narrowing of vessels and frictional forces within the vessels. While absolute dP/dt may not be an accurate measure of cardiac contractility, trends may be helpful.

Exercise caution when using dP/dt in patients with severe aortic stenosis, since the stenosis may reduce the coupling between the left ventricle and the afterload.

The dP/dt parameter, although predominantly determined by changes in LV contractility, may be impacted by afterload during periods of vasoplegic states (venoarterial decoupling). During these periods, dP/dt may not reflect changes in LV contractility.

By normalizing the arterial elastance by the ventricular elastance, their ratio becomes an index of the matching between the LV and the arterial system. When matching there is an optimal transfer of blood from the LV to the arterial system without loss of energy and with optimal stroke work [3, 8, 9].

$E_{a_{dyn}}$ has been shown to provide an indication of potential afterload responsiveness to increase MAP by giving volume in preload volume responsive mechanically ventilated patients [4] and spontaneously breathing patients [5]. Afterload responsiveness to increase MAP is greater potentially at values of $E_{a_{dyn}} > 0.8$ [4, 5, 8].

$E_{a_{dyn}}$ is not limited to patients that are mechanically ventilated because it is a computation of presented as the ratio of PPV/SVV [5, 8]. $E_{a_{dyn}}$ is best used in conjunction with stroke volume variation (in ventilated patients) and stroke volume or cardiac output assessment.

SVV or PPV, dP/dt, and $E_{a_{dyn}}$ share the property that one is seldom independent of one or the other. Giving volume to increase the preload and increase the stroke volume leads to an increase in cardiac output and arterial pressure; therefore, the afterload on the ventricle increases. Increasing afterload (increasing aortic pressure) by increasing systemic vascular resistance, will reduce the stroke volume. The resulting increased end-systolic volume, however, leads to a secondary increase in end-diastolic volume because more blood is left inside the ventricle following ejection and this extra blood is added to the venous return, thereby increasing ventricular filling, which increases contractility (Frank-Starling mechanism) and partially offsets the reduction in stroke volume caused by the initial increase in afterload.

SVV or PPV, dP/dt, and $E_{a_{dyn}}$ are intended as integrative decision support parameters to guide an interventional treatment of SV or SV and MAP.

To provide the performance of these parameters using NIBP monitored patients (ClearSight) compared with minimally-invasively monitored patients (FloTrac), the bias and limits of agreement (LoA) were calculated for SVV, PPV, and $E_{a_{dyn}}$. Results of this analysis with 95% confidence intervals are shown below in

table 14-7. 95% confidence intervals were calculated by accounting for the repeated measurements from the same test subject by using the Bland JM, Altman DG (2007) method. The Bland-Altman plots for these parameters are shown in figure 14-11.

Table 14-7 95% Confidence interval (CI) results for bias and limits of agreement (LoA)

Parameter	Bias [95% CI]	Lower LoA [95% CI]	Upper LoA [95% CI]
SVV (%)	-0.18 [-0.25, -0.11]	-3.03 [-3.52, -2.53]	2.66 [2.17, 3.16]
PPV (%)	-0.01 [-0.10, 0.08]	-3.78 [-4.40, -3.17]	3.76 [3.14, 4.38]
Ea _{dyn}	0.04 [0.04, 0.05]	-0.29 [-0.33, -0.25]	0.38 [0.34, 0.42]




Figure 14-11 Bland-Altman plots for SVV, PPV, and Ea_{dyn}

14.1.14 Clinical Validation with Hypotension Threshold in Minimally-Invasive Monitored Patients

Retrospective clinical validation studies were undertaken to assess the diagnostic performance of HPI to predict hypotensive and non-hypotensive events in minimally-invasive monitored surgical and non-surgical patients.

14.1.14.1 Surgical Patients

A retrospective clinical validation study assessed the diagnostic performance of HPI to predict hypotensive and non-hypotensive events in minimally-invasive monitored surgical patients. Table 14-8 provides the patient demographics of the 1141 surgical patients included in this study. The number of hypotensive event segments included for each MAP target and a summary of patient data lengths (in minutes) are provided below in table 14-9.

The 1141 surgical patients can be further stratified as described in table 14-10 below.

Table 14-8 Patient demographics (minimally-invasive monitored surgical patients, N=1141)

Description	Value
# of Patients	1141
Gender (Male/Female)	631/510
Age	58.0 ± 17.0
BSA (m ²)	1.9 ± 0.3
Height (cm)	169.1 ± 10.8
Weight (kg)	80.3 ± 22.2

Table 14-9 Analysis dataset length and number of hypotension events (minimally-invasive monitored surgical patients)

Proposed MAP targets	Number of hypotensive events	Median total data length (minutes)	Mean total data length (± Std Dev, minutes)
55 mmHg	6315	331.3	360.0 ± 158.2
60 mmHg	4763		
65 mmHg	6447		
70 mmHg	3858		
75 mmHg	4050		
80 mmHg	3740		
85 mmHg	3336		

Table 14-10 Surgical patient characteristics (minimally-invasive, N=1141)

Surgery Type	# patients (% of total)
abdominal	333 (27.1%)
cardiac	120 (9.7%)
cervical	83 (6.7%)
cranial	217 (17.6%)
extremity	126 (10.2%)
spinal	62 (5.0%)
thoracic	92 (7.5%)
other HRS	198 (16.1%)
TOTAL*	965 (100%)
*Note: Some patients may have multiple procedures done, thus the total number of procedures are more than the total number of patients	

14.1.14.2 Non-Surgical Patients

A retrospective clinical validation study assessed the diagnostic performance of HPI to predict hypotensive and non-hypotensive events in minimally-invasively monitored non-surgical patients. Table 14-11 provides the patient demographics of the 672 non-surgical patients included in this study. The number of hypotensive event segments included for each MAP target and a summary of patient data lengths (in minutes) are provided below in table 14-12.

The 672 non-surgical patients can be further stratified as described in table 14-13 below.

Table 14-11 Patient demographics (minimally-invasive monitored non-surgical patients, N=672)

Description	Value
# of Patients	672
Gender (Male/ Female)	430/242
Age	62.2 ± 15.8
BSA (m ²)	2.0 ± 0.3
Height (cm)	171.4 ± 11.3
Weight (kg)	88.0 ± 27.4

Table 14-12 Analysis dataset length and number of hypotension events (minimally-invasive monitored non-surgical patients)

Proposed MAP targets	Number of hypotensive events	Median total data length (minutes)	Mean total data length (± Std Dev, minutes)
55 mmHg	5772	1440.0	3321.8 ± 17900.3
60 mmHg	7125		
65 mmHg	11029		
70 mmHg	14845		
75 mmHg	10664		
80 mmHg	11531		
85 mmHg	15508		

Table 14-13 Non-Surgical patient characteristics (minimally-invasive, N=672)

Diagnosis	# patients (% of total)
cardiac ICU	269 (40.0%)
neuro ICU	17 (2.5%)
cardiac surgery floor	4 (0.6%)
medical intensive care unit	9 (1.3%)
surgical intensive care unit	83 (12.4%)
trauma intensive care unit	67 (10%)
cardiology floor	4 (0.6%)
general medical floor	22 (3.3%)
pulmonary floor	8 (1.2%)
surgical floor	138 (20.5%)
transplant floor	11 (1.6%)
unknown	40 (6.0%)
TOTAL	672 (100%)

14.1.14.3 Hypotension Threshold Clinical Validation Study Results – Minimally-Invasive Monitoring

The results of the receiver operating characteristic (ROC) analyses for surgical and non-surgical patients are summarized below in table 14-14 and table 14-15.

A hypotensive event is calculated by identifying a segment of at least 1 minute in length such that all data points in the section have a MAP less than the indicated Variable Map Value (55, 60, 65, 70, 75, 80, and 85). An event (positive) data point is chosen as the sample 5 minutes prior to the hypotensive event. If consecutive hypotension events are less than 5 minutes apart then a positive sample is defined as the first sample immediately following the preceding hypotension event.

A non-hypotensive event is calculated by identifying segments of data points such that the segment is at least 20 minutes away from any hypotensive events and all data points in that segment have MAP less than the indicated Variable Map Value (55, 60, 65, 70, 75, 80, and 85). One non-event (negative) data point is taken for each of the non-hypotensive event segments.

A true positive, as described in table 14-14 and table 14-15, is any event (positive) data point with HPI value greater than or equal to a chosen threshold. Sensitivity is the ratio of true positives to total number of events (positives) with a positive defined as a data point that is at most 5 minutes prior to a hypotensive event. A false negative is any positive data point with HPI value less than the threshold.

A true negative, as described in table 14-14 and table 14-15, is any negative (non-event) data point with HPI value less than a chosen threshold. Specificity is the ratio of true negatives to total number of non-events (negatives) with a negative defined as a data point that is at least 20 minutes away from any hypotensive event. A false positive is any negative data point with HPI value greater than or equal to the threshold.

Positive predictive value (PPV), negative predictive value (NPV), and area under the ROC curve (AUC) are also reported for each variable MAP target in table 14-14 and table 14-15.

Table 14-14 Clinical validation study results* (minimally-invasive monitored surgical patients)

HPI Threshold	Variable MAP Value (mmHg)	PPV [confidence interval]	NPV [confidence interval]	Specificity (%) [95% confidence interval]	Sensitivity (%) [95% confidence interval]	AUC
85	55	99.1 =(5583/5634) [97.7, 99.8]	88.4 =(5586/6318) [86.6, 90.0]	99.1 =(5586/5637) [97.9, 99.8]	88.4 =(5583/6315) [83.1, 91.7]	0.95 [0.93, 0.97]
	60	99.8 =(3958/3964) [99.6, 100.0]	86.5 =(5156/5961) [84.2, 88.5]	99.9 =(5156/5162) [99.8, 100.0]	83.1 =(3958/4763) [75.4, 88.2]	0.94 [0.92, 0.96]
	65	99.8 =(5346/5358) [99.5, 99.9]	76.8 =(3648/4749) [73.1, 80.1]	99.7 =(3648/3660) [99.3, 99.9]	82.9 =(5346/6447) [77.5, 87.1]	0.95 [0.93, 0.96]
	70	98.8 =(2551/2583) [97.8, 99.5]	81.6 =(5784/7091) [79.1, 83.7]	99.4 =(5784/5816) [99.0, 99.8]	66.1 =(2551/3858) [63.7, 68.6]	0.87 [0.86, 0.89]
	75	98.5 =(2715/2755) [97.8, 99.2]	78.7 =(4922/6257) [76.1, 81.2]	99.2 =(4922/4962) [98.8, 99.6]	67.0 =(2715/4050) [64.5, 69.4]	0.87 [0.86, 0.88]
	80	99.3 =(2590/2607) [98.8, 99.8]	78.0 =(4071/5221) [75.1, 80.6]	99.6 =(4071/4088) [99.2, 99.9]	69.3 =(2590/3740) [66.9, 71.5]	0.88 [0.86, 0.89]
	85	97.7 =(2204/2256) [96.1, 99.1]	73.6 =(3164/4296) [70.5, 76.5]	98.4 =(3164/3216) [97.3, 99.4]	66.1 =(2204/3336) [63.7, 68.4]	0.87 [0.85, 0.88]
*Note: Data on File at Edwards Lifesciences						

Table 14-15 Clinical validation study results* (minimally-invasive monitored non-surgical patients)

HPI Threshold	Variable MAP Value (mmHg)	PPV [confidence interval]	NPV [confidence interval]	Specificity (%) [95% confidence interval]	Sensitivity (%) [95% confidence interval]	AUC
85	55	98.7 =(5028/5095) [97.0, 99.8]	98.0 =(36308/ 37052) [97.5, 98.4]	99.8 =(36308/ 36375) [99.6, 100.0]	87.1 =(5028/5772) [84.7, 89.1]	0.97 [0.96, 0.97]
	60	96.1 =(5729/5963) [90.6, 99.4]	97.0 =(44955/ 46351) [96.3, 97.6]	99.5 =(44955/ 45189) [98.7, 99.9]	80.4 =(5729/7125) [76.9, 83.4]	0.95 [0.94, 0.96]
	65	99.0 =(9726/9828) [97.8, 99.8]	95.4 =(27312/ 28615) [94.1, 96.5]	99.6 =(27312/ 27414) [99.2, 99.9]	88.2 =(9726/ 11029) [85.8, 90.2]	0.98 [0.97, 0.98]
	70	99.0 =(13024/ 13162) [98.0, 99.6]	92.9 =(23939/ 25760) [90.7, 94.7]	99.4 =(23939/ 24077) [98.9, 99.8]	87.7 =(13024/ 14845) [84.8, 90.1]	0.98 [0.97, 0.98]
	75	96.8 =(8509/8793) [94.4, 98.6]	94.8 =(38946/ 41101) [93.2, 96.0]	99.3 =(38946/ 39230) [98.8, 99.7]	79.8 =(8509/ 10664) [75.0, 83.8]	0.96 [0.96, 0.97]
	80	95.8 =(9724/ 10154) [93.0, 98.1]	95.0 =(34611/ 36418) [94.0, 96.0]	98.8 =(34611/ 35041) [98.0, 99.4]	84.3 =(9724/ 11531) [81.6, 86.7]	0.96 [0.95, 0.97]
	85	96.0 =(13189/ 13741) [93.1, 98.3]	92.9 =(30359/ 32678) [91.4, 94.2]	98.2 =(30359/ 30911) [96.9, 99.3]	85.0 =(13189/ 15508) [82.6, 87.2]	0.96 [0.95, 0.97]

*Note: Data on File at Edwards Lifesciences

Table 14-16 provides the hypotensive event occurrence percentage and time-to-event data for a given HPI range for surgical patients in the clinical validation study. These data are presented using time windows that have been selected based upon how fast hypotensive events developed on average in surgical patients. Therefore based upon the clinical validation study data, table 14-16 presents data for surgical patients for a time-window of 15 minutes. This analysis is performed by taking samples in each patient from the validation dataset and looking forward in time for a hypotensive event within a 15-minute search window. Once a hypotensive event is found for a given sample then the time-to-event is noted, which is the time duration between the sample and the hypotensive event. The time-to-event statistic is the average event time of all samples that have an event within the search window.

Table 14-17 provides the hypotensive event occurrence percentage and time-to-event data for a given HPI range for non-surgical patients in the clinical validation study. These data are presented using time windows that have been selected based upon how fast hypotensive events developed on average in non-surgical patients. Therefore based upon the clinical validation study data, table 14-17 presents data for non-surgical patients for a time-window of 120 minutes. This analysis is performed by taking samples in each patient from the validation dataset and looking forward in time for a hypotensive event within a 120-minute search

window. Once a hypotensive event is found for a given sample then the time-to-event is noted, which is the time duration between the sample and the hypotensive event. The time-to-event statistic is the average event time of all samples that have an event within the search window.

The event rates, included in table 14-16 and table 14-17, are the ratio of the number of samples that have an event within the search window to the total number of samples. This is done for samples in each MAP target for each of the individual HPI ranges between 10 to 99 as shown in table 14-16 and table 14-17.

CAUTION The HPI parameter information provided in table 14-16 and table 14-17 is presented as general guidance and may not be representative of individual experience. A review of the patient's hemodynamics is recommended prior to initiating treatment.

Table 14-16 Event rate analysis (surgical minimally-invasive, N=1141)

HPI Range	55 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	60 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	65 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	70 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	75 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	80 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	85 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]
10-14	17.5 9 [3.7, 14]	11.0 9 [3.7, 13.7]	14.4 8 [3.3, 14]	22.9 8.7 [4, 13.7]	31.1 8.3 [3.7, 13.3]	24.8 7.7 [3.7, 13.3]	29.2 8.3 [3.7, 13.7]
15-19	19.9 7.8 [2.7, 13.7]	12.5 9 [3.7, 13.7]	18.2 8.3 [3.7, 13.7]	21.8 8.3 [3.7, 14]	30.9 8.3 [3.7, 13.7]	26.4 8 [3.3, 13.7]	28.1 8.3 [3.3, 13.7]
20-24	17.9 8.3 [3, 13.3]	15.1 8.3 [3.3, 14]	21.0 8.3 [3.7, 14]	26.2 8.3 [3.4, 13.7]	32.9 8.3 [3.3, 14]	28.1 8 [3.3, 13.4]	31.4 8.3 [3.3, 13.7]
25-29	21.6 8.3 [3, 13.7]	18.9 8 [3.3, 13.3]	24.2 8.7 [3.3, 13.3]	27.8 8.3 [3.3, 13.7]	30.3 8.3 [3.3, 13.7]	30.7 8 [3.3, 13.7]	33.4 8 [3, 13.7]
30-34	22.3 7.7 [2.7, 13.7]	23.4 7.3 [3, 13.7]	29.2 7.3 [2.7, 13]	32.9 8 [3, 13.7]	36.3 8.3 [3, 13.7]	30.2 7.7 [3, 13.7]	35.7 8 [3, 13.3]
35-39	24.1 7.3 [2.7, 13.3]	28.8 7.3 [2.7, 13.3]	34.9 6.7 [2.7, 12.7]	36.0 7.7 [2.7, 13.3]	39.5 7.7 [2.7, 13.7]	33.4 7.3 [3, 13.3]	38.2 7 [3, 13.3]
40-44	27.6 7 [2.3, 13]	35.0 7.3 [2.3, 13]	44.8 6.3 [2.3, 12.7]	41.7 7 [2.7, 13.3]	42.9 7 [2.3, 13]	37.1 7.3 [2.7, 13.3]	43.2 7.3 [2.7, 13.3]
45-49	30.0 6 [2, 13]	38.8 6.7 [2.3, 13.3]	47.8 6.7 [2.3, 13]	46.4 7 [2.7, 13.3]	48.6 6.7 [2.3, 13]	38.7 7.3 [2.7, 13.7]	46.9 6.7 [2.7, 13]
50-54	32.9 6.3 [2, 13]	42.3 6 [2, 13.3]	52.6 6 [2, 13.3]	48.9 6.7 [2.3, 12.7]	49.6 6.7 [2.3, 13]	42.3 7 [2.3, 13]	48.2 6.7 [2.3, 12.7]

Table 14-16 Event rate analysis (surgical minimally-invasive, N=1141) (continued)

HPI Range	55 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	60 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	65 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	70 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	75 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	80 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	85 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]
55-59	37.7 5.7 [1.7, 12.7]	46.3 5.7 [2, 12.7]	57.0 5.8 [2, 13]	52.1 6.7 [2.3, 13]	52.8 6.3 [2, 12.7]	44.2 6.3 [2, 13]	52.8 6.7 [2.3, 13]
60-64	40.2 6 [1.7, 12.7]	54.6 5.7 [1.7, 12.7]	64.6 5.7 [2, 12.3]	56.6 6.7 [2.3, 13]	58.3 6 [2, 12.7]	54.7 6.3 [2, 13]	55.5 6 [2, 12.7]
65-69	48.0 5.7 [1.7, 13]	61.9 4.7 [1.7, 11.3]	68.7 5 [1.7, 12.3]	63.1 6 [2, 12.7]	65.8 6 [2, 12.7]	59.4 6 [2, 13]	62.8 5.3 [2, 12.3]
70-74	60.7 5.3 [1.3, 12.7]	68.7 4.3 [1.7, 12]	79.5 4.7 [1.7, 12]	71.4 5.7 [1.7, 13]	73.6 5 [1.7, 12.3]	69.4 5.3 [1.7, 12.7]	70.6 5 [1.7, 12.3]
75-79	68.5 4.7 [1.3, 12]	78.1 4.3 [1.3, 11.3]	85.5 4.3 [1.3, 11.7]	77.4 5 [1.7, 12.3]	79.2 5 [1.3, 12]	73.5 5 [1.5, 12]	76.0 5 [1.3, 11.7]
80-84	78.7 4.3 [1, 11.7]	84.6 4.3 [1.3, 11]	88.8 4 [1.3, 11]	82.6 4.7 [1.3, 12]	82.6 4.7 [1.3, 12]	78.1 4.7 [1.3, 12]	81.4 4.7 [1.3, 11.3]
85-89	84.5 4 [1, 11.3]	90.2 4 [1, 11]	90.9 3.7 [1.3, 11.3]	85.8 4.3 [1.3, 11.7]	88.1 4 [1.3, 11.7]	86.1 4.7 [1.3, 12]	86.0 4 [1.3, 11.3]
90-94	92.9 3.7 [1, 11]	94.7 3.3 [1, 10.3]	94.5 3 [1, 10.3]	91.4 3.7 [1, 11.3]	90.7 3.3 [1, 11]	90.4 3.7 [1, 12]	88.2 3.3 [1, 10.7]
95-99	96.8 1.3 [0.3, 8.3]	97.3 1.3 [0.3, 8]	98.0 1.3 [0.3, 7.7]	96.9 1.3 [0.3, 8.7]	96.9 1.3 [0.3, 8.7]	96.4 1.3 [0.3, 9]	95.7 1.3 [0.3, 8.7]

Table 14-17 Event rate analysis (non-surgical minimally-invasive, N=672)

HPI Range	55 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	60 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	65 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	70 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	75 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	80 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	85 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]
10-14	20.5 50 [8.7, 101.3]	19.7 44.3 [9.3, 102.3]	9.6 49.3 [9.7, 106]	12.5 45 [9, 101.3]	16.6 47.2 [9, 103]	18.6 48.3 [10, 101.7]	24.7 41 [8.3, 101.2]
15-19	20.6 45.7 [10.3, 103.3]	20.8 44 [8.3, 100.7]	12.6 44.3 [10, 105.6]	17.2 45.7 [8, 104]	21.8 44.3 [9.3, 100]	21.1 50 [10, 104]	29.1 40.7 [8.7, 98]
20-24	22.7 47 [10.7, 104.7]	21.8 43.3 [9, 101.3]	16.8 40.3 [8.3, 102.1]	20.0 45.7 [8.7, 101.7]	24.7 43.7 [8.7, 100.3]	24.1 48.3 [9.7, 104]	31.4 43 [7.7, 101.3]

Table 14-17 Event rate analysis (non-surgical minimally-invasive, N=672) (continued)

HPI Range	55 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	60 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	65 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	70 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	75 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	80 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	85 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]
25-29	22.9 47 [10, 103.3]	24.8 40.3 [8, 101.7]	20.9 39.3 [7.1, 102]	24.6 40.7 [7.7, 100.7]	26.8 42.7 [8, 101]	28.1 44.7 [9.3, 101.7]	33.5 41.7 [7.3, 100.3]
30-34	24.9 46.3 [8, 103.3]	29.1 39.7 [7.3, 100.3]	24.0 37.3 [7, 101]	29.8 36 [6.3, 98]	30.8 37.7 [7, 98.7]	31.4 42 [8.7, 98.7]	37.5 38.3 [6.7, 97.7]
35-39	30.4 42.3 [6.7, 100.3]	32.0 37.7 [6, 100]	31.4 30 [5.7, 93.5]	35.9 29.3 [5.3, 98]	34.6 30.7 [5.3, 96]	35.8 41 [7, 97.3]	44.7 34.7 [5.7, 95.3]
40-44	35.9 36.7 [6.3, 100]	37.5 33.3 [5.3, 98]	38.4 27.3 [5, 90]	41.9 24.7 [4.7, 94.7]	41.1 25.5 [4.3, 92.3]	40.9 35.3 [6.3, 95.3]	51.1 32.3 [5.3, 93]
45-49	39.7 31 [5.3, 96.3]	41.9 26.3 [4.7, 96]	44.5 23.7 [4, 90]	47.8 19 [3.7, 88.3]	43.7 23.7 [4, 94]	42.6 30.7 [5.3, 92.3]	52.3 28 [5, 91]
50-54	42.0 29 [5, 94.7]	46.0 21.3 [4, 92.7]	48.2 19.7 [4, 91]	52.1 17.3 [3.3, 81.3]	47.6 19.7 [3.3, 91]	44.3 30 [5, 94]	52.5 24.7 [3.7, 93]
55-59	46.2 27 [4.7, 93.3]	51.6 18 [3.3, 88.3]	55.9 17 [3.7, 87.9]	62.7 15.7 [3, 78.3]	53.2 17 [3, 85]	47.5 27 [4.3, 93.3]	56.3 20 [3.3, 87.3]
60-64	49.6 20.3 [4, 89]	58.1 15.7 [3, 83.4]	63.3 12.3 [2.7, 72.3]	71.9 12.3 [2.7, 76.3]	60.1 14 [2, 80]	53.5 19 [3, 89.3]	63.3 16 [2.7, 83.3]
65-69	61.1 12.7 [3, 77.7]	66.9 10.3 [2.3, 70.3]	69.7 9 [2, 52.3]	78.3 8.3 [1.7, 51]	69.5 10.3 [1.7, 68.7]	61.6 12.3 [2.3, 79.3]	70.6 11.3 [1.7, 75.1]
70-74	71.4 9 [2, 50.3]	73.9 8 [1.7, 48.3]	81.7 7 [1.7, 25.3]	87.1 6.3 [1.3, 23.7]	76.5 8 [1.3, 52.3]	68.8 9 [1.7, 65.7]	78.1 8.7 [1.3, 62.7]
75-79	83.1 7 [1.7, 18.3]	81.1 6.3 [1.3, 27.7]	88.2 6 [1.3, 17]	93.8 5 [1, 16]	83.9 6.7 [1, 34.7]	76.2 7 [1.3, 54.8]	80.3 6.7 [1, 50]
80-84	90.0 6 [1.3, 16]	88.9 5.3 [1, 17.3]	92.9 5 [1, 15.7]	96.5 4.3 [1, 14.3]	88.4 5.3 [1, 18]	81.8 6 [1, 37]	84.3 5.3 [1, 34]
85-89	95.9 5 [1.3, 14.3]	94.8 4.7 [1, 15]	95.8 4 [1, 13.7]	98.2 3.7 [1, 13]	92.7 4.3 [1, 16]	87.1 5 [1, 18.7]	88.3 4.3 [1, 16.7]
90-94	99.3 3.3 [1, 12.3]	97.7 3.3 [1, 13.3]	98.4 2.7 [1, 11.3]	99.2 2.7 [0.7, 11]	96.7 3.3 [1, 13]	93.1 3.7 [1, 14.7]	92.6 3.3 [1, 14]
95-99	99.9 1.3 [0.3, 8.3]	99.7 1.3 [0.3, 9.3]	99.7 1.3 [0.3, 7.7]	99.9 1 [0.3, 7.3]	99.5 1.3 [0.3, 9]	98.8 1.3 [0.3, 9.3]	99.2 1.3 [0.3, 9]

14.1.15 Clinical Validation with Hypotension Threshold in Non-Invasively Monitored Patients

Retrospective clinical validation studies were undertaken to assess the diagnostic performance of HPI to predict hypotensive and non-hypotensive events in non-invasively monitored surgical and non-surgical patients.

14.1.15.1 Surgical Patients

A retrospective clinical validation study assessed the diagnostic performance of HPI to predict hypotensive and non-hypotensive events in non-invasively monitored surgical patients. Table 14-18 provides the patient demographics of the 927 surgical patients included in this study. The number of hypotensive event segments included for each MAP target and a summary of patient data lengths (in minutes) are provided below in table 14-19.

The 927 surgical patients can be further stratified as described in table 14-20 below.

Table 14-18 Patient demographics (non-invasively monitored surgical patients, N=927)

Description	Value
# of Patients	927
Gender (Male/Female)	468/459
Age	57.9 ± 13.9
BSA (m ²)	2.0 ± 0.3
Height (cm)	171.8 ± 12.2
Weight (kg)	86.6 ± 23.7

Table 14-19 Analysis dataset length and number of hypotension events (non-invasively monitored surgical patients)

Proposed MAP targets	Number of hypotensive events	Median total data length (minutes)	Mean total data length (± Std Dev, minutes)
55 mmHg	971	178.3	193.1 ± 104.2
60 mmHg	2219		
65 mmHg	2561		
70 mmHg	2113		
75 mmHg	2894		
80 mmHg	2440		
85 mmHg	3381		

Table 14-20 Surgical patient characteristics (non-invasive, N=927)

Surgery Type	# patients (% of total)
cardiac	110 (11.9%)
colorectal	15 (1.6%)
cranial	7 (0.8%)
ear	14 (1.5%)
esophageal	6 (0.6%)
eye	47 (5.1%)
facial	24 (2.6%)

Table 14-20 Surgical patient characteristics (non-invasive, N=927) (continued)

Surgery Type	# patients (% of total)
foot	2 (0.2%)
gastro-Intestinal	65 (7.0%)
gynecology	61 (6.6%)
hand	1 (0.1%)
leg	10 (1.1%)
liver	11 (1.2%)
nose	9 (1.0%)
orthopedic	13 (1.4%)
other Non-Cardiac	367 (39.6%)
pancreas	12 (1.3%)
renal	27 (2.9%)
skin	6 (0.6%)
spinal	1 (0.1%)
thoracic	3 (0.3%)
unknown	92 (9.9%)
urology	24 (2.6%)
TOTAL	927 (100%)

14.1.15.2 Non-Surgical Patients

A retrospective clinical validation study assessed the diagnostic performance of HPI to predict hypotensive and non-hypotensive events in non-invasively monitored non-surgical patients. Table 14-21 provides the patient demographics of the 424 non-surgical patients included in this study. The number of hypotensive event segments included for each MAP target and a summary of patient data lengths (in minutes) are provided below in table 14-22.

The 424 non-surgical patients can be further stratified as described in table 14-23 below.

Table 14-21 Patient demographics (non-invasively monitored non-surgical patients, N=424)

Description	Value
# of Patients	424
Gender (Male/ Female)	286/138
Age	61.8 ± 14.2
BSA (m ²)	2.0 ± 0.2
Height (cm)	174.5 ± 9.7
Weight (kg)	83.0 ± 19.4

Table 14-22 Analysis dataset length and number of hypotension events (non-invasively monitored non-surgical patients)

Proposed MAP targets	Number of hypotensive events	Median total data length (minutes)	Mean total data length ($\pm$ Std Dev, minutes)
55 mmHg	648	417.3	415.8 $\pm$ 199.6
60 mmHg	689		
65 mmHg	1672		
70 mmHg	1312		
75 mmHg	2868		
80 mmHg	4375		
85 mmHg	4826		

Table 14-23 Non-Surgical patient characteristics (non-invasive, N=424)

Diagnosis	# patients (% of total)
cardiac	211 (49.8%)
cerebral	2 (0.5%)
gastro-Intestinal	8 (1.9%)
hypertension	1 (0.2%)
hypotension	5 (1.2%)
hypothermia	1 (0.2%)
internal	8 (1.9%)
intestinal	1 (0.2%)
liver	2 (0.5%)
neurological	69 (16.3%)
orthopedic	1 (0.2%)
post-Surgical	4 (0.9%)
pulmonary	7 (1.7%)
renal	2 (0.5%)
respiratory	40 (9.4%)
sepsis	18 (4.2%)
shock	4 (0.9%)
trauma	8 (1.9%)
vascular	32 (7.5%)
TOTAL	424 (100%)

14.1.15.3 Hypotension Threshold Clinical Validation Study Results – Non-Invasive Monitoring

The results of the receiver operating characteristic (ROC) analyses for surgical and non-surgical patients are summarized below in table 14-24 and table 14-25.

A hypotensive event is calculated by identifying a segment of at least 1 minute in length such that all data points in the section have a MAP less than the indicated Variable Map Value (55, 60, 65, 70, 75, 80, and 85). An event (positive) data point is chosen as the sample 5 minutes prior to the hypotensive event. If consecutive hypotension events are less than 5 minutes apart then a positive sample is defined as the first sample immediately following the preceding hypotension event.

A non-hypotensive event is calculated by identifying segments of data points such that the segment is at least 20 minutes away from any hypotensive events and all data points in that segment have MAP less than the indicated Variable Map Value (55, 60, 65, 70, 75, 80, and 85). One non-event (negative) data point is taken for each of the non-hypotensive event segments.

A true positive, as described in table 14-24 and table 14-25, is any event (positive) data point with HPI value greater than or equal to a chosen threshold. Sensitivity is the ratio of true positives to total number of events (positives) with a positive defined as a data point that is at most 5 minutes prior to a hypotensive event. A false negative is any positive data point with HPI value less than the threshold.

A true negative, as described in table 14-24 and table 14-25, is any negative (non-event) data point with HPI value less than a chosen threshold. Specificity is the ratio of true negatives to total number of non-events (negatives) with a negative defined as a data point that is at least 20 minutes away from any hypotensive event. A false positive is any negative data point with HPI value greater than or equal to the threshold.

Positive predictive value (PPV), negative predictive value (NPV), and area under the ROC curve (AUC) are also reported for each variable MAP target in table 14-24 and table 14-25.

Table 14-24 Clinical validation study results* (non-invasively monitored surgical patients)

HPI Threshold	Variable MAP Value (mmHg)	PPV [confidence interval]	NPV [confidence interval]	Specificity (%) [95% confidence interval]	Sensitivity (%) [95% confidence interval]	AUC
85	55	97.2 =(693/713) [94.9, 99.1]	94.3 =(4610/4888) [93.0, 95.3]	99.6 =(4610/4630) [99.2, 99.9]	71.4 =(693/971) [67.0, 75.0]	0.88 [0.86, 0.90]
	60	97.9 =(1738/1775) [96.6, 98.9]	89.8 =(4244/4725) [88.2, 91.3]	99.1 =(4244/4281) [98.6, 99.6]	78.3 =(1738/2219) [75.9, 80.6]	0.91 [0.89, 0.92]
	65	98.3 =(2011/2046) [97.1, 99.2]	89.2 =(4533/5083) [87.4, 90.7]	99.2 =(4533/4568) [98.7, 99.6]	78.5 =(2011/2561) [75.7, 81.1]	0.90 [0.89, 0.92]
	70	96.7 =(1457/1506) [94.9, 98.4]	88.7 =(5157/5813) [87.2, 90.1]	99.1 =(5157/5206) [98.5, 99.5]	69 =(1457/2113) [66.4, 71.5]	0.86 [0.85, 0.88]
	75	98.4 =(2075/2109) [97.4, 99.2]	85.6 =(4868/5687) [83.8, 87.1]	99.3 =(4868/4902) [98.9, 99.6]	71.7 =(2075/2894) [69.6, 73.7]	0.87 [0.85, 0.88]
	80	99.2 =(1761/1775) [98.4, 99.8]	81.4 =(2963/3642) [78.5, 83.8]	99.5 =(2963/2977) [99.1, 99.9]	72.2 =(1761/2440) [69.7, 74.4]	0.87 [0.86, 0.88]
	85	99.5 =(2586/2599) [98.9, 99.9]	69 =(1773/2568) [64.9, 72.8]	99.3 =(1773/1786) [98.4, 99.8]	76.5 =(2586/3381) [74.1, 78.6]	0.88 [0.87, 0.89]

*Note: Data on File at Edwards Lifesciences

Table 14-25 Clinical validation study results* (non-invasively monitored non-surgical patients)

HPI Threshold	Variable MAP Value (mmHg)	PPV [confidence interval]	NPV [confidence interval]	Specificity (%) [95% confidence interval]	Sensitivity (%) [95% confidence interval]	AUC
85	55	99.2 =(486/490) [97.5, 100.0]	98.2 =(8963/9125) [97.7, 98.6]	100 =(8963/8967) [99.9, 100.0]	75 =(486/648) [67.9, 80.7]	0.88 [0.85, 0.91]
	60	99.4 =(541/544) [98.0, 100.0]	97.9 =(6931/7079) [97.2, 98.5]	100 =(6931/6934) [99.9, 100.0]	78.5 =(541/689) [73.4, 82.7]	0.90 [0.87, 0.92]
	65	99.8 =(1422/1425) [99.4, 100.0]	97 =(8012/8262) [95.9, 97.8]	100 =(8012/8015) [99.9, 100.0]	85 =(1422/1672) [82.0, 87.6]	0.92 [0.91, 0.94]
	70	99.2 =(1033/1041) [98.5, 99.8]	97.1 =(9367/9646) [96.5, 97.7]	99.9 =(9367/9375) [99.8, 100.0]	78.7 =(1033/1312) [74.7, 82.1]	0.9 [0.87, 0.91]
	75	98.2 =(2499/2544) [95.8, 99.7]	95.3 =(7449/7818) [94.1, 96.3]	99.4 =(7449/7494) [98.6, 99.9]	87.1 =(2499/2868) [84.8, 89.3]	0.94 [0.93, 0.95]
	80	98.8 =(3866/3913) [97.7, 99.6]	90.8 =(5048/5557) [88.5, 92.8]	99.1 =(5048/5095) [98.3, 99.7]	88.4 =(3866/4375) [86.1, 90.3]	0.94 [0.94, 0.95]
	85	99.5 =(4218/4241) [98.8, 99.9]	84.2 =(3238/3846) [79.8, 87.7]	99.3 =(3238/3261) [98.4, 99.9]	87.4 =(4218/4826) [85.0, 89.5]	0.94 [0.94, 0.95]
*Note: Data on File at Edwards Lifesciences						

Table 14-26 provides the hypotensive event occurrence percentage and time-to-event data for a given HPI range for surgical patients in the clinical validation study. These data are presented using time windows that have been selected based upon how fast hypotensive events developed on average in surgical patients. Therefore based upon the clinical validation study data, table 14-26 presents data for surgical patients for a time-window of 15 minutes. This analysis is performed by taking samples in each patient from the validation dataset and looking forward in time for a hypotensive event within a 15-minute search window. Once a hypotensive event is found for a given sample then the time-to-event is noted, which is the time duration between the sample and the hypotensive event. The time-to-event statistic is the average event time of all samples that have an event within the search window.

Table 14-27 provides the hypotensive event occurrence percentage and time-to-event data for a given HPI range for non-surgical patients in the clinical validation study. These data are presented using time windows that have been selected based upon how fast hypotensive events developed on average in non-surgical patients. Therefore based upon the clinical validation study data, table 14-27 presents data for non-surgical patients for a time-window of 120 minutes. This analysis is performed by taking samples in each patient from the validation dataset and looking forward in time for a hypotensive event within a 120-minute search window. Once a hypotensive event is found for a given sample then the time-to-event is noted, which is the time duration between the sample and the hypotensive event. The time-to-event statistic is the average event time of all samples that have an event within the search window.

The event rates, included in table 14-26 and table 14-27, are the ratio of the number of samples that have an event within the search window to the total number of samples. This is done for samples in each MAP target for each of the individual HPI ranges between 10 to 99 as shown in table 14-26 and table 14-27.

CAUTION The HPI parameter information provided in table 14-26 and table 14-27 is presented as general guidance and may not be representative of individual experience. A review of the patient's hemodynamics is recommended prior to initiating treatment.

Table 14-26 Event rate analysis (surgical non-invasive, N=927)

HPI Range	55 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	60 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	65 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	70 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	75 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	80 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	85 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]
10-14	13.1 9.3 [4.9, 14]	25.0 8.7 [4, 13]	9.1 8.7 [3.7, 14]	15.1 9 [4, 14]	22.8 8 [4, 13.7]	29.3 8.7 [3.3, 13.7]	34.0 7.7 [3.3, 13]
15-19	9.4 9 [4.3, 14]	25.0 8.7 [3.5, 12.9]	10.9 9 [3.7, 14]	14.0 8.3 [3.7, 13.3]	21.6 8.3 [3.7, 13.6]	26.3 8 [3.3, 13.3]	41.1 8 [3.3, 13.7]
20-24	9.2 8.7 [4, 14.1]	23.0 9 [4, 13.7]	11.2 8.7 [3.4, 13.7]	15.8 8.3 [3.7, 13.7]	20.8 9 [3.7, 14]	24.5 8 [3.3, 13.7]	42.7 7.3 [2.3, 13.3]
25-29	11.4 8.7 [3.7, 13.6]	24.5 8.3 [3.7, 14]	13.8 8 [3.3, 13.3]	17.2 8.3 [3.3, 13.7]	23.0 8.7 [3.7, 14]	24.4 7.7 [3, 13.3]	37.8 7.7 [3, 13.7]
30-34	9.5 9 [4, 13.7]	23.1 9 [4.3, 14]	16.0 8.3 [3.3, 13.5]	19.3 8 [3.3, 13.7]	25.6 8.3 [3.7, 14]	26.4 7.7 [3, 13.3]	41.1 7.3 [2.7, 13.3]
35-39	12.3 8 [3.3, 13.7]	27.8 8.3 [3, 13.3]	18.8 8 [3, 14]	21.7 8 [3, 13.7]	29.5 8.7 [3.3, 14]	28.3 7.7 [3, 13.7]	41.3 7.3 [2.7, 13.7]
40-44	16.0 8.3 [3.7, 13]	30.3 8.7 [3, 13.7]	23.1 7.7 [3, 13.3]	24.8 8 [3, 13.3]	36.1 8.3 [3, 13.7]	30.0 7.7 [3.3, 13.7]	41.3 6.7 [2.3, 13]
45-49	20.0 7.7 [2.7, 13.7]	32.2 8 [3, 13.3]	29.9 7.7 [2.6, 13.7]	29.4 7.7 [2.7, 13.3]	39.0 7.7 [2.7, 13.3]	35.0 7.7 [3, 13.3]	45.0 6.7 [2, 13]
50-54	21.3 7.7 [2.3, 13.7]	35.8 7.3 [2.7, 13.3]	33.9 7.3 [2.7, 13.3]	33.9 7.3 [2.3, 13.7]	42.5 7.7 [2.7, 13.3]	37.7 7 [2.5, 13]	49.4 7 [2.3, 13.3]
55-59	22.2 7.3 [2.3, 13.3]	38.8 8.3 [2.3, 13.3]	33.9 7.7 [2.3, 13.7]	34.4 7.3 [2.3, 13.7]	43.9 7.7 [2.7, 13]	41.2 7.3 [2.3, 13]	53.1 7 [2.3, 13.3]
60-64	23.9 6.7 [2.3, 13.3]	44.9 6.7 [2.3, 12.7]	40.4 7.3 [2.3, 13.3]	39.4 7.3 [2, 13.7]	45.0 7.7 [2.7, 13]	43.1 7 [2.3, 12.7]	55.6 6.7 [2.3, 13]

Table 14-26 Event rate analysis (surgical non-invasive, N=927) (continued)

HPI Range	55 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	60 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	65 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	70 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	75 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	80 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	85 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]
65-69	32.9 6.3 [2, 13.3]	45.6 6.7 [2, 12.3]	44.5 6.7 [2.3, 13]	42.0 6.7 [2, 13.3]	50.0 7.3 [2.3, 13]	45.9 7 [2.3, 13.3]	61.0 7 [2.3, 13.3]
70-74	37.7 6 [1.7, 12.3]	52.1 6.7 [2, 13.3]	49.4 6.7 [2, 13]	47.1 6.7 [2, 13.3]	51.2 7 [2, 12.8]	52.0 6.7 [2, 13.3]	67.1 7 [2, 13]
75-79	45.3 5.7 [1.7, 11.7]	58.1 6 [2, 12.7]	56.9 6.3 [1.7, 12.7]	55.1 6 [2, 12.7]	63.0 6.3 [2, 12.7]	62.5 6.3 [1.7, 13.3]	70.6 6.3 [2, 13]
80-84	58.8 5.3 [1.7, 11.7]	69.4 6 [1.7, 12.7]	61.3 6.2 [1.7, 13]	63.8 5.7 [1.7, 12.7]	71.2 6 [2, 12.7]	66.5 5.7 [1.7, 12.7]	76.8 5.3 [1.7, 13]
85-89	83.4 4.3 [1.3, 11.3]	86.0 4.7 [1.3, 12.3]	82.2 5 [1.3, 12.3]	81.6 5 [1.3, 12.3]	84.0 5 [1.3, 12.7]	83.3 4.7 [1.3, 12]	87.9 5 [1.3, 12]
90-94	95.3 3 [1, 10.7]	93.8 3.3 [1, 11]	93.4 3.3 [1, 10.7]	93.6 4 [1, 11.7]	92.5 3.7 [1, 11.7]	91.6 3.7 [1, 11.3]	92.9 3.7 [1, 11]
95-99	97.3 1 [0.3, 7.3]	96.9 1.3 [0.3, 7.7]	97.7 1.3 [0.3, 8]	97.0 1.3 [0.3, 8.7]	96.7 1.3 [0.3, 8.3]	96.5 1.3 [0.3, 8.3]	96.6 1.3 [0.3, 8]

Table 14-27 Event rate analysis (non-surgical non-invasive, N=424)

HPI Range	55 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	60 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	65 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	70 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	75 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	80 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	85 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]
10-14	16.1 63.3 [21, 103.7]	8.9 53.3 [10.1, 99.1]	14.6 60.3 [13.3, 107.5]	13.3 47.5 [9.3, 95.4]	17.7 26.5 [9, 95.5]	16.8 44.8 [11, 101.6]	28.7 35.7 [6.3, 95]
15-19	15.0 58 [14, 105.3]	8.7 46.3 [10, 97.7]	15.4 52 [11, 99]	15.1 44 [8.7, 105.5]	18.6 41 [8.4, 96.3]	18.6 32 [9, 91.8]	33.4 34.3 [8.6, 97.7]
20-24	12.7 54.3 [8.6, 98]	9.8 51 [10, 102.7]	15.7 42.5 [8.3, 102]	18.3 47 [8.5, 101.8]	21.8 44.3 [6.9, 98.3]	23.0 35.3 [9.9, 100.3]	31.0 28.5 [7.3, 76.7]

Table 14-27 Event rate analysis (non-surgical non-invasive, N=424) (continued)

HPI Range	55 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	60 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	65 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	70 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	75 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	80 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]	85 mmHg MAP target Event rate (%); Time-to-Event in minutes: Median [10 th percentile, 90 th percentile]
25-29	11.2 49.6 [8, 99.8]	43 [7.7, 103.9] 10.6	15.0 46.2 [11, 103.7]	21.7 42.7 [8.4, 100.9]	22.1 39 [6.8, 95.3]	25.2 32.7 [8.7, 97]	32.3 28.7 [7.7, 90.8]
30-34	11.1 40.7 [11.3, 99.6]	12.8 37.7 [8.7, 98.9]	17.2 41.3 [9.3, 102.3]	25.3 37.5 [8, 100]	27.4 35.7 [7.3, 94.3]	28.7 34.9 [6.7, 95.3]	34.4 26.4 [7.7, 93.1]
35-39	14.6 36.3 [8.7, 99.3]	13.7 39.8 [10.3, 104.4]	19.5 37.2 [6.7, 100.3]	29.5 36.6 [8.6, 96.7]	32.3 33.7 [6.3, 93.3]	29.1 30.5 [6.3, 84.8]	40.0 28 [7.7, 81]
40-44	17.0 34.8 [7, 102.1]	14.9 41 [6.3, 105.4]	22.8 26.3 [6.3, 96.4]	33.6 32.3 [7.7, 98]	39.2 29.7 [6, 98.4]	36.1 25.7 [5, 81.6]	45.1 22.3 [5, 87.1]
45-49	21.7 32.3 [6.3, 97.7]	18.5 27.7 [6.3, 94.7]	27.3 23.3 [5.7, 92.3]	34.0 29.2 [6, 91.3]	44.3 24 [4.7, 86.9]	25.9 [4.3, 85.3] 43.5	50.1 23 [4.7, 81.9]
50-54	25.1 27.3 [5.7, 92.6]	23.2 20 [5.7, 91.4]	33.2 27 [6, 94.4]	37.9 28.7 [6, 93.6]	49.9 21.3 [5.7, 86.3]	46.5 22.7 [4.3, 83.8]	49.8 19.7 [4.7, 85.7]
55-59	28.3 24 [5.4, 85.7]	25.9 16.7 [4.3, 78.3]	32.4 22.8 [5, 96.7]	43.5 28.2 [5.9, 99.4]	53.8 20.7 [5.3, 84.4]	50.5 21.7 [5.3, 82.3]	52.6 18 [3, 78]
60-64	28.2 18 [4.3, 82.8]	28.2 16 [4.7, 79.1]	34.6 16.3 [4.7, 89.7]	46.4 26 [4.3, 93.3]	55.4 20.7 [5.3, 85.3]	53.6 22.7 [4.7, 80.3]	58.1 16.3 [3.3, 81.2]
65-69	33.1 14.7 [3.7, 60]	29.3 13.8 [4, 80.7]	42.3 14.3 [4, 78.5]	51.3 20.3 [4.3, 90.9]	61.7 18 [4.3, 83.8]	60.3 20.3 [3.7, 80.7]	66.6 13.3 [2.7, 69.8]
70-74	38.2 13 [3.3, 58.1]	34.3 11.5 [3, 72.4]	46.0 12 [3, 61.3]	54.3 17.7 [3.3, 81.8]	70.0 14.7 [3, 79.3]	68.4 16 [2.7, 80.7]	66.6 12.3 [1.7, 57.3]
75-79	51.5 9.3 [2.3, 24.5]	49.8 9.3 [3, 30.6]	60.7 11 [2.7, 63.6]	64.5 12 [3, 60.7]	77.7 10 [2, 44.9]	74.3 10 [1.7, 59.8]	75.6 8.3 [1.3, 39.3]
80-84	73.9 8.3 [2.3, 18.7]	73.5 9.3 [2.3, 19.2]	79.6 8 [2, 19.3]	76.7 9 [2, 35.1]	82.9 8.3 [1.4, 22.4]	83.8 7.7 [1.3, 43.7]	80.9 6.7 [1, 25.7]
85-89	95.8 7.3 [1.7, 16.7]	93.1 8 [1.7, 17.3]	96.3 5.3 [1.3, 15.4]	95.7 6.7 [1.3, 17.2]	95.6 6 [1.3, 15.7]	94.3 5.7 [1, 17.3]	92.4 5 [1, 16]

Table 14-27 Event rate analysis (non-surgical non-invasive, N=424) (continued)

HPI Range	55 mmHg MAP target Event rate (%); Time-to-Event in minutes; Median [10 th percentile, 90 th percentile]	60 mmHg MAP target Event rate (%); Time-to-Event in minutes; Median [10 th percentile, 90 th percentile]	65 mmHg MAP target Event rate (%); Time-to-Event in minutes; Median [10 th percentile, 90 th percentile]	70 mmHg MAP target Event rate (%); Time-to-Event in minutes; Median [10 th percentile, 90 th percentile]	75 mmHg MAP target Event rate (%); Time-to-Event in minutes; Median [10 th percentile, 90 th percentile]	80 mmHg MAP target Event rate (%); Time-to-Event in minutes; Median [10 th percentile, 90 th percentile]	85 mmHg MAP target Event rate (%); Time-to-Event in minutes; Median [10 th percentile, 90 th percentile]
90-94	99.5 4.7 [1, 14.3]	100 5.7 [1, 16]	99.7 3.7 [1, 12.3]	99.7 4.3 [1, 15.3]	99.8 3.7 [1, 13]	99.7 3.7 [1, 13.7]	99.3 3 [1, 12]
95-99	100 1 [0.3, 9.7]	100 1.3 [0.3, 10.9]	99.8 1 [0.3, 8]	99.9 1.3 [0.3, 9.3]	99.9 1.3 [0.3, 8.7]	99.8 1 [0.3, 7.7]	99.7 1 [0.3, 8]

14.1.16 Additional Clinical Data

14.1.16.1 Study Design

A prospective, single-arm, open-label, multicenter study of the hypotension prevention and treatment in patients receiving arterial pressure monitoring with Acumen Hypotension Prediction Index Feature (HPI study) was undertaken to further understand the impact that the Acumen Hypotension Prediction Index (HPI) feature with its available patient hemodynamic data may have in the detection of hemodynamic instability and the reduction of intraoperative hypotension in non-cardiac surgery. The comparison group was a retrospective historical control group (N=22,109) with patient-level data from a non-profit academic consortium group, the Multicenter Perioperative Outcomes Group (MPOG), that collects perioperative data from hospitals across United States. All subjects in this study were treated with an arterial line.

The primary objective of the HPI study was to determine whether the use of the Acumen HPI feature to guide intraoperative hemodynamic management in non-cardiac surgery reduces the duration of intraoperative hypotension (IOH, defined as MAP < 65 mmHg for at least 1 minute) as compared with a historic retrospective control group. The duration of IOH was measured in the same way for the MPOG control cohort and the HPI study prospective cohort. All IOH events were measured and reported. For a subject with multiple IOH events, the events were individually measured and combined across the total surgery time for each patient to obtain a measure of the total duration of IOH. The only difference is that the data for the MPOG cohort were provided in one-minute intervals and for the prospective cohort were provided in 20-second intervals.

The HPI study was a single-arm, unblinded study conducted in 485 eligible subjects (460 pivotal subjects with an additional 25 roll-in cases) at 11 study sites in the United States. No more than 97 subjects (20% of the total population) were enrolled per site. The same sites that contributed to this historical control group were studied prospectively to determine if using the Acumen HPI feature to predict hypotension within 15 minutes of an actual event could reduce the mean duration of IOH by at least 25% [11].

Inclusion and Exclusion Criteria. Potential subjects were excluded from study participation if during the screening and enrollment process it was determined that the following inclusion and exclusion criteria were met. Table 14-28 and table 14-29 list the inclusion and exclusion criteria applied during the study. Due to the available data for the MPOG groups subjects, there are slight differences in the inclusion and exclusion criteria for the HPI and MPOG groups. Specifically, the differences between the inclusion criteria are the investigator determination of moderate- or high-risk non-cardiac surgery and the identification of planned

overnight hospitalization. The relevant specific differences between the two listed exclusion criteria are: patients who are confirmed to be pregnant/nursing, known clinically important intra-cardiac shunts, and known moderate to severe aortic and mitral valve disease.

Table 14-28 HPI prospective subject selection criteria

Inclusion criteria	Exclusion criteria
1 Written informed consent 2 Age ≥ 18 years 3 ASA physical status 3 or 4 4 Moderate- or high-risk non-cardiac surgery (for example, orthopedic, spine, urology, and general surgery) 5 Planned pressure monitoring with an arterial line 6 General anesthesia 7 Surgery duration expected to last ≥ 3 hours from induction 8 Planned overnight hospitalization	1 Participating in another (interventional) study 2 Contraindication to the invasive blood pressure monitoring 3 Patient who is confirmed to be pregnant and/or nursing mothers 4 Emergency surgery 5 Known clinically important intra-cardiac shunts 6 Patient in whom an intraoperative MAP target will be < 65 mmHg 7 Known aortic stenosis with valve area ≤ 1.5 cm² 8 Known moderate to severe aortic regurgitation 9 Known moderate to severe mitral regurgitation 10 Known moderate to severe mitral stenosis 11 Patient or surgical procedure type known as an SVV limitation (e.g. tidal volume < 8 mL/kg of theoretical ideal weight, spontaneous ventilation, persistent cardiac arrhythmia, known atrial fibrillation, open chest surgery, Heart Rate/Respiratory Rate (HR/RR) ratio < 3.6) 12 Current persistent atrial fibrillation 13 Known acute congestive heart failure 14 Craniotomy 15 Burn surgeries 16 Patients with intra-aortic balloon pump (IABP) or ventricular assist device(s) 17 Patient transfer from ICU requiring multiple vasoactive agents and known diagnosis of ongoing active sepsis

Table 14-29 MPOG historical control patient selection criteria

Inclusion criteria	Exclusion criteria
1 Receiving care at an Institution planning on participating in the prospective study of Hypotension Prediction Index software 2 Surgery date between January 1, 2017 to December 31, 2017 3 Adult patients 18 years of age or greater 4 Elective same day admission or inpatient 5 American Society of Anesthesiologists (ASA) physical status 3 or 4 6 General anesthesia 7 Blood pressure monitoring using an invasive arterial line monitoring for $> 75\%$ case (to account for arterial lines placed post induction) 8 Case duration (as defined as patient in room time to patient out of room time) ≥ 180 minutes	1 Baseline mean arterial pressure < 65 mmHg (A blood pressure measurement obtained in the immediate preoperative period, or the first valid blood pressure intraoperatively, was determined to be the baseline) 2 Use of more than one vasoactive infusion intraoperatively (phenylephrine, norepinephrine, vasopressin, dopamine, dobutamine, or epinephrine) 3 Emergency surgery 4 Cardiac (on or off pump), burn debridement, or intracranial surgery

The incidence of IOH in the MPOG group was 88% ($n=19,445/22,109$) and the dates of treatment were between January 1, 2017 and December 31, 2017. The dates of enrollment for the HPI group were May 16, 2019 to February 24, 2020. The secondary effectiveness endpoint was the determination of total area under

the curve of the time and MAP for all time periods for which MAP < 65 mmHg in each subject. This endpoint is correlated with the duration and a descriptive analysis of this endpoint was presented with the mean, standard deviation (SD), median, minimum and maximum.

The primary safety endpoint was the percentage of serious adverse events to include perioperative events, postoperative complications, and device-related serious adverse events. The secondary objective for this study (secondary safety endpoint) was to determine if the guidance provided by the Acumen HPI feature reduced a composite measure of complications as indicated below.

- Postoperative episodes of non-fatal cardiac arrest
- In-hospital death
- Stroke
- Acute Kidney Injury (AKI) within 30 days of the procedure
- Myocardial Injury in non-cardiac surgery (MINS) within 30 days of the procedure

14.1.16.2 Patient Demographics

Table 14-30 and table 14-31 provide a summary of the available patient demographic information for the prospective clinical cohort (HPI) and the historical control cohort (MPOG) as well as the procedure types undergone by the subjects in the HPI cohort.

Table 14-30 Patient demographics (MPOG study)

Description		HPI (Intent-to-treat)	HPI (Full analysis set)	MPOG (Full analysis set)
# of Patients		460	406*	22,109
Gender	Male	51.7 (n=238)	53.0 (n=215)	57.8 (n=12,779)
	Female	48.3 (n=222)	47.0 (n=191)	42.2 (n=9330)
Age (year)	Mean $\pm$ SD	63.0 $\pm$ 12.97	62.8 $\pm$ 13.0	65.3 $\pm$ 13.8
	Median (min - max)	65 (19 - 94)	65 (19 - 89)	65 (18 - 90)
BMI	Median (25 th and 75 th percentile)	28.09 (24.37, 32.81)	28.09 (24.41, 32.86)	28.1 (24.2, 32.9)
ASA score	II**	0.2 (n=1)	0.25 (n=1)	0.0 (n=0)
	III	91.5 (n=421)	92.1 (n=374)	80.83 (n=17,870)
	IV	8.0 (n=37)	7.6 (n=31)	19.17 (n=4239)
	Not Specified	0.2 (n=1)	0.0 (n=0)	0.0 (n=0)
Surgery duration (minutes, N=458)	Mean $\pm$ SD	338.1 $\pm$ 145.4	363.6 $\pm$ 134.0	355.2 $\pm$ 145.8
	Median (25 th and 75 th percentile)	315.5 (235, 416) (n=458)	336 (262, 430)	317 (245, 427)
*The Full Analysis Set (FAS) represents those subjects from the Intent-to-Treat (ITT) population that had a surgery duration of ≥ 3 hours. **ASA II subject was identified as a protocol deviation, though not excluded from ITT and FAS populations as this subject met the defined criteria (surgery >3 hours and hemodynamic monitoring data). This subject was included in the efficacy and safety analyses, although by inclusion/exclusion criteria should not have been enrolled in the study.				

Table 14-31 Procedure type (HPI)

Procedure type	% (n/N)
Spine surgery	18.5 (85/460)
Hepatectomy	13.7 (63/460)
Whipple	10.0 (46/460)
Major vascular	8.5 (39/460)
Other	8.5 (39/460)
Nephrectomy	5.7 (26/460)
Other genitourinary surgery	5.4 (25/460)
Cystectomy	5.0 (23/460)
Pancreatectomy	5.0 (23/460)
Renal transplant	4.3 (20/460)
Head & neck surgery	3.9 (18/460)
Complex combined oncologic surgery (including 2 or more distinct organs)	3.0 (14/460)
Exploratory laparotomy	3.0 (14/460)
Colectomy	2.8 (13/460)
Adrenalectomy	2.6 (12/460)
Gastrectomy	2.0 (9/460)
Other gastrointestinal surgery	2.0 (9/460)
Hip revision	1.7 (8/460)
Prostatectomy	1.7 (8/460)
HIPEC	1.3 (6/460)
Hysterectomy with debulking	1.3 (6/460)
Cholecystectomy	0.9 (4/460)
Reoperative orthopedic surgery	0.9 (4/460)
Splenectomy	0.9 (4/460)
Bariatric surgery	0.4 (2/460)
Liver transplant	0.4 (2/460)
Sigmoidectomy	0.4 (2/460)
Not specified	0.2 (1/460)

MPOG group surgery types were determined by Current Procedural Terminology (CPT) grouping. The MPOG group included head and neck; thorax extra- and intra-thoracic; spine and spinal cord; abdomen upper or lower; urology; gynecologic; male reproductive system; pelvis; hip/leg/foot; shoulder/arm/hand; radiologic; obstetrics; and other procedures.

Table 14-32 presents comparison of surgery types for the HPI and MPOG group surgery types as determined by CPT grouping.

Table 14-32 Surgery type by CPT grouping

Surgery type	HPI		MPOG	
	Number of Patients	Percentage of Total	Number of Patients	Percentage of Total
Head and neck	18	3.4	2024	10.2
Thorax surgery	0	0	3257	16.5
Spine surgery	85	16.2	3331	16.8
Upper abdomen	157	29.9	3838	19.4
Lower abdomen	40	7.6	1314	6.6
Urologic	114	21.7	2017	10.2
Gynecologic/obstetric	20	3.8	190	1.0
Orthopedic	12	2.3	2224	11.2
Major vascular	39	7.4	0	0
Other	40	7.6	1596	8.1
<i>Note: IOH duration by surgery type is not available for the MPOG population.</i>				

14.1.16.3 Study Results

Table 14-33 provides the results of the receiver operating characteristics (ROC) analysis for all HPI subjects with available data for analysis (N=482). The ROC analysis presented in table 14-33 is identical to the analysis performed for the clinical validation studies, presented earlier in table 14-12. For a detailed description of how hypotensive events, non-hypotensive events, sensitivity, and specificity are defined and calculated in table 14-33, see *Hypotension Threshold Clinical Validation Study Results – Minimally-Invasive Monitoring* on page 275.

Table 14-33 Receiver operating characteristics (ROC) for HPI subjects (N=482)*

HPI Threshold	PPV (%) [95% confidence interval]	NPV (%) [95% confidence interval]	Specificity (%) [95% confidence interval]	Sensitivity (%) [95% confidence interval]	AUC
85	98.4 (=821/834) [97.6, 99.3]	90.3 (=6782/7507) [89.7, 91.0]	99.8 (=6782/6795) [99.7, 99.9]	53.1 (=821/1546) [50.6, 55.6]	0.84
<i>*Note: Data on File at Edwards Lifesciences</i>					

Effectiveness. The HPI study was designed to evaluate the ability of the Acumen HPI feature, as a decision support tool, to reduce the duration of IOH by at least 25% in surgical patients that require advanced hemodynamic monitoring. An episode of intraoperative hypotension (IOH) was defined as a mean arterial pressure (MAP) below 65 for three (3) or more consecutive 20-second events for each subject, across all sites.

The primary effectiveness endpoint is a weighted average of site means and standard deviations combined in the same proportion of subjects that were included in the MPOG cohort. This weighted average and its properly computed standard deviation was compared to the estimates obtained from the subjects of the MPOG cohort.

The HPI study met its primary effectiveness endpoint. The HPI pivotal subjects of the full analysis set experienced a mean IOH duration of 11.97 ± 13.92 minutes compared with the MPOG historical control mean IOH of 28.20 ± 42.60 minutes. Table 14-34 demonstrates that this result was a reduction of 57.6% compared to the MPOG historical control ($p < 0.0001$). When considering instances where there were zero episodes of IOH experienced during surgery, there was a 65% reduction of IOH ($p < 0.0001$).

Table 14-34 Mean IOH duration – Primary effectiveness endpoint

Statistics	HPI (subject=406)	MPOG (subject=22,109)	p value
Sample size (n)	293	19,446	--
Total IOH minutes	3508	548,465	--
IOH mean (mins)**	11.97	28.20	<0.0001*
IOH STD	13.92	42.60	--
<i>Note: IOH estimated with stand method; STD estimated with pooled method (pivotal subject with IOH episode in test arm).</i> <i>Standard Method - IOH episode is defined with at least three consecutive observations having MAP < 65. FAS pivotal subjects, with at least 3-hour surgery time.</i> <i>*One-sided unequal variances t-test was used in analysis. Nominal alpha for the test is 0.025.</i> <i>**When the HPI cohort data are analyzed using 60-second interval the mean IOH duration increased slightly from 11.97 to 12.59 which remains statistically significantly different from the MPOG 28.20 IOH Mean with a p value <0.0001.</i>			

The results of the secondary effectiveness endpoint, determination of total area under the curve (AUC) of the time, and MAP for all time periods for which MAP < 65 mmHg in each Subject, are included in table 14-35.

Table 14-35 Intraoperative hypotension AUC - ITT, pivotal subjects

Study category	Subject	AUC mean (min*mmHg)	AUC SD (min*mmHg)	AUC median (min*mmHg)	AUC range (min*mmHg)	AUC Q3-Q1 (min*mmHg)
All pivotal subjects	457	46.38	82.75	16.67	833.00	54.00
All pivotal subjects with at least one episode	328	64.63	91.46	32.33	832.00	68.00
All pivotal subjects with ≥ 3 hours surgery duration	406	47.07	85.30	16.83	833.00	51.00
All pivotal subjects with ≥ 3 hours surgery duration and at least one IOH episode	293	65.23	94.36	32.00	832.00	62.67
All pivotal subjects with <3 hours surgery duration	51	40.89	58.94	12.33	291.00	71.33
All pivotal subjects with <3 hours surgery duration and at least one IOH episode	35	59.58	62.94	37.00	290.00	73.33
<i>Note: Standard Method - IOH episode is defined with at least three consecutive observations having MAP < 65. ITT pivotal subjects, with valid surgery time.</i>						

An analysis was undertaken to assess the effectiveness of HPI in the reduction of IOH when stratified by MAP level. The duration of IOH was compared between the HPI group and the MPOG group stratified by MAP level between 50 and 70 mmHg, using the standard calculation method. Table 14-36 shows that at all MAP levels, except for MAP < 50, the mean IOH duration in HPI Study Subjects was statistically significantly smaller than that reported for each MPOG MAP level.

Table 14-36 Effectiveness stratified by MAP level, HPI study versus MPOG historical control

MAP value	Statistic	HPI (subject=406)	MPOG (subject=22,109)	p value
MAP < 50	Sample size (n)	28	8555	--
	Total IOH minutes	97	35,790	--
	IOH mean (minutes)	3.45	4.20	0.1967
	IOH STD	3.56	13.10	--
MAP < 55	Sample size (n)	84	12,484	--
	Total IOH minutes	341	80,115	--
	IOH mean (minutes)	4.06	6.40	<0.0001
	IOH STD	4.30	15.40	--
MAP < 60	Sample size (n)	188	16,561	--
	Total IOH minutes	1098	212,362	--
	IOH mean (minutes)	5.84	12.80	<0.0001
	IOH STD	7.31	24.10	--
MAP < 65	Sample size (n)	293	19,446	--
	Total IOH minutes	3508	548,465	--
	IOH mean (minutes)	11.97	28.20	<0.0001
	IOH STD	13.92	42.60	--
MAP < 70	Sample size (n)	375	20,986	--
	Total IOH minutes	10,241	1,185,983	--
	IOH mean (minutes)	27.31	56.50	<0.0001
	IOH STD	28.79	70.40	--
<i>Note: Standard Method - IOH episode defined as at least three consecutive observations with MAP < MAP value defining IOH. FAS pivotal subjects with surgery duration at least 3 hours are included. Student's t-test was applied as specified in the SAP.</i>				

During the clinical study, the reduction in the duration of intraoperative hypotension was dependent upon clinical judgment as to when, what and how treatment was administered with guidance from the HPI parameter and HPI secondary screen. Intervention types included: colloid, crystalloid, blood products, vasopressors, and inotropes. Of particular interest was a comparison of frequency pattern of subjects and intervention by HPI threshold, meaning when the HPI parameter was predicting a hemodynamic instability

(HPI > 85). See table 14-37. These data suggest that HPI added value by providing an alert and providing insight through the secondary screen that allowed the clinician to implement more timely and appropriate interventions.

Table 14-37 Frequency pattern of subjects and intervention instances by HPI threshold

Intervention type	HPI group	Study subject				Intervention instance			
		N	n	n/N (%)	p value ^a	N	n	n/N (%)	p value ^b
Colloid	HPI>85	78	58	74.4	0.0004	134	87	64.9	<0.0001
	HPI≤85	78	36	46.2		134	47	35.1	
Crystalloid	HPI>85	163	134	82.2	<0.0001	360	250	69.4	<0.0001
	HPI≤85	163	80	49.1		360	110	30.6	
Blood products	HPI>85	24	18	75.0	0.0781	56	34	60.7	0.0245
	HPI≤85	24	12	50.0		56	22	39.3	
Vasopressor	HPI>85	307	277	90.2	<0.0001	1604	1156	72.1	<0.0001
	HPI≤85	307	189	61.6		1604	448	27.9	
Inotrope	HPI>85	87	72	82.8	<0.0001	187	131	70.1	<0.0001
	HPI≤85	87	39	44.8		187	56	30.0	

a, b: p value from logistic regression model with HPI ≤ 85 as the reference, a - subject, b - intervention instance. N = total subjects or total intervention instances, n = subjects or instances with intervention.

Safety. The Acumen HPI feature was shown to be safe when used in surgical patients that require advanced hemodynamic monitoring.

- There were no Subjects with events adjudicated to have any relationship to the Acumen HPI feature.
- There were no ADEs or SADEs adjudicated as related to the Acumen HPI feature.
- There were no unanticipated ADEs (0%) related to the HPI feature.
- There were no deaths that occurred whether related/unrelated to HPI feature.

The secondary safety endpoint is a descriptive statistic that was a composite of 30-day postoperative AEs in the completed cases (CC) population. Table 14-38 shows the components of the 30-day postoperative composite endpoint for the completed cases (CC) population. The results demonstrate that the composite event rate was 4.75% (composite events =19 [95% CI: 2.88, 7.32]), with one subject experiencing more than one of the individual composite elements. The safety data collected for the MPOG arm included mortality (375, 1.83%); AKI Stage 1 (2068, 9.35%); AKI Stage 2 (381, 1.72%); AKI Stage 3 (152, 0.69%); and, Myocardial Injury [MINS] (178, 0.81%).

Table 14-38 HPI Study - 30 days postoperative composite endpoint components - CC analysis population (pivotal subjects, n=400)

Analysis endpoint	AE event		POD post-surgery days		
	Events n (%)	95% CI	Mean	Median	Range
Postoperative non-fatal cardiac arrest	1 (0.25)	0.01, 1.38	2.00	2.00	2, 2
In-hospital death	0 (0.00)	0.00, 0.92	N/A	N/A	N/A
Stroke	0 (0.00)	0.00, 0.92	N/A	N/A	N/A
Acute kidney injury - overall	16 (4.00)	2.30, 6.41	5.94	1.00	0, 27
Acute kidney injury - stage 1	11 (2.75)	1.38, 4.87	6.82	1.00	0, 27
Acute kidney injury - stage 2	3 (0.75)	0.15, 2.18	6.33	7.00	2, 10
Acute kidney injury - stage 3	2 (0.50)	0.06, 1.79	0.50	0.50	0, 1
Myocardial injury (MINS)	3 (0.75)	0.15, 2.18	1.67	1.00	0, 4
CC = complete (evaluable) group, CI = confidence interval, post-surgery days (POD) = AESTDT-SGDT					

Analysis of the intent-to-treat population (n=460) yielded 3 (0.066%) instances of myocardial injury (MINS) and 17 (3.7%) incidents of acute kidney injury (AKI).

Length of stay in the hospital and the ICU for the HPI cohort is in table 14-39.

Table 14-39 Length of stay

Endpoint	n	Mean	Median	Range		95% exact CI	
				Min	Max	Lower	Upper
Hospital length of stay (LOS) days	455	6.8	5.3	0.3	50.5	6.2	7.3
ICU length of stay (LOS) days	151	2.7	2.0	0.1	27.0	2.2	3.1

14.1.16.4 Study Summary

These results demonstrate a substantial reduction in mean IOH that was consistent across most sites; most sites had a > 25% reduction in its mean duration of IOH, with all sites but one exceeding 35%; ranging from a 23% to 72% mean IOH reduction. The findings of the study showed a reduction of the duration of IOH to 11.97 minutes (SD 13.92), representing a 57.6% reduction ($p < 0.0001$). This reduction is clinically relevant, as IOH lasting at least 1-minute has been associated with perioperative complications and morbidity such as AKI, MINS and stroke [12].

Sensitivity analyses, including review of pooling of study sites, confounding factors and subjects excluded from the intent-to-treat cohort did not materially change this clinically relevant finding of reduction in mean intraoperative hypotension (IOH).

The results demonstrate that Acumen HPI feature was shown to be safe when used in surgical patients that require advanced hemodynamic monitoring, with no device-related adverse events. Additionally, the composite event rate of 4.75% (composite events = 19 [95% CI: 2.88, 7.32]) is low when considering that the subjects were ASA Physical Status 3 and 4 undergoing non-cardiac surgery.

In this unblinded prospective-to-historical comparison study design, IOH was demonstrated to be reduced with the use of the HPI software feature. This study has limitations secondary to potential bias associated with clinician awareness in the prospective arm and the comparison to a historical cohort.

14.1.16.5 Conclusion

The results of this study are robust and provide valid scientific evidence that the Acumen HPI feature is safe and provided a statistically and clinically significant reduction in mean IOH. Therefore, Acumen HPI is effective in detecting hemodynamic instability and substantially reducing the amount of intraoperative hypotension when used in surgical patients who require intraoperative hemodynamic monitoring during non-cardiac surgery.

14.1.17 References

- 1 De Hert et al, Evaluation of Left Ventricular Function in Anesthetized Patients Using Femoral Artery dP/dtmax. *Journal of Cardiothoracic and Vascular Anesthesia* 2006; 20(3): 325-330.
- 2 Tartiere et al, Non-invasive radial pulse wave assessment for the evaluation of left ventricular systolic performance in heart failure. *Eur Journal of Heart Failure* 2007; 9: 477-483.
- 3 Monge Garcia MI, Orduna PS, Cecconi M. Understanding arterial load. *Intensive Care Med* 2016; 42: 1625-1627.
- 4 Monge Garcia MI, Manuel Gracia Romero MG, Cano AG, Aya HD, Rhodes A, Grounds RM, Cecconi M. Dynamic arterial elastance as a predictor of arterial pressure response to fluid administration: a validation study. *Critical Care* 2014; 18: 626-637.
- 5 Cecconi M, Monge Garcia MI, Romero MG, Mellinshof J, Caliendo F, Grounds RM, Rhodes A. 2015. The Use of Pulse Pressure Variation and Stroke Volume Variation in Spontaneously Breathing Patients to Assess Dynamic Arterial Elastance and to Predict Arterial Pressure Response to Fluid Administration. *Anesth Analg* 2015; 120: 76-84.
- 6 Cannesson M, Le Manach Y, Hofer CK, Goarin JP, Lehot JJ, Vallet B, Tavernier B. Assessing the diagnostic accuracy of pulse pressure variations for the prediction of fluid responsiveness. A “gray zone” approach. *Anesthesiology* 2011; 115: 231-241.
- 7 Cannesson M, Musard H, Desebbe O, Boucau C, Simon R, Henaine R, Lehot JJ. The Ability of Stroke Volume Variations Obtained with Vigileo/FloTrac System to Monitor Fluid Responsiveness in Mechanically Ventilated. *Anesth Analg* 2009; 108: 513-517.
- 8 Pinsky MR. Protocolized Cardiovascular Management Based on Ventricular-arterial Coupling. In: *Functional Hemodynamic Monitoring. Update in Intensive Care and Emergency Medicine* (44). Springer-Verlag, Berlin, 2004; 381-395.
- 9 Sunagawa K, Maughan WL, Burkhoff D, Sagawa K. Left ventricular interaction with arterial load studied in isolated canine ventricle. *Am J Physiol, Heart Circ Physiol* 1983; 245: H773-H780.
- 10 Chantler PD, Lakatta EG, Najjar S. Arterial-ventricular coupling: mechanistic insights into cardiovascular performance at rest and during exercise. *J Appl Physiol* 2008; 105: 1342-1351.
- 11 Shah NJ, Mentz G, Kheterpal S. The incidence of intraoperative hypotension in moderate to high risk patients undergoing non-cardiac surgery: A retrospective multicenter observational analysis. *J Clin Anest.* 2020; 66: 109961.
- 12 Salmasi V, Maheshwari K, Yang D, Mascha EJ, Singh A, Sessler DI, Kurz A. Relationship between intraoperative hypotension, defined by either reduction from baseline or absolute thresholds, and acute kidney and myocardial injury after noncardiac surgery: A retrospective cohort analysis. *Anesthesiology.* 2017 Jan; 126(1): 47-65.

14.2 Assisted Fluid Management

The Acumen assisted fluid management (AFM) software feature provides clinical decision support for the management of patient fluids.

14.2.1 Introduction

WARNING The Assisted Fluid Management feature should not be used exclusively to treat the patient. A review of the patient's hemodynamics is recommended throughout the monitoring session to assess fluid responsiveness.

The assisted fluid management (AFM) feature cycles through various states during a session. Table 14-40 describes each of these states.

Table 14-40 AFM states

State	AFM dashboard notification	Definition
Prompted	Fluid Bolus Suggested / Test Bolus Suggested	A notification that has prompted the user to either (1) accept and inform the monitor that fluid administration has started or (2) decline the suggestion.
Not prompted	Fluid Not Suggested	Fluid is not suggested.
Decline	AFM Suggestions Suspended	An action by the user to decline the AFM prompt which places the AFM feature in a 5-minute quiet period with no new notifications.
Accepted	Bolus in Progress...	A fluid bolus that the user has accepted and elected to start. " Bolus in Progress... " may also appear after initiating a User Bolus .
Analysis declined		A fluid bolus that the user has declined to analyze and will not be presented to the AFM software for analysis.
Completed	Bolus Complete	A fluid bolus that the user has completed.
Analyzing	Bolus Complete; Analyzing Hemodynamic Response...	A fluid bolus that has been analyzed by AFM. It was delivered within the prescribed rate and volume limits and has the required information to assess the hemodynamic response to the fluid.

14.2.2 Principle of Operation

The AFM software feature has been designed to guide optimal intravenous fluid administration. It includes a rule-based algorithm to make fluid management suggestions by recognizing patterns of fluid responsiveness using a patient's hemodynamic data and past responses to fluid administration. Its inputs are:

- User settings (i.e., **Fluid Strategy** [desired change in stroke volume: 10%, 15% or 20%], **Surgery Mode** [**Open** or **Laparoscopic/Prone**]).
- Hemodynamic data from arterial pressure-based analysis (pulse rate [PR], mean arterial pressure [MAP], stroke volume [SV], stroke volume variation [SVV], systemic vascular resistance [SVR], and the rate of SV change over the past two minutes).
- Fluid delivery data (start time and stop time of the fluid bolus and the fluid bolus volume).

- Fluid responsiveness is derived from stroke volume changes as measured by the Acumen IQ sensor, and AFM fluid suggestions are derived from the predicted increase in stroke volume computed in part by measure of fluid responsiveness. This prediction is based upon a combination of the information derived from:
 - * **Patient population model.** This utilizes data on the relationship between percent increase in stroke volume (% Δ SV) and stroke volume variation (SVV) from patient responses to the administration of 500 mL fluid at different SVV levels (N = 413 patients).¹
 - * **Individual patient bolus history.** This utilizes the fluid administration response of the currently monitored patient.

The combined information allows the algorithm to determine a delta stroke volume by identifying boluses that were given in a similar hemodynamic state and aggregating their responses, taking into account systematic biases (i.e., the model is over-estimating or under-estimating the patient's actual response to fluid) and weighting the prediction by the quality of the information in the patient bolus history to provide a final prediction.

- The final prediction is compared to the chosen fluid strategy to determine if a fluid suggestion should be generated. If the predicted delta stroke volume is greater than the selected fluid strategy, then the output of the algorithm is a fluid suggestion prompt on the hemodynamic monitor. If the predicted stroke volume is not greater than the selected fluid strategy, the algorithm either does not output a fluid suggestion, or if there is limited information in the patient bolus history the algorithm may prompt a test bolus. For further information regarding possible AFM status, please refer to Table 14-41 on page 305.
- The fluid suggestions generated by the AFM software feature are focused on SV and CO and independent of MAP. Therefore, AFM may suggest fluid when a patient is normotensive. A full review of the patient's hemodynamic status is recommended prior to accepting an AFM recommendation or AFM test suggestion.

CAUTION The Assisted Fluid Management software feature relies on information provided by the clinician to accurately assess fluid responsiveness.

It is important to correctly identify **Surgery Mode** and **Fluid Strategy**. The selected **Surgery Mode** and **Fluid Strategy** influences AFM fluid suggestions. Selecting the incorrect **Surgery Mode** or **Fluid Strategy** can impact the frequency of AFM suggestions. It is also important that fluid administration information (volume and duration) is accurately entered into the system. See *Assisted Fluid Management Settings* on page 303 for more information about **Fluid Strategy** and **Surgery Mode**. See *Managing Fluids with the AFM Algorithm* on page 307 for more information about fluid administration.

If the AFM software feature estimates that a patient will be fluid responsive, it will provide a message suggesting fluid administration may improve the hemodynamic status of the patient. If the AFM software feature estimates that a patient will not be responsive to fluid, the system will not suggest fluid administration.

The AFM feature includes the display of relevant hemodynamic parameters and provides real-time tracking of current patient status and total fluid volume administered for each individual patient. The AFM feature is available when an Acumen IQ sensor is connected to a radial arterial catheter.

1. Cannesson M, Le Manach Y, Hofer CK, Goarin JP, Lehot JJ, Vallet B, Tavernier B. Assessing the diagnostic accuracy of pulse pressure variations for the prediction of fluid responsiveness: a "gray zone" approach. *Anesthesiology*. 2011 Aug;115(2):231-41.

CAUTION Fluid management suggestions provided by the AFM feature can be compromised by factors such as:

- Inaccurate FT-CO measurements
- Acute changes in FT-CO measurements secondary to vasoactive medication administration, patient repositioning or surgical interventions
- Bleeding at rates similar to, or greater than, the rate of fluid delivery
- Arterial line interference

Always review patient hemodynamic status before complying with AFM suggestions.

Accurate stroke volume variation (SVV) measurement is necessary for the AFM software feature to make fluid management suggestions. Patients must be:

- mechanically ventilated
- have a tidal volume of ≥ 8 mL/kg

NOTE When using both AFM and HPI smart alerts simultaneously, it is important to consider that AFM fluid recommendation behavior is based upon a prediction of fluid responsiveness, while HPI smart alert behaviors are based upon identification of potential underlying mechanism(s) to prevent or treat hypotension. As such, these two software features are considering different targets and patient hemodynamic conditions, and should be considered independently. Current patient hemodynamics should be reviewed prior to determining the most appropriate course of action. See *Acumen Hypotension Prediction Index (HPI) Software Feature* on page 253 for more information on that feature.

14.2.3 Help Screens for AFM

AFM help screens are available to support many common user questions. To access AFM help screens, touch the help icon on the AFM dashboard or on any of the Fluid Bolus prompts.



AFM help screens can also be accessed through the main help menu. Touch the settings icon



Help tab  **Help** → **Assisted Fluid Management** button.

The AFM help screens include content about getting started, using the AFM feature, and common questions about how the system works. On each AFM help screen, touch the question that interests you to see a brief answer. For additional information, contact your Edwards representative.

14.2.4 Starting or Restarting AFM

- 1 Touch the GDT tracking/AFM icon on the navigation bar.



- 2 Select the Assisted Fluid Management icon. The AFM dashboard will be displayed only on a graphical trends screen.



NOTE If Assisted Fluid Management is started during an active GDT tracking session, the user will be notified that this will end their current tracking session.

- 3 Set the desired AFM settings for **Surgery Mode (Laparoscopic/Prone or Open)**, **Fluid Strategy (10%, 15%, or 20%)**, and **Fluid Tracking (Fluid Meter or Manual)**. See *Assisted Fluid Management Settings* on page 303.
- 4 Enter the **Maximum Case Volume (Max Case Vol.)** on the keypad. Entering this value is required to start an AFM session.

The **Maximum Case Volume** provides the user with a target fluid volume based upon available information at the start of the case. A patient's fluid needs may change over the course of the case and therefore this value should be considered as a guide and not the absolute threshold between optimal and excessive fluid delivery.

During an active AFM session an alert is displayed on the status bar when the total fluid delivered through the AFM feature approaches (within 500 mL) or exceeds the pre-set **Maximum Case Volume** to guard against potential fluid overload. The **Maximum Case Volume** value does not limit the functionality of the AFM feature or influence AFM fluid suggestions. This value may be changed from AFM settings screen at any time during an active AFM session by touching the settings icon  on the AFM dashboard.

NOTE In the event of power loss during an AFM session, it must be re-initialized upon return of power. If monitoring with the same patient is resumed after powering back on the monitor, the history of boluses given to the current patient is cleared; however, the total volume delivered through the AFM feature and the **Maximum Case Volume** value remain.

- 5 Touch **Start AFM** icon  on the AFM dashboard.




Figure 14-12 AFM dashboard

14.2.5 AFM Dashboard Display

The AFM dashboard (shown in Figure 14-12) can be displayed while on a graphical trends screen and an AFM session is active. The AFM dashboard can be minimized at any time by touching the minimize icon



or the GDT tracking/AFM icon  on the navigation bar.

When the AFM dashboard is minimized, the fluid status icon is displayed on the navigation bar. To restore the AFM dashboard, touch the fluid status icon  on the navigation bar. See Table 14-41 on page 305.

14.2.6 Assisted Fluid Management Settings


Review all settings before starting an AFM session. An AFM session cannot be started without setting the **Maximum Case Volume**. To adjust settings related to the Assisted Fluid Management feature, touch the settings icon at the right edge of the AFM dashboard.



14.2.6.1 Fluid Strategy

It is important to correctly identify the **Fluid Strategy**. The selected fluid strategy influences AFM fluid suggestions. Selecting a **Fluid Strategy** that is not aligned to the clinician's fluid management strategy will lead to undesired fluid suggestions (e.g., clinician desires a restrictive fluid strategy but chooses **10% Fluid Strategy** in AFM settings) or a lack of fluid suggestions (e.g., clinician desires liberal fluid strategy but chooses **20% Fluid Strategy** in AFM settings).

For **Fluid Strategy**, select either **10%**, **15%**, or **20%**.




NOTE

Fluid Strategy can be used to adjust the AFM algorithm to be more liberal (**10%**) or restrictive (**20%**) in suggesting fluid. The default setting is **15%**. This percentage is the percent change in stroke volume in response to a 500 mL bolus of fluid. It is not necessary to administer a 500 mL bolus of fluid to use the AFM software feature. The percent change is adjusted to align with the volume of fluid delivered. A lower percentage indicates a lower threshold for suggesting fluid, and is therefore a more liberal setting.

14.2.6.2 Surgery Mode

On the **Surgery Mode** toggle button, select either **Open** or **Laparoscopic/Prone**.




NOTE

It is important to correctly identify the **Surgery Mode**. The selected surgery mode influences how AFM interprets SVV. Selecting the incorrect **Surgery Mode** can lead to inappropriate fluid suggestions. If the patient is undergoing a laparoscopic procedure or is in the prone position and **Open** is selected as the **Surgery Mode**, AFM may produce additional fluid suggestions. If the patient is undergoing an **Open** procedure and **Laparoscopic/Prone** is selected as the **Surgery Mode**, AFM may withhold fluid suggestions.

14.2.6.3 Fluid Tracking

On the **Fluid Tracking** toggle button, select either **Fluid Meter** or **Manual**.




During **Manual** mode, the user is responsible for entering the fluid bolus volume delivered. With a fluid meter, the user enters a target volume for the bolus and the fluid meter will track the start, end, and flow rate of fluid delivery after the user opens and closes the fluid line.

NOTE

By default, the AFM feature requires a fluid meter connection to initialize. Using the AFM feature in **Manual** mode is optional. For more information on changing this advanced setting, contact your Edwards representative.

14.2.6.4 Maximum Case Volume

The **Maximum Case Volume** provides the user with a target fluid volume delivery and is set by the clinician at the start of the case based upon available clinical data at that point. A patient's fluid needs may change over the course of the case and therefore this value should be considered as a guide and not the absolute threshold between optimal and excessive fluid delivery. During an active AFM session a visual notification popup is provided when the total fluid delivered through the AFM feature approaches (within 500 mL) or exceeds the pre-set **Maximum Case Volume** to guard against potential fluid overload. The **Maximum Case Volume** value does not limit the functionality of the AFM feature or influence AFM fluid suggestions. Entering this value is required to start an AFM session, and this value may be changed from the notification popup or through the AFM settings screen at any time during an active AFM session. To set the **Maximum Case Volume** when the AFM session has not been started, select the **Max Case Vol.** button and enter the volume for the AFM session on the keypad.



If the **Maximum Case Volume** has already been entered, the current **Maximum Case Volume** value will appear on the settings screen. To change the **Maximum Case Volume**, touch the button and enter the new value on the keypad.



NOTE If making a change to the **Maximum Case Volume**, the new value must be greater than the total volume displayed on the AFM dashboard.

Table 14-41 AFM fluid status icons

AFM fluid status icon in navigation bar display	AFM fluid status icon in AFM dashboard	Meaning
		Fluid is recommended. The estimated % change in stroke volume exceeds the threshold defined by the Fluid Strategy setting (10%, 15%, 20%). When the AFM algorithm recommends fluid, the final prediction is based on input from both the population model and the individual patient bolus history. This icon is also displayed as a shortcut on the SV parameter tile. Touch the icon to access the AFM dashboard.
		A test bolus is suggested. To learn about the patient's fluid responsiveness, a test bolus is suggested. When the AFM algorithm suggests a test bolus, the final prediction contains little to no input from the individual patient bolus history and relies primarily on the patient population model and will trigger a test bolus suggestion if SVV > 9% in Open Surgery Mode or SVV > 12% in Laparoscopic/Prone Surgery Mode. This icon is also displayed as a shortcut on the SV parameter tile. Touch the icon to access the AFM dashboard.
		Fluid is not recommended The AFM software feature will not suggest fluid (neither AFM recommendation nor test bolus) when specific physiology indicates that fluid is not recommended. This status display will appear when the AFM software feature has learned that the patient has not responded to fluid in this hemodynamic state in the past through the individual patient bolus history. If it does not have information in the Individual patient bolus history, it relies on SVV and will not suggest fluid if SVV ≤ 9% in Open Surgery Mode or SVV ≤ 12% in Laparoscopic/Prone Surgery Mode.

Table 14-41 AFM fluid status icons (continued)

AFM fluid status icon in navigation bar display	AFM fluid status icon in AFM dashboard	Meaning
		A bolus has completed. Review the information on the AFM dashboard and make an analysis decision.
		AFM Mode is paused/suspended. The AFM software feature will not suggest fluid in this state.
		A bolus has completed and is being analyzed. AFM algorithm is analyzing the hemodynamic response of a bolus. The estimated time left is displayed on the navigation bar and on the AFM dashboard. While the bolus is being analyzed by the algorithm, the User Bolus button will be unavailable and the user will not receive any fluid suggestions from the algorithm.
		A bolus is in progress. This icon will cycle through various fluid levels to indicate that a bolus is actively being administered (manually or with the fluid meter).

14.2.6.5 Live Flow Rate

When a bolus is not actively being measured, the user has the option to enable the Live Flow Rate to display on the AFM dashboard.




Figure 14-13 AFM dashboard with Live Flow Rate enabled

To enable the Live Flow Rate:

- 1 Touch the settings icon at the right edge of the AFM dashboard
- 2 Toggle **Live Flow Rate** from OFF to ON




Figure 14-14 Enabling Live Flow Rate

14.2.7 Managing Fluids with the AFM Algorithm

Once the AFM algorithm is initialized, the AFM feature will support fluid optimization in two ways: suggesting fluid or not suggesting fluid. An icon is displayed on the navigation bar or AFM dashboard to indicate the software's suggestion (see Table 14-41 on page 305).

To administer fluid when the AFM feature is not suggesting fluid, open the fluid line (**Fluid Meter**) or touch the **User Bolus** button (**Manual**).



When following an AFM fluid suggestion or selecting **User Bolus**, a prompt will appear and the fluid administration workflow will commence.

The fluid administration workflow is used to gather the fluid administration information used by AFM to analyze the hemodynamic response to the fluid bolus. The following workflows are followed for both an AFM fluid suggestion and a requested **User Bolus**. The following workflows outline steps for the user in **Fluid Meter** mode or **Manual** mode.

NOTE	By default, the AFM feature requires a fluid meter connection to initialize. Using the AFM feature in Manual mode is optional. For more information on changing this advanced setting, contact your Edwards representative.

14.2.7.1 Fluid Administration Workflow – Acumen IQ Fluid Meter

Use the following AFM software workflow when an Acumen IQ fluid meter is connected. The Acumen IQ fluid meter is a sterile, single-use device that tracks the flow rate of fluid delivered to a patient through the intravenous line to which it has an in line connection. For instructions on using the AFM software feature without a fluid meter, see *Fluid Administration Workflow – Manual Mode* on page 312. Refer to the directions for use provided with the Acumen IQ fluid meter for specific instructions on placement and use, and for

relevant warnings, cautions and notes. The Acumen IQ fluid meter is compatible with an Acumen AFM cable and HemoSphere technology module. The HemoSphere technology module fits into a standard module slot.




Figure 14-15 Acumen IQ fluid meter and Acumen AFM cable connection overview


Acumen IQ Fluid Meter Connection Steps

Refer to the Acumen IQ fluid meter instructions for use for full connection instructions.

- 1 Insert the HemoSphere technology module into the monitor. The module will click when properly engaged.
- 2 Refer to the Acumen IQ fluid meter instructions for use for detailed instructions on setup and in line connection of the fluid meter to the intravenous line.
- 3 Ensure proper orientation, then plug the Acumen AFM cable into the technology module.
- 4 Connect the Acumen IQ fluid meter to the end of the Acumen AFM cable indicated by ② in figure 14-15.

Acumen IQ Fluid Meter Fluid Administration Workflow

- 1 An audible chime is heard and the “**Fluid Bolus Suggested**” message appears on the AFM algorithm dashboard when the algorithm suggests a fluid bolus.



NOTE If 40 seconds have elapsed when the AFM algorithm does not recommend fluid for the patient, the “**Fluid Bolus Suggested**” message will be removed from the dashboard.


- 2 The fluid delivery message prompts the user to review patient hemodynamics and begin a fluid bolus if they agree with the suggestion. To decline the suggestion, touch the **Decline** icon . Fluid suggestions will be paused for five minutes. To proceed with administering a bolus, continue to step 3.
- 3 Touch the **Fluid Type** button to specify fluid.

CAUTION Use of any fluids not listed in the specified **Fluid Type** list or choosing the incorrect fluid type may result in measurement inaccuracies.

NOTE With a fluid meter connected, the **Fluid Type** must be specified.

NOTE It may be appropriate to decline an AFM algorithm suggestion if review of patient hemodynamics does not suggest administration of fluid or in surgical situations where it is inappropriate to administer fluid. Note that constantly declining bolus suggestions may limit the usefulness of the AFM algorithm to determine future fluid responsiveness. Touch the **Decline** icon  to decline the bolus suggestion.


- 4 Touch the **Target Bolus Vol.** button to enter the desired volume. This step is optional.




NOTE If an additional bolus for the same patient is desired, the previously entered Target Bolus Volume will be saved

NOTE The AFM software feature can only analyze fluid boluses that are of volumes between 100 and 500 mL and delivered at a rate between 1 and 10 L per hour. If analysis of the fluid bolus by the AFM feature is desired, ensure that the volume and rate of delivery are within the required ranges.

- 5 Open the fluid line to begin bolus delivery.
- 6 Once a bolus is started, the “**Bolus in Progress...**” message is displayed on the AFM dashboard and a meter appears to indicate the current volume of bolus delivered.



The color around the volume meter will turn green when the target volume has been reached.



- 7 Close the fluid line when the desired bolus volume has been delivered.


NOTE The bolus delivery rate is dependent upon stopping the bolus when fluid administration is complete. Incorrect bolus delivery rate can impact the accuracy of the hemodynamic response assessment to a fluid bolus and the reliability of the future AFM algorithm suggestions.

CAUTION The presence of confounding factors during bolus delivery may lead to an incorrect fluid recommendation by the AFM software. Therefore, boluses delivered in the presence of confounding factors should be discarded. Potential confounding factors include but are not limited to:

- Vasoactive agent was administered during bolus administration
- Additional fluid given after primary bolus administered
- Subject repositioning
- Ventilatory changes
- Surgical manipulation
- Arterial line interference
 - * External compression (i.e., leaning on A-line)
 - * ABG draw, fast flush

- * Overdamping of line
- Vascular clamping
- Additional line of fluid simultaneously opened during bolus administration
- Known acute hemorrhage during fluid administration
- Inaccurate FT-CO measurements

- 8 Verify if **Fluid Type** displayed on the AFM algorithm dashboard is correct. If incorrect, touch on the **Fluid Type** button to edit.



If changing the **Fluid Type**, verify that the **Bolus Volume** shown is still accurate. If necessary, adjust the volume by touching the **Bolus Volume** button.

NOTE The prompt to analyze the hemodynamic response after a fluid bolus times out after 90 seconds. If analysis is available (**YES** is selectable), this will automatically be chosen.

- 9 Upon fluid bolus completion, if the total volume delivered through the AFM algorithm is approaching (within 500 mL) or exceeding the **Maximum Case Volume**, the AFM algorithm session will pause and one of the following messages will appear:

A. AFM Paused (Total Tracked Volume is approaching the set Maximum Case Volume)

B. AFM Paused (Total Tracked Volume has exceeded the set Maximum Case Volume)

If one of these notifications appears, re-assess the **Maximum Case Volume** to ensure it meets the patient's fluid needs and end the AFM session if appropriate. The total volume delivered is available at all times on the AFM algorithm dashboard and the **Maximum Case Volume** can be reviewed or changed at any time through the AFM algorithm settings by touching the settings icon



on the AFM algorithm dashboard. For more information, see *Approaching/Exceeding Maximum Case Volume Workflow* on page 315.

NOTE If an additional AFM algorithm session for the same patient is desired after the previous session has ended, refer to *Starting or Restarting AFM* on page 301. All initial AFM settings, with the exception of the **Maximum Case Volume**, will be maintained. Refer to *Assisted Fluid Management Settings* on page 303 to access and modify these settings, as necessary.

- 10 Touch **YES** to accept the current bolus for analysis. Touch **NO** to exclude the current bolus from further analysis by the AFM algorithm.

If the user accepts the current bolus and the bolus volume and rate fits within the AFM algorithm's criteria, the bolus will be analyzed by the algorithm.



While the bolus is being analyzed by the algorithm, the **User Bolus** button will be unavailable and the user will not receive any fluid suggestions from the algorithm.

The AFM algorithm will only analyze fluid boluses within the following ranges:

Bolus Volume: 100 - 500 mL


Bolus Rate: 1 - 10 L/hr

14.2.7.2 Fluid Administration Workflow – Manual Mode

NOTE By default, the AFM feature requires a fluid meter connection to initialize. Using the AFM feature in **Manual** mode is optional. For more information on changing this advanced setting, contact your Edwards representative.

While in **Manual** mode, it is important that fluid administration information (volume and duration) is accurately entered into the system.

- 1 An audible chime is heard and “**Fluid Bolus Suggested**” message appears on the AFM dashboard when the algorithm suggests a fluid bolus.



NOTE If 40 seconds have elapsed when the AFM algorithm does not recommend fluid for the patient, the “**Fluid Bolus Suggested**” message will be removed from the dashboard.

- 2 The fluid delivery message prompts the user to review patient hemodynamics and begin a fluid bolus if they agree with the suggestion.


If a fluid bolus is started, touch green **Start Bolus** icon  to indicate the timing of the start of the bolus.

NOTE It may be appropriate to decline an AFM suggestion if review of patient hemodynamics does not suggest administration of fluid or in surgical situations where it is inappropriate to administer fluid. Note that constantly declining bolus suggestions may limit the usefulness of the AFM algorithm to determine future fluid responsiveness.

Touch the **Decline** icon  to decline the bolus suggestion.

NOTE The AFM software feature can only analyze fluid boluses that are of volumes between 100 and 500 mL and delivered at a rate between 1 and 10 L per hour. If analysis of the fluid bolus by the AFM feature is desired, ensure that the volume and rate of delivery are within the required ranges.

- Once a bolus is started, then the “**Bolus in Progress...**” message is displayed on the AFM dashboard.
When the bolus is completed, touch the red **Stop Bolus** button and the **Bolus Volume** keypad will display.




NOTE The bolus delivery rate is dependent upon stopping the bolus when fluid administration is complete. Incorrect bolus delivery rate can impact the accuracy of the hemodynamic response assessment to a fluid bolus and the reliability of the future AFM suggestions.

CAUTION The presence of confounding factors during bolus delivery may lead to an incorrect fluid recommendation by the AFM software. Therefore, boluses delivered in the presence of confounding factors should be discarded. Potential confounding factors include but are not limited to:

- Vasoactive agent was administered during bolus administration
- Additional fluid given after primary bolus administered
- Subject repositioning
- Ventilatory changes
- Surgical manipulation
- Arterial line interference
 - * External compression (i.e., leaning on A-line)
 - * ABG draw, fast flush
 - * Overdamping of line
- Vascular clamping

- Additional line of fluid simultaneously opened during bolus administration
- Known acute hemorrhage during fluid administration
- Inaccurate FT-CO measurements


4 Enter the fluid bolus volume on the **Bolus Volume** keypad and touch the enter key.



Precaution. When estimating the amount of fluid delivered and entering the information into the system for analysis, it is important to ensure that the fluid bolus volume entered into the system is as accurate as possible.

- If the bolus volume entered into the system is greater than what was actually given, it could be interpreted as less effective causing subsequent bolus suggestions to be suppressed if the patient returns to a similar hemodynamic state.
- If the bolus volume entered into the system is less than what was actually given, it could be interpreted as more effective causing subsequent bolus suggestions to be made if the patient returns to a similar hemodynamic state.

5 Verify if information on the AFM dashboard is correct. If incorrect, touch on the **End Time** or **Bolus Volume** button to edit.



NOTE The prompt to analyze the hemodynamic response after a fluid bolus times out after 90 seconds. If analysis is available (**YES** is selectable), this will automatically be chosen.

- 6 Upon fluid bolus completion, if the total volume delivered through the AFM algorithm is approaching (within 500 mL) or exceeding the **Maximum Case Volume**, the AFM session will pause and one of the following messages will appear:
- A. AFM Paused (Total Tracked Volume is approaching the set Maximum Case Volume)
 - B. AFM Paused (Total Tracked Volume has exceeded the set Maximum Case Volume)

If one of these notifications appears, re-assess the **Maximum Case Volume** to ensure it meets the patient's fluid needs and end the AFM session if appropriate. The total volume delivered is available at all times on the AFM dashboard and the **Maximum Case Volume** can be reviewed or changed at any time through the AFM settings by touching the settings icon  on the AFM dashboard. For more information, see *Approaching/Exceeding Maximum Case Volume Workflow* on page 315.

NOTE If an additional AFM session for the same patient is desired after the previous session has ended, refer to *Starting or Restarting AFM* on page 301. All initial AFM settings, with the exception of the **Maximum Case Volume**, will be maintained. Refer to *Assisted Fluid Management Settings* on page 303 to access and modify these settings, as necessary.

- 7 Touch **YES** to accept the current bolus for analysis. Touch **NO** to exclude the current bolus from further analysis by the AFM algorithm.

If the user accepts the current bolus and the bolus volume and rate fits within AFM algorithm's criteria, the bolus will be analyzed by the algorithm.



While the bolus is being analyzed by the algorithm, the **User Bolus** button will be unavailable and the user will not receive any fluid suggestions from the algorithm.

The AFM algorithm will only analyze fluid boluses within the following ranges:

Bolus Volume: 100-500 mL

Bolus Rate: 1-10 L/hr

14.2.7.3 Approaching/Exceeding Maximum Case Volume Workflow


Upon fluid bolus completion, if the total volume delivered through AFM is approaching (within 500 mL) or exceeding the **Maximum Case Volume**, the AFM session will pause. If one of the notifications listed below appears, re-assess the **Maximum Case Volume** to ensure it meets the patient's fluid needs and end the AFM session if appropriate. The AFM feature will remain paused until one of the two choices are made. The total volume delivered is available at all times on the AFM dashboard and the **Maximum Case Volume** can be reviewed or changed at any time through the AFM settings by touching the settings icon  on the AFM dashboard.

A. AFM Paused (Total Tracked Volume is approaching the set Maximum Case Volume)

If approaching the pre-set volume, touch:

- * **Change Maximum Case Volume** to enter a new value through the keypad if patient fluid needs have changed. A notification will appear again if the total volume delivered through AFM is approaching (within 500 mL) the **Maximum Case Volume**;
- or

- * **Acknowledge And Continue** to continue the AFM session without changing the **Maximum Case Volume**. If acknowledged, the next notification to appear will indicate that the **Maximum Case Volume** has been exceeded.




The AFM session will continue once a selection has been made. The session can also be ended through the AFM settings menu at any time as described in *Pausing and Ending an AFM Session* on page 317.

B. AFM Paused (Total Tracked Volume has exceeded the set Maximum Case Volume)

If exceeding the pre-set volume, touch:


- * **Change Maximum Case Volume** to enter a new volume amount if the decision is made to intentionally exceed the pre-set volume because patient fluid needs have changed and continue the AFM session;
- or
- * **End AFM Session** to discard the history of boluses given to the patient through the AFM feature and discontinue the AFM session as described in *Pausing and Ending an AFM Session* on page 317.



14.2.8 Fluid Bolus Information Popup

To review information on a previously delivered fluid bolus, view the **AFM Bolus** or **User Bolus** information popup. This popup contains the bolus volume, bolus start time, bolus duration, fluid type (**Fluid Meter** only), change in SV, and change in SVV from beginning to end of the bolus. To view this popup:

- 1 Touch the settings icon  → **Select Screens** tab  to navigate to any graphical trends screen.
- 2 Touch the blue shaded region on the graphical trends plot.



14.2.9 Pausing and Ending an AFM Session

An active AFM session can be paused at any time, causing the AFM algorithm to suspend new fluid suggestions. While AFM is paused, the AFM dashboard and past fluid boluses will continue to be displayed.

To pause the current AFM session, touch the AFM pause button in the AFM dashboard.




To resume AFM session after being paused, touch the AFM start button.



Each AFM session can be ended by the user. When an AFM session is ended, the history of boluses given to the current patient is cleared. The HemoSphere advanced monitor will end the AFM session if a new patient is selected or the user switches to a different monitoring technology. AFM is only available with a connected pressure cable and Acumen IQ sensor. When the AFM session ends, monitoring continues without AFM prompts and display features. To end the current AFM session, use the following steps:

- 1 Touch the settings icon  on the AFM dashboard.
- 2 Touch the stop button .
- 3 Confirm on the popup.



If a Fault occurs while an AFM session is active, AFM will be suspended until the Fault condition is cleared.

NOTE If an additional AFM session for the same patient is desired after the previous session has ended, refer to *Starting or Restarting AFM* on page 301. All initial AFM settings, with the exception of the **Maximum Case Volume**, will be maintained. Refer to *Assisted Fluid Management Settings* on page 303 to access and modify these settings, as necessary.

14.2.10 GDT Tracking During an AFM Session


By touching **Start AFM** on the AFM dashboard, a GDT tracking session is automatically started with the following settings:



Parameter	Target
SVV	≤ 12%

The GDT parameter and target are non-configurable during an AFM session. When the AFM session is paused or ended, the GDT tracking session is paused or ended as well. For additional information about the GDT Tracking feature, refer to *Enhanced Parameter Tracking* on page 324.

To view the current Time-In-Target value for $SVV \leq 12\%$, touch the target icon  on the AFM dashboard. This will display a dashboard of the GDT tracking session, including the Time-In-Target. To minimize this tab, touch on the target icon again.



14.2.11 Clinical Validation

A prospective, multicenter, clinical study with 330 subjects allocated to a single arm across 9 US clinical sites was carried out to evaluate the performance of the Acumen Assisted Fluid Management (AFM) software feature in its ability to predict a patient's fluid responsiveness.

NOTE This study was conducted using the **Manual** mode equivalent on a previous version of the graphical user interface software. There are differences in the graphical user interface of the AFM algorithm on previous user interfaces and the user interface presented here for the HemoSphere advanced monitor. Relevant differences have been noted where necessary.

Subjects included in the study were ≥ 18 years of age, with planned non-cardiac/non-thoracic surgery (e.g., abdominal surgery, combined abdominal/pelvic surgery, major peripheral vascular surgery) expected to last > 2 hours post-anesthesia induction and had an American Society of Anesthesiologists (ASA) Score of 3 or 4. Table 14-42 provides a summary of the subject demographics.

Table 14-42 Subject demographics

Type	AFM IDE study
# of Patients	330
Age	64.2 ± 12.9
BMI	26.3 ± 4.5
ASA 3	91.8%
ASA 4	8.2%

The primary objective of the study was to evaluate the performance of the AFM feature in its ability to predict a patient's fluid responsiveness. The primary objective is based upon the performance of the AFM feature and the clinical decision making that occurred during the clinical study. The validity of the fluid responsiveness was measured by reporting the number of recommendations followed by delivered boluses that did and did not have a stroke volume (SV) response meeting the set fluid strategy (for example, for 15% fluid strategy, 500 cc of fluid should increase the patient's stroke volume by 15% if the patient is fluid responsive).

The AFM software feature showed that 66.1% [62.1%, 69.7%] of the time a bolus was administered after an AFM recommendation (based primarily on the subject's previous SV response), there was an increase in stroke volume per set fluid strategy. Additionally, the AFM software feature showed that 60.5% [57.8, 63.2] of the time a bolus was administered after a test bolus suggestion (based primarily on SVV) there was an increase in stroke volume per set fluid strategy. (table 14-43).

Table 14-43 AFM response rates by bolus type

Type of bolus event	Mean response rate (%) [confidence interval]
AFM recommendation	66.1% [62.1, 69.7]
AFM test	60.5% [57.8, 63.2]

NOTE An AFM recommendation in this study is equivalent to a fluid bolus recommendation on the HemoSphere advanced monitor. An AFM test/test bolus is equivalent to a test bolus recommendation on the HemoSphere advanced monitor.

An analysis of the response rate at the subject level demonstrates that the mean response rate was 65.62% and the median [interquartile range] per-subject response is 75% [50%, 100%] with a range from 0% to 100%.

Out of the 330 subjects enrolled in the study, 307 subjects were assigned to the per-protocol pivotal cohort and included in the effectiveness evaluation for the primary endpoint. In the per-protocol pivotal cohort, 94% (289/307) and 54% (165/307) of the subjects received AFM test suggestions and AFM recommended suggestions, respectively, and 6% of the subjects (18/307) did not receive any AFM suggestions. Therefore, it should be noted that the primary effectiveness endpoint is based on the 54% that received AFM recommended boluses.

User boluses during the study were recorded whenever fluid was given outside of an AFM test or recommendation while the AFM feature was in use. When the clinician administered a user bolus, there was an increase in stroke volume 40.9% [37.4, 44.1] of the time. The user boluses were not given exclusively as part of a manually administered fluid management protocol.

A secondary analysis provided the AFM performance stratified by delivered bolus volume (see table 14-44). The results demonstrate that AFM performance can depend on the bolus volume used.

Table 14-44 AFM performance by bolus volume (mL)

Bolus volume (mL)	Mean response (%)	(2.5% LCL, 97.5% UCL)	Number of boluses	Number of subjects
≤100	77.26%	(72.60, 81.81)	147	76
>100-200	59.92%	(54.61, 65.13)	152	76
>200-250	57.73%	(50.63, 64.94)	79	49
>250-300	65.27%	(59.18, 69.39)	49	39
All Boluses	66.04%	(61.56, 71.13)	424	207

Accuracy of the AFM software feature was analyzed at the bolus level; this includes sensitivity and specificity, and positive and negative predictive values.

Sensitivity is the ratio of true positives to the total number of responders (positives). A true positive is any event with an increase in stroke volume per the predetermined fluid strategy when a bolus is given (within 5 minutes) after AFM recommendation. Sensitivity of the AFM feature was 77.7%.

Specificity is the ratio of true negatives to the total number of non-responders (negatives). In the context of the clinical study, a true negative is any bolus given outside of the AFM recommendations to which the patient did not respond. Specificity of the AFM feature was 40.6%.

Positive predictive value (PPV) is the probability that a patient will be responsive to a bolus suggested by the AFM algorithm. PPV of the AFM feature was 62.7%.

Negative predictive value (NPV) is the probability that a patient will be non-responsive to a bolus given outside of AFM recommendations. NPV of the AFM feature was 58.9%.

Table 14-45 Accuracy results of the AFM feature (bolus level)

Measurement	Value (%) [95% confidence interval]
PPV	62.7 [59.6, 65.3]
NPV	58.9 [54.4, 63.2]
Specificity	40.6 [37.1, 44.3]
Sensitivity	77.7 [74.9, 80.3]

14.2.11.1 Fluid Bolus Activity

The AFM software feature uses the current hemodynamic state and past response to fluid given in similar states to determine if a fluid recommendation should be generated. Therefore, it is possible to receive several AFM suggestions in a one-hour period. Post-hoc analysis of the clinical validation study determined that the number of recommendations can range from 0-6 AFM recommendations per hour, with no AFM recommendations for the majority of the time (see table 14-46 on page 320). It is also possible for an AFM suggestion to immediately follow the completion of a non-responsive fluid bolus if current hemodynamic state has changed since the prior non-responsive bolus.

Table 14-46 Frequency of AFM recommendations per hour**

AFM recommendations per hour	Frequency of occurrence*
0	73.8% (784/1062)
1	10.9% (116/1062)
2	6.7% (71/1062)
3	5.3% (56/1062)
4	2.4% (26/1062)
5	0.6% (6/1062)
6	0.3% (3/1062)
*The frequency of occurrence is based upon the number of hours with a given number of AFM recommendations divided by the total number of hours. **The frequency of AFM recommendations per hour is presented as general guidance and may not be representative of individual experience.	

As a clinical decision support system, AFM suggestions can be declined or discarded by the user. In the clinical validation study, 47% (1209/2550) of the total AFM suggestions were declined by the user which included 40% (324/803) of the AFM recommendations and 51% (885/1747) of AFM test suggestions. In addition, out of the 1341 AFM prompts that were accepted by the users, 13% (168/1341) were discarded which included 11% (52/479) of the AFM recommended boluses and 13% (116/862) of AFM test boluses.

NOTE For this study, the AFM feature utilized a prompt at the completion of the fluid bolus with the options of **DISCARD BOLUS** or **ACCEPT**. AFM functionality on the HemoSphere advanced monitor is identical, however the user required response is **YES** or **NO** to the prompt “**Analyze Hemodynamic Response?**”. A **NO** response results in a declined analysis. Therefore, current labeling for this workflow is “Analysis Declined” as opposed to “Discarded.” For reference, the term “Analysis Declined” is indicated alongside “Discarded” in this clinical validation study. Please refer to table 14-40 for additional explanation of the terms “declined” and “analysis declined.”

Although post-hoc analysis revealed no difference in performance based on compliance to AFM suggestions, the clinical validation study was not designed to directly address this question. Therefore, the AFM performance may be affected by the compliance to AFM suggestions. Table 14-47 includes a complete accounting of the fluid boluses in the clinical validation study.

Table 14-47 Complete accounting of fluid boluses

Bolus originator	Prompted	Suggestion declined	Accepted	Discarded (analysis declined)	Completed	Analyzed
AFM	2550	1209	1341	168	1173	1165
- Recommended	803	324	479	52	427	424
- Test	1747	885	862	116	746	741
User	606	14	592	81	511	508
Total	3156	1223	1933	249	1684	1673

During the clinical validation study, the boluses were discarded 13% of the time (analysis declined). The reasons for discarded boluses during the study are included in table 14-48.

Table 14-48 Reasons boluses were discarded (analysis declined) in the per protocol pivotal subjects

Fluid demographics Reasons bolus discarded (analysis declined)	% (n/N)
Administered vasoactive agent with fluids	35.0% (89/254)
Other	18.1% (46/254)
ABG draw / fast flush	11.8% (30/254)
Subject repositioning	11.8% (30/254)
Arterial line interference	10.2% (26/254)
Ventilatory changes	4.7% (12/254)
Additional fluid given after primary bolus administered	3.5% (9/254)
Overdamping of line	1.6% (4/254)
Surgical manipulation	0.8% (2/254)
Unknown	0.8% (2/254)
Additional line of fluid simultaneously opened up during bolus	0.4% (1/254)

Table 14-48 Reasons boluses were discarded (analysis declined) in the per protocol pivotal subjects (continued)

Fluid demographics Reasons bolus discarded (analysis declined)	% (n/N)
Known acute hemorrhage during fluid administration (blood loss >= 250cc in 7 min period)	0.4% (1/254)
Vascular clamping	0.4% (1/254)
Total	100.0% (254/254)
<i>*Note: More than one reason for discarding a bolus could be provided and as a result there are 254 reasons documented for 249 discarded boluses.</i> <i>Denominators are based on the total number of available data captured for each parameter.</i>	

During the clinical validation study, the AFM suggestions (recommendations and test) were declined 47% of the time. The reasons for decline identified during the study are provided in table 14-49.

Table 14-49 Reasons suggestions were declined in the per protocol pivotal subjects

Fluid demographics Reasons AFM prompt not accepted	% (n/N)
The subject is normotensive at this time	42.3% (592/1399)
Fluid is contraindicated by the procedure at present	7.2% (101/1399)
Clinician prefers to use a vasoactive agent instead at this time	7.0% (98/1399)
Clinician does not think subject will be fluid responsive	6.3% (88/1399)
Other	4.4% (62/1399)
This bolus recommendation is suspect based on recent bad data (i.e., artifact in BP signal)	3.6% (50/1399)
We are starting to close the case now	3.5% (49/1399)
Busy engaging in other tasks	3.5% (49/1399)
ABG / lab draw	2.7% (38/1399)
Clinician believes the hemodynamic changes are temporary and due to surgical manipulation	2.6% (36/1399)
Currently hypertensive	2.4% (34/1399)
Clinician is administering fluid (blood or other) outside of AFM	2.4% (34/1399)
Waiting for RBC administration	2.1% (29/1399)
There was a change in subject position and clinician would like to wait and see	1.9% (26/1399)
Fluid recently administered; now observing	1.9% (26/1399)
Subject recently received fluid but was not responsive	1.2% (17/1399)
Clinician hit decline to remove AFM popup prompt so that hemodynamics could be further reviewed before deciding on giving fluid	1.1% (15/1399)
Managing BP	1.1% (15/1399)
Questionable pressure tracing	1.0% (14/1399)
There was a brief period of arrhythmia and the clinician doesn't believe that the patient needs a bolus	0.8% (11/1399)
Clinician is concerned about dilutional anemia at this time	0.5% (7/1399)
Clinician mistakenly declined AFM recommendation	0.3% (4/1399)
There was an expected change with insufflation which is anticipated to be brief	0.2% (3/1399)
Clinician is concerned about right ventricular dysfunction	0.1% (1/1399)
We had a temporary change in ventilation strategy (i.e., recruitment maneuver)	0.1% (1/1399)
Total	100.0% (1399/1399)
<i>*Note: More than one reason for a declined AFM prompt could be provided and as a result there are 1399 reasons documented for 1223 declined boluses.</i> <i>Denominators are based on the total number of available data captured for each parameter.</i>	

In the clinical validation study, 66% of the AFM recommended boluses produced the desired change in SV that met the Fluid Strategy as reported in Table 14-43. However, a study limitation was that fluid was not delivered when the user declined an AFM recommendation and, as such, the SV responses of the declined AFM suggestions are unknown. If each declined AFM recommendation was categorized as a negative response, the response rate could be as low as 37%. Reasons for these declines included normotension, fluid contraindicated by the procedure at the present time, and clinician preference to use a vasopressor. The complete list of reasons and their prevalence are provided in Table 14-49.

14.2.12 Fluid Meter Only Mode

The Acumen IQ fluid meter can be connected (see figure 14-15 on page 308) to track fluid without initializing the AFM algorithm. In minimally-invasive mode, enabling the **“Fluid Meter Only Mode”** is an advanced setup feature. In invasive and non-invasive mode, proceed to step 5. For additional information, contact your Edwards representative.

- 1 Touch the **settings** icon  → **Settings** tab  **Settings** → **Advanced Setup** button and enter the **Secure User** password. All passwords are set during system initialization. Contact your hospital administrator or IT department for password.
- 2 Touch the **Engineering** button.
- 3 Touch the **AFM** button.
- 4 Toggle the **“Algorithm Suggestions (AFM)”** switch to **“Disabled”**.

NOTE Active AFM algorithm sessions will end when Fluid Meter Only Mode is entered. Total accumulated fluid volume delivered during all previous AFM algorithm sessions from the current patient session will continue to be tracked.

- 5 Connect the fluid meter by following connections steps outlined in *Fluid Administration Workflow – Acumen IQ Fluid Meter* on page 307.
- 6 Touch the **Assisted fluid management** icon .
- 7 Set the desired fluid meter settings for **Fluid Type** and **Maximum case volume** (optional).
- 8 Enter the target bolus volume (optional).
- 9 Open the fluid line to begin bolus delivery.
- 10 The total tracked volume will be recorded and displayed on the Assisted Fluid Management dashboard.

14.3 Enhanced Parameter Tracking

The HemoSphere advanced monitoring platform provides tools for performing **Goal Directed Therapy (GDT)**, enabling a user to track and manage key parameters in the optimal range. With enhanced parameter tracking, clinicians have the ability to create and monitor customized protocols.

14.3.1 GDT Tracking

14.3.1.1 Key Parameter and Target Selection

- 1 Touch the GDT tracking icon  on the navigation bar to access the GDT menu screen.



Figure 14-16 GDT Menu Screen - Key Parameter Selection

- 2 Touch the upper half of a **Parameter/Target** selection icon  and choose the desired parameter from the parameter panel. Up to four key parameters can be tracked.
- 3 Touch the lower half of the **Parameter/Target** selection icon  to enter a range value on the keypad. The selected operator (<, ≤, > or ≥) and value represent the upper or lower boundary during parameter tracking. Touch the enter key .



Figure 14-17 GDT Menu Screen - Target Selection

- 4 Touch any selected parameter to change it to a different available parameter or touch **None** on the parameter selection panel to remove it from tracking.
- 5 To view and select parameter/target settings from a previous GDT tracking session, touch the **Recents** tab.
- 6 Touch **OK** to begin GDT tracking.



Figure 14-18 GDT Active Tracking

14.3.1.2 Active GDT Tracking

During active GDT tracking, the plot area of the parameter trend graph within targeted range appears shaded in blue. See figure 14-18, “GDT Active Tracking,” on page 325.



GDT Tracking Control Panel. Touch the GDT Tracking button to pause or stop during active tracking. While tracking is paused, the plot area within target range on the parameter graph appears shaded in gray.



Time-In-Target Value. This is the primary output of enhanced parameter tracking. It is displayed below the **Time-In-Target** icon on the upper right corner of the parameter’s graphical trend plot. This value represents the accumulated percentage of time a parameter has been within target during an active tracking session.

Parameter Tile Target Indicator Colors. Table 14-50 defines clinical target indicator colors during GDT tracking.

Table 14-50 GDT Target Status Indicator Colors

Color	Indication
Blue	Tracked parameter is currently within the configured target range.
Black	Tracked parameter is currently outside of the configured target range.
Red	Tracked parameter is currently below the low alarm limit or above the high alarm limit.
Gray	Tracked parameter is unavailable, in a fault state, GDT tracking is paused, or a target has not been selected.

Auto Scale Trend Time. Upon initiating active GDT tracking, the graphical trend time is automatically scaled to fit all tracked data for the current session within the plot. The initial Graphical Trend time scale value is set to 15 minutes and increases as tracking time expands beyond 15 minutes. **Auto Scale Trend Time** can be disabled through the set scales popup menu while in GDT mode.

NOTE While viewing active GDT tracking on the Graphical Trend Screen, parameter selection menus are disabled.



14.3.1.3 Historical GDT



Touch the historical data icon to display recent GDT tracking sessions. A blue “**Viewing Historical GDT Session**” banner will appear at the top of the screen. Current parameter values are displayed on key parameter tiles while viewing a historical GDT session. Touch the scroll buttons to view different historical GDT sessions. Percent change measurements displayed on the trend screen represent percent changes between two historical values.

14.3.2 SV Optimization

During SV Optimization mode, the SV/SVI target range for GDT tracking is selected based on recent SV trends. This allows the user to identify the optimal SV value during active monitoring of fluid management.

- 1 Touch the GDT tracking icon  on the navigation bar.
- 2 Select **SV** or **SVI** as a key parameter.
- 3 Do NOT specify a target value in the lower half of the **Parameter/Target** selection icon  instead, touch **OK** to begin target selection on the trend graph.
- 4 Observe the SV trend while administering necessary fluid management to achieve an optimal value.
- 5 Touch the add target icon  on the right side of the SV/SVI trend graph. The trend line will turn blue.
- 6 Touch within the plot area to view a trend line value. A target value icon will appear along with an unlocked icon. A horizontal white dashed line will be displayed at 10% below the target cursor value. The area extending from this line to the top of the Y-axis will be shaded blue. 
- 7 If desired, touch the Exit Target Selection button  to return to monitoring of fluid management.
- 8 Touch the target value icon  to accept the displayed target range and initiate GDT tracking.
- 9 The edit target icon  can be touched at anytime after target selection to adjust the SV/SVI target value.
- 10 The GDT tracking icon  can be touched at anytime when GDT mode is active to end the GDT tracking session.

14.3.3 GDT Report Download

The Data Download screen allows a user to export GDT reports to a USB drive. See *Data Download* on page 158.

14.4 Fluid Responsiveness Test

With the **Fluid Responsiveness Test (FRT)**, clinicians have the ability to assess preload responsiveness. Preload responsiveness is assessed by tracking the changes in **SV**, **SVI**, **CO** or **CI** in response to a fluid challenge (**Passive Leg Raise** or **Fluid Bolus**).

To begin the test:

- 1 Touch the settings icon  → **Clinical Tools** tab .
- 2 Touch **Fluid Responsiveness Test** .




Figure 14-19 Fluid Responsiveness Test – New Test Screen

- 3 On the **New Test** tab (see Figure 14-19), touch the desired test type: **Passive Leg Raise** or **Fluid Bolus**.

Touch the question mark symbol for brief instructions on starting each test. For more detailed instructions, follow the steps below.

NOTE Interpretation of the Fluid Responsiveness Test (FRT) is directly correlated with the response time of the parameter being monitored. Response times of monitored parameters can vary depending on the monitoring mode and are dictated by the connected technology. Update rates for FRT selected parameters while in minimally-invasive mode are based on CO averaging time (see Table 6-4 on page 142).

14.4.1 Passive Leg Raise Test

The **Passive Leg Raise** is a sensitive non-invasive method for assessing a patient's fluid responsiveness. During this test, venous blood transferred from the lower body to the heart simulates a fluid challenge.




- 1 Touch and highlight **Passive Leg Raise** under the **New Test** tab. The **New Test** tab displays test configuration menu options.
- 2 Select the **Parameter** to be analyzed:

- **SV, SVI, CO, or CI** (**Minimally-Invasive** and **Non-Invasive** monitoring modes)
 - **SV_{20s}, SVI_{20s}, CO_{20s}, or CI_{20s}** (**Invasive** monitoring mode with PAP signal; see *20-Second Flow Parameters* on page 174)
- 3 Select the **Challenge Duration: 1 minute, 1 minute 30 sec, or 2 minutes** (**Minimally-Invasive** and **Non-Invasive** monitoring modes) or **3 minutes** (**Invasive** monitoring mode).
 - 4 Place the patient in a semi-recumbent position. Touch the **Start Baseline** button to begin the baseline measurement.


NOTE The baseline value is averaged from multiple readings. Ensure that the patient remains still and stays in the same position during this measurement period.

- 5 The **Baseline Measurement** screen will appear with a trend graph of the selected parameter and a countdown timer displaying the amount of time remaining for the baseline measurement.



NOTE To abort the baseline measurement, touch the **CANCEL** button and return to the **New Test** screen.

- 6 At the conclusion of the baseline measurement, the baseline value will appear below the trend graph. To remeasure the baseline value, touch **RESTART**.
- 7 To continue to the **Passive Leg Raise Measurement**, place the patient in supine position and touch the **START** button. Passively raise the patient's legs to a 45 degree angle within five seconds. A five second countdown clock will appear to indicate time remaining until the start of the challenge measurement.
- 8 A new countdown timer will appear starting at the selected **Challenge Duration** time. Ensure that the patient remains still during the measurement period.



NOTE Before sufficient measurements have been taken, the **CANCEL** button can be touched to abort the test. A confirmation popup window will appear. Touch **Cancel Test** to return to the test configuration screen (**New Test** tab).

After sufficient measurements have been taken, the **CANCEL** button is no longer available. To stop the test and analyze measured data before the full time of the test has been reached touch **END NOW**.

- 9 At the conclusion of the test, the change in the selected **Parameter** value as a response to the fluid challenge will be displayed. See Figure 14-20. Touch the return icon to perform another test, or the home icon to return to the main monitoring screen.



Figure 14-20 Fluid Responsiveness Test – Results Screen

14.4.2 Fluid Bolus Test


The **Fluid Bolus** test is a sensitive method for assessing a patient's fluid responsiveness. During this test, a fluid bolus is administered to the patient and preload responsiveness may be assessed by tracking the value of SV, SVI, CO, or CI.



- 1 Touch and highlight **Fluid Bolus** under the **New Test** tab. The **New Test** tab displays test configuration menu options.
- 2 Select the **Parameter** to be analyzed:
 - **SV, SVI, CO, or CI** (**Minimally-Invasive** and **Non-Invasive** monitoring modes)
 - **SV_{20s}, SVI_{20s}, CO_{20s}, or CI_{20s}** (**Invasive** monitoring mode with PAP signal; see *20-Second Flow Parameters* on page 174)
- 3 Select the **Challenge Duration: 5 minutes, 10 minutes, or 15 minutes**.
- 4 Touch the **Start Baseline** button to begin the baseline measurement.


NOTE The baseline value is averaged from multiple readings. Ensure that the patient remains still and stays in the same position during this measurement period.

- 5 The **Baseline Measurement** screen will appear with a trend graph of the selected parameter and a countdown timer displaying the amount of time remaining for the baseline measurement.



NOTE To abort the baseline measurement, touch the **CANCEL** button and return to the **New Test** screen.

- 6 At the conclusion of the baseline measurement, the baseline value will appear below the trend graph. To remeasure the baseline value, touch **RESTART**.
- 7 To continue to the **Fluid Bolus Measurement**, administer the fluid bolus and touch **START** when the bolus begins.
- 8 A new countdown timer will appear starting at the selected **Challenge Duration** time. Ensure that the patient remains still during the measurement period.



NOTE Before sufficient measurements have been taken, the **CANCEL** button can be touched to abort the test. A confirmation popup window will appear. Touch **Cancel Test** to return to the test configuration screen (**New Test** tab).

After sufficient measurements have been taken, the **CANCEL** button is no longer available. To stop the test and analyze measured data before the full time of the test has been reached touch **END NOW**.

- 9 At the conclusion of the test, change in the selected **Parameter** value as a response to the fluid challenge will be displayed. See Figure 14-20. Touch the return icon to perform another test, or the home icon to return to the main monitoring screen.

14.4.3 Historical Test Results


The user can view previous test results on the **Historical Results** tab. A list of all fluid responsiveness tests for the current patient is shown. Use the scroll buttons to highlight a specific test and touch the **Select** button to view a test summary. A popup window will appear listing the test configurations, key timestamped points and measured **Parameter** values.


14.5 Relative Change in Total Hemoglobin – Δ ctHb

The relative change in total hemoglobin (Δ ctHb) is a sub-parameter of StO_2 . A trending value, Δ ctHb is calculated from the sum of relative changes in oxygenated hemoglobin and deoxygenated hemoglobin ($\Delta\text{O}_2\text{Hb}$ and ΔHHb). Each connected tissue oximetry sensor site StO_2 measurement has its own Δ ctHb sub-parameter. Δ ctHb parameters are only available if the Δ ctHb parameter feature is enabled. Please contact your local Edwards representative for more information on enabling this advanced feature.

14.5.1 Δ ctHb Value Display


To display the value of Δ ctHb on the StO_2 parameter tile:




- 1 Touch anywhere in the StO_2 parameter tile → **Sensor Configuration** tab .
- 2 Toggle the Δ ctHb value button from **Off** to **On**.

14.5.2 Δ ctHb Trend Display


To display the trend of Δ ctHb on the StO_2 parameter trend graph:



- 1 Touch anywhere in the **StO₂** parameter tile → **Sensor Configuration** tab 
- 2 Toggle the Δ ctHb trend button from **Off** to **On**. The trend will be plotted in pink with a corresponding y-axis on the right side of the graph.

14.5.3 Reset Δ ctHb

To reset the baseline value of Δ ctHb to zero for all channels:

- 1 Touch the settings icon  → **Clinical Tools** tab  → **ctHb Tools** icon 
- 2 Touch the **Reset Δ ctHb** button.

14.5.4 Validation Methodology and Study Results

Table 14-51 summarizes the validation methodology and study results for relative change in hemoglobin (Δ ctHb).

Table 14-51 Clinical and blood bench validation study results for trending accuracy of relative change in hemoglobin (Δ ctHb)

Sensor size	Bland-Altman Bias $\pm$ Precision, RSME (A_{rms})	Method of evaluation*
large	$0.22 \pm 2.53 \mu\text{M}$ at 1 SD, $2.53 \mu\text{M}$	Under isovolumic hemodilution human study
	$-0.26 \pm 2.04 \mu\text{M}$ at 1 SD, $2.04 \mu\text{M}$	Under mild hypoxia human study
medium	$-1.10 \pm 5.27 \mu\text{M}$ at 1 SD, $5.39 \mu\text{M}$	Blood phantom study
small	$-0.02 \pm 5.96 \mu\text{M}$ at 1 SD, $5.96 \mu\text{M}$	Blood phantom study
	$-0.50 \pm 2.09 \mu\text{M}$ at 1 SD, $2.15 \mu\text{M}$	Under hemoglobin level desaturation blood phantom study
*Differential Pathlength factor = 5		

14.6 Aortic Stenosis Index (ASI) Algorithm

The Aortic Stenosis Index is calculated based on continuous reconstructed arterial blood pressure waveforms collected by a non-invasive continuous blood pressure device.

The ASI algorithm output is represented by a number from zero to one hundred, which determines the likelihood a patient has severe aortic stenosis (AS). If the ASI Algorithm output is greater than or equal to 20, it is recommended that the patient undergoes further evaluation to identify if they have severe AS.

CAUTION Do not rely on the lack of a suspected finding to rule out follow-up.

United States Federal law restricts the sale and use of this device to either by or on the order of a clinician.

This is NOT a diagnosis. The screening results offered by Edwards' ASI algorithm are meant only to provide clinical decision support to the clinician, who may use the result in conjunction with their own evaluation and clinical judgment. The recommendations are not diagnoses.

The quality of the ASI algorithm recommendation depends upon the quality of the inputted blood pressure waveform. Situations which may affect the inputted blood pressure causing invalid waveforms may include, but are not limited to, severe vasoconstriction, patient movement, or Raynaud's disease.

NOTE	The ASI algorithm may recommend further evaluation for patients with types of cardiovascular and valvular disorders other than severe AS.

NOTE	The ASI algorithm is not intended to be used with non-Edwards monitoring platforms.

Clinical validation studies (see *Performance Data - Clinical Testing Summary* on page 333) demonstrate that ASI algorithm is accurate and hence, may be useful across the typical range of variation of patients and clinical practice.

14.6.1 Indications For Use

The Aortic Stenosis Index (ASI) algorithm is intended to provide decision support to clinicians in their evaluation of patients. The algorithm analyzes patient arterial pressure waveforms for signs of structural heart abnormalities and screens for severe aortic stenosis.

The screening results provided by the ASI algorithm can be used by a clinician for decision support, in conjunction with their own evaluation and clinical judgment in consideration for referral for further diagnostic follow-up. The ASI algorithm is not intended as a sole means of diagnosis and should not be relied upon to make or confirm diagnosis. The ASI algorithm is for use in adult patients ≥ 18 years old in environments where health care is provided by clinicians.

14.6.2 Intended Use

The Aortic Stenosis Index (ASI) algorithm is intended to provide decision support to clinicians in their evaluation of patients and screen for severe aortic stenosis.

14.6.3 Aortic Stenosis Index (ASI) Parameter Display




Figure 14-21 Example ASI outputs and display user interface

The ASI algorithm calculates the likelihood of severe aortic stenosis using continuous reconstructed arterial blood pressure. It is able to provide an output as soon as 1 minute and may take up to ten (10) minutes from running the algorithm depending on the quality of blood pressure waveform signal. Figure 14-21 provides example displays of the user interface with possible results. The results include:

- Severe aortic stenosis detected, and additional evaluation recommended
- Severe aortic stenosis not detected

Table 14-52 provides a detailed explanation and interpretation of the ASI outputs and the recommended user actions.

Table 14-52 General interpretations and recommended actions for ASI algorithm

ASI Value	General Interpretation	Recommended User Action
ASI < 20	Likelihood of severe AS is lower than the threshold of 20; the patient will not be recommended to seek additional evaluation	None
ASI ≥ 20	Likelihood of severe aortic stenosis is greater than the threshold of 20; it is recommended the patient seek additional evaluation	Further evaluation recommended
ASI = -1	Inconclusive result	Clinician discretion

14.6.4 Performance Data - Clinical Testing Summary

The ASI algorithm has been tested using retrospective analysis on a proprietary database. The database contains 902 unique continuous blood pressure waveforms from 902 unique subjects. Subjects were from five (5) geographically diverse sites in North America and Europe. The subject population was 50% female and had an age range of 18 to 99, with average of 67.7 years old. Race/Ethnicity was recorded in 73.5% of the subjects and out of that subset, 74% were White, 17.2% were African American, and 18.9% were Hispanic.

Subjects were assessed and labeled based on their transthoracic echocardiogram reports (TTE) which contains valvular heart disorder diagnoses confirmed by a board-certified cardiologist. Subjects with a history of major cardiac surgeries such as heart transplantation, valve repair, valve prosthetics and device implantations were excluded. Among all 902 subjects, 195 patients had confirmed severe AS, 582 patients did not have any valvular heart disorders and 125 patients had at least one type of valvular heart disorder including higher than trace degree of mitral, aortic, pulmonary, or tricuspid regurgitation and/or any degree of stenosis in mitral valve and/or mild and moderate AS.

No study subjects included in the training datasets were included in the testing database. All algorithm testing was performed once the final ASI algorithm version was locked.

Table 14-53 demonstrates the result of the ASI performance in differentiating severe AS subjects from patients without severe AS.

NOTE Patients with known major cardiac anomalies, such as patients with transplanted hearts, prosthetic or repaired valves as well as patients with cardioverter defibrillator or ventricular assist device implantations were excluded from the algorithm development/training dataset. This is because the aforementioned conditions can severely alter the arterial pressure waveform morphology in unpredictable ways. The ASI algorithm is therefore not intended for patients with the conditions listed above.

Table 14-53 ASI performance in differentiating severe AS subjects from patients without severe AS

	Sensitivity (%)	Specificity (%)	NPV (%)	PPV (%)	AUC (%)
Severe AS vs. No Severe AS	95.9 (92.9, 98.9)	84.0 (81.2, 86.8)	98.8 (97.9 - 99.7)	60.3 (54.5, 66.1)	97.2 (96.1 -98.3)

Troubleshooting

Contents

On Screen Help	335
Monitor Status Lights	336
Pressure Cable Communication	337
ForeSight Oximeter Cable Sensor Communication	338
Pressure Controller Communication	339
HemoSphere Advanced Monitor Error Messages	340
HemoSphere Swan-Ganz Module Error Messages	344
Pressure Cable Error Messages	350
HemoSphere ClearSight Module Error Messages	355
Venous Oximetry Error Messages	362
Tissue Oximetry Error Messages	364

15.1 On Screen Help

The help topics outlined in this chapter and displayed on monitor help screens are associated with common error conditions. In addition to these error conditions, a list of unresolved anomalies and troubleshooting steps are available at eifu.edwards.com. This list is associated with the HemoSphere advanced monitor model number (HEM1) and software version indicated on the startup page (see *Start Up Procedure* on page 79). These issues are continually updated and compiled as a result of ongoing product improvements.

The main help screen allows the user to navigate to specific help for HemoSphere advanced monitoring platform issues. Faults, alerts and warnings notify the user of error conditions affecting parameter measurements. Faults are technical alarm conditions that suspend parameter measurement. The category help screen provides specific assistance for faults, warnings, alerts, and troubleshooting.

- 1 Touch the settings icon .
- 2 Touch the **Help** button to access the main help screen.
- 3 Touch the category help button corresponding to the technology for which help is needed: **Monitoring, Swan-Ganz Module, Pressure Cable, Venous Oximetry, 20-Second Flow, ClearSight module, Tissue Oximetry, or Assisted Fluid Management.**
- 4 Touch the type of help needed based on the message type: **Faults, Alerts, Warnings, or Troubleshooting.**

NOTE The help screens for 20-Second Flow do not list help categories for system messages. 20-Second Flow help screens contain information on how to monitor with 20-second parameters and how they are calculated.

The **Assisted Fluid Management** help screen also contains information on **Getting Started** and **Algorithm Help** in addition to **Faults, Alerts, and Warnings**.

- 5 A new screen appears with a list of the selected messages.
- 6 Touch a message or troubleshooting item from the list and touch **Select** to access information for that message or troubleshooting item. To view the full list, use the arrow buttons to move the selection highlight up or down the list. The next screen displays the message along with possible causes and suggested actions.
- 7 To display software versions and serial numbers for the monitor and connected technology

module(s)/cable(s) touch settings icon  → **Settings** tab  **Settings** → **Versions** button.

15.2 Monitor Status Lights

The HemoSphere advanced monitor has a visual alarm indicator to alert the user to alarm conditions. See *Alarm Priorities* on page 395 for more information on medium and high priority physiological alarm conditions. The monitor power button has an integrated LED to indicate the power status at all times.




Figure 15-1 HemoSphere advanced monitor LED indicators

① visual alarm indicator

② monitor power status

Table 15-1 HemoSphere advanced monitor visual alarm indicator


Alarm status	Color	Light pattern	Suggested action
High-priority physiological alarm	Red	Flashing ON/OFF	This physiological alarm condition needs immediate attention Refer to the status bar for specific alarm condition
High-priority technical faults and alerts	Red	Flashing ON/OFF	This alarm condition requires immediate attention and will remain active during an alarm pause If a particular technical alarm condition is unrecoverable, restart system If problem persists, contact Edwards Technical Support
Medium-priority technical faults and alerts	Yellow	Flashing ON/OFF	This alarm condition needs prompt attention Refer to the status bar for specific alarm condition
Medium-priority physiological alarm	Yellow	Flashing ON/OFF	This alarm condition needs prompt attention Refer to the status bar for specific alarm condition
Low-priority technical alert	Yellow	Solid ON	This alarm condition requires non-urgent attention Refer to the status bar for specific alarm condition

Table 15-2 HemoSphere advanced monitor power light

Monitor status	Color	Light pattern	Suggested action
Monitor power ON	Green	Solid ON	None
Monitor power OFF Monitor connected to AC mains Battery charging	Yellow	Flashing ON/OFF	Wait for battery to be charged before unplugging from AC mains.
Monitor power OFF Monitor connected to AC mains Battery not charging	Yellow	Solid ON	None
Monitor power OFF	No light	Solid OFF	None

15.3 Pressure Cable Communication

The pressure cable LED indicates the status of the pressure sensor or transducer.

**Figure 15-2 Pressure cable LED indicator****Table 15-3 Pressure cable communication light**


Condition	Color	Light Pattern	Suggested Action
No pressure sensor/transducer connected	No light	Solid OFF	None
Pressure sensor/transducer connected but not yet zeroed	Green	Flashing ON/OFF	Zero the pressure sensor to begin monitoring

Table 15-3 Pressure cable communication light (continued)

Condition	Color	Light Pattern	Suggested Action
Pressure sensor/transducer zeroed	No light	Solid OFF	None. The connected pressure sensor can actively monitor pressure signal
Pressure sensor/transducer medium priority technical alarm	Yellow	Flashing ON/OFF	Refer to the screen to ascertain the type of technical fault. Use the help menu or tables below for the appropriate suggested action

15.4 ForeSight Oximeter Cable Sensor Communication

The ForeSight oximeter cable LED indicates the status of the tissue oximetry sensor channels.

**Figure 15-3 ForeSight oximeter cable LED indicators****Table 15-4 ForeSight oximeter cable LED communication lights**

LED indicator	Color	Indication
Channel 1 status	White	No sensor connected
	Green	Sensor connected
Channel 2 status	White	No sensor connected
	Green	Sensor connected
Module status	Green	Channels are associated with port A on HemoSphere technology module
	Blue	Channels are associated with port B on HemoSphere technology module

CAUTION If any of the ForeSight oximeter cable LEDs fail to turn on, the cable must not be used until it has been serviced or replaced. Contact Edwards Technical Support. There is a risk that damaged parts could reduce the performance of the cable.

15.5 Pressure Controller Communication

The pressure controller lights indicate the status of the finger cuff(s) and heart reference sensor.




Figure 15-4 Pressure Controller LED Indicators

Table 15-5 Pressure controller communication lights*

Condition	Color	Light Pattern	Suggested Action
CUFF STATUS LIGHT			
No finger cuff connected	No light	Solid OFF	None
Finger cuff connected	Green	Solid ON	None. The connected cuff is detected, authenticated, and not expired.
Active monitoring	Green	Flashing ON/OFF	None. The connected finger cuff is actively monitoring.
Defective finger cuff connected Expired finger cuff connected Non-compatible Edwards finger cuff connected	Amber	Flashing ON/OFF	Verify that a compatible Edwards finger cuff has been used. Disconnect and reconnect the finger cuff. Replace the finger cuff with a compatible Edwards finger cuff. Restart the measurement. If the problem persists, contact Edwards Technical Support.
HEART REFERENCE SENSOR STATUS LIGHT			
No heart reference sensor connected	No light	Solid OFF	None
Heart reference sensor connected	Green	Solid ON	None. The system is ready to start a measurement.
Defective heart reference sensor connected Non Edwards heart reference sensor detected	Amber	Flashing ON/OFF	Verify that an Edwards heart reference sensor has been used. Disconnect and reconnect the heart reference sensor. Replace the heart reference sensor with a genuine heart reference sensor. Restart the measurement. If the problem persists, contact Edwards Technical Support.
*Finger cuff error may also be indicated by software. See table 15-22 on page 355.			

15.6 HemoSphere Advanced Monitor Error Messages

15.6.1 System/Monitoring Faults/Alerts

Table 15-6 System faults/alerts

Message	Possible causes	Suggested actions
Fault: Module Slot 1 – Hardware Failure	Module 1 is not inserted properly Connection points on slot or module are damaged	Reinsert the module Check for bent or broken pins Try switching to module slot 2 If problem persists, contact Edwards Technical Support
Fault: Module Slot 2 – Hardware Failure	Module 2 is not inserted properly Connection points on slot or module are damaged	Reinsert the module Check for bent or broken pins Try switching to module slot 1 If problem persists, contact Edwards Technical Support
Fault: L-Tech Module Slot – Hardware Failure	Large technology module is not inserted properly Connection points on slot or module are damaged	Reinsert the module Check for bent or broken pins If problem persists, contact Edwards Technical Support
Fault: Cable Port 1 – Hardware Failure	Cable is not inserted properly Connection points on cable or port are damaged	Reinsert the cable Check for bent or broken pins Try switching to cable port 2 If problem persists, contact Edwards Technical Support
Fault: Cable Port 2 – Hardware Failure	Cable is not inserted properly Connection points on cable or port are damaged	Reinsert the cable Check for bent or broken pins Try switching to cable port 1 If problem persists, contact Edwards Technical Support
Fault: Module Slot 1 – Software Failure	There is a software error with the module inserted in module slot 1	Contact Edwards Technical Support
Fault: Module Slot 2 – Software Failure	There is a software error with the module inserted in module slot 2	Contact Edwards Technical Support
Fault: L-Tech Module Slot – Software Failure	There is a software error with the module inserted in the large technology module slot	Contact Edwards Technical Support
Fault: Cable Port 1 – Software Failure	There is a software error with the cable inserted in cable port 1	Contact Edwards Technical Support
Fault: Cable Port 2 – Software Failure	There is a software error with the cable inserted in cable port 2	Contact Edwards Technical Support
Fault: Module Slot 1 – Communication Error	Module 1 is not inserted properly Connection points on slot or module are damaged	Reinsert the module Check for bent or broken pins Try switching to module slot 2 If problem persists, contact Edwards Technical Support
Fault: Module Slot 2 – Communication Error	Module 2 is not inserted properly Connection points on slot or module are damaged	Reinsert the module Check for bent or broken pins Try switching to module slot 1 If problem persists, contact Edwards Technical Support
Fault: L-Tech Module Slot – Communication Error	Large technology module is not inserted properly Connection points on slot or module are damaged	Reinsert the module Check for bent or broken pins If problem persists, contact Edwards Technical Support

Table 15-6 System faults/alerts (continued)

Message	Possible causes	Suggested actions
Fault: Cable Port 1 – Communication Error	Cable is not inserted properly Connection points on cable or port are damaged	Reinsert the cable Check for bent or broken pins Try switching to cable port 2 If problem persists, contact Edwards Technical Support
Fault: Cable Port 2 – Communication Error	Cable is not inserted properly Connection points on cable or port are damaged	Reinsert the cable Check for bent or broken pins Try switching to cable port 1 If problem persists, contact Edwards Technical Support
Fault: Monitor – Incompatible Software Version	Unsuccessful software upgrade or incompatible software version detected	Contact Edwards Technical Support
Fault: Module Slot 1 – Incompatible Software Version	Unsuccessful software upgrade or incompatible software version detected	Contact Edwards Technical Support
Fault: Module Slot 2 – Incompatible Software Version	Unsuccessful software upgrade or incompatible software version detected	Contact Edwards Technical Support
Fault: L-Tech Module Slot – Incompatible Software Version	Unsuccessful software upgrade or incompatible software version detected	Contact Edwards Technical Support
Fault: Cable Port 1 – Incompatible Software Version	Unsuccessful software upgrade or incompatible software version detected	Contact Edwards Technical Support
Fault: Cable Port 2 – Incompatible Software Version	Unsuccessful software upgrade or incompatible software version detected	Contact Edwards Technical Support
Fault: Second Swan-Ganz Module Detected	Multiple Swan-Ganz module connections detected	Disconnect one of the Swan-Ganz modules
Fault: Swan-Ganz Module Disconnected	HemoSphere Swan-Ganz module removed during monitoring HemoSphere Swan-Ganz module not detected Connection points on slot or module are damaged	Confirm that module is properly inserted Remove and re-insert the module Check module for bent or broken pins Try switching to other module slot If problem persists, contact Edwards Technical Support
Fault: Cable Port {0} – Pressure Cable Disconnected*	Pressure cable disconnected during monitoring Pressure cable not detected Bent or missing pressure cable connector pins	Confirm that pressure cable is connected Verify that connection between pressure cable and sensor/transducer is secure Check pressure cable connector for bent/ missing pins Disconnect and reconnect pressure cable Try switching to other cable port If problem persists, contact Edwards Technical Support
Fault: Second Oximetry Cable Detected	Multiple oximetry cable connections detected	Disconnect one of the oximetry cables
Fault: Oximetry Cable Disconnected	Oximetry cable connection at HemoSphere advanced monitor not detected Bent or missing oximetry cable connector pins	Verify secure oximetry cable / catheter connection Check oximetry cable connector for bent/ missing pins
Fault: HemoSphere ClearSight Module	Defective HemoSphere ClearSight module	Power cycle the system Replace HemoSphere ClearSight module If problem persists, contact Edwards Technical Support
Fault: HemoSphere ClearSight Module Disconnected	HemoSphere ClearSight module removed during monitoring HemoSphere ClearSight module not detected Connection points on slot or module are damaged	Confirm that module is properly inserted Remove and re-insert the module Check module for bent or broken pins If problem persists, contact Edwards Technical Support
Fault: Internal System Failure	Internal system malfunction	Power cycle the system If problem persists, contact Edwards Technical Support

Table 15-6 System faults/alerts (continued)

Message	Possible causes	Suggested actions
Fault: Battery Depleted	The battery is depleted and the system will shut down in 1 minute if not plugged in	Connect the HemoSphere advanced monitor to an alternate source of power to avoid loss of power and resume monitoring
Fault: System Temperature Too High - Shutdown Imminent	The internal temperature of the monitor is at a critically high level Monitor ventilation openings are obstructed	Reposition the monitor away from any heat sources Ensure that the monitor ventilation openings are unobstructed and clear of dust If problem persists, contact Edwards Technical Support
Fault: Pressure-Out – Hardware Failure	Pressure-out cable is not properly connected Connection points on cable or port are damaged	Reinsert the pressure-out cable Check for bent or broken pins If problem persists, contact Edwards Technical Support
Fault: HIS Connectivity Loss	There was a loss in HL7 communication Poor Ethernet connection Poor Wi-Fi connection Expired Secure Connection certificate Incorrect Secure Connection server name	Check ethernet connection Check Wi-Fi connection Check Secure Connection certificate Check Secure Connection server name If problem persists, contact Edwards Technical Support
Fault: Second CO Pressure Sensor Detected	Multiple pressure cables with CO sensor connections detected	Disconnect one of the pressure cable CO sensors
Fault: Wireless Module Failure	There was an internal hardware failure in the wireless module	Disable and re-enable wireless connection
Fault: CPO - Inputs Out of Display Range	MAP or CO value is out of the indicated display range so CPO cannot be calculated	Check analog input connections, if being used Wait for parameters to come back in range If problem persists, contact Edwards Technical Support
Fault: SVR - Inputs Out of Display Range	MAP, CVP, or CO value is out of the indicated display range so SVR cannot be calculated	Check analog input connections, if being used Wait for parameters to come back in range If problem persists, contact Edwards Technical Support
Alert: System Temperature Too High	The internal temperature of the monitor is reaching a critically high level Monitor ventilation openings are obstructed	Reposition the monitor away from any heat sources Ensure that the monitor ventilation openings are unobstructed and clear of dust If problem persists, contact Edwards Technical Support
Alert: System LED Indicators Inoperable	Visual alarm indicator hardware or communication error Visual alarm indicator malfunction	Power cycle the system If problem persists, contact Edwards Technical Support
Alert: System Buzzer Inoperable	Speaker hardware or software communication error Mainboard speaker malfunction	Power cycle the system If problem persists, contact Edwards Technical Support
Alert: Low Battery	The battery has less than 20% charge remaining or will be depleted within 8 minutes	Connect the HemoSphere advanced monitor to an alternate source of power to avoid loss of power and continue monitoring
Alert: Battery Disconnected	Previously inserted battery not detected Poor battery connection	Confirm battery is properly seated in the battery bay Remove and reinsert the battery pack Change HemoSphere battery pack If problem persists, contact Edwards Technical Support
Alert: Service Battery	Internal battery fault occurred Battery can no longer sustain the system adequately on a full charge	Power cycle the system If condition persists, replace the battery pack
Alert: Wireless Certificate Expires < 4 Weeks	Wireless certificate expires in less than 4 weeks	Navigate to Wireless Connectivity settings from the Advanced Setup menu and upload a valid certificate If problem persists, contact Edwards Technical Support

Table 15-6 System faults/alerts (continued)

Message	Possible causes	Suggested actions
Alert: Wireless Certificate Expired	Wireless certificate is expired	Navigate to Wireless Connectivity settings from the Advanced Setup menu and upload a valid certificate If problem persists, contact Edwards Technical Support
Alert: Transmit Pressure Not Active	Connection of new patient monitor pressure channel detected	Navigate to Zero & Waveform screen, and touch transmit pressure button (waveform icon) after zeroing patient monitor Disconnect the pressure-out cable
<i>*Note: {0} is the port number: 1 or 2.</i>		

15.6.2 System/Monitoring Warnings

Table 15-7 HemoSphere advanced monitor warnings

Message	Possible causes	Suggested actions
Battery Needs Conditioning	Gas gauge is not synched to actual battery capacity status	To ensure uninterrupted measurement, make certain the HemoSphere advanced monitor is connected to electrical outlet Condition the battery (ensure a measurement is not active): - Connect monitor to an electrical outlet to fully charge battery - Allow the battery to rest in fully charged state for at least two hours - Disconnect the monitor from electrical outlet and continue to run the system on battery power - The HemoSphere advanced monitor will power down automatically when the battery is fully depleted - Allow the battery to rest in fully depleted state for five hours or more - Connect monitor to an electrical outlet to fully charge battery If the condition battery message persists, replace battery pack
Service Battery	Internal battery fault occurred	Power cycle the system If condition persists, replace the battery pack

15.6.3 Numeric Keypad Errors

Table 15-8 Numeric keypad errors

Message	Possible causes	Suggested actions
Value out of range (xx-yy)	The entered value is either higher or lower than the allowed range	Displayed when the user enters a value that is out of range. The range is displayed as part of the notification replacing the xx and yy.
Value must be $\leq$ xx	The entered value is in range, but is higher than the high value setting such as the high scale setting (xx is the associated value)	Enter a lower value
Value must be $\geq$ xx	The entered value is in range, but is lower than the low value setting such as the low scale setting (xx is the associated value)	Enter a higher value
Incorrect password entered	The password entered is incorrect.	Enter the correct password
Please enter valid time	The time entered is invalid, i.e. 25:70	Enter the correct time in 12- or 24-hour format
Please enter valid date	The date entered is invalid, i.e. 33.13.009	Enter the correct date

15.6.4 Viewfinder Hub Connectivity Errors

Table 15-9 Viewfinder hub connectivity errors

Message	Possible causes	Suggested actions
Viewfinder Hub Connectivity Error – Viewfinder Hub	Issue with Viewfinder hub Incorrect server certificate Viewfinder hub pairing request rejected	Check Viewfinder hub server Contact your local IT If problem persists, contact Edwards Technical Support
Viewfinder Hub Connectivity Error – Viewfinder Hub Not Reachable	Wrong Viewfinder hub address or port Viewfinder hub not running on server	Verify and re-enter Viewfinder hub address and port Check Viewfinder hub server If problem persists, contact Edwards Technical Support
Viewfinder Hub Connectivity Error – HemoSphere Monitor	Client certificate invalid or unavailable	Contact Edwards Technical Support

15.7 HemoSphere Swan-Ganz Module Error Messages

15.7.1 CO Faults/Alerts

Table 15-10 HemoSphere Swan-Ganz module CO faults/alerts

Message	Possible causes	Suggested actions
Fault: CO – Blood Temp Out of Range (<31 °C or >41 °C)*	Monitored blood temperature is <31 °C or >41 °C	Verify proper catheter position in the pulmonary artery: - confirm wedge pressure balloon inflation volume of 1.25 - 1.50 mL - confirm appropriate catheter placement for patient's height, weight, and insertion site - consider chest x-ray for evaluation of proper placement Resume CO monitoring when blood temperature is within range
Fault: CO – Cardiac Output < 1.0 L/min*	Measured CO < 1.0 L/min	Follow hospital protocol to increase CO Resume CO monitoring
Fault: CO – Catheter Memory, Use Bolus Mode	Poor catheter thermal filament connection Patient CCO cable malfunction Catheter CO error Patient CCO cable is connected to cable test ports	Verify secure thermal filament connection Check catheter/ patient CCO cable thermal filament connections for bent/missing pins Perform patient CCO cable test Change patient CCO cable Use Bolus CO mode Replace catheter for CO measurement
Fault: CO – Catheter Verification, Use Bolus Mode	Patient CCO cable malfunction Catheter CO error Catheter connected is not an Edwards CCO catheter	Perform patient CCO cable test Change patient CCO cable Use Bolus CO mode Verify catheter is an Edwards CCO catheter
Fault: CO – Check Catheter and Cable Connections	Catheter thermal filament and thermistor connections not detected Patient CCO cable malfunction	Verify patient CCO cable and catheter connections Disconnect thermistor and thermal filament connections and check for bent/missing pins Perform patient CCO cable test Change patient CCO cable

Table 15-10 HemoSphere Swan-Ganz module CO faults/alerts (continued)

Message	Possible causes	Suggested actions
Fault: CO – Check Thermal Filament Connection	Catheter thermal filament connection not detected Patient CCO cable malfunction Catheter connected is not an Edwards CCO catheter	Verify that catheter thermal filament is connected securely to patient CCO cable Disconnect thermal filament connection and check for bent/missing pins Perform patient CCO cable test Change patient CCO cable Verify catheter is an Edwards CCO catheter Use Bolus CO mode
Fault: CO – Check Thermal Filament Position*	Flow around thermal filament may be reduced Thermal filament may be against vessel wall Catheter not in patient	Flush catheter lumens Verify proper catheter positions in the pulmonary artery: - confirm wedge pressure balloon inflation volume of 1.25 - 1.50 mL - confirm appropriate catheter placement for patient's height, weight, and insertion site - consider chest x-ray for evaluation of proper placement Resume CO monitoring
Fault: CO – Check Thermistor Connection	Catheter thermistor connection not detected Monitored blood temperature is <15 °C or >45 °C Patient CCO cable malfunction	Verify that catheter thermistor is connected securely to patient CCO cable Verify that blood temperature is between 15 - 45 °C Disconnect thermistor connection and check for bent/missing pins Perform patient CCO cable test Change patient CCO cable
Fault: CO – Signal Processor, Use Bolus Mode	Data processing error	Resume CO monitoring Power monitor off and on to restore system Use Bolus CO mode
Fault: CO – Thermal Signal Loss*	Thermal signal detected by monitor is too small to process Sequential compression device interference	Verify proper catheter position in the pulmonary artery: - confirm wedge pressure balloon inflation volume of 1.25 - 1.50 mL - confirm appropriate catheter placement for patient's height, weight, and insertion site - consider chest x-ray for evaluation of proper placement Temporarily turn off sequential compression device per hospital procedure Resume CO monitoring
Fault: Swan-Ganz Module	Electrocautery interference Internal system malfunction	Disconnect patient CCO cable during electrocautery use Remove and reinsert module to reset If problem persists, contact Edwards Technical Support
Alert: CO – Signal Adapting - Continuing	Large pulmonary artery blood temperature variations detected Patient's respiratory pattern may have changed Sequential compression device interference Catheter thermal filament not properly positioned	Allow more time for monitor to measure and display CO Verify proper catheter position in the pulmonary artery: - confirm wedge pressure balloon inflation volume of 1.25 - 1.50 mL - confirm appropriate catheter placement for patient's height, weight, and insertion site - consider chest x-ray for evaluation of proper placement Minimizing patient discomfort may reduce temperature variations Temporarily turn off sequential compression device per hospital procedure

Table 15-10 HemoSphere Swan-Ganz module CO faults/alerts (continued)

Message	Possible causes	Suggested actions
Alert: CO – Unstable Blood Temp. - Continuing	Large pulmonary artery blood temperature variations detected Sequential compression device interference	Wait for CO measurement to be updated Minimizing patient discomfort may reduce temperature variations Temporarily turn off sequential compression device per hospital procedure
Alert: CPO – Analog Input MAP Pressure Signal Loss	HemoSphere advanced monitor analog input port not configured to accept MAP Analog input interface cable connections not detected Inaccurate input signal External monitor malfunction	Verify correct voltage range and low/high voltage values on the HemoSphere advanced monitor for external monitor Verify cable connection between the monitoring platform and bedside monitor is secure Verify correct height/weight entries and units of measure for patient's BSA Check for signal at external monitor's analog output device Change external device module, if used
Alert: CPO - Configure MAP Analog Input for CPO Monitoring	HemoSphere advanced monitor analog input ports not configured to accept MAP signal	Use the analog input settings screen to configure analog input port 1 or 2 for external monitor MAP signal output
<i>*These are latching faults. Touch the silence icon to silence. To clear, restart monitoring.</i>		

15.7.2 EDV and SV Faults/Alerts

Table 15-11 HemoSphere Swan-Ganz module EDV and SV faults/alerts

Message	Possible causes	Suggested actions
Alert: EDV – Heart Rate Signal Missing	Patient's time-averaged heart rate out of range ($HR_{avg} < 30$ or > 200 bpm) No heart rate detected ECG interface cable connection not detected	Wait until average heart rate is within range Select appropriate lead configuration to maximize heart rate triggers Verify cable connection between the HemoSphere advanced monitor and bedside monitor is secure Change ECG interface cable
Alert: EDV – Exceeding HR Threshold Limit	Patient's time-averaged heart rate out of range ($HR_{avg} < 30$ or > 200 bpm)	Wait until average heart rate is within range Select appropriate lead configuration to maximize heart rate triggers Verify cable connection between the HemoSphere advanced monitor and bedside monitor is secure Change ECG interface cable
Alert: EDV – Signal Adapting - Continuing	Patient's respiratory pattern may have changed Sequential compression device interference Catheter thermal filament not properly positioned	Allow more time for monitor to measure and display EDV Temporarily turn off sequential compression device per hospital procedure Verify proper catheter position in the pulmonary artery: - confirm wedge pressure balloon inflation volume of 1.25 - 1.50 mL - confirm appropriate catheter placement for patient's height, weight, and insertion site - consider chest x-ray for evaluation of proper placement
Alert: SV – Heart Rate Signal Missing	Patient's time-averaged heart rate out of range ($HR_{avg} < 30$ or > 200 bpm) No heart rate detected ECG interface cable connection not detected	Wait until average heart rate is within range Select appropriate lead configuration to maximize heart rate triggers Verify cable connection between HemoSphere advanced monitor and bedside monitor is secure Change ECG interface cable

15.7.3 iCO Faults/Alerts

Table 15-12 HemoSphere Swan-Ganz module iCO faults/alerts

Message	Possible causes	Suggested actions
Fault: iCO – Check Injectate Probe Connection	Injectate temperature probe not detected Injectate temperature probe malfunction Patient CCO cable malfunction	Verify connection between patient CCO cable and injectate temperature probe Change injectate temperature probe Change patient CCO cable
Fault: iCO – Check Thermistor Connection	Catheter thermistor connection not detected Monitored blood temperature is <15 °C or >45 °C Patient CCO cable malfunction	Verify that catheter thermistor is connected securely to patient CCO cable Verify that blood temperature is between 15 – 45 °C Disconnect thermistor connection and check for bent/missing pins Change patient CCO cable
Fault: iCO – Injectate Volume Not Valid	In-line probe injectate volume must be 5 mL or 10 mL	Change injectate volume to 5 mL or 10 mL Use a bath type probe for an injectate volume of 3 mL
Fault: iCO – Injectate Temperature Out of Range, Check Probe	Injectate temperature < 0 °C, > 30 °C or > BT Injectate temperature probe malfunction Patient CCO cable malfunction	Verify injectate fluid temperature Check injectate probe connections for bent/missing pins Change injectate temperature probe Change patient CCO cable
Fault: iCO – Blood Temperature Out of Range	Monitored blood temperature is <31 °C or >41 °C	Verify proper catheter position in the pulmonary artery: - confirm wedge pressure balloon inflation volume of 1.25 - 1.50 mL - confirm appropriate catheter placement for patient's height, weight, and insertion site - consider chest x-ray for evaluation of proper placement Resume bolus injections when blood temperature is within range
Alert: iCO – Unstable Baseline	Large pulmonary artery blood temperature variations detected	Allow more time for blood temperature baseline to stabilize Use Manual mode
Alert: iCO – Curve Not Detected	No bolus injection detected for >4 minutes (Automatic mode) or 30 seconds (Manual mode)	Restart Bolus CO monitoring and proceed with injections
Alert: iCO – Extended Curve	Thermodilution curve slow to return to baseline Injectate port in introducer sheath Possible cardiac shunt	Verify correct injection technique Verify proper catheter position in the pulmonary artery: - confirm wedge pressure balloon inflation volume of 1.25 - 1.50 mL - confirm appropriate catheter placement for patient's height, weight and insertion site - consider chest x-ray for evaluation of proper placement Ensure injectate port location is outside of the introducer sheath Use "iced" injectate and/or 10 mL injectate volume to create a larger thermal signal
Alert: iCO – Irregular Curve	Thermodilution curve has multiple peaks	Verify correct injection technique Verify proper catheter position in the pulmonary artery: - confirm wedge pressure balloon inflation volume of 1.25 - 1.50 mL - confirm appropriate catheter placement for patient's height, weight, and insertion site - consider chest x-ray for evaluation of proper placement Use "iced" injectate and/or 10 mL injectate volume to create a larger thermal signal

Table 15-12 HemoSphere Swan-Ganz module iCO faults/alerts (continued)

Message	Possible causes	Suggested actions
Alert: iCO – Warm Injectate	Injectate temperature within 8 °C of blood temperature Injectate temperature probe malfunction Patient CCO cable malfunction	Use cooler injectate fluid Change injectate temperature probe Change patient CCO cable

15.7.4 SVR Faults/Alerts

Table 15-13 HemoSphere Swan-Ganz module SVR faults/alerts

Message	Possible causes	Suggested actions
Alert: SVR – Analog Input MAP Pressure Signal Loss	HemoSphere advanced monitor analog input port not configured to accept MAP Analog input interface cable connections not detected Inaccurate input signal External monitor malfunction	Verify correct voltage range and low/high voltage values on HemoSphere advanced monitor for external monitor Verify cable connection between the HemoSphere advanced monitor and bedside monitor is secure Verify correct height/weight entries and units of measure for patient's BSA Check for signal at external monitor's analog output device Change external device module, if used
Alert: SVR – Configure MAP Analog Input for SVR Monitoring	HemoSphere advanced monitor analog input port is not configured to accept MAP signal	Use the analog input settings screen to configure analog input port 1 or 2 for external monitor MAP signal output

15.7.5 20-Second Parameters Faults/Alerts

Table 15-14 HemoSphere Swan-Ganz module 20s parameters faults/alerts

Message	Possible causes	Suggested actions
Fault: 20s Parameters – Poor PA Signal Quality	Pulmonary artery pressure waveform is inadequate to measure 20s parameters accurately Poor pressure waveform over extended period of time Integrity of pressure monitoring line is compromised Pressure waveform has shifted or is measuring negative signals due to change in phlebostatic axis or other related movement impacting pressure signal	Verify proper catheter position in the pulmonary artery: - confirm wedge pressure balloon inflation volume of 1.25 - 1.50 mL - confirm appropriate catheter placement for patient's height, weight, and insertion site - consider chest x-ray for evaluation of proper placement Make sure the pulmonary artery pressure line is not kinked Make sure there are no loose connections Perform Square Wave Test to assess the frequency response of the system Re-zero pulmonary artery pressure transducer
Fault: 20s Parameters – Software Failure	There is a software error with the 20s parameters	Power cycle the system Re-zero pulmonary artery pressure transducer If problem persists, contact Edwards Technical Support
Alert: 20s Parameters – Negative PA Pressure Detected	Pulmonary artery pressure waveform is inadequate to measure 20s parameters accurately Pressure transducer is not aligned with the patient's phlebostatic axis Integrity of pressure monitoring line is compromised	Verify proper catheter position in the pulmonary artery: - confirm wedge pressure balloon inflation volume of 1.25 - 1.50 mL - confirm appropriate catheter placement for patient's height, weight, and insertion site - consider chest x-ray for evaluation of proper placement Confirm the pressure transducer is aligned with the patient's phlebostatic axis Zero the pressure transducer on the HemoSphere advanced monitor to re-zero the transducer and confirm pressure cable connection

15.7.6 General Troubleshooting

Table 15-15 HemoSphere Swan-Ganz module general troubleshooting

Message	Possible causes	Suggested actions
Connect HemoSphere Swan-Ganz module for CO monitoring	Connection to the HemoSphere Swan-Ganz module has not been detected	Insert the HemoSphere Swan-Ganz module into slot 1 or slot 2 of the monitor Remove and re-insert module
Connect patient CCO cable for CO monitoring	Connection between the HemoSphere Swan-Ganz module and patient CCO cable has not been detected	Verify connection between patient CCO cable and the inserted HemoSphere Swan-Ganz module Disconnect patient CCO cable and check for bent/missing pins Change patient CCO cable
Connect thermistor for CO monitoring	Connection between patient CCO cable and catheter thermistor has not been detected Patient CCO cable malfunction	Verify that catheter thermistor is connected securely to patient CCO cable Disconnect thermistor connection and check for bent/missing pins Perform patient CCO cable test Change patient CCO cable
Connect thermal filament for CO monitoring	Connection between patient CCO cable and catheter thermal filament has not been detected Patient CCO cable malfunction Catheter connected is not an Edwards CCO catheter	Verify that catheter thermal filament is connected securely to patient CCO cable Disconnect thermal filament connection and check for bent/missing pins Perform patient CCO cable test Change patient CCO cable Verify catheter is an Edwards CCO catheter
Connect MAP analog input for SVR monitoring	Analog input interface cable connections not detected	Verify cable connection between the monitoring platform and bedside monitor is secure Check for signal at external monitor's analog output device
Configure MAP analog input for SVR monitoring	HemoSphere advanced monitor analog input port is not configured to accept MAP signal	Use the analog input settings screen to configure analog input port 1 or 2 for external monitor MAP signal output
Connect ECG Input for EDV or SV monitoring	ECG interface cable connection not detected	Verify cable connection between the HemoSphere advanced monitor and bedside monitor is secure Change ECG interface cable
Connect pressure cable for 20s parameter monitoring	Connection between the HemoSphere advanced monitor and pressure cable has not been detected	Verify connection between pressure cable and monitor Disconnect pressure cable and check for bent/missing pins Change pressure cable
Connect pulmonary artery pressure sensor for 20s parameter monitoring	CO _{20s} , CI _{20s} , SV _{20s} or SVI _{20s} is configured as a key parameter Connection between the pressure cable and a pulmonary artery pressure sensor has not been detected	Verify connection between pressure cable and monitor Disconnect pressure cable and check for bent/missing pins Change pressure cable
Zero pulmonary artery pressure for 20s parameter monitoring	The pulmonary artery pressure signal was not zeroed prior to monitoring	Touch the "Zero & Waveform" icon on the navigation bar
CI > CO	Incorrect patient BSA BSA < 1	Verify units of measure and values for patient's height and weight.
CO ≠ iCO	Incorrectly configured bolus information Faulty thermistor or injectate probe Unstable baseline temperature affecting bolus CO measurements	Verify that computation constant, injectate volume, and catheter size have been correctly selected Use "iced" injectate and/or 10 mL injectate volume to create a large thermal signal Verify correct injection technique Change injectate temperature probe

Table 15-15 HemoSphere Swan-Ganz module general troubleshooting (continued)

Message	Possible causes	Suggested actions
SVR > SVRI	Incorrect patient BSA BSA <1	Verify units of measure and values for patient's height and weight
HemoSphere Advanced Monitor HR _{avg} ≠ External Monitor HR	External monitor not optimally configured for ECG signal output External monitor malfunction ECG interface cable malfunction Elevated patient heart rate HemoSphere advanced monitor uses up to 3 minutes of HR data to calculate HR _{avg}	Stop CO monitoring and verify heart rate is the same for HemoSphere advanced monitor and external monitor Select appropriate lead configuration to maximize heart rate triggers and minimize atrial spike sensing Verify signal output from external monitoring device Wait for patient's HR to stabilize Change ECG interface cable
HemoSphere Advanced Monitor Display of MAP and CVP ≠ External Monitor	HemoSphere advanced monitoring platform configured incorrectly Inaccurate input signal External monitor malfunction	Verify correct voltage range and low/high voltage values on HemoSphere advanced monitor for external monitor Confirm correct units of measure for analog input port voltage values (mmHg or kPa) Verify correct height/weight entries and units of measure for patient's BSA Check for signal at external monitor's analog output device Change analog input interface cable

15.8 Pressure Cable Error Messages

15.8.1 General Pressure Cable Faults/Alerts

Table 15-16 HemoSphere pressure cable general faults/alerts

Message	Possible causes	Suggested actions
Fault: Cable Port {0} –Pressure Cable*	Internal system malfunction	Disconnect and reconnect pressure cable Reposition the cable away from any heat sources or insulating surfaces If the cable body feels warm, allow it to cool before operating again Power monitor off and on to restore platform If problem persists, contact Edwards Technical Support
Fault: Cable Port {0} – Pressure Sensor*	Cable or sensor malfunction Damaged or defective sensor	Disconnect sensor and check for bent/missing contacts Change pressure sensor Change pressure cable If problem persists, contact Edwards Technical Support
Fault: Cable Port {0} – Pressure Sensor Disconnected*	Pressure sensor disconnected during monitoring Cable connections not detected Edwards pressure cable or sensor malfunction Internal system malfunction	Verify catheter connection Verify pressure cable and sensor and check for missing pins Change Edwards pressure cable Change Edwards CO/pressure sensor If problem persists, contact Edwards Technical Support
Fault: Cable Port {0}– Incompatible Pressure Sensor*	A non-Edwards sensor has been detected Cable or sensor malfunction Internal system malfunction	Verify that an Edwards pressure sensor has been used Disconnect sensor and check for bent/missing contacts Change pressure sensor Change pressure cable If problem persists, contact Edwards Technical Support

Table 15-16 HemoSphere pressure cable general faults/alerts (continued)

Message	Possible causes	Suggested actions
Fault: Cable Port {0} – Pressure Waveform Not Stable*	Arterial waveform is inadequate to measure blood pressure accurately Integrity of pressure monitoring line is compromised Systolic pressure too high or diastolic pressure too low Fluid line is being flushed	Assess Edwards pressure monitoring system starting from patient leading to pressure bag Check the arterial waveform for severe hypotension, severe hypertension, and motion artifact Make sure the arterial catheter is not kinked or clotted Make sure all arterial pressure lines are patent and stopcocks are properly positioned Make sure Edwards pressure sensor/transducer is aligned with the patient's phlebostatic axis Zero the Edwards pressure sensor/transducer on HemoSphere advanced monitor and confirm pressure cable connection Make sure the pressure bag is inflated and flush bag is at least ¼ full. Perform Square Wave Test to assess the Edwards continuous CO system frequency response
Alert: Cable Port {0} – Release Pressure Cable Zero Button*	The pressure cable zero button has been depressed for more than 10 seconds Pressure cable malfunction	Release the pressure cable zero button Check that the button releases properly Replace the pressure cable
*Note: {0} is the port number: 1 or 2.		

15.8.2 Arterial Pressure Faults/Alerts

Table 15-17 HemoSphere pressure cable ART faults/alerts

Message	Possible causes	Suggested actions
Fault: Check Arterial Waveform	Arterial waveform is inadequate to measure blood pressure accurately Poor pressure waveform over extended period of time Integrity of pressure monitoring line is compromised Systolic pressure too high or diastolic pressure too low	Assess Edwards pressure monitoring system starting from patient leading to pressure bag Check the arterial waveform for severe hypotension, severe hypertension, and motion artifact Make sure the arterial catheter is not kinked or clotted Make sure all arterial pressure lines are patent and stopcocks are properly positioned Make sure Edwards pressure sensor/transducer is aligned with the patient's phlebostatic axis Zero the Edwards pressure sensor/transducer on HemoSphere advanced monitor and confirm pressure cable connection Make sure the pressure bag is inflated and flush bag is at least ¼ full Perform Square Wave Test to assess the Edwards pressure monitoring system frequency response

Table 15-17 HemoSphere pressure cable ART faults/alerts (continued)

Message	Possible causes	Suggested actions
Fault: Arterial Waveform Compromised	Edwards pressure cable or sensor malfunction Internal system malfunction Patient condition results in a low pulse pressure Integrity of pressure monitoring line is compromised Pressure transducer is not aligned with the patient's phlebostatic axis	Assess Edwards pressure monitoring system starting from patient leading to pressure bag Check the arterial waveform for severe hypotension, severe hypertension, and motion artifact Make sure the arterial catheter is not kinked or clotted Make sure all arterial pressure lines are patent and stopcocks are properly positioned Make sure Edwards pressure sensor/transducer is aligned with the patient's phlebostatic axis Zero the Edwards pressure sensor/transducer on HemoSphere advanced monitor and confirm pressure cable connection Make sure the pressure bag is inflated and flush bag is at least ¼ full Enter Non-Pulsatile Mode Perform Square Wave Test to assess the Edwards pressure monitoring system frequency response Verify Edwards pressure cable and sensor and check for missing pins Change Edwards pressure cable Change Edwards CO/pressure sensor If problem persists, contact Edwards Technical Support
Fault: Arterial Pressure Disconnected	Arterial pressure low and non-pulsatile Arterial catheter disconnected Cable connections not detected Edwards pressure cable or sensor malfunction Internal system malfunction	Verify arterial catheter connection Verify connection between pressure cable and sensor and check for missing pins Change pressure cable Change pressure sensor If problem persists, contact Edwards Technical Support
Alert: Pulse Pressure Low	Integrity of pressure monitoring line is compromised Patient condition results in a low pulse pressure	Assess Edwards pressure monitoring system starting from patient leading to pressure bag Check the arterial waveform for severe hypotension, severe hypertension, and motion artifact Make sure the arterial catheter is not kinked or clotted Make sure all arterial pressure lines are patent and stopcocks are properly positioned Make sure Edwards pressure sensor/transducer is aligned with the patient's phlebostatic axis Zero the Edwards pressure sensor/transducer on HemoSphere advanced monitor and confirm pressure cable connection Make sure the pressure bag is inflated and flush bag is at least ¼ full Perform Square Wave Test to assess the Edwards pressure monitoring system frequency response

Table 15-17 HemoSphere pressure cable ART faults/alerts (continued)

Message	Possible causes	Suggested actions
Alert: Arterial Pressure Waveform Not Stable	Arterial waveform is inadequate to measure blood pressure accurately Integrity of pressure monitoring line is compromised Systolic pressure too high or diastolic pressure too low Fluid line is being flushed	Assess Edwards pressure monitoring system starting from patient leading to pressure bag Check the arterial waveform for severe hypotension, severe hypertension, and motion artifact Make sure the arterial catheter is not kinked or clotted Make sure all arterial pressure lines are patent and stopcocks are properly positioned Make sure Edwards pressure sensor/transducer is aligned with the patient's phlebostatic axis Zero the Edwards pressure sensor/transducer on HemoSphere advanced monitor and confirm pressure cable connection Make sure the pressure bag is inflated and flush bag is at least ¼ full Perform Square Wave Test to assess the Edwards pressure monitoring system frequency response

15.8.3 Assisted Fluid Management Faults/Alerts

Table 15-18 HemoSphere pressure cable AFM faults/alerts

Message	Possible causes	Suggested actions
Fault: Assisted Fluid Management	Data processing error while initializing Assisted Fluid Management algorithm Internal system malfunction Integrity of pressure monitoring line is compromised	Assess arterial waveform and continuous CO system Restart AFM session If problem persists, contact Edwards Technical Support
Fault: AFM Cable	Internal system malfunction	Disconnect and reconnect Acumen AFM Cable Replace Acumen AFM Cable If problem persists, contact Edwards Technical Support
Fault: AFM Cable – Incompatible Software Version	Unsuccessful software upgrade or incompatible software version detected	Contact Edwards Technical Support
Fault: Second AFM Cable Detected	Multiple Acumen AFM Cable connections detected	Disconnect one of the Acumen AFM Cables
Fault: AFM Cable Disconnected	Acumen AFM Cable has become disconnected	Connect Acumen AFM Cable to HemoSphere Technology Module Continue AFM in Manual Fluid Tracking mode
Fault: Fluid Meter Disconnected	Acumen IQ Fluid Meter has become disconnected	Connect Acumen IQ Fluid Meter to Acumen AFM Cable Continue AFM in Manual Fluid Tracking mode
Fault: Fluid Meter	Damaged or defective Acumen IQ Fluid Meter	Disconnect Acumen IQ Fluid Meter and check for bent / missing contacts Replace Acumen IQ Fluid Meter If problem persists, contact Edwards Technical Support
Fault: Incompatible Fluid Meter	Non Edwards fluid meter in use Damaged or defective Acumen IQ Fluid Meter	Verify that an Edwards fluid meter is being used Disconnect and reconnect Acumen IQ Fluid Meter Replace fluid meter with genuine Acumen IQ Fluid Meter If problem persists, contact Edwards Technical Support
Alert: AFM – Exceeded Maximum Case Volume	Tracked volume has exceeded configured Maximum Case Volume	Set a new Maximum Case Volume limit End the AFM session
Alert: AFM – Detected Flow Rate Too High	Tracked bolus flow rate through fluid meter has exceeded 8.0 L/hr	Reduce bolus flow rate to below 8.0 L/hr Continue AFM session in Manual fluid tracking mode
Alert: AFM – Bolus Detected During Initialization	Fluid bolus detected during initialization of AFM session	Close bolus line and retry AFM initialization

Table 15-18 HemoSphere pressure cable AFM faults/alerts (continued)

Message	Possible causes	Suggested actions
Alert: Fluid Meter Detected	AFM is in Manual fluid tracking mode but Acumen IQ Fluid Meter is connected	Disconnect Acumen IQ Fluid Meter Select to continue AFM in Fluid Meter mode
Alert: AFM – Bolus Detected During AFM Analysis	Additional fluid bolus detected during ongoing AFM bolus analysis	When possible, deliver fluids after bolus analysis is complete

Table 15-19 HemoSphere pressure cable AFM warnings

Message	Possible causes	Suggested actions
AFM – Approaching Maximum Case Volume	Tracked volume is within 500 mL of configured Maximum Case Volume	Acknowledge and continue AFM session Set a new Maximum Case Volume limit

15.8.4 SVR Faults/Alert

Table 15-20 HemoSphere pressure cable SVR faults/alerts

Message	Possible causes	Suggested actions
Alert: SVR – Analog Input CVP Pressure Signal Loss	HemoSphere advanced monitor analog input port not configured to accept CVP Analog input interface cable connection not detected Inaccurate input signal External monitor malfunction	Verify correct voltage range and low/high voltage values on HemoSphere advanced monitor for external monitor Verify cable connection between HemoSphere advanced monitor and bedside monitor is secure Verify correct height/weight entries and units of measure for patient's BSA Check for signal at external monitor's analog output device Change external device module, if used

15.8.5 General Troubleshooting

Table 15-21 HemoSphere pressure cable general troubleshooting

Message	Possible causes	Suggested actions
Connect pressure cable for CO or pressure monitoring	Connection between the HemoSphere advanced monitor and pressure cable has not been detected	Verify connection between pressure cable and monitor Disconnect pressure cable and check for bent/missing pins Change pressure cable
Connect CO pressure sensor for CO monitoring	A CO-dependent key parameter is configured Connection between the pressure cable and CO pressure sensor has not been detected The incorrect pressure sensor type is connected	Verify connection between pressure cable and catheter Verify that the pressure sensor connected is for CO monitoring Disconnect pressure cable and check for missing pins Change Edwards CO sensor Change pressure cable
Connect pressure sensor for arterial pressure monitoring	An arterial pressure-dependent key parameter is configured Connection between the pressure cable and an arterial pressure sensor has not been detected	Verify connection between pressure cable and catheter Disconnect pressure cable and check for missing pins Change Edwards pressure sensor Change pressure cable
Connect pressure sensor for pulmonary artery monitoring	MPAP is configured as a key parameter Connection between the pressure cable and a pulmonary artery pressure sensor has not been detected	Verify connection between pressure cable and catheter Disconnect pressure cable and check for missing pins Change Edwards pressure sensor Change pressure cable

Table 15-21 HemoSphere pressure cable general troubleshooting (continued)

Message	Possible causes	Suggested actions
Connect pressure sensor for CVP monitoring	CVP is configured as a key parameter Connection between the pressure cable and a central venous pressure sensor has not been detected	Verify connection between pressure cable and catheter Disconnect pressure cable and check for missing pins Change Edwards pressure sensor Change pressure cable
Zero arterial pressure for CO monitoring	The arterial pressure signal was not zeroed prior to CO monitoring	Touch the “Zero & Waveform” icon on the navigation bar or from the Clinical Actions Menu to zero pressure
Zero pressure for arterial pressure monitoring	The arterial pressure signal was not zeroed prior to monitoring	Touch the “Zero & Waveform” icon on the navigation bar or from the Clinical Actions Menu to zero pressure
Zero pressure for pulmonary artery monitoring	The pulmonary artery pressure signal was not zeroed prior to monitoring	Touch the “Zero & Waveform” icon on the navigation bar or from the Clinical Actions Menu to zero pressure
Zero pressure for CVP monitoring	The central venous pressure signal was not zeroed prior to monitoring	Touch the “Zero & Waveform” icon on the navigation bar or from the Clinical Actions Menu to zero pressure
Connect CVP analog input or enter CVP value for SVR monitoring	CVP cable connection not detected No CVP value entered	Verify cable connection between the HemoSphere advanced monitor and bedside monitor is secure Change CVP cable Enter CVP value
Configure CVP analog input or enter CVP for SVR monitoring	HemoSphere advanced monitor analog input port not configured to accept CVP signal No CVP value entered	Use the analog input settings screen to configure analog input ports 1 or 2 for external monitor CVP signal output Enter CVP value
CI > CO	Incorrect patient BSA BSA <1	Verify units of measure and values for patient's height and weight.
SVR > SVRI	Incorrect patient BSA BSA <1	Verify units of measure and values for patient's height and weight

15.9 HemoSphere ClearSight Module Error Messages

15.9.1 Faults/Alerts

Table 15-22 HemoSphere ClearSight module faults/alerts

Message	Possible causes	Suggested actions
Fault: Finger Cuff #1 – BP Measurement Error Fault: Finger Cuff #2 – BP Measurement Error	Blood pressure measurement failed due to movement or poor measurement conditions	Apply finger cuff to a different finger Resize finger cuff and replace finger cuff with different size Restart measurement [†]
Fault: Finger Cuff #1 – Sensor Light Out of Range Fault: Finger Cuff #2 – Sensor Light Out of Range	Light signal too high	Warm the hand Apply finger cuff to a different finger Resize finger cuff and replace finger cuff with different size Restart measurement [†]
Fault: Finger Cuff #1 – No Signal Detected – Low Perfusion Fault: Finger Cuff #2 – No Signal Detected – Low Perfusion	No measurable Plethysmogram detected on startup Possibly contracted arteries	Warm the hand Apply finger cuff to a different finger Restart measurement

Table 15-22 HemoSphere ClearSight module faults/alerts (continued)

Message	Possible causes	Suggested actions
Fault: Finger Cuff #1 – No Pressure Waveforms Detected Fault: Finger Cuff #2 – No Pressure Waveforms Detected	The system failed to detect pressure waveforms Pressure pulsations in finger diminished due to pressure applied to the upper arm, elbow or wrist	Check if the blood flow in the arm of the patient is free of obstructions Check the blood pressure waveforms Reapply finger cuff(s) Restart measurement
Fault: Insufficient Pressure Build Up in Cuff #1 Fault: Insufficient Pressure Build Up in Cuff #2	Finger cuff air tube kinked Finger cuff leaking Cable between HemoSphere ClearSight module and pressure controller kinked or leaking Defective pressure controller Defective HemoSphere ClearSight module	Check finger cuff Check cable between HemoSphere ClearSight module and pressure controller Replace finger cuff Replace pressure controller Replace HemoSphere ClearSight module Restart measurement
Fault: Finger Cuff Disconnected	Previously connected finger cuff(s) not detected	Disconnect and reconnect Edwards finger cuff(s) Replace finger cuff(s) Restart measurement
Fault: Accumulated Single Cuff Monitoring Has Reached the Duration Limit	Cumulative measurement time on the same finger exceeded maximum duration of 8 hours	Remove Cuff from finger Place the Cuff on another finger and press 'Continue' on the Popup Restart Measurement
Fault: Finger Cuff #1 Has Expired. Replace Cuff	Finger cuff #1 has exceeded maximum use time	Replace finger cuff #1 Restart measurement
Fault: Finger Cuff #2 Has Expired. Replace Cuff	Finger cuff #2 has exceeded maximum use time	Replace finger cuff #2 Restart measurement
Fault: Invalid Finger Cuff #1 Connected	Non Edwards finger cuff #1 detected Defective finger cuff #1 connected	Verify that an Edwards finger cuff has been used Disconnect and reconnect Edwards finger cuff #1 Replace finger cuff #1 with a genuine Edwards cuff Restart measurement If problem persists, contact Edwards Technical Support
Fault: Invalid Finger Cuff #2 Connected	Non Edwards finger cuff #2 detected Defective finger cuff #2 connected	Verify that an Edwards finger cuff has been used Disconnect and reconnect Edwards finger cuff #2 Replace finger cuff #2 with a genuine Edwards cuff Restart measurement If problem persists, contact Edwards Technical Support
Fault: Finger Cuff #1 or Finger Cuff Connector Error	Finger cuff #1 is defective Cuff connector on pressure controller is damaged or defective	Disconnect and reconnect Edwards finger cuff #1 Replace finger cuff #1 Replace pressure controller Restart measurement If problem persists, contact Edwards Technical Support
Fault: Finger Cuff #2 or Finger Cuff Connector Error	Finger cuff #2 is defective Cuff connector on pressure controller is damaged or defective	Disconnect and reconnect Edwards finger cuff #2 Replace finger cuff #2 Replace pressure controller Restart measurement If problem persists, contact Edwards Technical Support
Fault: HRS Value Out of Physiological Range	Heart end of HRS is loose and may no longer be at heart level HRS detached from finger cuff HRS incorrectly calibrated HRS is defective	Verify HRS placement. Finger end should be attached to finger cuff and heart end should be placed at phlebostatic axis Vertically align the two ends of HRS and calibrate Replace HRS Restart measurement If problem persists, contact Edwards Technical Support

Table 15-22 HemoSphere ClearSight module faults/alerts (continued)

Message	Possible causes	Suggested actions
Fault: HRS Disconnected	Heart reference sensor (HRS) disconnected during monitoring HRS connection not detected	Verify HRS connection Disconnect and reconnect Edwards HRS Replace HRS If problem persists, contact Edwards Technical Support
Fault: HRS Detected	Measurement without HRS chosen but HRS is connected	Disconnect HRS Or select to measure with HRS
Fault: Invalid HRS Connected	Non Edwards HRS detected HRS is defective	Verify that an Edwards HRS has been used Disconnect and reconnect Edwards HRS Replace HRS with a genuine Edwards HRS Restart measurement If problem persists, contact Edwards Technical Support
Fault: HRS or HRS Connector Error	HRS is defective HRS connector on pressure controller is damaged	Disconnect and reconnect Edwards HRS Replace HRS Replace pressure controller Restart measurement If problem persists, contact Edwards Technical Support
Fault: HRS Has Expired. Replace HRS	HRS has expired as it is past useful life	Disconnect and reconnect Edwards HRS Replace HRS Restart Measurement If problem persists, contact Edwards Technical Support
Fault: Pressure Controller Disconnected	Pressure controller connection not detected	Disconnect and reconnect Edwards pressure controller Replace pressure controller If problem persists, contact Edwards Technical Support
Fault: Invalid Pressure Controller Connected	Incompatible pressure controller detected Non Edwards pressure controller detected Defective pressure controller connected	Verify that an Edwards pressure controller has been used Disconnect and re-connect Edwards pressure controller Replace pressure controller with a genuine Edwards pressure controller If problem persists, contact Edwards Technical Support
Fault: Pressure Controller Communication Error	Unresponsive pressure controller Poor connection between pressure controller and HemoSphere ClearSight module Pressure controller authentication failure Defective pressure controller Defective HemoSphere ClearSight module	Disconnect and reconnect Edwards pressure controller Power cycle the system Replace pressure controller Replace HemoSphere ClearSight module If problem persists, contact Edwards Technical Support
Fault: Pressure Controller Error	Defective pressure controller Poor connection between Edwards pressure controller and HemoSphere ClearSight module	Disconnect and reconnect Edwards pressure controller Replace pressure controller If problem persists, contact Edwards Technical Support
Fault: Pressure Controller Power Failure	Defective HemoSphere ClearSight module Defective Edwards pressure controller	Disconnect and reconnect Edwards pressure controller Replace pressure controller Replace HemoSphere ClearSight module If Problem Persists, contact Edwards Technical Support
Fault: Incompatible Pressure Controller Software	Unsuccessful software upgrade or incompatible software version detected	Replace pressure controller with a genuine Edwards pressure controller If problem persists, contact Edwards Technical Support
Fault: Continuous Monitoring Has Reached the 72 Hour Limit	Continuous measurement on the same hand exceeded maximum duration of 72 hours	Place the cuffs on fingers of opposite hand and resume monitoring

Table 15-22 HemoSphere ClearSight module faults/alerts (continued)

Message	Possible causes	Suggested actions
Fault: Air Supply Error	Kinked or damaged pressure controller cable Damaged finger cuff System malfunction Defective HemoSphere ClearSight module Defective pressure controller	Verify that connection between pressure controller and HemoSphere ClearSight module is not kinked or damaged Power cycle the system Replace pressure controller Replace HemoSphere ClearSight module Replace finger cuff If problem persists, contact Edwards Technical Support
Fault: CO – Check Arterial Waveform	Arterial waveform is inadequate to measure CO accurately Poor pressure waveform over extended period of time Systolic pressure too high or diastolic pressure too low	Assess noninvasive system starting from patient leading to finger cuff and HemoSphere ClearSight module Check the arterial waveform for severe hypotension, severe hypertension, and motion artifact Make sure the heart end of Edwards HRS is aligned with the patient's phlebostatic axis Confirm electrical connections of cables Apply finger cuff to a different finger Resize finger cuff and replace finger cuff with different size [†]
Fault: CO – Arterial Waveform Compromised	The system failed to detect pressure waveforms Pressure pulsations in finger diminished due to pressure applied to the upper arm, elbow, or wrist	Check if the blood flow in the arm of the patient is free of obstructions Make sure the heart end of Edwards HRS is aligned with the patient's phlebostatic axis Check the blood pressure waveforms Reapply finger cuff(s) Restart measurement If problem persists, contact Edwards Technical Support
Fault: Cuff Disconnected During Double Cuff Monitoring	Previously connected finger cuff(s) not detected	Disconnect and reconnect Edwards finger cuff(s) Replace finger cuff(s) Restart measurement
Fault: Second Cuff Connected During Single Cuff Monitoring	A second finger cuff connection is detected	Disconnect one of the finger cuffs and restart measurement Restart measurement in double cuff monitoring mode
Alert: CO – Pulse Pressure Low	Poor pressure waveform over extended period of time Patient condition results in a low pulse pressure	Assess noninvasive system starting from patient leading to finger cuff and HemoSphere ClearSight module Check the arterial waveform for severe hypotension, severe hypertension, and motion artifact Make sure the heart end of Edwards HRS is aligned with the patient's phlebostatic axis Apply finger cuff to a different finger Resize finger cuff and replace finger cuff with different size If problem persists, contact Edwards Technical Support [†]
Alert: CO – Pressure Waveform Not Stable	Arterial waveform is inadequate to measure CO accurately Poor pressure waveform over extended period of time Systolic pressure too high or diastolic pressure too low	Assess noninvasive system starting from patient leading to finger cuff and HemoSphere ClearSight module Check the arterial waveform for severe hypotension, severe hypertension, and motion artifact Make sure the heart end of Edwards HRS is aligned with the patient's phlebostatic axis Confirm electrical connections of cables Apply finger cuff to a different finger Resize finger cuff and replace finger cuff with different size If problem persists, contact Edwards Technical Support [†]

Table 15-22 HemoSphere ClearSight module faults/alerts (continued)

Message	Possible causes	Suggested actions
Alert: Cuff Pressure Release Mode – Monitoring Suspended	Finger cuff pressure has been released	Monitoring will automatically resume when the Countdown Clock on the status bar reaches 00:00 To resume monitoring, touch the countdown clock and select “Postpone Release”
Alert: SVV – Check Blood Pressure Waveform	Arterial waveform is inadequate to measure SVV accurately Poor pressure waveform over extended period of time Frequent Physiocalcs within waveform Systolic pressure too high or diastolic pressure too low	Assess noninvasive system starting from patient leading to finger cuff and HemoSphere ClearSight module Check the arterial waveform for severe hypotension, severe hypertension, and motion artifact Make sure the heart end of Edwards HRS is aligned with the patient’s phlebostatic axis Confirm electrical connections of cables Apply finger cuff to a different finger Resize finger cuff and replace finger cuff with different size†
Alert: Finger Cuff #1 – BP Measurement Error – Restarting Alert: Finger Cuff #2 – BP Measurement Error – Restarting	Blood pressure measurement failed due to movement or poor measurement conditions	Allow system to automatically resolve issue Apply finger cuff to a different finger Resize finger cuff and replace finger cuff with different size†
Alert: Finger Cuff #1 – No Pressure Waveforms Detected Alert: Finger Cuff #2 – No Pressure Waveforms Detected	The system failed to detect pressure waveforms Pressure pulsations in finger diminished due to pressure applied to the upper arm, elbow or wrist	Allow system to automatically resolve issue Check if the blood flow in the arm of the patient is free of obstructions Check the blood pressure waveforms Reapply finger cuff(s)
Alert: HRS Value Out of Physiological Range	Heart end of HRS is loose and may no longer be at heart level HRS detached from finger cuff HRS incorrectly calibrated HRS is defective	Verify HRS placement. Finger end should be attached to finger cuff and heart end should be placed at phlebostatic axis Vertically align the two ends of HRS and calibrate Replace HRS Restart measurement If problem persists, contact Edwards Technical Support
Alert: No HRS Connected – Verify Patient Positioning Alert: Current Offset: Finger {0} {1} Above Heart* Alert: Current Offset: Finger at Heart Level Alert: Current Offset: Finger {0} {1} Below Heart*	The patient positioning mode is “Patient Sedated and Stationary” and an HRS is not connected	Verify that the displayed offset is still accurate If the patient has been re-positioned, update the offset value on the “Zero & Waveform” screen
Alert: HemoSphere ClearSight Module Service Required	HemoSphere ClearSight module service time is overdue	Replace HemoSphere ClearSight module Contact Edwards Technical Support
Alert: Updated BP Calibration Might Be Required	Updated calibration may be required due to changes to hemodynamic state	Perform new calibration Keep Calibration Clear BP Calibration
Alert: Calibrate HRS	HRS not calibrated or previous calibration failed	Ensure HRS is connected and calibrate the HRS to start measurement
Alert: Check Pressure Controller Cable	Pressure controller cable is bent	Allow system to automatically resolve issue Ensure the pressure controller cable is not kinked in any areas If problem persists, contact Edwards Technical Support

*Note: {0} {1} is the specified distance where {0} is the value and {1} is the unit of measurement (CM or IN)

†Cuff sizing may not be applicable to all cuffs

Table 15-23 HemoSphere ClearSight warnings

Message	Possible causes	Suggested actions
HRS Out of Range	HRS pressure offset exceeded limit during the calibration process HRS is defective	Vertically align the two ends of HRS Calibrate HRS Replace HRS
HRS Calibration Unsuccessful – No Movement Detected	Prior to calibration, no HRS movement detected HRS is defective Defective pressure controller	Move heart end of HRS up and down. Next, keep both ends at same level, wait 1-2 seconds, and then calibrate while keeping both ends steady Replace HRS and calibrate HRS If problem persists, contact Edwards Technical Support
HRS Calibration Unsuccessful – Excessive Movement Detected	During calibration, HRS movement detected Defective pressure controller	Move heart end of HRS up and down. Next, keep both ends at same level, wait 1-2 seconds, and then calibrate while keeping both ends steady Replace HRS and calibrate HRS If problem persists, contact Edwards Technical Support
Unstable Arterial Pressure	System detecting large variability in the arterial pressure due to physiological or artificial noise	Ensure no external or artificial noise is interfering with arterial pressure measurements Stabilize arterial pressure
BP Calibration Unavailable	Insufficient monitoring data has been collected Blood pressure values from past 1 minute are too variable for a reliable calibration Non-physiological noise or artifacts are detected in the pressure signal	Allow for additional monitoring time and try again Stabilize arterial pressure Ensure no external or artificial noise is interfering with arterial pressure measurements
Finger Cuff #1 – No Signal Detected – Low Perfusion – Restarting Finger Cuff #2 – No Signal Detected – Low Perfusion – Restarting	No measurable Plethysmogram detected on startup Possibly contracted arteries	Allow system to automatically resolve issue Warm the hand Apply finger cuff to a different finger
Connect HemoSphere ClearSight module for CO or pressure monitoring	Connection to the HemoSphere ClearSight module has not been detected	Insert HemoSphere ClearSight module into the large technology module slot of the monitor Remove and re-insert module
Finger Cuff #1 – Sensor Light out of Range – Restarting Finger Cuff #2 – Sensor Light out of Range – Restarting	Light signal too high	Allow system to automatically resolve issue Warm the hand Apply finger cuff to a different finger Resize finger cuff and replace finger cuff with different size [†]
Insufficient Pressure Build Up in Cuff #1 – Restarting Insufficient Pressure Build Up in Cuff #2 – Restarting	Finger cuff air tube kinked Finger cuff leaking Cable between HemoSphere ClearSight module and pressure controller kinked or leaking Defective pressure controller Defective HemoSphere ClearSight module	Check finger cuff Check cable between HemoSphere ClearSight module and pressure controller Replace finger cuff Replace pressure controller Replace HemoSphere ClearSight module Restart measurement
Severe Vasoconstriction	Very small arterial volume pulsations detected, possibly contracted arteries	Allow system to automatically resolve issue Warm the hand Apply finger cuff to a different finger Resize finger cuff and replace finger cuff with different size [†]

Table 15-23 HemoSphere ClearSight warnings (continued)

Message	Possible causes	Suggested actions
Moderate Vasoconstriction	Very small arterial volume pulsations detected, possibly contracted arteries	Allow system to automatically resolve issue Warm the hand Apply finger cuff to a different finger Resize finger cuff and replace finger cuff with different size [†]
Finger Cuff #1 – Pressure Waveform Oscillations Detected Finger Cuff #2 – Pressure Waveform Oscillations Detected	Possibly contracted arteries Finger cuff too loose	Allow system to automatically resolve issue Warm the hand Apply finger cuff to a different finger Resize finger cuff and replace finger cuff with different size [†]
Connect Pressure Controller	Pressure Controller not connected. Defective Pressure Controller connected.	Connect pressure controller. Replace pressure controller. If problem persists, contact Edwards Technical Support.
Finger Cuff #1 Expiration in < 5 minutes	Finger cuff #1 approaching maximum use time	Replace finger cuff #1 to ensure uninterrupted measurement
Finger Cuff #2 Expiration in < 5 minutes	Finger cuff #2 approaching maximum use time	Replace finger cuff #2 to ensure uninterrupted measurement
Finger Cuff #1 Has Expired	Finger cuff #1 has exceeded maximum use time	Replace finger cuff #1 Restart measurement
Finger Cuff #2 Has Expired	Finger cuff #2 has exceeded maximum use time	Replace finger cuff #2 Restart measurement
Connect Finger Cuff	No finger cuff(s) detected. Defective finger cuff(s) connected.	Connect finger cuff(s) Replace finger cuff(s)
Finger Cuff #1 Approaching Maximum Use Time	Finger cuff #1 approaching maximum use time	Replace finger cuff #1 to ensure uninterrupted measurement
Finger Cuff #2 Approaching Maximum Use Time	Finger cuff #2 approaching maximum use time	Replace finger cuff #2 to ensure uninterrupted measurement
Connect HRS	HRS connection not detected	Connect HRS Replace HRS
HRS Expires in <2 weeks	HRS will expire in less than 2 weeks	Replace HRS to prevent delay in start of monitoring
HRS Expires in <4 weeks	HRS will expire in less than 4 weeks	Replace HRS to prevent delay in start of monitoring
HemoSphere ClearSight Module Service Required	HemoSphere ClearSight module service time is approaching	Replace HemoSphere ClearSight module Contact Edwards Technical Support
[†] Cuff sizing may not be applicable to all cuffs		

Table 15-24 HemoSphere ClearSight general troubleshooting

Message	Possible causes	Suggested actions
Pressure Difference: ClearSight BP vs. Other BP	HRS detached from finger cuff or phlebostatic axis HRS not properly calibrated Possibly contracted arteries (due to cold fingers) Finger cuff too loose Other BP measurement device not zeroed Other BP measurement sensor incorrectly applied	Verify HRS placement - The finger end should be attached to finger cuff and heart end should be placed at phlebostatic axis In case of invasive BP reference, HRS heart end and the transducer should be at the same level Calibrate HRS Warm the hand Reapply finger cuff (to a different finger) or replace Finger Cuff with proper size Re-zero other BP measurement device Remove and reapply other BP measurement sensor [†]
Connect Acumen IQ Cuff for HPI	Acumen IQ cuff is not detected and HPI or HPI key parameter is configured	Connect Acumen IQ cuff Replace Acumen IQ cuff

Table 15-24 HemoSphere ClearSight general troubleshooting

Message	Possible causes	Suggested actions
Connect Acumen IQ Cuff In CUFF 1 for HPI	CUFF 1 connection is not an Acumen IQ cuff and HPI or HPI key parameter is configured	Replace ClearSight cuff for Acumen IQ cuff in CUFF 1
Connect Acumen IQ Cuff in CUFF 2 for HPI	CUFF 2 connection is not an Acumen IQ cuff and HPI or HPI key parameter is configured	Replace ClearSight cuff for Acumen IQ cuff in CUFF 2
Connect HRS for HPI	HRS is not detected and HPI or HPI key parameter is configured	Connect HRS Replace HRS
[†] Cuff sizing may not be applicable to all cuffs		

15.10 Venous Oximetry Error Messages

15.10.1 Venous Oximetry Faults/Alerts

Table 15-25 Venous oximetry faults/alerts

Message	Possible causes	Suggested actions
Fault: Venous Oximetry – Light Range	Poor oximetry cable/catheter connection Debris or film obstructing oximetry cable/catheter connector lens Oximetry cable malfunction Catheter kinked or damaged	Verify secure oximetry cable / catheter connection Clean oximetry cable / catheter connectors with 70% isopropyl alcohol and swab, let air-dry and recalibrate Change oximetry cable and recalibrate Replace catheter if damage is suspected and recalibrate
Fault: Venous Oximetry – Red/IR Transmit	Debris or film obstructing oximetry cable / catheter connector lens Oximetry cable malfunction	Clean oximetry cable / catheter connectors with 70% isopropyl alcohol and swab, let air dry and recalibrate Power monitor off and on to restore platform Change oximetry cable and recalibrate
Fault: Venous Oximetry – Value Out of Range	Incorrectly entered ScvO ₂ /SvO ₂ , HGB or Hct values Incorrect HGB units of measure Calculated ScvO ₂ /SvO ₂ value is outside of the 0-99% range	Verify correctly entered ScvO ₂ /SvO ₂ , HGB, and Hct values Verify correct HGB units of measure Obtain updated ScvO ₂ /SvO ₂ lab values and recalibrate
Fault: Venous Oximetry – Input Signal Unstable	Poor oximetry cable/catheter connection Debris or film obstructing oximetry cable/catheter connector lens Oximetry cable malfunction Catheter kinked or damaged	Verify secure oximetry cable /catheter connection Clean oximetry cable /catheter connectors with 70% isopropyl alcohol and swab, let air-dry and recalibrate Change oximetry cable and recalibrate Replace catheter if damage is suspected and recalibrate
Fault: Venous Oximetry – Signal Processing Malfunction	Oximetry cable malfunction	Power monitor off and on to restore platform Change oximetry cable and recalibrate If problem persists, contact Edwards Technical Support
Fault: Oximetry Cable Memory	Oximetry cable memory malfunction	Disconnect and then reconnect the cable Change oximetry cable and recalibrate
Fault: Oximetry Cable Temperature	Oximetry cable malfunction	Power monitor off and on to restore platform Change oximetry cable and recalibrate If the cable is wrapped in fabric or sitting on an insulating surface such as a pillow, place it on a smooth surface that allows it to readily dissipate heat If the cable body feels warm, allow it to cool before operating again If problem persists, contact Edwards Technical Support

Table 15-25 Venous oximetry faults/alerts (continued)

Message	Possible causes	Suggested actions
Fault: Oximetry Cable Malfunction	Internal system malfunction	Power monitor off and on to restore platform If problem persists, contact Edwards Technical Support
Alert: Venous Oximetry – Poor Signal Quality	Low blood flow at catheter tip or catheter tip against vessel wall Significant change in HGB/Hct values Catheter tip clotted Catheter kinked or damaged Catheter is not connected to oximetry cable	If the cable is wrapped in fabric or sitting on an insulating surface such as a pillow, place it on a smooth surface that allows it to readily dissipate heat If the cable body feels warm, allow it to cool before operating again Verify proper catheter position (for SvO ₂ , verify proper catheter position in the pulmonary artery): - Confirm wedge pressure balloon inflation volume of 1.25-1.50 ml (for SvO₂ only) - Confirm appropriate catheter placement for patient's height, weight, and insertion site - Consider chest x-ray for evaluation of proper placement Aspirate then flush distal lumen per hospital protocol Update HGB/Hct values using Update function Check catheter for kinking and recalibrate Replace catheter if damage is suspected and recalibrate Ensure catheter is connected to oximetry cable

15.10.2 Venous Oximetry Warnings

Table 15-26 Venous oximetry warnings

Message	Possible causes	Suggested actions
In vitro Calibration Error	Poor oximetry cable and catheter ScvO ₂ /SvO ₂ connection Calibration cup wet Catheter kinked or damaged Oximetry cable malfunction Catheter tip is not in catheter calibration cup	Verify secure oximetry cable /catheter connection Straighten any visible kinks; replace catheter if damage is suspected Change oximetry cable and recalibrate Verify catheter tip is securely seated in calibration cup Perform In vivo calibration
Warning: Unstable Signal	Changing ScvO ₂ /SvO ₂ , HGB/Hct, or unusual hemodynamic values	Stabilize patient per hospital protocol and perform In vivo calibration
Warning: Wall Artifact or Wedge Detected	Low blood flow at catheter tip Catheter tip clotted Catheter tip wedged in vessel or against vessel wall	Aspirate then flush distal lumen per hospital protocol. Verify proper catheter position (for SvO ₂ , verify proper catheter position in the pulmonary artery): - confirm wedge pressure balloon inflation volume of 1.25-1.50 ml (For SvO₂ only) - confirm appropriate catheter placement for patient's height, weight, and insertion site - consider chest x-ray for evaluation of proper placement Perform In vivo calibration.

15.10.3 Venous Oximetry General Troubleshooting

Table 15-27 Venous oximetry general troubleshooting

Message	Possible causes	Suggested actions
Oximetry Cable Not Calibrated — Select Venous Oximetry to Calibrate	Oximetry cable has not been calibrated (in vivo or in vitro) Recall venous oximetry data function has not been performed Oximetry cable malfunction	Run in-vitro calibration Run in-vivo calibration Recall calibration values
Patient Data in oximetry cable more than 24 hours old — Recalibrate	Last oximetry cable calibration >24 hours old Date and time on Edwards' monitors at facility differ	Perform in vivo calibration Synchronize date and time on all Edwards' monitors at facility
Connect oximetry cable for venous oximetry monitoring	Oximetry cable connection at HemoSphere monitoring platform not detected Bent or missing oximetry cable connector pins	Verify secure oximetry cable connection Check oximetry cable connector for bent/missing pins

15.11 Tissue Oximetry Error Messages

15.11.1 Tissue Oximetry Faults/Alerts

Table 15-28 Tissue oximetry faults/alerts

Message	Possible causes	Suggested actions
Fault: Second Technology Module Detected	Multiple technology module connections detected	Remove one of the technology modules from the monitor slots
Fault: StO ₂ – Technology Module Disconnected	HemoSphere technology module removed during monitoring HemoSphere technology module not detected Connection points on slot or module are damaged	Confirm that module is properly inserted Remove and re-insert the module Check module for bent or broken pins Try switching to other module slot If problem persists, contact Edwards Technical Support
Fault: StO ₂ – ForeSight Oximeter Cable A Disconnected	FSOC A has become disconnected	Connect FSOC to port A of the inserted HemoSphere technology module
Fault: StO ₂ – ForeSight Oximeter Cable B Disconnected	FSOC B has become disconnected	Connect FSOC to port B of the inserted HemoSphere technology module
Fault: StO ₂ {0} – Sensor Disconnected*	Edwards Sensor on the indicated channel has become disconnected	Connect Edwards Sensor
Fault: StO ₂ – Technology Module	Internal system malfunction	Remove and re-insert module to reset If problem persists, contact Edwards Technical Support
Fault: StO ₂ – ForeSight Oximeter Cable A	FSOC A is defective	If condition persists, contact Edwards to replace the FSOC
Fault: StO ₂ – ForeSight Oximeter Cable B	FSOC B is defective	If condition persists, contact Edwards to replace the FSOC
Fault: StO ₂ – ForeSight Oximeter Cable A Communication Error	The technology module has lost communication with the indicated FSOC	Reconnect the module Check for bent or broken pins Try switching FSOC to other port of technology module If problem persists, contact Edwards Technical Support

Table 15-28 Tissue oximetry faults/alerts (continued)

Message	Possible causes	Suggested actions
Fault: StO ₂ – ForeSight Oximeter Cable B Communication Error	The technology module has lost communication with the indicated FSOC	Reconnect the module Check for bent or broken pins Try switching FSOC to other port of technology module If problem persists, contact Edwards Technical Support
Fault: StO ₂ – ForeSight Oximeter Cable A Incompatible Software Version	Unsuccessful software upgrade or incompatible software version detected	Contact Edwards Technical Support
Fault: StO ₂ – ForeSight Oximeter Cable B Incompatible Software Version	Unsuccessful software upgrade or incompatible software version detected	Contact Edwards Technical Support
Fault: StO ₂ {0} – Faulty Sensor*	Sensor is defective or Non-Edwards sensor in use	Replace with Edwards sensor
Fault: StO ₂ {0} – Ambient Light Too High*	Sensor is not in correct contact with patient	Check that sensor is in direct contact with skin Apply a light blocker or drape over the sensor to limit exposure to light
Fault: StO ₂ {0} – Sensor Temperature High*	Temperature under sensor is > 45 °C (Adult Mode) or > 43 °C (Pediatric/ Neonatal Mode)	Cooling of patient or environment may be required
Fault: StO ₂ {0} – Signal Level Too Low*	Insufficient light detected from patient Tissue under the sensors may have conditions such as excessive skin pigmentation, elevated hematocrit, birth marks, hematoma, or scar tissue A large (adult) sensor is being used on a pediatric patient (<18 years of age)	Verify that sensor is well adhered to patient's skin Move sensor to a location where SQI is 3 or 4 In the case of edema, remove the sensor until tissue condition returns to normal Replace large sensor with medium or small sensor in pediatric patients (<18 years of age)
Fault: StO ₂ {0} – Signal Level Too High*	Very unusual condition that is likely caused by optical shunting, where most of the light emitted is directed to the detectors Certain non-physiological materials, anatomical characteristics or scalp edema may trigger this message	Check that sensor is in direct contact with skin and that the clear liner has been removed
Fault: StO ₂ {0} – Check Tissue Under Sensor*	Tissue under sensor may have fluid accumulation/edema	Check patient for edema under sensor When tissue condition returns to normal range (e.g., patient is no longer edematous) the Sensor may be reapplied
Fault: StO ₂ {0} – Stool Interference High*	The sensor is interrogating primarily stool versus perfused tissue and StO ₂ cannot be measured	Move the sensor to a location where the relative amount of intestinal tissue is less, such as the flank
Fault: StO ₂ {0} – Sensor Off*	Computed StO ₂ not in valid range or sensor placed on an inappropriate object Low sensor temperature Poorly adhered or detached sensor Ambient light	Sensor may need to be repositioned
Fault: StO ₂ {0} – Not Physiological*	The measured value is out of physiological range Sensor malfunction	Verify correct placement of sensor Check sensor connection
Fault: StO ₂ {0} – Incorrect Sensor Size*	The sensor size is incompatible with either the Patient Mode or body location	Use a different sensor size (Refer to Sensor Instructions for Use for sensor size table) Change the Patient Mode or body location on the tile configuration menu accordingly

Table 15-28 Tissue oximetry faults/alerts (continued)

Message	Possible causes	Suggested actions
Fault: StO ₂ {0} – Algorithm Fault*	A processing error has occurred in the calculation of StO ₂ for the indicated channel	Disconnect and reconnect the indicated sensor channel Replace the FSOC Replace the technology module If problem persists, contact Edwards Technical Support
Fault: ΔctHb {0} – Out Of Range*	ΔctHb went outside display range	Reset ctHb to re-baseline all applicable channels
Alert: StO ₂ {0} – Unstable Signal*	Interference from outside source	Move sensor away from interfering source
Alert: StO ₂ {0} – Reduce Ambient Light*	Ambient light approaching maximum value	Check that sensor is in direct contact with skin Apply a light blocker or drape over the sensor to limit exposure to light
Alert: StO ₂ {0} – Stool Interference*	Stool Interference is approaching maximum acceptable level The sensor is interrogating some perfused tissue to make a StO ₂ measurement, but there is also a high concentration of stool in the sensor's interrogation path	Consider moving the sensor to a different abdominal location with less stool interference
Alert: StO ₂ {0} – Sensor Temperature Low*	Temperature under sensor < -10 °C	Warming of patient or environment may be required
Alert: StO ₂ {0} – Configure location for tissue oximetry sensor*	An anatomical location on the patient has not been configured for the connected sensor	Use the tissue oximetry configuration menu to select a body location for the indicated sensor channel
Alert: ΔctHb {0} – Reset Failed*	One of the connected channels produced a fault or alert during Reset	Check the information bar or event review screen for any faults or alerts associated with the tissue oximetry sensors Follow suggested actions for given faults or alerts
*Note: {0} is the sensor channel. The channel options are A1 and A2 for ForeSight cable A and B1 and B2 for ForeSight cable B. FSOC indicates ForeSight oximeter cable. The following components may have alternative labeling conventions: ForeSight oximeter cable (FSOC) may also be labeled as FORE-SIGHT ELITE tissue oximeter module (FSM) . HemoSphere technology module may also be labeled as HemoSphere tissue oximetry module. ForeSight sensors or ForeSight Jr sensors may also be labeled as FORE-SIGHT ELITE tissue oximetry sensors.		

15.11.2 Tissue Oximetry General Troubleshooting

Table 15-29 Tissue oximetry general troubleshooting

Message	Possible causes	Suggested actions
Connect Technology Module for StO ₂ Monitoring	Connection between the HemoSphere advanced monitor and technology module has not been detected	Insert the HemoSphere technology module into slot 1 or slot 2 of the monitor Remove and re-insert module
Connect ForeSight Oximeter Cable A for StO ₂ Monitoring	Connection between the HemoSphere technology module and FSOC at the indicated port has not been detected	Connect a FSOC to the indicated port of the HemoSphere technology module Reconnect the FSOC
Connect ForeSight Oximeter Cable B for StO ₂ Monitoring	Connection between the HemoSphere technology module and FSOC at the indicated port has not been detected	Connect a FSOC to the indicated port of the HemoSphere technology module Reconnect the FSOC
Connect Tissue Oximetry Sensor for StO ₂ Monitoring – {0}*	Connection between the FSOC and tissue oximetry sensor has not been detected on the channel for which StO ₂ has been configured	Connect a tissue oximetry sensor to the indicated channel Reconnect the tissue oximetry sensor on the indicated channel
StO ₂ {0} - Sensor Temperature Below Expected Range	Temperature under sensor < 28°C	Verify correct placement of sensor If patient is cooled intentionally, no action is required

Table 15-29 Tissue oximetry general troubleshooting (continued)

Message	Possible causes	Suggested actions
<i>*Note: {0} is the sensor channel. The channel options are A1 and A2 for ForeSight cable A and B1 and B2 for ForeSight cable B. FSOC indicates ForeSight oximeter cable.</i> <i>The following components may have alternative labeling conventions:</i> <i>ForeSight oximeter cable (FSOC) may also be labeled as FORE-SIGHT ELITE tissue oximeter module (FSM).</i> <i>HemoSphere technology module may also be labeled as HemoSphere tissue oximetry module.</i> <i>ForeSight sensors or ForeSight Jr sensors may also be labeled as FORE-SIGHT ELITE tissue oximetry sensors.</i>		

Appendix A

Specifications and Device Characteristics

Contents

Essential Performance Characteristics	368
HemoSphere Advanced Monitor Characteristics and Specifications	370
HemoSphere Battery Pack Characteristics and Specifications	372
HemoSphere Swan-Ganz Module Characteristics and Specifications	373
HemoSphere Pressure Cable Characteristics and Specifications	375
HemoSphere Oximetry Cable Characteristics and Specifications	376
HemoSphere Tissue Oximetry Characteristics and Specifications	377
HemoSphere ClearSight Module Characteristics and Specifications	378
Acumen AFM Cable Characteristics and Specifications	380

A.1 Essential Performance Characteristics

Under normal and single fault conditions either the essential performance listed in table A-1 below is provided or failure to provide this performance is readily identifiable by the user (e.g., no display of parameter values, technical alarm, distorted waveforms or delay in parameter value update, complete failure of the monitor, etc.).

Table A-1 represents the minimum performance when operating under non-transient electromagnetic phenomena, such as radiated and conducted RF, according to IEC 60601-1-2. Table A-1 also identifies the minimum performance for transient electromagnetic phenomena, such as electrical fast transients and surges, according to IEC 60601-1-2.

Table A-1 HemoSphere advanced monitor essential performance – transient and non-transient electromagnetic phenomena

Module or cable	Parameter	Essential Performance
General: all monitoring modes and parameters		No interruption of current monitoring mode. No unexpected reboots or halting of operation. No spontaneous triggering of events that require user interaction to initiate. Patient connections provide defibrillator protection. Following exposure to defibrillation voltages, the system shall return to an operational state within 10 seconds. After the transient electromagnetic phenomena, the system shall return to an operational state within 30 seconds. If Swan-Ganz continuous cardiac output (CO) was active during the event, the system will automatically re-initiate monitoring. The system shall exhibit no loss of any stored data following the transient electromagnetic phenomena. When used with HF Surgical Equipment, the monitor shall return to operational mode within 10 seconds without loss of stored data after exposure to the field produced by the HF Surgical Equipment.
HemoSphere Swan-Ganz module	Continuous Cardiac Output (CO), and associated parameters, both indexed and non-indexed (SV, SVR, RVEF, EDV)	Monitors the filament surface temperature and time at temperature. If a time and temperature threshold is exceeded (above 45 °C), monitoring halts and alarm triggered. Measurement of blood temperature within specified accuracy (± 0.3 °C). Alarm if blood temperature outside of monitoring range. Alarm if CO and related parameters outside of alarm ranges. Alarm delay based on a variable averaging time. Typical averaging time is 57 seconds.
	intermittent cardiac output (iCO) and associated parameters, both indexed and non-indexed (SV, SVR)	Measurement of blood temperature within specified accuracy (± 0.3 °C). Alarm if blood temperature outside monitoring range.
HemoSphere Swan-Ganz module and pressure cable	20-second flow parameters (CO _{20s} , CI _{20s} , SV _{20s} , SVI _{20s})	Alarm if 20-second parameters outside of alarm ranges. Alarm delay based on a 20 second averaging time.
HemoSphere pressure cable	arterial blood pressure (SYS, DIA, MAP), central venous blood pressure (CVP), pulmonary artery blood pressure (MPAP)	Measurement of blood pressure within specified accuracy ($\pm 4\%$ or ± 4 mmHg, whichever is greater). Alarm if blood pressure outside of alarm ranges. Alarm delay of 7 seconds based on averaging time of 2 seconds and 5 consecutive seconds outside of alarm ranges. The device supports detection of invasive pressure transducer and transducer cable fault. The device supports detection of disconnected catheter.

Table A-1 HemoSphere advanced monitor essential performance – transient and non-transient electromagnetic phenomena (continued)

Module or cable	Parameter	Essential Performance
HemoSphere ClearSight module	non-invasive blood pressure (SYS, DIA, MAP)	Measurement of blood pressure within specified accuracy ($\pm 1\%$ of full scale with a maximum of ± 3 mmHg). Alarm if blood pressure outside alarm ranges. Alarm delay of approximately 10 seconds based on averaging window of 5 heartbeats (at 60 bpm this would be 5 seconds but will vary based on heart rate) and 5 consecutive seconds outside of alarm ranges.
HemoSphere oximetry cable	oxygen saturation (mixed venous SvO ₂ or central venous ScvO ₂)	Measurement of oxygen saturation within specified accuracy ($\pm 2\%$ oxygen saturation). Alarm if oxygen saturation outside of alarm ranges. Alarm delay of 7 seconds based on averaging time of 2 seconds outside of alarm ranges.
HemoSphere tissue oximetry module with ForeSight oximeter cable	tissue oxygen saturation (StO ₂)	The ForeSight oximeter cable shall recognize attached sensor and issue an appropriate equipment status if inoperable or disconnected. When a sensor is properly positioned on the patient and connected to the ForeSight oximeter cable, the ForeSight oximeter cable shall measure StO₂ values within system specifications (refer to table A-18 on page 378) and correctly output values to HemoSphere technology module. In response to a defibrillation event, the ForeSight oximeter cable shall not be electrically damaged. In response to an external noise event, the values may continue to report as pre-event values or may be reported as indeterminate value (dashed). The ForeSight oximeter cable shall automatically recover and resume reporting appropriate values within 20 seconds after the noise event.
HemoSphere technology module with Acumen AFM cable	fluid delivery tracking (flow rate)	When used with a compatible fluid meter, measurement of flow rate within specified accuracy ($\pm 20\%$ or ± 1 mL/min, whichever is greater). During transient electromagnetic phenomena, flow rate values may continue to report as pre-event values. The Acumen AFM cable shall automatically recover and resume reporting appropriate values within 30 seconds after the noise event.

A.2 HemoSphere Advanced Monitor Characteristics and Specifications

Table A-2 HemoSphere advanced monitor physical and mechanical characteristics

HemoSphere advanced monitor		
Weight	10 $\pm$ 0.2 lb (4.5 $\pm$ 0.1 kg)	
Dimensions	Height	11.7 in (297 mm)
	Width	12.4 in (315 mm)
	Depth	5.56 in (141 mm)
Footprint	Width	10.6 in (269 mm)
	Depth	4.8 in (122 mm)
Ingress protection	IPX1	
Display	Active Area	12.1 in (307 mm)
	Resolution	1024 $\times$ 768 LCD

Table A-2 HemoSphere advanced monitor physical and mechanical characteristics (continued)

Operating system	Windows 10 IoT
Speaker count	1

Table A-3 HemoSphere advanced monitor environmental specifications

Environmental specification		Value
Temperature	Operational	10 to 32.5 °C
	Non-operational/storage*	-18 to 45 °C
Relative humidity	Operational	20 to 90% non-condensing
	Non-operational/storage	90% non-condensing at 45 °C
Altitude	Operational	0 to 10,000 ft (3048 m)
	Non-operational/storage	0 to 20,000 ft (6096 m)

*Note: Battery capacity starts to degrade with extended exposure above 35 °C.

Table A-4 HemoSphere advanced monitor transportation environmental specifications

Environmental specification	Value
Temperature*	-18 to 45 °C
Relative humidity*	20 to 90% RH non-condensing
Altitude	maximum of 20,000 ft (6096 m) for up to 8 hours
Standard	ASTM D4169, DC13

*Note: Pre-conditioning temperature and humidity

NOTE Unless otherwise stated, all compatible HemoSphere advanced monitor accessories, components, and cables have the environmental specifications listed in table A-3 and table A-4.

MRI Information. Do not use the HemoSphere advanced monitor or platform modules and cables in an MR environment. The HemoSphere advanced monitoring platform, including all modules and cables, is MR unsafe since the device contains metallic components, which can experience RF-induced heating in the MRI environment.

**Table A-5 HemoSphere advanced monitor technical characteristics**

Input/Output	
Touch screen	Projective capacitive touch
RS-232 serial port (1)	Edwards proprietary protocol; Maximum data rate = 57.6 kilo baud
USB ports (2)	one USB 2.0 (rear) and one USB 3.0 (side)
RJ-45 Ethernet port	One
HDMI port	One
Analog inputs (2)	Input voltage range: 0 to 10V; Selectable full-scale: 0 to 1V, 0 to 5V, 0 to 10V; >100 kΩ input impedance; 1/8 in. stereo jack; Bandwidth: 0 to 5.2 Hz; Resolution: 12 bits ±1 LSB of full scale

Table A-5 HemoSphere advanced monitor technical characteristics (continued)

Input/Output (continued)	
Pressure output (1)	DPT pressure out signal is compatible with monitors and accessories intended to interface with Edwards minimally-invasive pressure transducers Post-zero minimum patient monitor display range:-20 mmHg to 270 mmHg
ECG monitor input	ECG sync line conversion from ECG signal: 1V/mV; Input voltage range ± 10 V full scale; Resolution = ± 1 BPM; Accuracy = $\pm 10\%$ or 5 BPM of the input, whichever is greater; Range = 30 to 200 BPM; 1/4 in. stereo jack, tip at positive polarity; analog cable Pacemaker pulse rejection capabilities. Instrument rejects all pacemaker pulses having amplitudes from ± 2 mV to ± 5 mV (assumes 1V/mV ECG sync line conversion) and pulse widths from 0.1 ms to 5.0 ms, both with normal and ineffective pacing. Pacemaker pulses with overshoot of $\leq 7\%$ of pulse amplitude (Method A of EN 60601-2-27:2014, subclause 201.12.1.101.13) and overshoot time constants from 4 ms to 100 ms are rejected. Maximum T-wave rejection capability. Maximum T-wave amplitude that can be rejected by instrument: 1.0 mV (assumes 1V/mV ECG sync line conversion). Irregular Rhythm. Figure 201.101 of EN 60601-2-27:2014. * Complex A1: Ventricular bigeminy, system displays 80 BPM * Complex A2: Slow alternating ventricular bigeminy, system displays 60 BPM * Complex A3: Rapid alternating ventricular bigeminy, system displays 60 BPM * Complex A4: Bidirectional systoles, system displays 104 BPM
HR _{avg} display	CO Monitoring Off. Averaging time: 57 seconds; Update rate: Per beat; Response time: 40 seconds for step increase from 80 to 120 BPM, 29 seconds for step decrease from 80 to 40 BPM. CO Monitoring On. Averaging time: Time between CO measurements (3 to 21 minutes); Update rate: Approximately 1 minute; Response time: 175 seconds for step increase from 80 to 120 BPM, 176 seconds for step decrease from 80 to 40 BPM.
Electrical	
Rated supply voltage	100 to 240 Vac; 50/60 Hz
Rated input	1.5 to 2.0 Amps
Fuses	T 2.5AH, 250V; High breaking capacity; Ceramic
Alarm	
Sound pressure level	45 to 85 dB(A)
Wireless	
Type	connection to Wi-Fi networks that are compliant to 802.11b/g/n, minimum

A.3 HemoSphere Battery Pack Characteristics and Specifications

Table A-6 HemoSphere battery pack physical characteristics

HemoSphere battery pack	
Weight	1.1 lb (0.5 kg)

Table A-6 HemoSphere battery pack physical characteristics

Dimensions	Height	1.38 in (35 mm)
	Width	3.15 in (80 mm)
	Depth	5.0 in (126 mm)

Table A-7 HemoSphere battery pack environmental specifications

Environmental specification		Value
Temperature	Operational	10 to 37 °C
	Recommended storage	21 °C
	Maximum long term storage	35 °C
	Minimum long term storage	0 °C
Relative humidity	Operational	5 to 95% non-condensing at 40 °C

Table A-8 HemoSphere battery pack technical characteristics

Specification	Value
Output voltage (nominal)	12.8 V
Maximum discharge current	5 A
Cells	4 x LiFePO ₄ (lithium iron phosphate)

A.4 HemoSphere Swan-Ganz Module Characteristics and Specifications

Table A-9 HemoSphere Swan-Ganz module physical characteristics

HemoSphere Swan-Ganz module		
Weight	approximately 1.0 lb (0.45 kg)	
Dimensions	Height	1.36 in (3.45 cm)
	Width	3.53 in (8.96 cm)
	Depth	5.36 in (13.6 cm)
Ingress protection	IPX1	
Applied part classification	Type CF defibrillation proof	

NOTE For HemoSphere Swan-Ganz module environmental specifications, see table A-3, *HemoSphere advanced monitor environmental specifications*, on page 371.

Table A-10 HemoSphere Swan-Ganz module parameter measurement specifications

Parameter	Specification	
Continuous Cardiac Output (CO)	Range	1 to 20 L/min
	Reproducibility ¹	±6% or 0.1 L/min, whichever is greater
	Average response time ²	<10 mins (for CCO catheters) <14 mins (for CCO volumetric catheters)
	Maximum thermal filament surface temperature	48 °C
Intermittent (Bolus) Cardiac Output (iCO)	Range	1 to 20 L/min
	Reproducibility ¹	±3% or 0.1 L/min, whichever is greater

Table A-10 HemoSphere Swan-Ganz module parameter measurement specifications (continued)

Parameter	Specification	
Blood Temperature (BT)	Range	15 to 45 °C (59 to 113 °F)
	Accuracy	±0.3 °C
Injectate Temperature (IT)	Range	0 to 30 °C (32 to 86 °F)
	Accuracy	±1 °C
Average Heart Rate for EDV/ RVEF Determination (HR _{avg})	Acceptable input range	30 to 200 bpm
Continuous Right Ventricular Ejection Fraction (RVEF)	Range	10 to 60%
	Reproducibility ¹	±6% or 3 efu, whichever is greater
¹ Coefficient of variation — measured using electronically generated data		
² 90% change under conditions of stable blood temperature		

NOTE The expected useful life of the HemoSphere Swan-Ganz module is 5 years from date of purchase. If your equipment experiences a malfunction, please contact Technical Support or your local Edwards representative for further assistance.

Table A-11 HemoSphere Swan-Ganz module 20-second flow parameter measurement specifications¹

Parameter	Specification	
CO _{20s}	Range	1 to 20 L/min
	Update rate	20 ±1 seconds
CI _{20s}	Range	0 to 20 L/min/m ²
	Update rate	20 ±1 seconds
SV _{20s}	Range	0 to 300 mL/b
	Update rate	20 ±1 seconds
SVI _{20s}	Range	0 to 200 mL/b/m ²
	Update rate	20 ±1 seconds
¹ 20-second flow parameters only available when monitoring pulmonary artery pressure with a connected HemoSphere pressure cable and TruWave DPT. For more information on these parameters, see "20-Second Flow Parameters" on page 174.		

A.5 HemoSphere Pressure Cable Characteristics and Specifications

Table A-12 HemoSphere pressure cable physical characteristics

HemoSphere pressure cable		
Weight	approximately 0.64 lb (0.29 kg)	
Dimensions	Length	10 ft (3.0 m)
Ingress protection	IPX4	
Applied part classification	Type CF defibrillation proof	

NOTE For HemoSphere pressure cable specifications, see table A-3, *HemoSphere advanced monitor environmental specifications*, on page 371.

Table A-13 HemoSphere pressure cable parameter measurement specifications

Parameter	Specification	
FloTrac cardiac output (CO)	Display range	1.0 to 20 L/min
	Reproducibility ¹	±6% or 0.1 L/min, whichever is greater
Blood pressure ²	Live pressure display range	-34 to 312 mmHg
	MAP/DIA/SYS display range	0 to 300 mmHg
	CVP display range	0 to 50 mmHg
	MPAP display range	0 to 99 mmHg
	Accuracy	±4% or ±4 mmHg, whichever is greater, from -30 to 300 mmHg
	Bandwidth	1-10Hz
Pulse rate (PR)	Accuracy ³	A _{rms} ≤3 bpm
¹ Coefficient of variation - measured using electronically generated data. ² Parameter specifications compliant with IEC 60601-2-34 standards. Testing performed under laboratory conditions. ³ Accuracy tested under laboratory conditions.		

NOTE The expected useful life of the HemoSphere pressure cable is 5 years from date of purchase. If your equipment experiences a malfunction, please contact Technical Support or your local Edwards representative for further assistance.

A.6 HemoSphere Oximetry Cable Characteristics and Specifications

Table A-14 HemoSphere oximetry cable physical characteristics

HemoSphere oximetry cable		
Weight	approximately 0.54 lb (0.24 kg)	
Dimensions	Length	9.6 ft (2.9 m)
Ingress protection	IPX4	
Applied part classification	Type CF defibrillation proof	

NOTE For HemoSphere oximetry cable environmental specifications, see table A-3, *HemoSphere advanced monitor environmental specifications*, on page 371.

Table A-15 HemoSphere oximetry cable parameter measurement specifications

Parameter	Specification	
ScvO ₂ /SvO ₂ Oximetry (Oxygen Saturation)	Range	0 to 99%
	Precision ¹	±2% at 30 to 99%
	Update rate	2 seconds
¹ Precision tested under laboratory conditions.		

NOTE The expected useful life of the HemoSphere oximetry cable is 3 years from date of purchase. If your equipment experiences a malfunction, please contact Technical Support or your local Edwards representative for further assistance.

A.7 HemoSphere Tissue Oximetry Characteristics and Specifications

Table A-16 HemoSphere technology module physical characteristics

HemoSphere technology module		
Weight	approximately 1.0 lb (0.4 kg)	
Dimensions	Height	1.4 in (3.5 cm)
	Width	3.5 in (9.0 cm)
	Depth	5.4 in (13.6 cm)
Ingress protection	IPX1	
Applied part classification	Type BF defibrillation proof	

NOTE For HemoSphere technology module and ForeSight oximeter cable environmental specifications, see table A-3, *HemoSphere advanced monitor environmental specifications*, on page 371.

Table A-17 ForeSight oximeter cable physical characteristics

ForeSight oximeter cable characteristics		
Weight	mounting clip	0.1 lb (0.05 kg)
	case, cables, and clip	2.3 lb (1.0 kg)
Dimensions	technology module cable length	15 ft (4.6 m) ¹
	sensor cable length (2)	4.9 ft (1.5 m) ¹
	cable housing (H × W × D)	6.0 in (15.24 cm) × 3.75 in (9.52 cm) × 2.75 in (6.00 cm)
	mounting clip (H × W × D)	2.4 in (6.2 cm) × 1.75 in (4.47 cm) × 3.2 in (8.14 cm)
Ingress protection	IPX4	
Applied part classification	Type BF defibrillation proof	
¹ The length of the technology module and sensor cables are nominal lengths.		

Table A-18 HemoSphere technology module with ForeSight oximeter cable parameter measurement characteristics

Parameter	Sensor	Specification
StO ₂ (all locations)	all sensor sizes	display range: 0 to 99% update rate: 2 seconds
Cerebral StO ₂	large sensors	$A_{rms}^* < 3.4\% \text{ StO}_2$
	small/medium sensors	$A_{rms}^* < 6.1\% \text{ StO}_2$
Somatic StO ₂	large sensors	$A_{rms}^* < 4.3\% \text{ StO}_2$
	small/medium sensors	$A_{rms}^* < 6.1\% \text{ StO}_2$
<i>*Note 1: A_{rms} from 50 to 85% StO₂. See "Interpreting StO₂ Values" on page 248 for more information</i> <i>Note 2: Measurements are statistically distributed, and therefore about two-thirds of tissue oximeter equipment measurements are expected to fall within $+A_{rms}$ of the reference measurement over the measurement range</i>		

NOTE The expected useful life of the HemoSphere technology module and ForeSight oximeter cable is 5 years from date of purchase. If your equipment experiences a malfunction, please contact Technical Support or your local Edwards representative for further assistance.

A.8 HemoSphere ClearSight Module Characteristics and Specifications

Table A-19 HemoSphere ClearSight module physical characteristics

HemoSphere ClearSight module		
Weight	approximately 2 lbs (0.9 kg)	
Dimensions	Height	5.1 in (13 cm)
	Width	5.6 in (14 cm)
	Depth	3.9 in (10 cm)
Ingress protection	IPX1	
Applied part classification	Type BF	

Table A-20 HemoSphere Smart pressure controller physical characteristics

HemoSphere Smart pressure controller		
Weight	housing	approximately 0.7 lbs (0.27 kg)
Dimensions	cable length	14.8 +/- 0.2 ft (4.5 +/- 0.06 m)
Ingress protection	IP44	
Applied part classification	Type BF defibrillation proof	

Table A-21 HemoSphere ClearSight module environmental specifications

Environmental specification		Value
Temperature	Operational	10 to 37 °C
	Non-operational/storage	-18 to 45 °C
Relative humidity	Operational	20 to 85% non-condensing
	Non-operational/storage	20 to 90% non-condensing at 45 °C
Altitude	Operational	0 to 9,483 feet (3000 m)
	Non-operational/storage	0 to 19,685 feet (6000 m)

Table A-22 HemoSphere ClearSight module parameter measurement specifications

Parameter	Specification	
Arterial blood pressure	Display range	0 to 300 mmHg
	Accuracy ¹	Bias systolic pressure (SYS) $\leq \pm 5.0$ mmHg Bias diastolic pressure (DIA) $\leq \pm 5.0$ mmHg Precision (1 σ) systolic pressure (SYS) $\leq \pm 8.0$ mmHg Precision (1 σ) diastolic pressure (DIA) $\leq \pm 8.0$ mmHg
Pulse Rate (PR)	Accuracy ³	$A_{rms} \leq 3$ bpm
Finger cuff pressure	Range	0 to 300 mmHg
	Accuracy	1% of full scale (max 3 mmHg), zeroing automatically
Cardiac output (CO)	Display range	1.0 to 20.0 L/min
	Accuracy	Bias $\leq \pm 0.6$ L/min or $\leq 10\%$ (whichever is greater) Precision (1 σ) $\leq \pm 20\%$ over the range of cardiac output from 2 to 20 L/min
	Reproducibility ²	$\pm 6\%$
	Update rate	20 seconds


¹Accuracy tested under laboratory conditions compared to a calibrated pressure gauge

²Coefficient of variation – measured using electronically generated data

³Accuracy tested under laboratory conditions compared to ECG heart rate

Table A-23 Edwards finger cuff characteristics

Finger cuff	
Maximum weight	11 g (0.02 lbs)
LED spectral irradiance	See figure A-1
Max optical output	0.013 mWatts
Max variation of output over treatment area	50%

**Figure A-1 Spectral Irradiance and location of light emission aperture****NOTE**

The expected useful life of the HemoSphere ClearSight module is 5 years from date of purchase. If your equipment experiences a malfunction, please contact Technical Support or your local Edwards representative for further assistance.

A.9 Acumen AFM Cable Characteristics and Specifications

Table A-24 Acumen AFM cable physical characteristics

Acumen AFM cable		
Weight	approximately 0.6 lbs (0.3 kg)	
Dimensions	Length	15 ft (4.6 m)
Ingress protection	IPX4	
Applied part classification	Type BF defibrillation proof	

Table A-25 Acumen AFM cable operational environmental specifications

Environmental specifications	Value
Temperature	10 to 37 °C
Relative humidity	20 to 90% non-condensing
Altitude	0 to 10,000 feet (3048 m)

Table A-26 Acumen AFM cable transportation environmental specifications

Environmental specifications	Value
Temperature*	18 to 45 °C
Relative humidity*	20 to 90% non-condensing at 45 °C
Altitude	0 to 20,000 feet (6096 m)
*Note: Pre-conditioning temperature and humidity	

NOTE The expected useful life of the Acumen AFM cable is 5 years from date of purchase. If your equipment experiences a malfunction, please contact Technical Support or your local Edwards representative for further assistance.

Table A-27 HemoSphere technology module with Acumen AFM cable parameter measurement specifications

Parameter	Specification	
Flow rate	range	0 to 8000 mL/hr
	accuracy	±20% or ±60 mL/hr, whichever is greater
Bolus volume	range	100 to 500 mL
	accuracy	±9%*
*Accuracy tested under laboratory conditions		

Appendix B

Accessories

Contents

Accessories List.	381
Additional Accessories Description.	382

B.1 Accessories List

WARNING Only use approved HemoSphere advanced monitor accessories, cables and or components that have been supplied and labeled by Edwards. Using unapproved accessories, cables and or components may affect patient safety and measurement accuracy.

Table B-1 HemoSphere advanced monitor components

Description	Model number
HemoSphere advanced monitor	
HemoSphere advanced monitor	HEM1
HemoSphere battery pack	HEMBAT10
HemoSphere expansion module	HEMEXPM10
HemoSphere L-tech expansion module	HEMLTECHM10
HemoSphere advanced monitor roll stand	HEMRLSTD1000
HemoSphere Swan-Ganz monitoring	
HemoSphere Swan-Ganz module	HEMSGM10
Patient CCO cable	70CC2
Edwards Swan-Ganz/Swan-Ganz Jr catheters	*
In-line temperature probe (CO-SET+ closed injectate delivery system)	93522
Bath temperature injectate probe	9850A
HemoSphere pressure cable monitoring	
HemoSphere pressure cable	HEMPSC100
Edwards FloTrac, FloTrac Jr, or Acumen IQ sensor	*
Edwards TruWave pressure monitoring transducer	*

Table B-1 HemoSphere advanced monitor components

Description	Model number
HemoSphere venous oximetry monitoring	
HemoSphere oximetry cable	HEMOXSC100
HemoSphere oximetry cradle	HEMOXCR1000
Edwards oximetry catheter	*
HemoSphere tissue oximetry monitoring	
HemoSphere technology module (May also be labeled as HemoSphere tissue oximetry module)	HEMTOM10
ForeSight oximeter cable (May also be labeled as FORE-SIGHT ELITE oximeter module)	HEMFSM10
ForeSight Jr sensors (size: non-adhesive small and small) (May also be labeled as FORE-SIGHT ELITE oximetry sensors)	*
ForeSight sensors (sizes: medium and large) (May also be labeled as FORE-SIGHT ELITE oximetry sensors)	*
HemoSphere ClearSight module monitoring	
HemoSphere ClearSight module	HEMCMSM10
Pressure controller kit	PC2K HEMPC2K

Table B-1 HemoSphere advanced monitor components

Description	Model number
Pressure controller	PC2 HEMPC
Smart pressure controller kit	PC1QAK
Smart pressure controller	PC1Q
Pressure controller band multi pack	PC2B
Pressure controller cuff connector caps multi pack	PC2CCC
Pressure controller cover	PCCVR
Smart pressure controller cable clip	PC1QACC
Smart pressure controller clip band	PC1QACB
Smart pressure controller plug	PC1QAP
Heart reference sensor	HRS
HemoSphere ClearSight module upgrade (HEMCSM10, PC2K/HEMPC2K, HRS, and ClearSight software)	HEMCSMUPG
ClearSight, ClearSight Jr, VitaWave, VitaWave Plus, Acumen IQ, and Acumen IQ Plus finger cuff	*
HemoSphere advanced monitor cables	
Mains power cord	*
Analog pressure cable	**
Analog ECG monitor cables	**
Acumen AFM cable	AAFMC
Acumen IQ fluid meter	AIQFM
Pressure-out cable	HEMDPT1000

Table B-1 HemoSphere advanced monitor components

Description	Model number
Additional HemoSphere Accessories	
HemoSphere advanced monitor operator's manual	***
HemoSphere advanced monitor service manual	***
HemoSphere advanced monitor quick start guide (contains HemoSphere advanced monitor operator's manual)	HEMQG1000
<i>* Please contact your Edwards representative for model and ordering information.</i> <i>**Edwards Lifesciences analog cables are bedside monitor specific; they are available for a family of bedside monitor companies such as Philips (Agilent), GE (Marquette) and Spacelabs (OSI Systems). Please contact your Edwards representative for specific model and ordering information.</i> <i>***Please contact your Edwards representative for the most current version.</i>	

B.2 Additional Accessories Description

B.2.1 Roll Stand

The HemoSphere advanced monitor roll stand is intended for use with the HemoSphere advanced monitor. Follow included instructions for roll stand assembly and warnings. Place the assembled roll stand on the floor, ensuring that all wheels are in contact with the floor, and securely mount the monitor to the roll stand plate as indicated in the directions.

B.2.2 Oximetry Cradle

The HemoSphere oximetry cradle is a reusable accessory intended to properly secure the HemoSphere oximetry cable while monitoring with the HemoSphere advanced monitoring platform. Follow included instructions for proper cradle mounting directions.

B.2.3 Pressure Controller Cover

The pressure controller cover secures the heart reference sensor into the pressure controller. The pressure controller cover is intended for limited reuse. The operator shall assess whether reuse is appropriate. When reused, follow the platform cleaning instruction listed in *Cleaning the Monitor and Modules* on page 400. Replace if damaged.

To apply the pressure controller cover:

- 1 Ensure the heart reference sensor (HRS) is attached prior to attaching the pressure controller cover to the pressure controller.
- 2 Place the pressure controller cover's back mounting notch around the pressure controller cable. See step 1 in figure B-1.
- 3 Snap the pressure controller cover over the pressure controller, making sure that the pressure controller cover does not interfere with the heart reference sensor (HRS) connection. See step 2 in figure B-1.




Figure B-1 Applying pressure controller cover

- 4 To remove the pressure controller cover, pull upwards from the front tab. This is indicated by the arrows symbol . Do not remove the pressure controller cover from the side by HRS connection indicated by the do not remove symbol .

CAUTION Do not pinch any heart reference sensor tubes or wires under the pressure controller cover during application. Be careful the only wire between the back mounting notch is the pressure controller cable.

Do not lift PCCVR from any other point than the front tab.

Appendix C

Equations for Calculated Patient Parameters

This section describes the equations used to calculate continuous and intermittent patient parameters displayed on the HemoSphere advanced monitor.

NOTE Patient parameters are calculated to more decimal places than are displayed on the screen. For example, a screen CO value of 2.4 may actually be a CO of 2.4492. Consequently, attempts to verify the accuracy of the monitor's display using the following equations may produce results that are slightly different from the data computed by the monitor.

For all calculations that include SvO₂, ScvO₂ will be substituted when the user selects the ScvO₂.

Subscript SI = Standard International Units

Table C-1 Cardiac and oxygenation profile equations

Parameter	Description and formula	Units
BSA	Body Surface Area (DuBois formula) $BSA = 71.84 \times (WT^{0.425}) \times (HT^{0.725}) / 10,000$ where: WT – Patient Weight, kg HT – Patient Height, cm	m ²
CaO ₂	Arterial Oxygen Content $CaO_2 = (0.0138 \times HGB \times SaO_2) + (0.0031 \times PaO_2) \text{ (mL/dL)}$ $CaO_2 = [0.0138 \times (HGB_{SI} \times 1.611) \times SaO_2] + [0.0031 \times (PaO_{2SI} \times 7.5)] \text{ (mL/dL)}$ where: HGB – Total Hemoglobin, g/dL HGB _{SI} – Total Hemoglobin, mmol/L SaO ₂ – Arterial O ₂ Saturation, % PaO ₂ – Partial Pressure of Arterial Oxygen, mmHg PaO _{2SI} – Partial Pressure of Arterial Oxygen, kPa	mL/dL

Table C-1 Cardiac and oxygenation profile equations (continued)

Parameter	Description and formula	Units
CvO ₂	Venous Oxygen Content $\text{CvO}_2 = (0.0138 \times \text{HGB} \times \text{SvO}_2) + (0.0031 \times \text{PvO}_2) \text{ (mL/dL)}$ $\text{CvO}_2 = [0.0138 \times (\text{HGB}_{\text{SI}} \times 1.611) \times \text{SvO}_2] + [0.0031 \times (\text{PvO}_{2\text{SI}} \times 7.5)] \text{ (mL/dL)}$ where: HGB – Total Hemoglobin, g/dL HGB _{SI} – Total Hemoglobin, mmol/L SvO ₂ – Venous O ₂ Saturation, % PvO ₂ – Partial Pressure of Venous Oxygen, mmHg PvO _{2SI} – Partial Pressure of Venous Oxygen, kPa and PvO ₂ can be entered by the user in Invasive monitoring mode and is assumed to be 0 during all other monitoring modes	mL/dL
Ca-vO ₂	Arteriovenous Oxygen Content Difference $\text{Ca-vO}_2 = \text{CaO}_2 - \text{CvO}_2 \text{ (mL/dL)}$ where: CaO ₂ – Arterial Oxygen Content (mL/dL) CvO ₂ – Venous Oxygen Content (mL/dL)	mL/dL
CI	Cardiac Index $\text{CI} = \text{CO} / \text{BSA}$ where: CO – Cardiac Output, L/min BSA – Body Surface Area, m ²	L/min/m ²
CPI	Cardiac Power Index $\text{CPI} = \text{MAP} \times \text{CI} \times 0.0022$	W/m ²
CPO	Cardiac Power Output $\text{CPO} = \text{CO} \times \text{MAP} \times K$ where: cardiac power output (CPO) (W) was calculated as $\text{MAP} \times \text{CO} / 451$ K is the conversion factor (2.22×10^{-3}) into watts MAP in mmHg CO L/min	W
DO ₂	Oxygen Delivery $\text{DO}_2 = \text{CaO}_2 \times \text{CO} \times 10$ where: CaO ₂ – Arterial Oxygen Content, mL/dL CO – Cardiac Output, L/min	mL O ₂ /min
DO ₂ I	Oxygen Delivery Index $\text{DO}_2\text{I} = \text{CaO}_2 \times \text{CI} \times 10$ where: CaO ₂ – Arterial Oxygen Content, mL/dL CI – Cardiac Index, L/min/m ²	mL O ₂ /min/m ²
dP/dt	Systolic slope calculated as maximal first derivative of arterial pressure waveform with respect to time $\text{dP/dt} = \max(\text{P}[n+1] - \text{P}[n]) / \text{ts}, \text{ for } n=0 \text{ to } N=1$ where: P[n] – current sample of the arterial pressure signal, mmHg ts – sampling time interval, second N – total number of samples in a given cardiac cycle	mmHg/sec

Table C-1 Cardiac and oxygenation profile equations (continued)

Parameter	Description and formula	Units
Ea _{dyn}	Dynamic Arterial Elastance $Ea_{dyn} = PPV/SVV$ where: SVV – Stroke Volume Variation, % PPV – Pulse Pressure Variation, %	none
EDV	End Diastolic Volume $EDV = SV/EF$ where: SV – Stroke Volume (mL) EF – Ejection Fraction, % (efu)	mL
EDVI	End Diastolic Volume Index $EDVI = SVI/EF$ where: SVI – Stroke Volume Index (mL/m ²) EF – Ejection Fraction, % (efu)	mL/m ²
ESV	End Systolic Volume $ESV = EDV - SV$ where: EDV – End Diastolic Volume (mL) SV – Stroke Volume (mL)	mL
ESVI	End Systolic Volume Index $ESVI = EDVI - SVI$ where: EDVI – End Diastolic Volume Index (mL/m ²) SVI – Stroke Volume Index (mL/m ²)	mL/m ²
LVSWI	Left Ventricular Stroke Work Index $LVSWI = SVI \times (MAP - PAWP) \times 0.0136$ $LVSWI = SVI \times (MAP_{SI} - PAWP_{SI}) \times 0.0136 \times 7.5$ where: SVI – Stroke Volume Index, mL/beat/m ² MAP – Mean Arterial Pressure, mmHg MAP _{SI} – Mean Arterial Pressure, kPa PAWP – Pulmonary Artery Wedge Pressure, mmHg PAWP _{SI} – Pulmonary Artery Wedge Pressure, kPa	g-m/m ² /beat
O ₂ El	Oxygen Extraction Index $O_2El = \{(SaO_2 - SvO_2) / SaO_2\} \times 100 (\%)$ where: SaO ₂ – Arterial O ₂ Saturation, % SvO ₂ – Mixed Venous O ₂ Saturation, %	%
O ₂ ER	Oxygen Extraction Ratio $O_2ER = (Ca-vO_2 / CaO_2) \times 100 (\%)$ where: CaO ₂ – Arterial Oxygen Content, mL/dL Ca-vO ₂ – Arteriovenous Oxygen Content Difference, mL/dL	%

Table C-1 Cardiac and oxygenation profile equations (continued)

Parameter	Description and formula	Units
PPV	Pulse Pressure Variation $PPV = 100 \times (PP_{max} - PP_{min}) / \text{mean}(PP)$ where: PP – Pulse Pressure, mmHg calculated as: $PP = SYS - DIA$ SYS – systolic pressure DIA – diastolic pressure	%
PVR	Pulmonary Vascular Resistance $PVR = \{(MPAP - PAWP) \times 80\} / CO$ $PVR = \{(MPAP_{SI} - PAWP_{SI}) \times 60\} / CO$ where: MPAP – Mean Pulmonary Artery Pressure, mmHg MPAP _{SI} – Mean Pulmonary Artery Pressure, kPa PAWP – Pulmonary Artery Wedge Pressure, mmHg PAWP _{SI} – Pulmonary Artery Wedge Pressure, kPa CO – Cardiac Output, L/min	dyne-s/cm ⁵ (kPa-s/L) _{SI}
PVRI	Pulmonary Vascular Resistance Index $PVRI = \{(MPAP - PAWP) \times 80\} / CI$ $PVRI = \{(MPAP_{SI} - PAWP_{SI}) \times 60\} / CI$ where: MPAP – Mean Pulmonary Artery Pressure, mmHg MPAP _{SI} – Mean Pulmonary Artery Pressure, kPa PAWP – Pulmonary Artery Wedge Pressure, mmHg PAWP _{SI} – Pulmonary Artery Wedge Pressure, kPa CI – Cardiac Index, L/min/m ²	dyne-s-m ² /cm ⁵ (kPa-s-m ² /L) _{SI}
RVSWI	Right Ventricular Stroke Work Index $RVSWI = SVI \times (MPAP - CVP) \times 0.0136$ $RVSWI = SVI \times (MPAP_{SI} - CVP_{SI}) \times 0.0136 \times 7.5$ where: SVI – Stroke Volume Index, mL/beat/m ² MPAP – Mean Pulmonary Artery Pressure, mmHg MPAP _{SI} – Mean Pulmonary Artery Pressure, kPa CVP – Central Venous Pressure, mmHg CVP _{SI} – Central Venous Pressure, kPa	g-m/m ² /beat
StO ₂	Tissue Oxygen Saturation $StO_2 = [HbO_2 / (HbO_2 + Hb)] \times 100$ where: HbO ₂ – Oxygenated Hemoglobin Hb – De-Oxygenated Hemoglobin	%
SV	Stroke Volume $SV = (CO / PR) \times 1000$ where: CO – Cardiac Output, L/min PR – Pulse rate, beats/min	mL/beat

Table C-1 Cardiac and oxygenation profile equations (continued)

Parameter	Description and formula	Units
SVI	Stroke Volume Index $SVI = (CI/PR) \times 1000$ where: CI – Cardiac Index, L/min/m ² PR – Pulse rate, beats/min	mL/beat/m ²
SVR	Systemic Vascular Resistance $SVR = \{(MAP - CVP) \times 80\} / CO$ (dyne-sec/cm ⁵) $SVR = \{(MAP_{SI} - CVP_{SI}) \times 60\} / CO$ where: MAP – Mean Arterial Pressure, mmHg MAP _{SI} – Mean Arterial Pressure, kPa CVP – Central Venous Pressure, mmHg CVP _{SI} – Central Venous Pressure, kPa CO – Cardiac Output, L/min	dyne-s/cm ⁵ (kPa-s/L) _{SI}
SVRI	Systemic Vascular Resistance Index $SVRI = \{(MAP - CVP) \times 80\} / CI$ $SVRI = \{(MAP_{SI} - CVP_{SI}) \times 60\} / CI$ where: MAP – Mean Arterial Pressure, mmHg MAP _{SI} – Mean Arterial Pressure, kPa CVP – Central Venous Pressure, mmHg CVP _{SI} – Central Venous Pressure, kPa CI – Cardiac Index, L/min/m ²	dyne-s-m ² /cm ⁵ (kPa-s-m ² /L) _{SI}
SVV	Stroke Volume Variation $SVV = 100 \times (SV_{max} - SV_{min}) / \text{mean}(SV)$	%
VO ₂	Oxygen Consumption $VO_2 = Ca-vO_2 \times CO \times 10$ (mL O ₂ /min) where: Ca-vO ₂ – Arteriovenous Oxygen Content Difference, mL/dL CO – Cardiac Output, L/min	mL O ₂ /min
VO _{2e}	Estimated Oxygen Consumption Index when ScvO ₂ is being monitored $VO_{2e} = Ca-vO_2 \times CO \times 10$ (mL O ₂ /min) where: Ca-vO ₂ – Arteriovenous Oxygen Content Difference, mL/dL CO – Cardiac Output, L/min	mL O ₂ /min
VO _{2I}	Oxygen Consumption Index VO_2 / BSA	mL O ₂ /min/m ²

Table C-1 Cardiac and oxygenation profile equations (continued)

Parameter	Description and formula	Units
VO ₂ le	Estimated Oxygen Consumption Index VO ₂ e / BSA	mL O ₂ /min/m ²
VQI	Ventilation Perfusion Index $VQI = \frac{\{1.38 \times HGB \times (1.0 - (SaO_2/100)) + (0.0031 \times PAO_2)\}}{\{1.38 \times HGB \times (1.0 - (SvO_2/100)) + (0.0031 \times PAO_2)\}} \times 100$ $VQI = \frac{\{1.38 \times HGB_{SI} \times 1.611344 \times (1.0 - (SaO_2/100)) + (0.0031 \times PAO_2)\}}{\{1.38 \times HGB_{SI} \times 1.611344 \times (1.0 - (SvO_2/100)) + (0.0031 \times PAO_2)\}} \times 100$ where: HGB – Total Hemoglobin, g/dL HGB_{SI} – Total Hemoglobin, mmol/L SaO₂ – Arterial O₂ Saturation, % SvO₂ – Mixed Venous O₂ Saturation, % PAO₂ – Alveolar O₂ Tension, mmHg and: $PAO_2 = ((PBAR - PH_2O) \times FiO_2) - PaCO_2 \times (FiO_2 + (1.0 - FiO_2)/0.8)$ where: FiO₂ – Fraction of Inspired Oxygen PBAR – 760 mmHg PH₂O – 47 mmHg PaCO₂ – 40 mmHg	%

Appendix D

Monitor Settings and Defaults

Contents

Patient Data Input Range	.390
Trend Scale Default Limits	.390
Parameter Display and Configurable Alarm/Target Ranges	.391
Alarm and Target Defaults	.393
Alarm Priorities	.395
Language Default Settings	.396

D.1 Patient Data Input Range

Table D-1 Patient information

Parameter	Minimum	Maximum	Available units
Gender	M (Male) / F (Female)	N/A	N/A
Age	2	120	years
Height	12 in / 30 cm	98 in / 250 cm	inches (in) or cm
Weight	2 lbs / 1.0 kg	881 lbs / 400.0 kg	lbs or kg
BSA	0.08	5.02	m ²
ID	0 digits	40 characters	None

D.2 Trend Scale Default Limits

Table D-2 Graphical trend parameter scale defaults

Parameter	Units	Minimum default value	Maximum default value	Setting increment
CO/iCO/sCO	L/min	0.0	12.0	1.0
CI/iCI/sCI	L/min/m ²	0.0	12.0	1.0
CPO	W	0.0	9.99	0.01
CPI	W/m ²	0.0	9.99	0.01
SV	mL/b	0	160	20

Table D-2 Graphical trend parameter scale defaults

Parameter	Units	Minimum default value	Maximum default value	Setting increment
SVI	mL/b/m ²	0	80	20
SVV	%	0	50	10
ScvO ₂ /SvO ₂	%	0	99	10
StO ₂	%	1	99	10
SVR/iSVR	dyne-s/cm ⁵	500	1500	100
SVRI/iSVRI	dyne-s-m ² /cm ⁵	500	3000	200
EDV/sEDV	mL	0	800	25
EDVI/sEDVI	mL/m ²	0	400	25
RVEF/sRVEF	%	0	100	10
SYS _{ART}	mmHg	80	160	5
SYS _{PAP}	mmHg	0	55	5
DIA _{ART}	mmHg	50	110	5
DIA _{PAP}	mmHg	0	35	5
MAP	mmHg	50	130	5
MPAP	mmHg	0	45	5
PPV	%	0	50	10
PR	bpm	40	130	5
dP/dt	mmHg/sec	0	2000	100
Ea _{dyn}	none	0.2	1.5	0.1
HPI	none	0	100	10
ΔctHb	none	-20	20	5

NOTE The HemoSphere advanced monitor will not accept a setting of an upper scale setting that is less than the lower scale setting. Nor will it accept a lower scale setting that is higher than the upper scale setting.

D.3 Parameter Display and Configurable Alarm/Target Ranges

Table D-3 Configurable parameter alarm and display ranges

Parameter	Units	Display Range	Configurable Alarm/Target Range	Setting increment
CO	L/min	1.0 to 20.0	1.0 to 20.0	1.0
iCO	L/min	0.0 to 20.0	0.0 to 20.0	1.0

Table D-3 Configurable parameter alarm and display ranges (continued)

Parameter	Units	Display Range	Configurable Alarm/Target Range	Setting increment
sCO	L/min	1.0 to 20.0	1.0 to 20.0	1.0
CO _{20s}	L/min	1.0 to 20.0	1.0 to 20.0	1.0
CI	L/min/m ²	0.0 to 20.0	0.0 to 20.0	1.0
iCI	L/min/m ²	0.0 to 20.0	0.0 to 20.0	1.0
sCI	L/min/m ²	0.0 to 20.0	0.0 to 20.0	1.0
CI _{20s}	L/min/m ²	0.0 to 20.0	0.0 to 20.0	1.0
CPO	W	0.0 to 9.99	0.0 to 9.99	0.01
CPI	W/m ²	0.0 to 9.99	0.0 to 9.99	0.01
SV	mL/b	0 to 300	0 to 300	1.0
SV _{20s}	mL/b	0 to 300	0 to 300	1.0
SVI	mL/b/m ²	0 to 200	0 to 200	1.0
SVI _{20s}	mL/b/m ²	0 to 200	0 to 200	1.0
SVR	dyne-s/cm ⁵	0 to 5000	0 to 5000	1.0
SVRI	dyne-s-m ² /cm ⁵	0 to 9950	0 to 9950	1.0
iSVR	dyne-s/cm ⁵	0 to 5000	0 to 5000	1.0
iSVRI	dyne-s-m ² /cm ⁵	0 to 9950	0 to 9950	1.0
SVV	%	0 to 99	0 to 99	1.0
Venous oximetry (ScvO ₂ / SvO ₂)	%	0 to 99	0 to 99	1.0
Tissue oximetry (StO ₂) [*]	%	0 to 99	0 to 99	1.0
ΔctHb [*]	none	0 to 20	N/A [^]	N/A
EDV	mL	0 to 800	0 to 800	1.0
sEDV	mL	0 to 800	0 to 800	1.0
EDVI	mL/m ²	0 to 400	0 to 400	1.0
sEDVI	mL/m ²	0 to 400	0 to 400	1.0
RVEF	%	0 to 100	0 to 100	1.0
sRVEF	%	0 to 100	0 to 100	1.0
CVP [*]	mmHg	0 to 50	0 to 50	1.0
MAP [*]	mmHg	0 to 300	10 to 300	1.0
ART/PAP/CVP [*] (live pressure waveform display)	mmHg	-34 to 312	0 to 300	1.0
MPAP [*]	mmHg	0 to 99	0 to 99	1.0
SYS _{ART} [*]	mmHg	0 to 300	10 to 300	1.0
SYS _{PAP} [*]	mmHg	0 to 99	0 to 99	1.0
DIA _{ART} [*]	mmHg	0 to 300	10 to 300	1.0

Table D-3 Configurable parameter alarm and display ranges (continued)

Parameter	Units	Display Range	Configurable Alarm/Target Range	Setting increment
DIA _{PAP} *	mmHg	0 to 99	0 to 99	1.0
PPV	%	0 to 99	0 to 99	1.0
PR	bpm	0 to 220	0 to 220	1.0
HPI	none	0 to 100	N/A†	N/A
dP/dt	mmHg/sec	0 to 3000	0 to 3000	1.0
Ea _{dyn}	none	0.0 to 3.0	N/A^	N/A
HR _{avg}	bpm	0 to 220	0 to 220	1.0

*Parameter is available in Non-Pulsatile mode. Blood pressure parameters are only available in Non-Pulsatile mode with minimally-invasive and invasive monitoring. ART waveform, MAP, DIA_{ART}, and SYS_{ART} are not available in Non-Pulsatile mode during non-invasive monitoring.

†Parameter alarm range for HPI is non-configurable.

^Ea_{dyn} and $\Delta ctHb$ are non-alarming parameters. Ranges shown here are for display only.

D.4 Alarm and Target Defaults

Table D-4 Parameter alarm red zone and target defaults

Parameter	Units	EW default lower alarm (red zone) setting	EW default lower target setting	EW default upper target setting	EW default upper alarm (red zone) setting
CI/iCI/sCI/CI _{20s}	L/min/m ²	1.0	2.0	4.0	6.0
SVI/SVI _{20s}	mL/b/m ²	20	30	50	70
SVRI/iSVRI	dyne-s-m ² /cm ⁵	1000	1970	2390	3000
SVV	%	0	0	13	20
ScvO ₂ /SvO ₂	%	50	65	75	85
StO ₂	%	50	60	85	90
EDVI/sEDVI	mL/m ²	40	60	100	200
RVEF/sRVEF	%	20	40	60	60
DO ₂ I	mL O ₂ /min/m ²	300	500	600	800
VO ₂ I/VO ₂ Ie	mL O ₂ /min/m ²	80	120	160	250
CVP	mmHg	2	2	8	10
SYS _{ART}	mmHg	90	100	130	150
SYS _{PAP}	mmHg	10	14	23	34
DIA _{ART}	mmHg	60	70	90	100
DIA _{PAP}	mmHg	0	4	13	16

Table D-4 Parameter alarm red zone and target defaults

Parameter	Units	EW default lower alarm (red zone) setting	EW default lower target setting	EW default upper target setting	EW default upper alarm (red zone) setting
MAP	mmHg	60	70	100	120
MPAP	mmHg	5	9	18	25
HR _{avg}	bpm	60	70	100	120
HGB	g/dL	7.0	11.0	17.0	19.0
	mmol/L	4.3	6.8	10.6	11.8
SaO ₂	%	90	94	100	100
PPV	%	0	0	13	20
PR	bpm	60	70	100	120
HPI	none	0	N/A	N/A	85
dP/dt	mmHg/sec	380	480	1300	1800

NOTE Non-indexed ranges are based on indexed ranges and entered BSA values.

D.5 Alarm Priorities

Table D-5 Parameter alarms, faults, and alerts priorities

Physiologic parameter (alarms)/ message type	Lower physiological alarm (red zone) priority	Upper physiological alarm (red zone) priority	Message type priority
CO/CI/sCO/sCI/CO _{20s} /CI _{20s}	High	Medium	
CPO/CPI	Medium	N/A	
SV/SVI/SV _{20s} /SVI _{20s}	High	Medium	
SVR/SVRI	Medium	Medium	
SVV	Medium	Medium	
ScvO ₂ /SvO ₂	High	Medium	
StO ₂	High	N/A	
EDV/EDVI/sEDV/sEDVI	Medium	Medium	
RVEF/sRVEF	Medium	Medium	
SYS _{ART} /SYS _{PAP}	High	High	
DIA _{ART} /DIA _{PAP}	High	High	
MAP	High	High	
PR	High	High	
MPAP	Medium	Medium	
CVP	Medium	Medium	
PPV	Medium	Medium	
Fault			Medium/High
Alert			Low

NOTE

The alarm signal generation delay is parameter dependent. For oximetry associated parameters, the delay is less than 2 seconds after the parameter is out of range continuously for 5 or more seconds. For HemoSphere Swan-Ganz module continuous CO and associated parameters, the delay is less than 360 seconds, although typical delay due to parameter calculation is 57 seconds. For HemoSphere pressure cable continuous CO and associated FloTrac system parameters, the delay is 2 seconds for 5 second parameter averaging (after the parameter is out of range continuously for 5 or more seconds for a total of 7 seconds), and 20 seconds for 20 second and 5 minute parameter averaging (see table 6-4 on page 142). For HemoSphere pressure cable with TruWave DPT measured parameters, the delay is 2 seconds, after the parameter is out of range continuously for 5 or more seconds (total of 7 seconds). For HemoSphere ClearSight module non-invasive continuous CO and associated hemodynamic parameters, the delay is 20 seconds. For real-time blood pressure waveform display while monitoring with the HemoSphere ClearSight module, the delay is 5 heartbeats after the parameter is out of range continuously for 5 or more seconds.

The parameter value will flash at a higher frequency for a high priority physiological

alarm as compared to a medium physiological alarm. If medium and high priority alarms are sounding at the same time, the physiological high priority alarm tone will be heard. If a low priority alarm is active and a medium or higher priority alarm is generated, the low priority alarm visual indicator will be replaced by the higher priority alarm visual indicator.

Most technical faults are medium priority. Alerts and other system messages are low priority.

D.6 Language Default Settings

Table D-6 Language default settings

Language	Default display units				Time format	Date format	CO trend averaging time
	PaO ₂	HGB	Height	Weight			
English (US)	mmHg	g/dL	in	lbs	12 hour	MM/DD/YYYY	20 seconds
English (UK)	kPa	mmol/L	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Français	kPa	mmol/L	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Deutsch	kPa	mmol/L	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Italiano	kPa	mmol/L	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Español	kPa	mmol/L	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Svenska	kPa	mmol/L	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Nederlands	kPa	mmol/L	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Ελληνικά	kPa	mmol/L	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Português	kPa	mmol/L	cm	kg	24 hour	DD.MM.YYYY	20 seconds
日本語	mmHg	g/dL	cm	kg	24 hour	MM/DD/YYYY	20 seconds
中文	kPa	mmol/L	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Čeština	kPa	mmol/l	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Polski	kPa	mmol/l	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Suomi	kPa	mmol/l	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Norsk	kPa	mmol/L	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Dansk	kPa	mmol/L	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Eesti	mmHg	mmol/L	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Lietuvių	mmHg	g/dl	cm	kg	24 hour	DD.MM.YYYY	20 seconds
Latviešu	kPa	mmol/L	cm	kg	24 hour	DD.MM.YYYY	20 seconds
<i>Note: Temperature defaults to Celsius for all languages.</i>							

NOTE Languages listed above are for reference only and may not be available for selection.

Computation Constants

Contents

Computation Constant Values	397

E.1 Computation Constant Values

In iCO mode, the HemoSphere Swan-Ganz module computes cardiac output employing either a bath probe setup or an in-line temperature probe using the computation constants listed in the following tables. The HemoSphere Swan-Ganz module automatically senses the type of injectate temperature probe being used, and the corresponding injectate temperature, catheter size, and injectate volume define the computation constant to be used.

NOTE	The computation constants given below are nominal and generally applicable to the specified catheter sizes. For computation constants specific to the catheter being used, refer to the catheter directions for use.

Model-specific computation constants are entered manually in the setup menu for the iCO mode.

Table E-1 Computation constants for bath temperature probe

Injectate temperature range* (°C)	Injectate volume (mL)	Catheter size (French)				
		8	7.5	7	6	5.5
Room temp. 22.5–27 °C	10	0.612	0.594	0.595	0.607	0.616
	5	0.301	0.283	0.287	0.304	0.304
	3	0.177	0.159	0.165	0.180	0.180
Room temp. 18–22.5 °C	10	0.588	0.582	0.578	0.597	0.606
	5	0.283	0.277	0.274	0.297	0.298
	3	0.158	0.156	0.154	0.174	0.175
Cold (iced) 5–18 °C	10	0.563	0.575	0.562	0.573	0.581
	5	0.267	0.267	0.262	0.278	0.281
	3	0.148	0.150	0.144	0.159	0.161
Cold (iced) 0–5 °C	10	0.564	0.564	0.542	0.547	0.555
	5	0.262	0.257	0.247	0.259	0.264
	3	0.139	0.143	0.132	0.144	0.148

**To optimize cardiac measurement, it is recommended that the temperature of the injectate correspond to one of the temperature ranges listed in the catheter's directions for use.*

Table E-2 Computation constants for in-line temperature probe

Injectate temperature range* (°C)	Injectate volume (mL)	Catheter size (French)				
		8	7.5	7	6	5.5
Room temp. 22.5–27 °C	10	0.601	0.599	0.616	0.616	0.624
	5	0.294	0.301	0.311	0.307	0.310
Room temp. 18–22.5 °C	10	0.593	0.593	0.603	0.602	0.612
	5	0.288	0.297	0.295	0.298	0.304
Cold (iced) 5–18 °C	10	0.578	0.578	0.570	0.568	0.581
	5	0.272	0.286	0.257	0.276	0.288
Cold (iced) 0–5 °C	10	0.562	0.563	0.537	0.533	0.549
	5	0.267	0.276	0.217	0.253	0.272

**To optimize cardiac measurement, it is recommended that the temperature of the injectate correspond to one of the temperature ranges listed in the catheter's directions for use.*

Appendix F

System Care, Service and Support

Contents

General Maintenance	399
Cleaning the Monitor and Modules	400
Cleaning the Platform Cables	400
Service and Support	404
Edwards Lifesciences Regional Headquarters	405
Monitor Disposal	406
Preventive Maintenance	406
Testing of Alarm Signals	407
Warranty	407

F.1 General Maintenance

The HemoSphere advanced monitor contains no user-serviceable parts, and should be repaired only by qualified service representatives. Hospital biomedics or service technician can refer to the HemoSphere advanced monitor service manual for information on maintenance and recurrent testing. This appendix provides instructions for cleaning the monitor and monitor accessories and contains information on how to contact your local Edwards representative for support and information on repair and/or replacement.

WARNING The HemoSphere advanced monitor contains no user-serviceable parts. Removing the cover or any other disassembly will expose you to hazardous voltages.

CAUTION Clean and store the instrument and accessories after each use.

CAUTION The HemoSphere advanced monitor modules and platform cables are electrostatic discharge (ESD) sensitive. Do not attempt to open cable or module housing or use if the housing has been damaged.

F.2 Cleaning the Monitor and Modules

WARNING **Shock or fire hazard!** Do not immerse the HemoSphere advanced monitor, modules, or platform cables in any liquid solution. Do not allow any fluids to enter the instrument.

The HemoSphere advanced monitor and modules can be cleaned using a lint-free cloth dampened with cleaning agents that are based on the following chemical content:

- 70% isopropyl alcohol
- 2% glutaraldehyde
- 10% bleach solution (sodium hypochlorite)
- quaternary ammonium solution

Do not use any other cleaning agents. Unless otherwise stated, these cleaning agents are approved for all HemoSphere advanced monitoring accessories, cables and modules.

NOTE Once inserted, modules do not need to be removed unless maintenance or cleaning is necessary. If it is necessary to remove platform modules, store them in a cool, dry place in original packaging to prevent damage.

CAUTION Do not pour or spray liquid on any portion of the HemoSphere advanced monitor, accessories, modules, or cables.

Do not use any disinfecting solution other than the types specified.

DO NOT:

- Allow any liquid to come in contact with the power connector
- Allow any liquid to penetrate connectors or openings in the monitor case or modules

If any liquid does come in contact with any of the above mentioned items, DO NOT attempt to operate the monitor. Disconnect power immediately and call your Biomedical Department or local Edwards representative.

F.3 Cleaning the Platform Cables

Platform cables, such as the pressure-out cable, can be cleaned using the cleaning agents listed above in section F.2 and the following methods.

CAUTION Conduct periodic inspections of all cables for defects. Do not coil cables tightly when storing.

- 1 Moisten a lint-free cloth with disinfectant and wipe the surfaces.
- 2 Follow the disinfectant wipe with rinsing wipes using cotton gauze moistened with sterile water. Use sufficient rinsing wipes to remove all residual disinfectant.
- 3 Dry the surface with a clean dry cloth.

Store platform cables in a cool, dry place in original packaging to prevent damage. Additional instructions specific to certain cables are listed in the following sub-sections.

CAUTION Do not use any other cleaning agents, spray, or pour cleaning solution directly on platform cables.
Do not steam, radiate, or EO sterilize platform cables.
Do not immerse platform cables.

F.3.1 Cleaning the HemoSphere Oximetry Cable

Use the cleaning agents listed above in section F.2 to clean the oximetry cable housing and the connecting cable. The fiber-optic interface of the oximetry cable must be kept clean. The optical fibers within the oximetry catheter fiber optic connector mate with the optical fibers in the oximetry cable. Moisten a lint-free cotton-tipped applicator with sterile alcohol and apply gentle pressure to clean the optical fibers recessed within the front of the oximetry cable housing.

CAUTION Do not steam, radiate, or EO sterilize the HemoSphere oximetry cable.
Do not immerse the HemoSphere oximetry cable.

F.3.2 Cleaning the Patient CCO Cable and Connector

The patient CCO cable contains electrical and mechanical components and is therefore subject to normal use wear and tear. Visually inspect the cable insulation jacket, strain relief and connectors before each use. If any of the following conditions are present, discontinue use of the cable.

- Broken insulation
 - Frays
 - Connector pins are recessed or bent
 - Connector is chipped and/or cracked
- 1 The patient CCO cable is not protected against fluid ingress. Wipe the cable with a damp, soft cloth using 10% bleach and 90% water solution as needed.
 - 2 Air dry the connector.

CAUTION If any electrolytic solution, for example Ringer's lactate solution, is introduced into the cable connectors while they are connected to the monitor, and the monitor is turned on, the excitation voltage can cause electrolytic corrosion and rapid degradation of the electrical contacts.

Do not immerse any cable connectors in detergent, isopropyl alcohol or glutaraldehyde.

Do not use a hot air gun to dry cable connectors.

- 3 Please contact Technical Support or your local Edwards representative for further assistance.

F.3.3 Cleaning the HemoSphere Pressure Cable

The HemoSphere pressure cable can be cleaned using the cleaning agents listed in section F.2 and methods specified for platform cables at the start of this section (section F.3). Disconnect the pressure cable from the monitor to air dry the transducer connector. To blow dry the transducer connector, use clean, dry wall air, canned air, or CO₂ aerosol for at least two minutes. If left to dry under room conditions, allow the connector to dry for two days before using.

CAUTION If any electrolytic solution, for example Ringer's lactate solution, is introduced into the cable connectors while they are connected to the monitor, and the monitor is turned on, the excitation voltage can cause electrolytic corrosion and rapid degradation of the electrical contacts.

Do not immerse any cable connectors in detergent, isopropyl alcohol or glutaraldehyde.

Do not use a hot air gun to dry cable connectors.

Device contains electronics. Handle with care.

F.3.4 Cleaning the ForeSight Oximeter Cable

Regular cleaning and preventive maintenance of the ForeSight oximeter cable is an important function that should be performed routinely to ensure safe and efficient cable operation. The cable does not require calibration, but following maintenance intervals are recommended:

The cable should be tested upon installation and every six (6) months thereafter. Please contact Edwards Technical Support for more information.

WARNING Do not, under any circumstances, perform any cleaning or maintenance of the ForeSight oximeter cable while the cable is being used to monitor a patient. The cable must be turned off and the HemoSphere advanced monitor power cord disconnected, or the cable must be disconnected from the monitor and the sensors removed from the patient.

WARNING Before starting cleaning or maintenance of any sort, check the ForeSight oximeter cable, cable connections, ForeSight sensors, and other accessories for damage. Check the cables for bent or broken prongs, cracks, or fraying. If any damage is noted, the cable must not be used until it has been inspected and serviced or replaced. Contact Edwards Technical Support.

There is a risk of serious injury or death if this procedure is not followed.

The following cleaning agents are recommended to clean the ForeSight oximeter cable:

- Aspeti-Wipe
- 3M Quat #25
- Metrex CaviCide
- Phenolic germicidal detergent solution (per manufacturer's recommendations)
- Quaternary ammonium germicidal detergent solution (per manufacturer's recommendations)

See the product directions for use and labeling for detailed information on active ingredients and any disinfecting claims.

The ForeSight oximeter cable is designed to be cleaned using wipes or towelettes designed for that purpose. When all surfaces have been cleaned, wipe the entire surface of the cable using a soft cloth dampened with fresh water to remove any trace residue.

The sensor cables may be cleaned using wipes or towelettes designed for that purpose. They may be cleaned by wiping from the ForeSight oximeter cable housing end towards the sensor connections.

F.3.5 Cleaning the Heart Reference Sensor and Pressure Controller

The heart reference sensor (HRS) and pressure controller can be cleaned using the following disinfectants:

- 70% isopropyl alcohol solution
 - 10% sodium hypochlorite water solution
- 1 Moisten a clean cloth with disinfectant and wipe the surfaces.
 - 2 Dry the surface with a clean, dry cloth.

CAUTION Do not disinfect the heart reference sensor or pressure controller by autoclave or gas sterilization.

Do not immerse the pressure controller, heart reference sensor, or any cable connectors in fluid.

Clean and store the heart reference sensor after each use.

F.3.5.1 Removing the Pressure Controller Band




Figure F-1 Removing pressure controller from band

To remove the pressure controller from the pressure controller band, pull the sleeve slightly outwards (see step 1 in figure F-1) and tilt the pressure controller to remove it from the sleeve (see step 2 in figure F-1). The pressure controller band is intended for limited reuse. The operator shall assess whether reuse is appropriate. When reused, follow the platform cleaning instruction listed in *Cleaning the Monitor and Modules* on page 400. Replace if damaged.

F.3.6 Cleaning the Smart Pressure Controller

The Smart pressure controller can be cleaned using the following disinfectant wipes, or equivalent:

- Clorox HealthCare™ Bleach Germicidal Wipes, PDI Sani-Cloth® Bleach Germicidal Disposable wipes
 - PDI Super Sani-Cloth® Germicidal Disposable wipes, 70% Isopropyl alcohol solution
 - Metrex Cavicide1™ / CaviWipes1™
 - Clorox HealthCare™ Hydrogen Peroxide Cleaner Disinfectant wipe, Hydrogen Peroxide solution (3%)
- 1** Grab unused wipe from container or moisten a clean cloth with disinfectant and wipe the surfaces.
 - 2** Dry the surface with a clean, dry cloth.

CAUTION Do not disinfect the Smart pressure controller by autoclave or gas sterilization.
Do not immerse the Smart pressure controller, or any cable connectors in fluid.
Do not use any other cleaning agents aside from those listed.

F.4 Service and Support

See chapter 15: *Troubleshooting* for diagnosis and remedies. If this information does not solve the problem, contact Edwards Lifesciences.

Edwards provides HemoSphere advanced monitor operations support:

- Within the United States and Canada, call 1.800.822.9837.

- Outside the United States and Canada, contact your local Edwards Lifesciences representative.
- E-mail operational support questions to tech_support@edwards.com.

Have the following information before you call:

- The HemoSphere advanced monitor's serial number, located on the rear panel;
- The text of any error message and detailed information as to the nature of the problem.

F.5 Edwards Lifesciences Regional Headquarters

USA: Edwards Lifesciences LLC
One Edwards Way
Irvine, CA 92614 USA
949.250.2500
800.424.3278
www.edwards.com

China: Edwards (Shanghai) Medical
Products Co., Ltd.
Unit 2602-2608, 2 Grand Gateway,
3 Hong Qiao Road, Xu Hui
District
Shanghai, 200030
China
Phone 86.21.5389.1888

Switzerland: Edwards Lifesciences S.A.
Route de l'Etraz 70
1260 Nyon, Switzerland
Phone 41.22.787.4300

India: Edwards Lifesciences (India) Pvt.
Ltd.
Techniplex II, 7th floor,
Unit no 1 & 2, off. S.V.Road
Goregaon west-Mumbai
400062
India
Phone +91.022.66935701 04

Japan: Edwards Lifesciences Ltd.
Nittochi Nishi-Shinjuku Bldg.
6-10-1, Nishi-Shinjuku, Shinjuku-ku,
Tokyo 160-0023 Japan
Phone 81.3.6894.0500

Australia: Edwards Lifesciences Pty Ltd
Unit 2 40 Talavera Road
North Ryde
NSW 2113
PO Box 137, North Ryde BC
NSW 1670
Australia
Phone +61(2)8899 6300

Brazil: Edwards Lifesciences
Avenida das Nações Unidas, 14.401 –
Parque da Cidade
Torre Sucupira – 17º. Andar – cj. 171
Chácara Santo Antonio – São Paulo/SP
CEP: 04794-000
Brazil
Phone 55.11.5567.5200

F.6 Monitor Disposal

To avoid contaminating or infecting personnel, the environment or other equipment, make sure the HemoSphere advanced monitor and/or cables are disinfected and decontaminated appropriately in accordance with your country's laws for equipment containing electrical and electronic parts prior to disposal.

For single use parts and accessories, where not otherwise specified, follow local regulations regarding disposal of hospital waste.

F.6.1 Battery Recycling

Replace the HemoSphere battery pack when it no longer holds a charge. After removal, follow your local recycling guidelines.

CAUTION Recycle or dispose of the lithium-ion battery in accordance to all federal, state, and local laws.

F.7 Preventive Maintenance

Periodically examine the HemoSphere advanced monitor exterior for general physical condition. Make sure the housing is not cracked, broken or dented, and that everything is present. Make sure there is no sign of spilled liquids or signs of abuse.

Routinely inspect cords and cables for fraying and cracks, and make sure there are no exposed conductors. In addition, check that the enclosure door at catheter connection point of the oximetry cable moves freely and latches properly.

F.7.1 Battery Maintenance

F.7.1.1 Battery Conditioning

The battery pack may require periodic conditioning. This feature should only be performed by trained hospital staff or technicians. Refer to the HemoSphere advanced monitor service manual for conditioning instructions.

WARNING **Explosion Hazard!** Do not open battery, dispose of in fire, store at high temperature or short circuit. It may ignite, explode, leak or get hot, causing serious personal injury or death.

F.7.1.2 Battery Storage

The battery pack can remain stored in the HemoSphere advanced monitor. Refer to *HemoSphere Advanced Monitor Characteristics and Specifications* on page 370 for environmental specifications for storage.

NOTE Long term storage at high temperatures may decrease life of battery pack.

F.7.2 HemoSphere ClearSight Module Maintenance

Do not pull on the pressure controller cable when unplugging it from the HemoSphere ClearSight module. If it is necessary to remove the module from the HemoSphere advanced monitor, press the release button to unlatch and slide module out. It is recommended to send the HemoSphere ClearSight module to a qualified Edwards Service Center for routine service and preventive maintenance checks every two years. Additional testing includes a visual inspection, a software inspection, safety testing and functional testing. For more information on the testing contact your local Edwards Lifesciences representative.

F.7.3 HRS Preventive Maintenance

The finger component of the heart reference sensor (HRS) may be damaged if subjected to moderate to significant surface impact. Although the likelihood of damage is small, the resulting displayed values would be biased by the difference in height from the heart to the finger cuff. Even though this damage cannot be seen by looking at the heart reference sensor, it is possible to confirm whether the damage has occurred by following the below procedure prior to each use:

- 1 Connect the heart reference sensor to the pressure controller connected to the HemoSphere advanced monitor and go to the zeroing screen.
- 2 As instructed in *Calibrate the Heart Reference Sensor* on page 218, bring the two ends of the heart reference sensor level with each other.
- 3 Observe the value shown on the zeroing screen.
- 4 Raise one end of the heart reference sensor 6 inches (15 cm) above the other end.
- 5 Observe that the value shown has changed by at least 5 mmHg.
- 6 Reverse the ends such that the other end is now 6 inches (15 cm) above the first end.
- 7 Observe the value shown changed in the opposite direction by at least 5 mmHg from the original value.

If the value does not change as described, then the heart reference sensor may have been damaged. Contact your local Technical Support office as indicated on the inside cover or *Service and Support* on page 404. A replacement unit shall be provided. If the value does change, the heart reference sensor is functioning normally and can be used for hemodynamic monitoring.

F.8 Testing of Alarm Signals

Each time the HemoSphere advanced monitor is powered on, a self test is automatically performed. As a part of the self test, an alarm tone will sound. This indicates that the audible alarm indicators are functioning correctly. For further testing of individual measurement alarms, periodically adjust alarm limits and check that the appropriate alarm behavior is observed.

F.9 Warranty

Edwards Lifesciences (Edwards) warrants that the HemoSphere advanced monitor is fit for the purposes and indications described in the labeling for a period of one (1) year from the date of purchase when used in accordance with the directions for use. Unless equipment is used in accordance with such instructions, this warranty is void and of no effect. No other express or implied warranty exists, including any warranty of merchantability or fitness for a particular purpose. This warranty does not include cables, batteries, probes, or oximetry cables used with the HemoSphere advanced monitor. Edwards' sole obligation and purchaser's exclusive remedy for breach of any warranty shall be limited to repair or replacement of the HemoSphere advanced monitor at Edwards' option.

Edwards shall not be liable for proximate, incidental, or consequential damages. Edwards shall not be obligated under this warranty to repair or replace a damaged or malfunctioning HemoSphere advanced monitor if such damage or malfunction is caused by the customer's use of catheters other than those manufactured by Edwards.

Guidance and Manufacturer's Declaration

Contents

Electromagnetic Compatibility.	409
Instructions for Use	410
Wireless Technology Information	418

G.1 Electromagnetic Compatibility

Reference: *IEC/EN 60601-1-2:2007 and IEC 60601-2-49:2011-02*
 IEC/EN 60601-1-2:2014-02 and IEC 60601-2-49:2011-02

The HemoSphere advanced monitor is intended for use in the electromagnetic environment specified in this appendix. The customer or the user of the HemoSphere advanced monitor should assure that it is used in such an environment. When connected to the HemoSphere advanced monitor, all accessory cables listed in table B-1 on page 381 comply with the EMC standards listed above.

G.2 Instructions for Use

Medical electrical equipment needs special precautions regarding EMC and needs to be installed and put into service according to the EMC information provided in the following information and tables.

WARNING Use of accessories, transducers, and cables other than those specified or provided by the manufacturer of this equipment could result in increased electromagnetic emissions or decreased electromagnetic immunity of this equipment and result in improper operation.

No modification of the HemoSphere advanced monitor is allowed.

Portable and mobile RF communication equipment and other sources of electromagnetic disturbance such as diathermy, lithotripsy, RFID, electromagnetic ant-theft systems and metal detectors can potentially affect all electronic medical equipment, including the HemoSphere advanced monitor. Guidance on maintaining appropriate separation between communications equipment and the HemoSphere advanced monitor is provided in table G-3. The effects of other RF emitters are unknown and may interfere with the function and safety of the HemoSphere monitoring platform.

CAUTION The instrument has been tested and complies with the limits of IEC 60601-1-2. These limits are designed to provide reasonable protection against harmful interference in a typical medical installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to other devices in the vicinity. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to other devices which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving device.
 - Increase the separation between the equipment.
 - Consult the manufacturer for help.
-

NOTE The EMISSIONS characteristics of this equipment make it suitable for use in industrial areas and hospitals (CISPR 11 class A). If it is used in a residential environment (for which CISPR 11 class B is normally required) this equipment might not offer adequate protection to radio-frequency communication services. The user might need to take mitigation measures, such as relocating or re-orienting the equipment.

Table G-1 Electromagnetic emissions

Guidance and Manufacturer's Declaration - Electromagnetic Emissions		
The HemoSphere advanced monitor is intended for use in the electromagnetic environment specified below. The customer or user of the HemoSphere advanced monitor should assure that it is used in such an environment.		
Emissions	Compliance	Description
RF emissions CISPR 11	Group 1	The HemoSphere advanced monitor uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference with nearby electronic equipment.
RF emissions CISPR 11	Class A	The HemoSphere advanced monitor is suitable for use in all establishments other than domestic and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Harmonic emissions IEC 61000-3-2	Class A	
Voltage fluctuation/ Flicker emissions IEC 61000-3-3	Complies	

Table G-2 Guidance and Manufacturer's Declaration - Immunity to RF wireless communications equipment

Test Frequency	Band ¹	Service ¹	Modulation ²	Maximum Power	Distance	Immunity Test Level
MHz	MHz			W	Meters	(V/m)
The HemoSphere advanced monitor is intended for use in the electromagnetic environment specified below. The customer or user of the HemoSphere advanced monitor should ensure that it is used in such an environment.						
385	380 - 390	TETRA 400	Pulse modulation ² 18 Hz	1.8	0.3	27
450	430 - 470	GMRS 460, FRS 460	FM ³ ± 5 kHz deviation 1 kHz sine	2	0.3	28
710 745 780	704 - 787	LTE Band 13, 17	Pulse modulation ² 217 Hz	0.2	0.3	9
810 870 930	800 - 960	GSM 800/900, TETRA 800, iDEN 820, CDMA 850, LTE Band 5	Pulse modulation ² 18 Hz	2	0.3	28
1720 1845 1970	1700 - 1900	GSM 1800; CDMA 1900; GSM 1900; DECT; LTE Band 1, 3, 4, 25; UMTS	Pulse modulation ² 217 Hz	2	0.3	28

Table G-2 Guidance and Manufacturer's Declaration - Immunity to RF wireless communications equipment

Test Frequency	Band ¹	Service ¹	Modulation ²	Maximum Power	Distance	Immunity Test Level
MHz	MHz			W	Meters	(V/m)
The HemoSphere advanced monitor is intended for use in the electromagnetic environment specified below. The customer or user of the HemoSphere advanced monitor should ensure that it is used in such an environment.						
2450	2400 - 2570	Bluetooth, WLAN, 802.11 b/g/n, RFID 2450, LTE Band 7	Pulse modulation ² 217 Hz	2	0.3	28
5240 5500 5785	5100 - 5800	WLAN 802.11a/n	Pulse modulation ² 217 Hz	0.2	0.3	9
<i>Note: If necessary to achieve the IMMUNITY TEST LEVEL, the distance between the transmitting antenna and the ME EQUIPMENT or ME SYSTEM may be reduced to 1 m. The 1 m test distance is permitted by IEC 61000-4-3.</i>						
¹ For some services, only the uplink frequencies are included.						
² The carrier shall be modulated using a 50% duty cycle square wave signal.						
³ As an alternative to FM modulation, 50% pulse modulation at 18 Hz may be used because while it does not represent actual modulation, it would be worst case.						

Table G-3 Recommended Separation Distances between Portable and Mobile RF Communications Equipment and the HemoSphere advanced monitor

The HemoSphere advanced monitor is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. To help prevent electromagnetic interference, maintain a minimum distance between portable and mobile RF communications equipment (transmitters) and the HemoSphere advanced monitor as recommended below, according to the maximum output power of the communications equipment.				
Transmitter Frequency	150 kHz to 80 MHz	80 to 800 MHz	800 to 2500 MHz	2.5 to 5.0 GHz
Equation	$d = 1.2 \sqrt{P}$	$d = 1.2 \sqrt{P}$	$d = 2.3 \sqrt{P}$	$d = 2.3 \sqrt{P}$
Rated Maximum Output Transmitter Power (watts)	Separation Distance (meters)	Separation Distance (meters)	Separation Distance (meters)	Separation Distance (meters)
0.01	0.12	0.12	0.24	0.24
0.1	0.37	0.37	0.74	0.74
1	1.2	1.2	2.3	2.3
10	3.7	3.8	7.4	7.4
100	12	12	23	23
For transmitters rated at a maximum output power not listed above, the recommended separation distance d can be estimated using the equation in the corresponding column, where P is the maximum output power rating of the transmitter in watts according to the transmitter manufacturer. Note 1: At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies. Note 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects, and people.				

Table G-4 In band wireless coexistence – Threshold of Interference (Tol) and Threshold of Communication (ToC) between HemoSphere advanced monitor (EUT) in Invasive mode and external devices

Test Specifications ¹	Threshold of interference (Tol) or Threshold of Communication (ToC) Results					Extrapolated Interference Thresholds based upon the Intended Signal located 3 m away from the HemoSphere advanced monitor							
	Unintended Type and min level	EUT Intended Frequency (EUT)	Frequency of Unintended Signal (MHz)	Unintended Signal Level at EUT (dBm)	I/U Ratio (Tol or ToC)	EIRP (W)	Distance (m)	EIRP (W)	Distance (m)	EIRP (W)	Distance (m)	EIRP (W)	Distance (m)
A (Tol)	Tier 3 / 802.11n 64 qam 20MHzAdj Channel 20dBm (TRP/ EIRP)	2437	2412	25.57	3.85	10	15.80	1	5.00	0.1	1.58	0.01	0.50
A (ToC)		2437	2412	47.56	-18.14	10	1.26	1	0.40	0.1	0.13	0.01	0.04
B (Tol)		5200	5180	32.19	-15.81	10	7.37	1	2.33	0.1	0.74	0.01	0.23
B (ToC)		5200	5180	38.53	-22.15	10	3.55	1	1.12	0.1	0.36	0.01	0.11
C (Tol)		5765	5745	28.17	-12.15	10	11.70	1	3.70	0.1	1.17	0.01	0.37
C (ToC)		5765	5745	30.21	-14.19	10	9.25	1	2.93	0.1	0.93	0.01	0.29
¹ Test Specifications [Threshold of interference (Tol) or Threshold of Communication (ToC) results]:													
A. 2.4 GHz; Ch 6, 2437 MHz – Invasive Mode													
B. 5 GHz, 20 MHz; Ch 40, (5190-5210 MHz) – Invasive Mode													
C. 5 GHz, 20 MHz; Ch 153, (5755-5775 MHz) – Invasive Mode													

Table G-5 In band wireless coexistence – Threshold of Interference (Tol) and Threshold of Communication (ToC) between HemoSphere advanced monitor (EUT) in Non-invasive mode and external devices

Test Specifications ¹	Threshold of interference (Tol) or Threshold of Communication (ToC) Results					Extrapolated Interference Thresholds based upon the Intended Signal located 3 m away from the HemoSphere advanced monitor							
	Unintended Type and min level	EUT Intended Frequency (EUT)	Frequency of Unintended Signal (MHz)	Unintended Signal Level at EUT (dBm)	I/U Ratio (Tol or ToC)	EIRP (W)	Distance (m)	EIRP (W)	Distance (m)	EIRP (W)	Distance (m)	EIRP (W)	Distance (m)
A (Tol)	Tier 3 / 802.11n 64 qam 20MHz Adj Channel 20dBm (TRP/ EIRP)	2437	2412	24.06	3.05	10	18.80	1	5.94	0.1	1.88	0.01	0.59
A (ToC)		2437	2412	47.96	-20.85	10	1.20	1	0.38	0.1	0.12	0.01	0.04
B (Tol)		5200	5180	36.19	-18.7	10	4.65	1	1.47	0.1	0.47	0.01	0.15
B (ToC)		5200	5180	36.19	-18.7	10	4.65	1	1.47	0.1	0.47	0.01	0.15
C (Tol)		5765	5745	28.18	-12.1	10	11.69	1	3.70	0.1	1.17	0.01	0.37
C (ToC)		5765	5745	32.34	-16.26	10	7.24	1	2.29	0.1	0.72	0.01	0.23

¹**Test Specifications [Threshold of interference (Tol) or Threshold of Communication (ToC) results]:**

A. 2.4 GHz; Ch 6, 2437 MHz – Non-Invasive Mode

B. 5 GHz, 20 MHz; Ch 40, (5190-5210 MHz) – Non-Invasive Mode

C. 5 GHz, 20 MHz; Ch 153, (5755-5775 MHz) – Non-Invasive Mode

Table G-6 Electromagnetic Immunity (ESD, EFT, Surge, Dips and Magnetic Field)

Immunity Test	IEC 60601-1-2 Test Level	Compliance Level	Electromagnetic Environment - Guidance
The HemoSphere advanced monitor is intended for use in the electromagnetic environment specified below. The customer or user of the HemoSphere advanced monitor should assure that it is used in such an environment.			
Electrostatic discharge (ESD) IEC 61000-4-2	±8 kV contact	±8 kV	Floors should be wood, concrete, or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
	±15 kV air	±15 kV	
Electrical fast transient/burst IEC 61000-4-4	±2 kV for power supply lines	±2 kV for power supply lines	Mains power quality should be that of a typical commercial and/or hospital environment.
	±1 kV for 1 kV for input/output lines > 3 meters	±1 kV for 1 kV for input/output lines > 3 meters	
Surge IEC 61000-4-5	±1 kV line(s) to line(s)	±1 kV line(s) to line(s)	
	±2 kV line(s) to earth	±2 kV line(s) to earth	
Voltage dips, short interruptions and voltage variations on power supply AC input lines IEC 61000-4-11	0% U_T (100% dip in U_T) for 0.5 cycle (0°, 45°, 90°, 135°, 180°, 225°, 270°, and 315°)	0% U_T	Mains power quality should be that of a typical commercial or hospital environment. If the HemoSphere advanced monitor user requires continued operation during power mains interruptions, it is recommended that the HemoSphere advanced monitor be powered by an uninterruptible power supply or battery.
	0% U_T (100% dip in U_T) for 1 cycle (single phase at 0°)	0% U_T	
	70% U_T (30% dip in U_T) for 25/30 cycles (single phase at 0°)	70% U_T	
	Interrupt: 0% U_T (100% drop in U_T) for 250/300 cycles	0% U_T	
Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	30 A(rms)/m	30 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
<i>Note: U_T is the AC mains voltage prior to application of the test level.</i>			

Table G-7 Electromagnetic Immunity (RF Radiated and Conducted)

Immunity Test	IEC 60601-1-2 Test Level	Compliance Level	Electromagnetic Environment - Guidance
The HemoSphere advanced monitor is intended for use in the electromagnetic environment specified below. The customer or user of the HemoSphere advanced monitor should assure that it is used in such an environment.			
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz	3 Vrms	Portable and mobile RF communication equipment should be used no closer to any part of the HemoSphere advanced monitor, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.
Conducted RF IEC 61000-4-6	6 Vrms (ISM band) 150 kHz to 80 MHz	6 Vrms	Recommended Separation Distance $d = [1.2] \times \sqrt{P}$; 150 kHz to 80 MHz $d = [1.2] \times \sqrt{P}$; 80 MHz to 800 MHz $d = [2.3] \times \sqrt{P}$; 800 MHz to 2500 MHz
Radiated RF IEC 61000-4-3	3 V/m 80 to 2700 MHz	3 V/m	Where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey, ^a should be less than the compliance level in each frequency range. ^b Interference may occur in the vicinity of equipment with the following symbol: 
^aField strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast, and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the HemoSphere advanced monitor is used exceeds the applicable RF compliance level above, the HemoSphere advanced monitor should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the HemoSphere advanced monitor. ^bOver the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m. Note 1: At 80 MHz and 800 MHz, the higher frequency range applies. Note 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			

G.3 Wireless Technology Information

The HemoSphere advanced monitor contains wireless communication technology that provides Wi-Fi connectivity. HemoSphere advanced monitor wireless technology supports IEEE 802.11a/b/g/n with a fully integrated security supplicant providing 802.11i/WPA2 authentication, data encryption.

Technical details of the wireless technology implemented in the HemoSphere advanced monitor are provided in the following table.

Table G-8 HemoSphere advanced monitor wireless information

Feature	Description
Wi-Fi standards	IEEE 802.11a, 802.11b, 802.11g, 802.11n
Wi-Fi media	Direct Sequence-Spread Spectrum (DSSS) Complementary Code Keying (CCK) Orthogonal Frequency Divisional Multiplexing (OFDM)
Wi-Fi Media Access Protocol	Carrier sense multiple access with collision avoidance (CSMA/CA)
Wi-Fi Data Rates Supported	802.11a (OFDM): 6, 9, 12, 18, 24, 36, 48, 54 Mbps 802.11b (DSSS, CCK): 1, 2, 5.5, 11 Mbps 802.11g (OFDM): 6, 9, 12, 18, 24, 36, 48, 54 Mbps 802.11n (OFDM, HT20, MCS 0-7): 6.5, 13, 19.5, 26, 39, 52, 58.5, 72.2 Mbps 7.2, 14.4, 21.7, 28.9, 43.3, 57.8, 65 Mbps
Modulation	BPSK at 1, 6, 6.5, 7.2 and 9 Mbps QPSK at 2, 12, 13, 14.4, 18, 19.5 and 21.7 Mbps CCK at 5.5 and 11 Mbps 16-QAM at 24, 26, 28.9, 36, 39 and 43.3 Mbps 64-QAM at 48, 52, 54, 57.8, 58.5, 65, and 72.2 Mbps
802.11n Spatial Streams	1X1 SISO (Single Input, Single Output)
2.4 GHz Frequency Bands	ETSI: 2.4 GHz to 2.483 GHz FCC: 2.4 GHz to 2.483 GHz MIC: 2.4 GHz to 2.495 GHz KC: 2.4 GHz to 2.483 GHz
2.4 GHz Operating Channels	ETSI: 13 (3 non-overlapping) FCC: 11 (3 non-overlapping) MIC: 14 (4 non-overlapping) KC: 13 (3 non-overlapping)
5 GHz Frequency Bands	ETSI: 5.15 GHz to 5.35 GHz FCC: 5.15 GHz to 5.35 GHz 5.47 GHz to 5.725 GHz 5.47 GHz to 5.725 GHz 5.725 GHz to 5.825 GHz MIC: 5.15 GHz to 5.35 GHz KC: 5.15 GHz to 5.25 GHz 5.47 GHz to 5.725 GHz 5.725 GHz to 5.825 GHz
5 GHz Operating Channels	ETSI: 19 non-overlapping FCC: 24 non-overlapping MIC: 19 non-overlapping KC: 19 non-overlapping

Table G-8 HemoSphere advanced monitor wireless information (continued)

Feature	Description
Maximum Transmit Power <i>Note: Maximum transmits power varies according to individual country regulations. All values nominal, ± 2 dBm. At 2.4 GHz, a single spatial stream and 20 MHz channel bandwidth is supported.</i>	802.11a 6 Mbps 15 dBm (31.623 mW) 54 Mbps 12 dBm (19.953 mW) 802.11b 1 Mbps 16 dBm (39.81 mW) 11 Mbps 16 dBm (39.81 mW) 802.11g 6 Mbps 16 dBm (39.81 mW) 54 Mbps 12 dBm (25.12 mW) 802.11n (2.4 GHz) 6.5 Mbps (MCS0) 16 dBm (39.81 mW) 65 Mbps (MCS7) 12 dBm (15.85 mW) 802.11n (5 GHz HT20) 6.5 Mbps (MCS0) 15 dBm (31.62 mW) 65 Mbps (MCS7) 12 dBm (15.85 mW)
Typical Receiver Sensitivity <i>Note: All values nominal, ± 3 dBm. Variant by channels.</i>	802.11a 6 Mbps -90 dBm 54 Mbps -73 dBm (PER $\leq 10\%$) 802.11b 1 Mbps -89 dBm 11 Mbps -82 dBm (PER $\leq 8\%$) 802.11g 6 Mbps -85 dBm 54 Mbps -68 dBm (PER $\leq 10\%$) 802.11n (2.4 GHz) MCS0 Mbps -86 dBm MCS7 Mbps -65 dBm 802.11n (5 GHz HT20) MCS0 Mbps -90 dBm MCS7 Mbps -70 dBm
Security	Standards IEEE 802.11i (WPA2) Encryption Advanced Encryption Standard (AES, Rijndael Algorithm) Encryption Key Provisioning Pre-Shared (PSK) Dynamic 802.1X Extensible Authentication Protocol Types EAP-FAST, EAP-TLS, EAP-TTLS PEAP-GTC, PEAP-MSCHAPv2, PEAP-TLS LEAP FIPS 140-2 Mode Operation restricted to WPA2-AES with EAP-TLS, and WPA2-PSK/AES

Table G-8 HemoSphere advanced monitor wireless information (continued)

Feature	Description
Compliance	ETSI Regulatory Domain EN 300 328 EN 55022:2006 Class B EN 300 328 v1.8.1 (BT 2.1) EN 55024:1998 +A1:2001, A2:2003 EN 301 489-1 EN 61000-3-2:2006 EN 301 489-17 EN 61000-3-3:1995 +A1:2001, A2:2005 EN 301 893 EU 2002/95/EC (RoHS) EN 60950-1 FCC Regulatory Domain (Certification ID: SQG-WB45NBT) FCC 15.247 DTS – 802.11b/g (Wi-Fi): 2.4 GHz & 5.8 GHz FCC 15.407 UNII – 802.11a (Wi-Fi): 2.4 GHz & 5.4 GHz FCC Part 15 Class B UL 60950 Industry Canada (Certification ID: 3147A-WB45NBT) RSS-210 – 802.11a/b/g/n (Wi-Fi) – 2.4 GHz, 5.8 GHz, 5.2 GHz, and 5.4 GHz ICES-003, Class B MIC (Japan) (Certification ID:  R 201-140137) STD-T71 Article 2 Item 19, Category WW (2.4 GHz Channels 1-13) Article 2 Item 19-2, Category GZ (2.4GHz Channel 14) Article 2 Item 19-3, Category XW (5150-5250 W52 & 5250-5350 W53) KC (Korea) (Certification ID: MSIP-CRM-LAI-WB45NBT) NCC (Taiwan) (Certification ID:  CCAM18LP0760T) acma (Australia) (Certification ID: ABN 75 082 447 194) ANATEL (Brazil) (Certification ID: 05725-17-10188) China (Certification ID: 2018AJ0489(M))
Certifications	Wi-Fi Alliance 802.11a, 802.11b, 802.11g , 802.11n WPA authentication WPA2 authentication Cisco Compatible Extensions (Version 4) FIPS 140-2 Level 1 Linux 3.8 running on 45 Series Wi-Fi Module with ARM926 (ARMv5TEJ) - OpenSSL FIPS Object Module v2.0 (validation certificate #1747)
Antenna Type	PCB Dipole
Antenna Dimensions	36 mm x 12 mm x 0.1 mm

G.3.1 Quality of Service for Wireless and Wired Technology

Quality of service (QoS) is specified in terms of total data loss for a normal connection where the HemoSphere advanced monitor is operating at medium wireless signal strength or higher (table 8-1), with good network connection. HemoSphere advanced monitor wireless data transmission has been validated to have less than 5% total data loss under these conditions. HemoSphere advanced monitor wireless technology has an effective range of 150 feet, line of sight and 75 feet, non-line of sight. The effective range might be affected due to the presence of other wireless emitters.

The HemoSphere advanced monitor supports data transmission via a wired or wireless connection. All data transmitted is expected to be acknowledged by the receiving system. Data is resent if not sent successfully. The HemoSphere advanced monitor automatically tries to re-establish any HIS or Viewfinder hub connections that are interrupted. If a pre-existing connection cannot be reestablished, the HemoSphere advanced monitor alerts the user with an audible alert and message (**Alert: HIS Connectivity Loss** [see table 15-6] or Viewfinder hub connectivity error messages [see table 15-9 on page 344]).

CAUTION The wireless Quality of Service (QoS) may be influenced by the presence of other devices that create radio frequency interference (RFI). Such RFI devices may include electrocautery equipment, cellular telephones, wireless PC and tablets, pagers, RFID, MRI, or other electrically powered devices. When used in the presence of potential RFI devices, consideration should be taken to maximize separation distances and to observe for any potential signs of interference such as loss of communication or reduced Wi-Fi signal strength.

G.3.2 Wireless Security Measures

The wireless signals are secured using industry standard wireless security protocols (table G-8). Wireless security standards WEP and WPA have been shown to be vulnerable to intrusions and are not recommended. Edwards recommends securing wireless data transmission by enabling IEEE 802.11i (WPA2) security and FIPS mode. Edwards also recommends implementing network security measures like virtual LANs with firewalls to further secure HemoSphere advanced monitoring platform data in transit to the HIS.

G.3.3 Troubleshooting Wireless Coexistence Issues

The instrument has been tested and complies with the limits of IEC 60601-1-2. If you experience communication issues with HemoSphere advanced monitor wireless technology, ensure a minimum distance between portable and mobile RF communications equipment (transmitters) and the HemoSphere advanced monitor are maintained. Refer to table G-3 for additional details on separation distances.

G.3.4 Federal Communication Commission (FCC) Interference Statements

IMPORTANT NOTE To comply with FCC RF exposure compliance requirements, the antenna used for this transmitter must be installed to provide a separation distance of at least 20 cm from all persons and must not be co-located or operating in conjunction with any other antenna or transmitter.

Federal Communication Commission Interference Statement

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one of the following measures:

- 1 Reorient or relocate the receiving antenna.
- 2 Increase the separation between the equipment and receiver.
- 3 Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- 4 Consult the dealer or an experienced radio/TV technician for help.

FCC CAUTION Any changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate this equipment.

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

This device is restricted to *indoor* use when operated in the 5.15 to 5.25 GHz frequency range.

FCC requires this product to be used indoors for the frequency range 5.15 to 5.25 GHz to reduce the potential for harmful interference to co-channel Mobile Satellite systems.

This device does not permit operations on channels 116-128 (5580 – 5640 MHz) for 11na and 120-128 (5600-5640 MHz) for 11a which overlap the 5600 - 5650 MHz band.

IMPORTANT NOTE FCC Radiation Exposure Statement:
This equipment complies with FCC radiation exposure limits set forth for an uncontrolled environment. This equipment should be installed and operated with minimum distance 20 cm between the radiator & your body.

G.3.5 Industry Canada Statements

RF Radiation Hazard Warning

To ensure compliance with FCC and Industry Canada RF exposure requirements, this device must be installed in a location where the antennas of the device will have a minimum distance of at least 20 cm from all persons. Using higher gain antennas and types of antennas not certified for use with this product is not allowed. The device shall not be co-located with another transmitter.

Maximum Antenna Gain – If the integrator configures the device such that the antenna is detectable from the host product.

This radio transmitter (IC ID: 3147A-WB45NBT) has been approved by Industry Canada to operate with the antenna types listed below with the maximum permissible gain and required antenna impedance for each antenna type indicated. Antenna types not included in this list, having a gain greater than the maximum gain indicated for that type, are strictly prohibited for use with this device.

“To reduce potential radio interference to other users, the antenna type and its gain should be so chosen that the equivalent isotropically radiated power (EIRP) is not more than that required for successful communication.”

“This device has been designed to operate with an antenna having a maximum gain of [4] dBi. Antenna having a higher gain is strictly prohibited per regulations of Industry Canada. The required antenna impedance is 50 ohms.”

This device complies with Industry Canada license-exempt RSS standard(s). Operation is subject to the following two conditions: (1) this device may not cause interference, and (2) this device must accept any interference, including interference that may cause undesired operation of the device.

G.3.6 European Union Radio Equipment Directive (RED) Statements

This device complies with the essential requirements of the 2014/53/EU – Radio Equipment Directive (RED). The following test methods have been applied in order to prove presumption of conformity with the essential requirements of the 2014/53/EU – Radio Equipment Directive (RED):

- **EN 62368-1:2014/A11:2017**
Safety requirements for audio/video, information, and technology equipment
- **EN 300 328 V2.2.2: (2019-07)**
Electromagnetic compatibility and Radio Spectrum Matters (ERM); Wideband Transmission systems; Data transmission equipment operating in the 2,4 GHz ISM band and using spread spectrum modulation techniques; Harmonized EN covering essential requirements under article 3.2 of the R&TTE Directive
- **EN 62311:2008 | EN 50665:2017 | EN 50385:2017**
RF exposure
- **EN 301 489-1 V2.2.0 (2017-03)**
Electromagnetic compatibility and Radio Spectrum Matters (ERM); ElectroMagnetic Compatibility (EMC) standard for radio equipment and services; Part 1: Common technical requirements
- **EN 301 489-17 V3.2.0 (2017-03)**
Electromagnetic compatibility and Radio spectrum Matters (ERM); ElectroMagnetic Compatibility (EMC) standard for radio equipment and services; Part 17: Specific conditions for 2,4 GHz wideband transmission systems and 5 GHz high performance RLAN equipment
- **EN 301 893 V2.1.1 (2017-05)**
Electromagnetic compatibility and Radio spectrum Matters (ERM); Broadband Radio Access Networks (BRAN); Specific conditions for 5 GHz high performance RLAN equipment
- **EU 2015/863 (RoHS 3)**
Declaration of Compliance – EU Directive 2015/863; Reduction of Hazardous Substances (RoHS)

This device is a 2.4 GHz wideband transmission system (transceiver), intended for use in all EU member states and EFTA countries, except in France and Italy where restrictive use applies.

In Italy the end-user should apply for a license at the national spectrum authorities in order to obtain authorization to use the device for setting up outdoor radio links and/or for supplying public access to telecommunications and/or network services.

This device may not be used for setting up outdoor radio links in France and in some areas the RF output power may be limited to 10 mW EIRP in the frequency range of 2454 – 2483.5 MHz. For detailed information the end-user should contact the national spectrum authority in France.

Hereby, Edwards Lifesciences, declares that this monitor is in compliance with the essential requirements and other relevant provisions of Directive 2014/53/EU.

Glossary

Acumen Hypotension Prediction Index (HPI)

The likelihood that the patient may be trending toward a hypotensive event (MAP < 65 mmHg for at least one minute).

Alarms

Audible and visual indicators that notify operator that a measured patient parameter is outside the alarm limits.

Alarm Limits

Maximum and minimum values for monitored patient parameters.

Analog Input Cable

Cable that transfers data to the HemoSphere advanced monitor from another monitor.

Baseline Blood Temperature

Blood temperature that serves as the basis for cardiac output measurements.

Blood Pressure (BP)

Blood pressure measured with HemoSphere pressure cable.

Blood Temperature (BT)

Temperature of the blood in the pulmonary artery when the catheter is properly positioned.

Body Surface Area (BSA)

The calculated surface area of a human body.

Bolus (iCO) Mode

Functional state of the HemoSphere Swan-Ganz module in which cardiac output is measured by the bolus thermodilution method.

Bolus Injection

A known volume of iced or room temperature fluid, which is injected into a port on the pulmonary artery catheter and serves as the indicator for measuring cardiac output.

Button

A screen image with text that, when touched, initiates an action or provides access to a menu.

Cardiac Index (CI)

Cardiac output adjusted for body size.

Cardiac Output (CO)

Volume of blood ejected per minute from the heart into the systemic circulation measured in liters per minute.

Central Venous Oxygen Saturation (ScvO₂)

Percentage of hemoglobin saturated with oxygen in the venous blood as measured in the superior vena cava (SVC). Displayed as ScvO₂.

Central Venous Pressure (CVP)

The average pressure in the superior vena cava (right atrium) as measured by an external monitor. Indicates venous return to the right side of the heart.

Computation Constant

A constant used in the cardiac output equation that accounts for density of blood and injectate, injectate volume, and indicator loss in the catheter.

Default Settings

Initial operating conditions assumed by the system.

Dynamic Arterial Elastance (Ea_{dyn})

Dynamic arterial elastance is the ratio of pulse pressure variation and stroke volume variation (PPV/SVV). It is an estimate of arterial elastance.

End-Diastolic Volume (EDV)

The volume of blood in the right ventricle at the end of diastole.

End-Diastolic Volume Index (EDVI)

Right heart end diastolic volume adjusted for body size.

Estimated Oxygen Consumption (VO_{2e})

An expression of the estimated rate at which oxygen is used by tissues, usually given in mL/min of oxygen consumed in 1 hour by 1 milligram dry weight of tissue. Calculated with ScvO₂.

FloTrac Arterial Pressure Autocalibrated Cardiac Output (FT-CO)

CO continuously calculated from the arterial blood pressure waveform.

Heart Rate (HR)

Number of ventricular contractions per minute. HR data from an external monitor using an analog input is averaged over time and displayed as HRavg.

Hematocrit (Hct)

Percentage of blood volume that contain red blood cells.

Hemoglobin (HGB)

Oxygen carrying component of red blood cells. Volume of red blood cells measured in grams per deciliter.

Icon

A screen image that represents a specific screen, platform status, or menu item. When enabled and touched, icons initiate an action or provide access to a menu.

Injectate

Fluid used for iCO (bolus thermodilution cardiac output) measurement.

Intermittent Cardiac Index (iCI)

Intermittent cardiac output adjusted according to body size.

Intermittent Cardiac Output (iCO)

Intermittent measurement of blood ejected per minute from the heart into the systemic circulation measured through thermodilution.

Intervention

Steps taken to change a patient's condition.

Mean Arterial Pressure (MAP)

Average systemic arterial blood pressure as measured by an external monitor.

Mixed Venous Oxygen Saturation (SvO₂)

Percentage of hemoglobin saturated with oxygen in the venous blood as measured in the pulmonary artery. Displayed as SvO₂.

Oxygen Consumption (VO₂)

An expression of the rate at which oxygen is used by tissues, usually given in mL/min of oxygen consumed in 1 hour by 1 milligram dry weight of tissue. Calculated with SvO₂.

Oxygen Delivery (DO₂)

Amount of oxygen in milliliters per minute (mL/min) delivered to the tissues.

Oxygen Delivery Index (DO₂I)

Amount of oxygen in milliliters per minute (mL/min/m²) delivered to the tissues, adjusted for body size.

Oximetry (Oxygen Saturation, ScvO₂/SvO₂)

Percentage of hemoglobin saturated with oxygen in the blood.

Patient CCO Cable Test

Test to verify the integrity of the patient CCO cable.

Phlebostatic Axis

Reference axis in the patient that passes through the patient's right atrium in any anatomical plane.

Physiocal

A physiological calibration procedure used to obtain accurate blood pressure readings from the artery of the finger.

Plethysmograph Sensor

A device built into the finger cuff that measures fluctuations of volume within the finger artery.

Pressure Controller (PC2/HEMPC)

The unit worn on the patient's wrist that connects the heart reference sensor and compatible Edwards finger cuffs to the HemoSphere ClearSight module.

Pulse Rate (PR)

Number of arterial blood pressure pulses per minute.

Right Ventricular Ejection Fraction (RVEF)

Percentage of blood volume ejected from the right ventricle during systole.

Sensitivity

The ability of a test to correctly identify those with the condition (true positive rate).

Mathematically defined as:

$$(\text{number of true positives} / [\text{number of true positives} + \text{number of false negatives}]) \times 100$$

Signal Quality Indicator (SQI)

The oximetry signal quality based on the catheter condition and positioning in the vessel.

Specificity

The ability of a test to correctly identify those without the condition (true negative rate).

Mathematically defined as:

$$(\text{number of true negatives} / [\text{number of true negatives} + \text{number of false positives}]) \times 100$$

STAT Value

A fast estimate of CO/CI, EDV/EDVI, and RVEF values.

Stroke Volume (SV)

Amount of blood ejected from the ventricles with each contraction.

Stroke Volume Index (SVI)

Stroke volume adjusted for body size.

Stroke Volume Variation (SVV)

Stroke volume variation is the percent difference between maximum and minimum stroke volume.

Systemic Vascular Resistance (SVR)

A derived measure of impedance to blood flow from left ventricle (afterload).

Systemic Vascular Resistance Index (SVRI)

Systemic vascular resistance adjusted for body size.

Systolic Slope (dP/dt)

The maximum upslope of the arterial pressure waveform measured from a peripheral artery.

Thermal Filament

Area on the CCO thermodilution catheter that transfers small amounts of energy into the blood to serve as indicator for trending cardiac output continuously.

Thermistor

Temperature sensor near the tip of the pulmonary artery catheter.

Thermodilution (TD)

A variant of the indicator dilution technique using temperature change as the indicator.

USB

Universal Serial Bus.

Volume Clamp Method

Arterial blood volume is kept constant using the signal from the photo-plethysmograph and a rapidly changing pressure in the air bladder.

Washout Curve

Indicator dilution curve produced by a bolus injection. Cardiac output is inversely related to the area under this curve.

Caution: Federal (USA) law restricts this device to sale by or on the order of a physician. See instructions for use for full prescribing information.

Edwards, Edwards Lifesciences, and the stylized E logo are trademarks of Edwards Lifesciences Corporation. Acumen, Acumen AFM, Acumen HPI, Acumen IQ, AFM, CCombo, CCombo V, ClearSight, ClearSight Jr, CO-Set, CO-Set+, FloTrac, FloTrac Jr, FORE-SIGHT, ForeSight, FORE-SIGHT ELITE, ForeSight Jr, HemoSphere, HPI, PediaSat, PhysioCal, Swan, Swan-Ganz, Swan-Ganz Jr, Time-In-Target, TruWave, and VitaWave are trademarks of Becton, Dickinson and Company. All other trademarks are the property of their respective owners.

© 2026 Becton, Dickinson and Company. All rights reserved. A/W Part No 10061041002/B

Edwards Lifesciences • One Edwards Way, Irvine CA 92614 USA • [edwards.com](https://www.edwards.com)



Edwards